

DECLARATION OF CONFORMITY

for

Pendulum

Frequency Counters CNT-85 & CNT-85R

Pendulum Instruments AB
Karlsbodavägen 39, Box 20020
SE-161 02 Bromma
Sweden

Statement of Conformity

Based on test results using appropriate standards,
the product is in conformity with
Electromagnetic Compatibility Directive 89/336/EEC
Low Voltage Directive 73/23/EEC

Sample tests

Standards used:

EN 61010-1 (1997)

Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use

EN 61326/A1 (1998)

EMC Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use

The tests have been performed in a
typical configuration.

This Conformity is indicated by the symbol **CE**,
i.e. "Conformité Européenne".