

1865 Specified Accuracy

The resistance accuracy of the 1865 instrument is specified as:

$$\pm[0.45\% + \{(R_x/V_x)(0.0005 * F_s + 2*10^{-12} \text{ A}) + 30\Omega/R_x\} * 100\%]$$

R_x: Measured Resistance in ohms (Ω)

V_x: Programmed voltage in volts (V)

F_s: Full scale current range in amperes (A)

The full-scale current ranges (F_s) are 1mA, 100uA, 10uA, 1uA, 100nA, 10nA and 1nA.

To make it easier to determine the specified accuracy three graphs are illustrated in Figures 1, 2 and 3 for voltage levels of 1000V, 100V and 10V respectively.




Figure 1: Resistance Accuracy at 1000V

Quickly scanning over the three graphs, a basic resistance accuracy of 0.5% is valid over the 1MΩ to 500GΩ range at 1000V; over the 100kΩ to 50GΩ range at 100V and over the 100kΩ to 1GΩ range at 10V.

More Pictures




Figure 2: Resistance Accuracy at 100V




Figure 3: Resistance Accuracy at 10V

For complete product specifications on the 1865 Megohmmeter/IR Tester or any of QuadTech's products, visit us at <http://www.quadtech.com/resources/dataindex.html>. Do you have an application specific testing need? Call us at 1-800-253-1230 or email applications at jkramer@quadtech.com and we'll work with you on a custom solution. Put QuadTech to the test because we're committed to solving your testing requirements.

The information presented here is subject to change and is intended for general information only

©QuadTech, Incorporated

Telephone: 1- 800-253-1230, Website: <http://www.quadtech.com>

Printed in U.S.A.

P/N 035035

January 2001