

Annex to declaration of accreditation (scope of accreditation)  
 Normative document: EN ISO/IEC 17025:2005  
 Registration number: **K 018**

of **Intermes B.V.**  
**Calibration Laboratory**

This annex is valid from: **08-11-2018** to **30-11-2020**

Replaces annex dated: **01-11-2017**

**Location(s) where activities are performed under accreditation**

**Head Office**

Joseph Strumpeterstraat 10  
 7559 SG  
 Hengelo  
 The Netherlands

Location	Abbreviation/ location code
Joseph Strumpeterstraat 10 7559 SG Hengelo The Netherlands	HLO

HCS code	Measured quantity, Instrument, Measure	Range	CMC <sup>1</sup>	Remarks	Location
DM 0 0	DIMENSIONAL QUANTITIES				
DM 1 0	Gauge blocks				HLO
	Gauge blocks, steel	(0.5 - 100) mm (0.02 - 4) inch	0.06 μm + 1.2•10 <sup>-6</sup> •l	Central length, fixed sizes	
	Gauge blocks, tungsten carbide	(0.5 - 100) mm (0.02 - 4) inch	0.06 μm + 0.7•10 <sup>-6</sup> •l	Central length, fixed sizes	
	Gauge blocks, ceramic	(0.5 - 100) mm (0.02 - 4) inch	0.06 μm + 1.0•10 <sup>-6</sup> •l	Central length, fixed sizes	

This annex has been approved by the Board of the Dutch Accreditation Council, on its behalf,

J.A.W.M. de Haas  
 Director of Operations

<sup>1</sup> Calibration and Measurement Capability (CMC): Demonstrated measurement uncertainty, with coverage probability of 95%, in a given measurement point or measurement range. Measurement uncertainty, *U*, is calculated according to EA-4/02 "Expression of the Uncertainty of Measurement in Calibration".

of **Intermes B.V.**  
**Calibration Laboratory**

This annex is valid from: **08-11-2018 to 30-11-2020**

Replaces annex dated: **01-11-2017**

HCS code	Measured quantity, Instrument, Measure	Range	CMC <sup>1</sup>	Remarks	Location
	Gauge blocks, steel / tungsten carbide / ceramic		$0.05 \mu\text{m} + 0.1 \cdot 10^{-6} \cdot l$	Length variation	
	Step gauge	up to 1000 mm	$1.2 \mu\text{m} + 6.0 \cdot 10^{-6} \cdot l$		
DM 2 0	Line scales, distances				HLO
	Rulers (all models)	up to 300 mm	$0.8 \mu\text{m} + 4 \cdot 10^{-6} \cdot l$		
		up to 600 mm	$1.1 \mu\text{m} + 4 \cdot 10^{-6} \cdot l$		
		up to 3000 mm	$6 \mu\text{m} + 5 \cdot 10^{-6} \cdot l$		
		up to 100 m	$6 \mu\text{m} + 6 \cdot 10^{-6} \cdot l$		
DM 3 0	Length measuring instruments				HLO
	1D-measuring machines			Laser interferometer; machine equipped with	
		up to 20 m	$0.15 \mu\text{m} + 0.7 \cdot R + 1.0 \cdot 10^{-6} \cdot l$	Zerodur scales;(1)	
		up to 20 m	$0.15 \mu\text{m} + 0.7 \cdot R + 1.3 \cdot 10^{-6} \cdot l$	Glass scales;(1)	
		up to 20 m	$0.15 \mu\text{m} + 0.7 \cdot R + 1.6 \cdot 10^{-6} \cdot l$	Steel scales;(1)	
		up to 400 mm	$0.3 \mu\text{m} + 0.7 \cdot R + 3 \cdot 10^{-6} \cdot l$	Optical systems (1)	
		up to 700 mm	$0.05 \mu\text{m} + 0.5 \cdot 10^{-6} \cdot l + S$	Using special gauge blocks (1)	
	Handheld tools for external measurements	(0 - 200) mm	$0.45 \mu\text{m} + 0.5 \cdot R + 25 \cdot 10^{-6} \cdot l$	e.g. vernier, micrometer, (1)	
		(200 - 2000) mm	$4 \mu\text{m} + 0.5 \cdot R + 5 \cdot 10^{-6} \cdot l$		

of **Intermes B.V.**  
**Calibration Laboratory**

This annex is valid from: **08-11-2018** to **30-11-2020**

Replaces annex dated: **01-11-2017**

HCS code	Measured quantity, Instrument, Measure	Range	CMC <sup>1</sup>	Remarks	Location
	Handheld tools for internal measurements (2-point)	(0 - 200) mm	$0.7 \mu\text{m} + 0.5 \cdot R + 25 \cdot 10^{-6} \cdot l$	e.g. vernier, internal micrometer, (1)	
		(200 - 400) mm	$5 \mu\text{m} + 0.5 \cdot R + 4 \cdot 10^{-6} \cdot l$		
	Handheld tools for internal measurements (2- and 3-point)	(1 - 250) mm	$1.5 \mu\text{m} + 0.5 \cdot R + 25 \cdot 10^{-6} \cdot l$	e.g. internal micrometers (1)	
	Handheld tools for height- and depth measurements	(0 - 200) mm	$0.7 \mu\text{m} + 0.5 \cdot R + 25 \cdot 10^{-6} \cdot l$	e.g. (depth) vernier, (1)	
		(200 - 500) mm	$4 \mu\text{m} + 0.5 \cdot R + 5 \cdot 10^{-6} \cdot l$		
	Linear displacement sensor	up to 200 mm	$0.05 \mu\text{m} + 0.7 \cdot R + 2.5 \cdot 10^{-6} \cdot l + S$	e.g. dial gauge, ... (1)	
		(200 - 300) mm	$0.7 \mu\text{m} + 0.7 \cdot R + 3.5 \cdot 10^{-6} \cdot l$	e.g. dial gauge, ...	
	Height gauge	up to 1500 mm	$0.8 \mu\text{m} + 0.7 \cdot R + 2.5 \cdot 10^{-6} \cdot l$	(1)	
	Inside micrometer	up to 300 mm	$0.7 \mu\text{m} + 0.7 \cdot R + 2.5 \cdot 10^{-6} \cdot l$	(1)	
		(300 - 1000) mm	$0.4 \mu\text{m} + 0.7 \cdot R + 2 \cdot 10^{-6} \cdot l$	(1)	
		(1000 - 3000) mm	$0.4 \mu\text{m} + 0.7 \cdot R + 2 \cdot 10^{-6} \cdot l$		
	Film thickness gauge	up to 25 mm	$0.6 \mu\text{m} + 0.7 \cdot R + 22 \cdot 10^{-6} \cdot l$		
	Laser distance meter	up to 25 m	$0.5 \text{ mm} + 40 \cdot 10^{-6} \cdot L + 0,6 \cdot R$		
DM 4 0	Diameter, length				HLO
	Setting rings and ring gauges	Ø (1 - 4) mm	$1.2 \mu\text{m} + 6 \cdot 10^{-6} \cdot l$		
		Ø (4 - 200) mm	$1.0 \mu\text{m} + 2 \cdot 10^{-6} \cdot l$		

of **Intermes B.V.**  
**Calibration Laboratory**

This annex is valid from: **08-11-2018** to **30-11-2020**

Replaces annex dated: **01-11-2017**

HCS code	Measured quantity, Instrument, Measure	Range	CMC <sup>1</sup>	Remarks	Location
		Ø (200 - 500) mm	1.2 µm + 6•10 <sup>-6</sup> •/		
	Pin gauge	up to Ø 300 mm	0.8 µm + 2.5•10 <sup>-6</sup> •/		
	Plug gauge	up to Ø 300 mm	0.8 µm + 2.5•10 <sup>-6</sup> •/		
	Thread wires	up to Ø 20 mm	0.8 µm + 2.5•10 <sup>-6</sup> •/		
	Other external diameters	up to Ø 100 mm	0.5 µm + (1+6•ΔT)•10 <sup>-6</sup> •/	(1)	
		up to Ø 300 mm	0.8 µm + 2.5•10 <sup>-6</sup> •/		
	Other internal diameters	Ø (1 - 4) mm	1.2 µm + 6•10 <sup>-6</sup> •/		
		Ø (4 - 200) mm	1.0 µm + 2•10 <sup>-6</sup> •/		
		Ø (200 - 500) mm	1.2 µm + 6•10 <sup>-6</sup> •/		
	Feeler gauges	up to 5 mm	0.8 µm + 2.5•10 <sup>-6</sup> •/		
	Setting gauges for micrometers	up to 300 mm	0.8 µm + 2.5•10 <sup>-6</sup> •/	(1)	
		(300 - 1000) mm	0.4 µm + 2•10 <sup>-6</sup> •/	(1)	
		(1000 - 3000) mm	0.4 µm + 2•10 <sup>-6</sup> •/		
	Other distances for parallel faces	up to 300 mm	0.8 µm + 2.5•10 <sup>-6</sup> •/	(1)	
		(300 - 1000) mm	0.4 µm + 2•10 <sup>-6</sup> •/	(1)	
		(1000 - 3000) mm	0.4 µm + 2•10 <sup>-6</sup> •/		
DM 5 0	Form error				HLO
	Roundness in- and externally	Ø (1 - 500 mm)	0.05 µm + 0.01•A	A=roundness deviation	
	Roundness testers and other instruments for measuring roundness		0.04 µm + 0.5•R	(1)	
	Knife edge straight edge	up to 100 mm	0.25 µm	(1)	
		(100 - 300) mm	0.6 µm	(1)	
		(300 - 500) mm	0.7 µm	(1)	

of **Intermes B.V.**  
**Calibration Laboratory**

This annex is valid from: **08-11-2018** to **30-11-2020**

Replaces annex dated: **01-11-2017**

HCS code	Measured quantity, Instrument, Measure	Range	CMC <sup>1</sup>	Remarks	Location
	Straight edge	up to 10 m	$0.4 \mu\text{m} + 0.25 \cdot 10^{-6} \cdot l$	(1)	
	Surface plate	Up to 6 x 10 m <sup>2</sup>	$0.2 \mu\text{m} + 1.5 \cdot 10^{-6} \cdot l$	l = longest side of the surface plate (1)	
	Flick standard (roundness standard)		0.15 $\mu\text{m}$		
DM 6 0	Roughness				HLO
	Surface texture measuring instruments	Ra up to 5 $\mu\text{m}$	$0.01 \mu\text{m} + 0.02 \cdot A + 0.5 \cdot R+S$	A = Ra-value of reference (1)	
		Rz up to 10 $\mu\text{m}$	$0.01 \mu\text{m} + 0.05 \cdot A + 0.5 \cdot R+S$	A = Rz-value of reference (1)	
		Rt Rmax up to 10 $\mu\text{m}$	$0.01 \mu\text{m} + 0.05 \cdot A + 0.5 \cdot R+S$	A = Rt Rmax-value of reference (1)	
	Roughness standards	Ra up to 10 $\mu\text{m}$	$0.015 \mu\text{m} + 0.045 \cdot A$	A = measured Ra-value	
		Rz up to 15 $\mu\text{m}$	$0.025 \mu\text{m} + 0.07 \cdot A$	A = measured Rz-value	
		Rt (Rmax) up to 15 $\mu\text{m}$	$0.025 \mu\text{m} + 0.07 \cdot A$	A = measured Rt (Rmax)-value	
	Groove depth (-standaard)	up to 6 mm	$0.05 \mu\text{m} + 0.007 \cdot A$	A = measured profile height	
DM 7 0	Thread quantities external			e.g. screw plug gauge	HLO
	Pitch	up to 10 mm	2 $\mu\text{m}$		
	Profile angle	up to 180°	$(0.2 + 9/L)$ arcmin		
	Single pitch diameter	$\varnothing(1 - 300)$ mm	$\alpha = 30^\circ: \pm (6.0 - 7.5) \mu\text{m}$	According to Euramet/CG-10/V.02, method 1a or 1b	
		$\varnothing(1 - 300)$ mm	$\alpha = 55^\circ 60^\circ: \pm (3.2 - 4.1) \mu\text{m}$		
		$\varnothing(1 - 300)$ mm	$\alpha = 90^\circ: \pm (2.6 - 3.4) \mu\text{m}$		

of **Intermes B.V.**  
**Calibration Laboratory**

This annex is valid from: **08-11-2018** to **30-11-2020**

Replaces annex dated: **01-11-2017**

HCS code	Measured quantity, Instrument, Measure	Range	CMC <sup>1</sup>	Remarks	Location
	Pitch diameter	Ø(1 - 300) mm	$\alpha = 30^\circ: \pm (6.0 - 7.5) \mu\text{m}$	According to Euramet/CG-10/V.02, method 2a, 2b or 3	
		Ø(1 - 300) mm	$\alpha = 55^\circ 60^\circ: \pm (3.2 - 4.1) \mu\text{m}$		
		Ø(1 - 300) mm	$\alpha = 90^\circ: \pm (2.6 - 3.4) \mu\text{m}$		
DM 7 0	Thread quantities internal			e.g. screw ring gauge	
	Pitch	up to 10 mm	2 $\mu\text{m}$		
	Profile angle	up to 180°	(0.2 + 9/L) arcmin	Measurement on cast	
	Simple pitch diameter	Ø(4 - 200) mm	$\alpha = 30^\circ: \pm (9 - 14) \mu\text{m}$	According to Euramet/CG-10/V.02, method 1a or 1b	
		Ø(4 - 200) mm	$\alpha = 55^\circ 60^\circ: \pm (3.6 - 7) \mu\text{m}$		
		Ø(4 - 200) mm	$\alpha = 90^\circ: \pm (3.1 - 6.2) \mu\text{m}$		
	Pitch diameter	Ø(4 - 200) mm	$\alpha = 30^\circ: \pm (9 - 14) \mu\text{m}$	According to RvA-I-4.05, Euramet/CG-10/V.02, method 2a, 2b or 3	
		Ø(4 - 200) mm	$\alpha = 55^\circ 60^\circ: \pm (3.6 - 7) \mu\text{m}$		
		Ø(4 - 200) mm	$\alpha = 90^\circ: \pm (3.1 - 6.2) \mu\text{m}$		
DM 8 0	Combined instruments				HLO
	1D-, 2D- en 3D-measuring machines	up to 20 m	$0.15 \mu\text{m} + 0.7 \bullet R + 1.0 \bullet 10^{-6} \bullet l$	Laser interferometer, Zerodur scales (1)	
		up to 20 m	$0.15 \mu\text{m} + 0.7 \bullet R + 1.3 \bullet 10^{-6} \bullet l$	Laser interferometer, glass scales (1)	
		up to 20 m	$0.15 \mu\text{m} + 0.7 \bullet R + 1.6 \bullet 10^{-6} \bullet l$	Laser interferometer, steel scales (1)	

of **Intermes B.V.**  
**Calibration Laboratory**

This annex is valid from: **08-11-2018** to **30-11-2020**

Replaces annex dated: **01-11-2017**

HCS code	Measured quantity, Instrument, Measure	Range	CMC <sup>1</sup>	Remarks	Location
		up to 400 mm	$0.3 \mu\text{m} + 0.7 \cdot R + 2.3 \cdot 10^{-6} \cdot l$	Optical systems (1)	
		up to 700 mm	$0.05 \mu\text{m} + 0.5 \cdot 10^{-6} \cdot l + S$	Using special gauge blocks (1)	
	Deviation of nominal displacement	up to 20 m	$0.15 \mu\text{m} + 0.7 \cdot R + 1.0 \cdot 10^{-6} \cdot l$	Laser interferometer, Zerodur scales (1)	
		up to 20 m	$0.15 \mu\text{m} + 0.7 \cdot R + 1.3 \cdot 10^{-6} \cdot l$	Laser interferometer, glass scales (1)	
		up to 20 m	$0.15 \mu\text{m} + 0.7 \cdot R + 1.6 \cdot 10^{-6} \cdot l$	Laser interferometer, steel scales (1)	
		up to 400 mm	$0.3 \mu\text{m} + 0.7 \cdot R + 2.3 \cdot 10^{-6} \cdot l$	Optical systems (1)	
		up to 700 mm	$0.05 \mu\text{m} + 0.5 \cdot 10^{-6} \cdot l + S$	Using special gauge blocks (1)	
	Deviations transverse to the translation directions	up to 0,5 mm	$0.1 \mu\text{m} + 3 \cdot 10^{-6} \cdot l + 0.005 \cdot A$	A = measured deviation; measuring length up to 3000 mm (1)	
	Rotational deviations around the translation direction	up to 400 arcsec	$0.5 \text{ arcsec} + 0.0035 \cdot H$	H = measured angle; only horizontal translations (1)	
		up to 2000 $\mu\text{m}/\text{m}$	$2.5 \mu\text{m}/\text{m} + 0.0035 \cdot H$		
		up to 400 arcsec	$1.6 \text{ arcsec} + 0.007 \cdot H$	Up to 2000 mm translation; ceramic straight edge and 2 displacement sensors (1)	
		up to 2000 $\mu\text{m}/\text{m}$	$8 \mu\text{m}/\text{m} + 0.007 \cdot H$		
	Other rotational deviations	up to 7200 arcsec	$0.5 \text{ arcsec} + 0.0016 \cdot H$	H = measured angle, translation up to 20 m (1)	
	Translation deviation along a rotational axis		0.025 $\mu\text{m}$	(1)	
	Parallelism of a rotation and a translation	translation up to 500 mm	1arcsec	(1)	

of **Intermes B.V.**  
**Calibration Laboratory**

This annex is valid from: **08-11-2018** to **30-11-2020**

Replaces annex dated: **01-11-2017**

HCS code	Measured quantity, Instrument, Measure	Range	CMC <sup>1</sup>	Remarks	Location
	Squareness of 2 translations	up to 500 x 500 mm <sup>2</sup>	1arcsec	Ceramic square and displacement sensor (1)	
		up to 500 x 500 mm <sup>3</sup>	0.6 arcsec	Ceramic square and measurement system op measuring machine; reversal method (1)	
	Squareness of a rotation and a translation	translation up to 150 mm	0.07 µm	(1)	
		translation up to 300 mm	0.7 µm	(1)	
DM 8 1	Tools, products				HLO
	Surface profiles	up to 6 x 120 mm <sup>2</sup>	0.05 µm + 0.007•A	A = measured profile height	
	Roughness	Ra: up to 10 µm	0.015 µm + 0.045•A	A = measured Ra-value	
		Rz: up to 15 µm	0.025 µm + 0.07•A	A = measured Rz-value	
		Rt, Rmax: up to 15 µm	0.025 µm + 0.07•A	A = measured Rt, Rmax-value	
	Straightness	up to 6 x 120 mm <sup>2</sup>	0.05 µm + 0.007•A	A = measured profile height	
		up to 100 mm	0.25 µm		
		(100 - 300) mm	0.6 µm		
		(300 - 500) mm	0.7 µm		
		up to 1000 mm	1.3 µm + 2•10 <sup>-6</sup> •l		
		up to 10 m	0.5 µm + 0.5•10 <sup>-6</sup> •l	(1)	
	Roundness external	up to Ø500 mm	0.05 µm + 0.01•A	A = measured roundness	
	Roundness internal	Ø(0,7 - 500) mm	0.05 µm + 0.01•A	A = measured roundness	

of **Intermes B.V.**  
**Calibration Laboratory**

This annex is valid from: **08-11-2018** to **30-11-2020**

Replaces annex dated: **01-11-2017**

HCS code	Measured quantity, Instrument, Measure	Range	CMC <sup>1</sup>	Remarks	Location
	Cylindricity	up to Ø500 and up to height 100 mm	$0.5 \mu\text{m} + 1.1 \cdot 10^{-6} \cdot H + 0.01 \cdot A$	A = measured cylindricity H = height cylinder	
		up to Ø500 and up to height 500 mm	$1.1 \mu\text{m} + 2 \cdot 10^{-6} \cdot H + 0.01 \cdot A$		
	Coaxiality and concentricity	up to Ø500 and up to height 500 mm	$0.1 \mu\text{m} + 0.02 \cdot A$	A = measured coaxiality / concentricity	
	Flatness	up to Ø60 mm	0.04 $\mu\text{m}$		
		up to Ø145 mm	0.06 $\mu\text{m}$		
		up to Ø300 mm	0.6 $\mu\text{m}$		
		up to 6 x 10 m <sup>2</sup>	$0.2 \mu\text{m} + 1.5 \cdot 10^{-6} \cdot l$	(1)	
	Angles between sides or planes	up to 180°	$(0.2 + 9/A) \text{ arcmin}$	A = leg length; leg length up to 200 mm	
	Diameter external	up to Ø 300 mm	$0.8 \mu\text{m} + 2.5 \cdot 10^{-6} \cdot l$		
		up to Ø 100 mm	$0.5 \mu\text{m} + (1 + 6 \cdot \Delta T) \cdot 10^{-6} \cdot l$	(1)	
		Ø(300 - 500) mm	$1.2 \mu\text{m} + 6 \cdot 10^{-6} \cdot l$		
	Diameter internal	Ø(1 - 4) mm	$1.2 \mu\text{m} + 6 \cdot 10^{-6} \cdot l$		
		Ø(4 - 200) mm	$1.0 \mu\text{m} + 2 \cdot 10^{-6} \cdot l$		
		Ø(200 - 500) mm	$1.2 \mu\text{m} + 6 \cdot 10^{-6} \cdot l$		
DM 8 1	Tools, products	Distance of 2 parallel planes			HLO
	External	up to 300 mm	$0.8 \mu\text{m} + 2.5 \cdot 10^{-6} \cdot l$		
		up to 100 mm	$0.5 \mu\text{m} + (1 + 6 \cdot \Delta T) \cdot 10^{-6} \cdot x$	(1)	
		(300 - 3000)mm	$0.4 \mu\text{m} + 2 \cdot 10^{-6} \cdot l$		
	Internal	Ø(1 - 4) mm	$1.2 \mu\text{m} + 6 \cdot 10^{-6} \cdot l$		
		Ø(4 - 200) mm	$1.0 \mu\text{m} + 2 \cdot 10^{-6} \cdot l$		
		Ø(200 - 500) mm	$1.2 \mu\text{m} + 6 \cdot 10^{-6} \cdot l$		

of **Intermes B.V.**  
**Calibration Laboratory**

This annex is valid from: **08-11-2018** to **30-11-2020**

Replaces annex dated: **01-11-2017**

HCS code	Measured quantity, Instrument, Measure	Range	CMC <sup>1</sup>	Remarks	Location
DM 8 1	Tools, products	Thread external			HLO
	Pitch	up to 10 mm	2 µm		
	Profile angle	up to 180°	(0.2 + 9/L) arcmin		
	Simple pitch diameter	Ø(1 - 300) mm	α = 30°: ± (6.0 – 7.5) µm	According to Euramet/CG-10 V 2.0, method 1a or 1b	
		Ø(1 - 300) mm	α = 55° 60°: ± (3.2 – 4.1) µm		
		Ø(1 - 300) mm	α = 90°: ± (2.6 – 3.4) µm		
	Pitch diameter	Ø(1 - 300) mm	α = 30°: ± (6.0 – 7.5) µm	According to Euramet/CG-10 V 2.0, method 2a, 2b or 3	
		Ø(1 - 300) mm	α = 55° 60°: ± (3.2 – 4.1) µm		
		Ø(1 - 300) mm	α = 90°: ± (2.6 – 3.4) µm		
	Simple pitch diameter	Ø(4 - 100) mm	α = 30°: ± (9 – 14) µm	According to Euramet/CG-10 V 2.0, method 1a or 1b	
		Ø(4 - 100) mm	α = 55° 60°: ± (3.6 – 7) µm		
		Ø(4 - 100) mm	α = 90°: ± (3.1 – 6) µm		
	Pitch diameter	Ø(4 - 100) mm	α = 30°: ± (9 – 14) µm	According to Euramet/CG-10 V 2.0, method 2a, 2b or 3	
		Ø(4 - 100) mm	α = 55° 60°: ± (3.6 – 7) µm		
		Ø(4 - 100) mm	α = 90°: ± (3.1 – 6) µm		
DM 9 0	Angle measurement				HLO
	Angle gauge block	0° - 180°	2 arcsec		
	Cylindrical square	up to Ø300 mm, up to height 500 mm	0.9 µm + 2.1•10 <sup>-6</sup> •/+ 0.02•A	A = measured squareness	

of **Intermes B.V.**  
**Calibration Laboratory**

This annex is valid from: **08-11-2018** to **30-11-2020**

Replaces annex dated: **01-11-2017**

HCS code	Measured quantity, Instrument, Measure	Range	CMC <sup>1</sup>	Remarks	Location
	Square	up to 500 mm leg length	$0.7 \mu\text{m} + 2.2 \times 10^{-6} \cdot l + 0.02 \cdot A$	A = measured squareness	
	Angle plate	90°	0.5 arcsec		
	Autocollimator	up to 12.5 mm/m	$0.5 \mu\text{m}/\text{m} + 0.001 \cdot H + 0.7 \cdot R$		
		up to 2600 arcsec	$0.1 \text{ arcsec} + 0.001 \cdot H + 0.7 \cdot R$		
	Spirit level	up to 12.5 mm/m	$0.5 \mu\text{m}/\text{m} + 0.001 \cdot H + 0.7 \cdot R$		
		up to 2600 arcsec	$0.1 \text{ arcsec} + 0.001 \cdot H + 0.7 \cdot R$		
DM 9 1	Angle measurement				HLO
	Leveling instruments		0.01 mm/m		
DM 9 2	Angle measurement				HLO
	Polygon	up to 360°	0.5 arcsec		
	Pentagon prism	90°	0.5 arcsec		
DM 9 3	Angle measurement				HLO
	Deviation of the nominal rotation	360°	$0.9 \text{ arcsec} + 0.7 \cdot R$	f.i. rotary heads and rotary table (1)	
DM 9 4	Angle measurement				HLO
	Clinometer	up to 360°	5 arcsec	(1)	

HCS code	Measured quantity, Instrument, Measure	Frequency	CMC <sup>1</sup>	Remarks	Location
LF 0 0	DC / LF				
LF 1 0	Direct voltage				HLO
	0 mV – 200mV		$1.0 \cdot 10^{-5} \cdot U$ , minimum 0.1 $\mu\text{V}$	Measuring (1)	
	0.2 V – 2 V		$7 \cdot 10^{-6} \cdot U$	Measuring (1)	

of **Intermes B.V.**  
**Calibration Laboratory**

This annex is valid from: **08-11-2018** to **30-11-2020**

Replaces annex dated: **01-11-2017**

HCS code	Measured quantity, Instrument, Measure	Frequency	CMC <sup>1</sup>	Remarks	Location
	2 V – 20 V		$5 \cdot 10^{-6} \cdot U$	Measuring (1)	
	20 V – 200 V		$7 \cdot 10^{-6} \cdot U$	Measuring (1)	
	200 V – 1000 V		$5 \cdot 10^{-6} \cdot U$	Measuring (1)	
	0 mV – 220 mV		$2.0 \cdot 10^{-5} \cdot U$ , minimum 1.5 $\mu$ V	Generate (1)	
	0.22 V – 2,2 V		$7 \cdot 10^{-6} \cdot U$	Generate (1)	
	2.2 V – 22 V		$1.0 \cdot 10^{-5} \cdot U$	Generate (1)	
	22 V – 220 V		$1.5 \cdot 10^{-5} \cdot U$	Generate (1)	
	220 V – 1100 V		$1.0 \cdot 10^{-5} \cdot U$	Generate (1)	
LF 2 0	Direct current				HLO
	1 $\mu$ A – 200 $\mu$ A		$1 \cdot 10^{-4} \cdot I$ , minimum 0.5 nA	Measuring (1)	
	200 $\mu$ A – 20 mA		$3 \cdot 10^{-5} \cdot I$	Measuring (1)	
	20 mA – 200 mA		$5 \cdot 10^{-5} \cdot I$	Measuring (1)	
	0.2 A – 2 A		$2.0 \cdot 10^{-4} \cdot I$	Measuring (1)	
	2 A – 20 A		$2.0 \cdot 10^{-4} \cdot I$	Measuring (1)	
	0 $\mu$ A – 220 mA		$1.0 \cdot 10^{-4} \cdot I$ , minimum 0.5 nA	Generate compliance < 0,5 V (1)	
	0.22 A – 2.2 A		$1 \cdot 10^{-4} \cdot I$	Generate compliance < 0.5 V (1)	
	2.2 A – 20 A		$2.0 \cdot 10^{-4} \cdot I$	Generate compliance < 2 V (1)	
	20 A – 1000 A		$5 \cdot 10^{-3} \cdot I$	Generate, with coils	
LF 3 0	Alternating voltage				HLO
	10 mV – 100 mV	20 Hz – 20 kHz	$1.2 \cdot 10^{-3} \cdot U$	Measuring (1)	
	10 mV – 100 mV	20 kHz – 100 kHz	$4 \cdot 10^{-3} \cdot U$	Measuring (1)	

of **Intermes B.V.**  
**Calibration Laboratory**

This annex is valid from: **08-11-2018** to **30-11-2020**

Replaces annex dated: **01-11-2017**

HCS code	Measured quantity, Instrument, Measure	Frequency	CMC <sup>1</sup>	Remarks	Location
	100 mV – 200 mV	20 Hz – 20 kHz	$3 \cdot 10^{-4} \cdot U$	Measuring (1)	
	100 mV – 200 mV	20 kHz – 100 kHz	$4 \cdot 10^{-4} \cdot U$	Measuring (1)	
	0.2 V – 2 V	20 Hz – 10 kHz	$1.0 \cdot 10^{-4} \cdot U$	Measuring (1)	
	0.2 V – 2 V	10 kHz – 100 kHz	$5 \cdot 10^{-4} \cdot U$	Measuring (1)	
	2 V – 20 V	20 Hz – 10 kHz	$1.5 \cdot 10^{-4} \cdot U$	Measuring (1)	
	2 V – 20 V	10 kHz – 100 kHz	$3 \cdot 10^{-4} \cdot U$	Measuring (1)	
	20 V – 200 V	20 Hz – 10 kHz	$1.0 \cdot 10^{-4} \cdot U$	Measuring (1)	
	20 V – 200 V	10 kHz – 100 kHz	$2.0 \cdot 10^{-4} \cdot U$	Measuring (1)	
	200 V – 1000 V	55 Hz – 10 kHz	$1.5 \cdot 10^{-4} \cdot U$	Measuring (1)	
	200 V – 1000 V	10 kHz – 30 kHz	$1.5 \cdot 10^{-3} \cdot U$	Measuring (1)	
	1 kV – 100 kV	50 Hz	$1.0 \cdot 10^{-3} \cdot U$	Measuring (1)	
	2.2 mV – 22 mV	40 Hz – 20 kHz	$5 \cdot 10^{-4} \cdot U$	Generate (1)	
	22 mV – 220 V	40 Hz – 20 kHz	$1.0 \cdot 10^{-4} \cdot U$	Generate (1)	
	220 V – 1100 V	40 Hz – 1 kHz	$1.0 \cdot 10^{-4} \cdot U$	Generate (1)	
LF 4 0	Alternating current				HLO
	10 $\mu$ A – 100 $\mu$ A	55 Hz – 1 kHz	$4 \cdot 10^{-3} \cdot I$	Measuring (1)	
	100 $\mu$ A – 2 mA	55 Hz – 1 kHz	$1.5 \cdot 10^{-4} \cdot I$	Measuring (1)	
	2 mA – 20 mA	55 Hz – 1 kHz	$5 \cdot 10^{-4} \cdot I$	Measuring (1)	
	20 mA – 200 mA	55 Hz – 1 kHz	$1.0 \cdot 10^{-4} \cdot I$	Measuring (1)	
	0.2 A – 2 A	55 Hz – 300 Hz	$3 \cdot 10^{-4} \cdot I$	Measuring (1)	
	0.2 A – 2 A	300 Hz – 1 kHz	$2.0 \cdot 10^{-4} \cdot I$	Measuring (1)	
	2 A – 20 A	55 Hz – 1 kHz	$4 \cdot 10^{-4} \cdot I$	Measuring (1)	
	20 A – 600 A	50 Hz	$6 \cdot 10^{-4} \cdot I$	Measuring (1)	
	100 $\mu$ A – 220 mA	40 Hz – 1 kHz	$2.0 \cdot 10^{-4} \cdot I$	Generate (1)	
	0.22 A – 2.2 A	40 Hz – 1 kHz	$3 \cdot 10^{-4} \cdot I$	Generate (1)	
	2.2 A – 20 A	40 Hz – 440 Hz	$1.0 \cdot 10^{-3} \cdot I$	Generate (1)	
	20 A – 1000 A	45 – 60 Hz	$5 \cdot 10^{-3} \cdot I$	Generate, with coils	

of **Intermes B.V.**  
**Calibration Laboratory**

This annex is valid from: **08-11-2018** to **30-11-2020**

Replaces annex dated: **01-11-2017**

HCS code	Measured quantity, Instrument, Measure	Frequency	CMC <sup>1</sup>	Remarks	Location
	20 A – 200 A	60 – 440 Hz	$7.5 \cdot 10^{-3} \cdot I$	Generate, with coils	
LF 6 1	Resistance				HLO
	100 $\mu\Omega$ - 1 m $\Omega$		$3 \cdot 10^{-4} \cdot R$	Measuring (1)	
	1 m $\Omega$ - 100 m $\Omega$		$1.5 \cdot 10^{-4} \cdot R$	Measuring (1)	
	100 m $\Omega$ - 1 $\Omega$		$5 \cdot 10^{-5} \cdot R$	Measuring (1)	
	1 $\Omega$ – 2 $\Omega$		$3.0 \cdot 10^{-5} \cdot R$	Measuring (1)	
	2 $\Omega$ – 2 k $\Omega$		$1.0 \cdot 10^{-5} \cdot R$	Measuring (1)	
	2 k $\Omega$ – 20 k $\Omega$		$5 \cdot 10^{-6} \cdot R$	Measuring (1)	
	20 k $\Omega$ – 2 M $\Omega$		$1.0 \cdot 10^{-5} \cdot R$	Measuring (1)	
	2 M $\Omega$ - 20 M $\Omega$		$3 \cdot 10^{-5} \cdot R$	Measuring (1)	
	20 M $\Omega$ - 200 M $\Omega$		$1.0 \cdot 10^{-4} \cdot R$	Measuring (1)	
	200 M $\Omega$ – 2 G $\Omega$		$1.5 \cdot 10^{-3} \cdot R$	Measuring (1)	
	0 $\Omega$		70 $\mu\Omega$	Generate (1)	
	100 $\mu\Omega$ , 1 m $\Omega$ , 10 m $\Omega$		$1 \cdot 10^{-4} \cdot R$	Generate (1)	
	100 m $\Omega$		$4 \cdot 10^{-5} \cdot R$	Generate (1)	
	1 $\Omega$ , 1.9 $\Omega$		$8 \cdot 10^{-5} \cdot R$	Generate (1)	
	10 $\Omega$		$2.5 \cdot 10^{-5} \cdot R$	Generate (1)	
	19 $\Omega$ , 100 $\Omega$ , 190 $\Omega$ , 1 k $\Omega$ , 1.9 k $\Omega$ , 10 k $\Omega$ , 19 k $\Omega$ , 100 k $\Omega$ , 190 k $\Omega$		$2.0 \cdot 10^{-5} \cdot R$	Generate (1)	
	1 M $\Omega$ , 1.9 M $\Omega$		$3 \cdot 10^{-5} \cdot R$	Generate (1)	
	10 M $\Omega$		$4 \cdot 10^{-5} \cdot R$	Generate (1)	
	19 M $\Omega$ , 100 M $\Omega$		$6 \cdot 10^{-5} \cdot R$	Generate (1)	

of **Intermes B.V.**  
**Calibration Laboratory**

This annex is valid from: **08-11-2018** to **30-11-2020**

Replaces annex dated: **01-11-2017**

HCS code	Measured quantity, Instrument, Measure	Frequency	CMC <sup>1</sup>	Remarks	Location
LF 6 5	LF Capacity				HLO
	2 nF, 10 nF, 20 nF, 200 nF	1 kHz	$1.0 \cdot 10^{-3} \cdot C$	Generate (1) only sine-shaped signals	

HCS code	Measured quantity, Instrument, Measure	Range	CMC <sup>1</sup>	Remarks	Location
PV 1 0	Gas pressure				
PV 1 1	Absolute pressure	(750 - 1150) hPa a	0.3 hPa	By comparison to a reference barometer	HLO
		(0.01 – 1.1) MPa a	$0.3 \text{ hPa} +  25 \cdot 10^{-5} \cdot (p - 100 \text{ kPa}) $		
		(1.1 – 60.1) MPa a	$ 1 \cdot 10^{-3} \cdot (p - 0.1 \text{ MPa}) $	By comparison with digital pressure indicators (1)	
PV 1 2	Relative pressure	(-1.5 - -90) kPa g	$25 \cdot 10^{-5} \cdot p_e$		HLO
		(1.5 - 1000) kPa g	$25 \cdot 10^{-5} \cdot p_e$		
		(1 - 60) MPa g	$1 \cdot 10^{-3} \cdot p_e$	By comparison with digital pressure indicators (1)	
PV 2 0	Liquid pressure				
PV 2 1	Absolute pressure	(120 - 300) kPa a	$0.03 \text{ hPa} + 1 \cdot 10^{-3} \cdot (p - 100 \text{ kPa})$	By comparison with digital pressure indicators (1)	HLO
		(0.3 – 70.1) MPa a	$0.3 \text{ hPa} + 25 \cdot 10^{-5} \cdot (p - 0.1 \text{ MPa})$		
PV 2 2	Relative pressure	( 20 - 200) kPa g	$1 \cdot 10^{-3} \cdot p_e$	By comparison with	HLO

of **Intermes B.V.**  
**Calibration Laboratory**

This annex is valid from: **08-11-2018** to **30-11-2020**

Replaces annex dated: **01-11-2017**

HCS code	Measured quantity, Instrument, Measure	Range	CMC <sup>1</sup>	Remarks	Location
		(0.2 - 70) MPa g	$25 \cdot 10^{-5} \cdot p_e$	digital pressure indicators (1)	
TE 1 0	Resistance thermometer with and without readout	-20 °C up to 250 °C	0.10 °C	By comparison in liquid block baths with external reference probe (1)	HLO
TE 3 0	Thermocouple with and without readout	-20 °C up to 250 °C	0.10 °C	By comparison in liquid block baths with external reference probe (1)	HLO
TF 2 0	Relative time				HLO
	Electronic chronometers	24 h	0.1 s / 24 h	Direct measurement	
	Mechanical chronometers	24 h	5 s / 24 h	Direct measurement	
TF 2 1	Time and Frequency				HLO
	10 Hz – 225 MHz		$3 \cdot 10^{-6} \cdot f$	Measure	
TF 2 2	Time interval				HLO
	1 μs – 1000 s		$3 \cdot 10^{-6} \cdot t$	Measuring; period applicable to repetitive signals	
	$6 \text{ min}^{-1} - 100.000 \text{ min}^{-1}$		$3 \cdot 10^{-6} \cdot n$	By comparison with frequency references with n = number of revolutions min-1	
TQ 0 0	Torque				HLO
	Torque wrenches	0.1 - 1350 Nm	$1.5 \cdot 10^{-2} \cdot M + 0.5R$	(1)	

Annex to declaration of accreditation (scope of accreditation)  
Normative document: EN ISO/IEC 17025:2005  
Registration number: **K 018**

of **Intermes B.V.**  
**Calibration Laboratory**

This annex is valid from: **08-11-2018** to **30-11-2020**

Replaces annex dated: **01-11-2017**

Remarks:

R = reading accuracy of the instrument

Temperature conditions for electrical calibrations is nominal 23 °C; temperature conditions for geometrical and torque calibrations is nominal 20 °C  
, temperature conditions for pressure and temperature calibrations is nominal 21 °C

$p_e = p - p_{amb}$ :  $p_e$  is overpressure,  $p_{amb}$  is ambient pressure

This list of calibrations is , unless otherwise stated, applicable for calibrations performed inside the IntermeS laboratory.

(1) Calibrations performed at customers' premises.