

Annex to declaration of accreditation (scope of accreditation)
 Normative document: EN ISO/IEC 17025:2005
 Registration number: **K 048**

of **Minerva Meettechniek B.V.**

This annex is valid from: **20-09-2017** to **01-01-2021**

Replaces annex dated: **29-09-2016**

Location(s) where activities are performed under accreditation

Head Office

Chrysantstraat 1
 3812 WX
 Amersfoort
 Nederland

Location	Abbreviation/ location code
Chrysantstraat 1 3812 WX Amersfoort Nederland	AM

HCS code	Measured quantity, Instrument, Measure	Range	CMC ¹	Remarks	Location
MW 1 0	Mass	100 mg - 11 kg	$1,0 \cdot 10^{-5} \cdot m_c + 0,05 \text{ mg}$	Mass piece density $\geq 6400 \text{ kg/m}^3$	AM
	True Mass	100 mg - 11 kg	$1,1 \cdot 10^{-5} \cdot m + 0,05 \text{ mg}$		
PV 1 1	Absolute pressure	0 kPa - 15 kPa	$2,9 \cdot 10^{-5} \cdot p + 0,008 \text{ Pa}$	Nitrogen determination of effective area by means of cross-floating calibration of secondary standards and pressure devices	AM
		5 kPa - 190 kPa	$1,8 \cdot 10^{-5} \cdot p + 0,5 \text{ Pa}$		
		25 kPa - 2,5 MPa	$2,8 \cdot 10^{-5} \cdot p + 0,5 \text{ Pa}$		
		50 kPa - 5 MPa	$3,0 \cdot 10^{-5} \cdot p + 0,5 \text{ Pa}$		

This annex has been approved by the Board of the Dutch Accreditation Council, on its behalf,

J.A.W.M. de Haas
 Director of Operations

¹ Calibration and Measurement Capability (CMC): Demonstrated measurement uncertainty, with coverage probability of 95%, in a given measurement point or measurement range. Measurement uncertainty, U , is calculated according to EA-4/02 "Evaluation of the Uncertainty of Measurement in Calibration".

of **Minerva Meettechniek B.V.**

This annex is valid from: **20-09-2017** to **01-01-2021**

Replaces annex dated: **29-09-2016**

HCS code	Measured quantity, Instrument, Measure	Range	CMC ¹	Remarks	Location
		300 kPa - 20 MPa	$3,1 \cdot 10^{-5} \cdot (p - p_{amb}) + 5,5 \text{ Pa}$		
		1 MPa - 70 MPa	$4,0 \cdot 10^{-5} \cdot (p - p_{amb}) + 19 \text{ Pa}$		
PV 1 2	Gauge pressure	0 kPa - 15 kPa	$3,0 \cdot 10^{-5} \cdot p_e + 0,005 \text{ Pa}$	Nitrogen determination of effective area by means of cross-floating calibration of secondary standards and pressure devices	AM
		5 kPa - 190 kPa	$1,8 \cdot 10^{-5} \cdot p_e + 0,12 \text{ Pa}$		
		25 kPa - 2,5 MPa	$2,8 \cdot 10^{-5} \cdot p_e + 0,06 \text{ Pa}$		
		50 kPa - 5 MPa	$3,0 \cdot 10^{-5} \cdot p_e + 0,12 \text{ Pa}$		
		200 kPa - 20 MPa	$3,1 \cdot 10^{-5} \cdot p_e + 3,8 \text{ Pa}$		
		1 MPa - 70 MPa	$4,0 \cdot 10^{-5} \cdot p_e + 19 \text{ Pa}$		
PV 2 1	Absolute pressure	600 kPa - 50 MPa	$3,1 \cdot 10^{-5} \cdot (p - p_{amb}) + 31 \text{ Pa}$	Oil determination of effective area by means of cross-floating calibration of secondary standards and pressure devices	AM
		2 MPa - 200 MPa	$4,2 \cdot 10^{-5} \cdot (p - p_{amb}) + 50 \text{ Pa}$		
		5 MPa - 500 MPa	$5,8 \cdot 10^{-5} \cdot (p - p_{amb}) + 0,1 \text{ kPa}$		
PV 2 2	Gauge pressure	500 kPa - 50 MPa	$3,1 \cdot 10^{-5} \cdot p_e + 30 \text{ Pa}$	Oil determination of effective area by means of cross-floating calibration of secondary standards and pressure devices	AM
		2 MPa - 200 MPa	$4,2 \cdot 10^{-5} \cdot p_e + 50 \text{ Pa}$		
		5 MPa - 500 MPa	$5,8 \cdot 10^{-5} \cdot p_e + 0,1 \text{ kPa}$		

of **Minerva Meettechniek B.V.**

This annex is valid from: **20-09-2017** to **01-01-2021**

Replaces annex dated: **29-09-2016**

HCS code	Measured quantity, Instrument, Measure	Range	CMC ¹	Remarks	Location
	Differential pressure on elevated line pressure	8 MPa (max. line pressure)	$1 \cdot 10^{-6} \cdot p_e + 5,6 \cdot 10^{-5} \cdot \Delta p + 13 \text{ Pa}$	$\Delta p =$ Differential pressure	
PV 3 1	Under atmospheric pressure	-8 ... -98 kPa	$2,8 \cdot 10^{-5} \cdot p_e + 0,12 \text{ Pa}$	Negative gauge pressure determination of effective area by means of cross-floating calibration of secondary standards and pressure devices	AM

Remarks:

- This annex is applicable to calibrations carried out in the own laboratory.
- The calibrations are carried out at an ambient temperature of 20 °C (nominal).
- $p_e = p - p_{amb}$; p_e is the gauge pressure, p_{amb} is the ambient pressure.
- The accreditation for mass measurements is restricted to calibrations of weights related to pressure balances.
- For a weight with a temperature of 20 °C, the conventional mass is the mass of a reference weight with a density of 8000 kg/m³, which is in balance in air with a density of 1,2 kg/m³.

Annex to declaration of accreditation (scope of accreditation)
Normative document: EN ISO/IEC 17025:2005
Registration number: **K 048**

of **Minerva Meettechniek B.V.**

This annex is valid from: **20-09-2017** to **01-01-2021**

Replaces annex dated: **29-09-2016**

HCS code	Measured quantity, Range	Frequency	CMC ²	Remarks	Location
LF 1 1	DC Voltage				AM
	0 - 10 V		$2,4 \cdot 10^{-5} \cdot U + 50 \mu\text{V}$		
LF 2 1	DC Current				AM
	0 - 100 mA		$5,0 \cdot 10^{-4} \cdot I + 5 \mu\text{A}$		

Remarks:

- This annex is applicable to calibrations carried out in the own laboratory.
- The calibrations are carried out at an ambient temperature of 20 °C (nominal).

² Calibration and Measurement Capability (CMC): Demonstrated measurement uncertainty, with coverage probability of 95%, in a given measurement point or measurement range. Measurement uncertainty, U , is calculated according to EA-4/02 "Evaluation of the Uncertainty of Measurement in Calibration".