

Annex to declaration of accreditation (scope of accreditation)
 Normative document: EN ISO/IEC 17025:2005
 Registration number: **K 050**

of **Tradinco Instrumenten en Apparaten B.V.**
Tradinco Calibration Laboratory (TCL)

This annex is valid from: **31-01-2018** to **30-11-2020**

Replaces annex dated: **17-01-2018**

Location(s) where activities are performed under accreditation

Head Office

Radonstraat 250
 2718 TB
 Zoetermeer
 The Netherlands

Location	Abbreviation/ location code
Radonstraat 250 2718 TB Zoetermeer The Netherlands	ZOE

HCS code	Measured quantity, Range	Frequency	CMC ¹	Remarks	Location
LF 0 0	DC/LF Quantities				
LF 1 1	Direct voltage				ZOE
	10 mV to 100 mV		$5.5 \cdot 10^{-5} \cdot U$	Measuring	
	100 mV to 1 V		$2.0 \cdot 10^{-5} \cdot U$		
	1 V to 1000 V		$1.5 \cdot 10^{-5} \cdot U$		
	1 mV to 10 mV		$3.5 \cdot 10^{-3} \cdot U$	Generating	
	10 mV to 100 mV		$3.5 \cdot 10^{-4} \cdot U$		
	100 mV to 330 mV		$1.0 \cdot 10^{-4} \cdot U$		

¹ Calibration and Measurement Capability (CMC): Demonstrated measurement uncertainty, with coverage probability of 95%, in a given measurement point or measurement range. Measurement uncertainty, *U*, is calculated according to EA-4/02 "Evaluation of the Uncertainty of Measurement in Calibration".

This annex has been approved by the Board of the
 Dutch Accreditation Council, on its behalf,

J.A.W.M. de Haas
 Director of Operations

Annex to declaration of accreditation (scope of accreditation)
 Normative document: EN ISO/IEC 17025:2005
 Registration number: **K 050**

of **Tradinco Instrumenten en Apparaten B.V.**
Tradinco Calibration Laboratory (TCL)

This annex is valid from: **31-01-2018** to **30-11-2020**

Replaces annex dated: **17-01-2018**

HCS code	Measured quantity, Range	Frequency	CMC ¹	Remarks	Location
	330 mV to 1020 V		$9.0 \cdot 10^{-5} \cdot U$		
LF 2 1	Direct current				ZOE
	100 µA to 10 mA		$1.5 \cdot 10^{-4} \cdot I$	Measuring	
	10 mA to 100 mA		$1.8 \cdot 10^{-4} \cdot I$		
	100 mA to 1 A		$3.5 \cdot 10^{-4} \cdot I$		
	100 µA to 300 µA		$6.0 \cdot 10^{-4} \cdot I$	Generating	
	300 µA to 1 mA		$3.0 \cdot 10^{-4} \cdot I$		
	1 mA to 3.3 mA		$1.7 \cdot 10^{-4} \cdot I$		
	3.3 mA to 330 mA		$2.0 \cdot 10^{-4} \cdot I$		
	330 mA to 2.2 A		$4.1 \cdot 10^{-4} \cdot I$		
	2.2 A to 11 A		$6.7 \cdot 10^{-4} \cdot I$		
LF 3 1	Alternating voltage				ZOE
	100 mV to 100 V	40 Hz to 1 kHz	$3.0 \cdot 10^{-4} \cdot U$	Measuring	
	100 mV to 100 V	1 kHz to 100 kHz	$1.7 \cdot 10^{-3} \cdot U$		
	100 V to 500 V	40 Hz to 10 kHz	$5.0 \cdot 10^{-4} \cdot U$		
	500 V to 1000 V	40 Hz to 10 kHz	$1.6 \cdot 10^{-3} \cdot U$		
	33 mV to 330 mV	10 Hz to 45 Hz	$2.2 \cdot 10^{-3} \cdot U + 51 \mu V$	Generating	
		45 Hz to 10 kHz	$6.0 \cdot 10^{-4} \cdot U + 30 \mu V$		
		10 kHz to 20 kHz	$9.0 \cdot 10^{-4} \cdot U + 22 \mu V$		
		20 kHz to 50 kHz	$1.5 \cdot 10^{-3} \cdot U + 41 \mu V$		
		50 kHz to 100 kHz	$2.1 \cdot 10^{-3} \cdot U + 0.15 mV$		

Annex to declaration of accreditation (scope of accreditation)

Normative document: EN ISO/IEC 17025:2005

Registration number: **K 050**

of **Tradinco Instrumenten en Apparaten B.V.**
Tradinco Calibration Laboratory (TCL)

This annex is valid from: **31-01-2018** to **30-11-2020**

Replaces annex dated: **17-01-2018**

HCS code	Measured quantity, Range	Frequency	CMC ¹	Remarks	Location
		100 kHz to 500 kHz	$6.1 \cdot 10^{-3} \cdot U + 0.29 \text{ mV}$		
	330 mV to 3.3 V	10 Hz to 45 Hz	$1.3 \cdot 10^{-3} \cdot U + 0.34 \text{ mV}$		
		45 Hz to 10 kHz	$3.0 \cdot 10^{-4} \cdot U + 0.17 \text{ mV}$		
		10 kHz to 20 kHz	$7.0 \cdot 10^{-4} \cdot U + 0.07 \text{ mV}$		
		20 kHz to 50 kHz	$1.3 \cdot 10^{-3} \cdot U + 0.29 \text{ mV}$		
		50 kHz to 100 kHz	$2.1 \cdot 10^{-3} \cdot U + 1.5 \text{ mV}$		
		100 kHz to 500 kHz	$4.4 \cdot 10^{-3} \cdot U + 2.9 \text{ mV}$		
	3.3 V to 33 V	10 Hz to 45 Hz	$1.3 \cdot 10^{-3} \cdot U + 3.4 \text{ mV}$		
		45 Hz to 10 kHz	$4.0 \cdot 10^{-4} \cdot U + 1.5 \text{ mV}$		
		10 kHz to 20 kHz	$7.0 \cdot 10^{-4} \cdot U + 2.4 \text{ mV}$		
		20 kHz to 50 kHz	$1.7 \cdot 10^{-3} \cdot U + 4.6 \text{ mV}$		
		50 kHz to 100 kHz	$2.1 \cdot 10^{-3} \cdot U + 16 \text{ mV}$		
	33 V to 330 V	45 Hz to 1 kHz	$5.0 \cdot 10^{-4} \cdot U + 15 \text{ mV}$		
		1 kHz to 10 kHz	$7.0 \cdot 10^{-4} \cdot U + 15 \text{ mV}$		
		10 kHz to 20 kHz	$8.0 \cdot 10^{-4} \cdot U + 33 \text{ mV}$		
	330 V to 1020 V	45 Hz to 1 kHz	$4.0 \cdot 10^{-4} \cdot U + 0.17 \text{ V}$		
		1 kHz to 10 kHz	$1.8 \cdot 10^{-3} \cdot U + 0.46 \text{ V}$		
LF 4 1	Alternating current				ZOE
	100 µA to 100 mA	40 Hz to 5 kHz	$1.7 \cdot 10^{-3} \cdot I$	Measuring	
	100 mA to 1 A	40 Hz to 1 kHz	$3.0 \cdot 10^{-3} \cdot I$		
	29 µA to 0.33 mA	10 Hz to 20 Hz	$2.1 \cdot 10^{-3} \cdot I + 0.17 \text{ µA}$	Generating	
		20 Hz to 45 Hz	$1.0 \cdot 10^{-3} \cdot I + 0.17 \text{ µA}$		
		45 Hz to 1 kHz	$1.1 \cdot 10^{-3} \cdot I + 0.22 \text{ µA}$		
		1 kHz to 5 kHz	$3.5 \cdot 10^{-3} \cdot I + 0.14 \text{ µA}$		
		5 kHz to 10 kHz	$1.1 \cdot 10^{-2} \cdot I + 0.16 \text{ µA}$		
	0.33 mA to 3.3 mA	10 Hz to 20 Hz	$1.8 \cdot 10^{-3} \cdot I + 0.27 \text{ µA}$		

of **Tradinco Instrumenten en Apparaten B.V.**
Tradinco Calibration Laboratory (TCL)

This annex is valid from: **31-01-2018 to 30-11-2020**

Replaces annex dated: **17-01-2018**

HCS code	Measured quantity, Range	Frequency	CMC ¹	Remarks	Location
		20 Hz to 45 Hz	$8.0 \cdot 10^{-4} \cdot I + 0.91 \mu\text{A}$		
		45 Hz to 1 kHz	$9.0 \cdot 10^{-4} \cdot I + 0.47 \mu\text{A}$		
		1 kHz to 5 kHz	$1.8 \cdot 10^{-3} \cdot I + 0.41 \mu\text{A}$		
		5 kHz to 10 kHz	$5.2 \cdot 10^{-3} \cdot I + 0.82 \mu\text{A}$		
	3.3 mA to 33 mA	10 Hz to 20 Hz	$1.7 \cdot 10^{-3} \cdot I + 6.1 \mu\text{A}$		
		20 Hz to 45 Hz	$9.0 \cdot 10^{-4} \cdot I + 4.7 \mu\text{A}$		
		45 Hz to 1 kHz	$8.0 \cdot 10^{-4} \cdot I + 4.1 \mu\text{A}$		
		1 kHz to 5 kHz	$1.8 \cdot 10^{-3} \cdot I + 3.6 \mu\text{A}$		
		5 kHz to 10 kHz	$5.2 \cdot 10^{-3} \cdot I + 5.2 \mu\text{A}$		
	33 mA to 330 mA	10 Hz to 20 Hz	$1.7 \cdot 10^{-3} \cdot I + 61 \mu\text{A}$		
		20 Hz to 45 Hz	$9.0 \cdot 10^{-4} \cdot I + 47 \mu\text{A}$		
		45 Hz to 1 kHz	$8.0 \cdot 10^{-4} \cdot I + 41 \mu\text{A}$		
		1 kHz to 5 kHz	$1.8 \cdot 10^{-3} \cdot I + 36 \mu\text{A}$		
		5 kHz to 10 kHz	$5.2 \cdot 10^{-3} \cdot I + 52 \mu\text{A}$		
	330 mA to 2.2 A	10 Hz to 45 Hz	$1.7 \cdot 10^{-3} \cdot I + 0.63 \text{ mA}$		
		45 Hz to 1 kHz	$8.0 \cdot 10^{-4} \cdot I + 0.65 \text{ mA}$		
		1 kHz to 5 kHz	$6.5 \cdot 10^{-3} \cdot I + 0.41 \text{ mA}$		
	2.2 A to 11 A	45 Hz to 65 Hz	$4.0 \cdot 10^{-4} \cdot I + 4.1 \text{ mA}$		
		65 Hz to 500 Hz	$8.0 \cdot 10^{-4} \cdot I + 3.8 \text{ mA}$		
		500 Hz to 1 kHz	$2.8 \cdot 10^{-3} \cdot I + 7.4 \text{ mA}$		
LF 6 2	DC Resistance				ZOE
	1 Ω to 10 Ω		$7.0 \cdot 10^{-5} \cdot R$	4-wire measuring	
	10 Ω to 1 k Ω		$4.0 \cdot 10^{-5} \cdot R$	4-wire	
	1 k Ω to 100 k Ω		$3.0 \cdot 10^{-5} \cdot R$	4-wire	
	100 k Ω to 1 M Ω		$4.0 \cdot 10^{-5} \cdot R$	4-wire	
	1 M Ω to 10 M Ω		$8.5 \cdot 10^{-5} \cdot R$	4-wire	
	10 M Ω to 100 M Ω		$8.5 \cdot 10^{-4} \cdot R$	4-wire	

Annex to declaration of accreditation (scope of accreditation)
 Normative document: EN ISO/IEC 17025:2005
 Registration number: **K 050**

of **Tradinco Instrumenten en Apparaten B.V.**
Tradinco Calibration Laboratory (TCL)

This annex is valid from: **31-01-2018** to **30-11-2020**

Replaces annex dated: **17-01-2018**

HCS code	Measured quantity, Range	Frequency	CMC ¹	Remarks	Location
	100 MΩ to 1 GΩ		$8.5 \cdot 10^{-3} \cdot R$	2-wire	
	1 Ω to 11 Ω		$5.5 \cdot 10^{-3} \cdot R$	4-wire generating	
	11 Ω to 33 Ω		$9.5 \cdot 10^{-4} \cdot R$	4-wire	
	33 Ω to 110 Ω		$4.0 \cdot 10^{-4} \cdot R$	4-wire	
	110 Ω to 110 kΩ		$3.0 \cdot 10^{-4} \cdot R$	4-wire	
	110 kΩ to 3.3 MΩ		$3.0 \cdot 10^{-4} \cdot R$	2-wire	
	3.3 MΩ to 11 MΩ		$7.0 \cdot 10^{-4} \cdot R$	2-wire	
	11 MΩ to 33 MΩ		$1.0 \cdot 10^{-3} \cdot R$	2-wire	
	33 MΩ to 330 MΩ		$4.6 \cdot 10^{-3} \cdot R$	2-wire	
TF 0 0	Time and frequency				
TF 2 1	Frequency				ZOE
	0.01 Hz to 225 MHz		$10 \cdot 10^{-6} \cdot f$	Measuring	
	1 Hz to 300 Hz		$2.5 \cdot 10^{-5} \cdot f + 1 \text{ mHz}$	Generating	
	300 Hz to 2 MHz		$2.5 \cdot 10^{-5} \cdot f$		

HCS code	Measured quantity, Instrument, Measure	Range	CMC ¹	Remarks	Location
TE 9 0	Simulators/ Display Units				
TE 9 1	Resistance thermometer			Electrical Calibration	ZOE
	-200 °C to 0 °C		0.05 °C	Generating and Measuring	
	0 °C to 630 °C		0.10 °C		

Annex to declaration of accreditation (scope of accreditation)
 Normative document: EN ISO/IEC 17025:2005
 Registration number: **K 050**

of **Tradinco Instrumenten en Apparaten B.V.**
Tradinco Calibration Laboratory (TCL)

This annex is valid from: **31-01-2018 to 30-11-2020**

Replaces annex dated: **17-01-2018**

HCS code	Measured quantity, Instrument, Measure	Range	CMC ¹	Remarks	Location
	630 °C to 800 °C		0.30 °C		
TE 9 2	Thermocouples			Electrical Calibration	ZOE
	Type B	600 °C to 800 °C	0.51 °C	Generating and Measuring	
		800 °C to 1000 °C	0.40 °C		
		1000 °C to 1550 °C	0.34 °C		
		1550 °C to 1820 °C	0.34 °C		
	Type C	0 °C to 150 °C	0.30 °C		
		150 °C to 650 °C	0.26 °C		
		650 °C to 1000 °C	0.29 °C		
		1000 °C to 1800 °C	0.49 °C		
		1800 °C to 2316 °C	0.92 °C		
	Type E	-250 °C to -100 °C	0.49 °C		
		-100 °C to -25 °C	0.15 °C		
		-25 °C to 350 °C	0.13 °C		
		350 °C to 650 °C	0.16 °C		
		650 °C to 1000 °C	0.20 °C		
	Type J	-210 °C to -100 °C	0.26 °C		
		-100 °C to -30 °C	0.15 °C		
		-30 °C to 150 °C	0.13 °C		
		150 °C to 760 °C	0.17 °C		
		760 °C to 1200 °C	0.22 °C		
	Type K	-200 °C to -100 °C	0.32 °C		
		-100 °C to -25 °C	0.17 °C		
		-25 °C to 120 °C	0.15 °C		
		120 °C to 1000 °C	0.26 °C		
		1000 °C to 1372 °C	0.38 °C		

Annex to declaration of accreditation (scope of accreditation)
 Normative document: EN ISO/IEC 17025:2005
 Registration number: **K 050**

of **Tradinco Instrumenten en Apparaten B.V.**
Tradinco Calibration Laboratory (TCL)

This annex is valid from: **31-01-2018** to **30-11-2020**

Replaces annex dated: **17-01-2018**

HCS code	Measured quantity, Instrument, Measure	Range	CMC ¹	Remarks	Location
	Type L	-200 °C to -100 °C	0.33 °C		
		-100 °C to 800 °C	0.24 °C		
		800 °C to 900 °C	0.17 °C		
	Type N	-200 °C to -100 °C	0.41 °C		
		-100 °C to -25 °C	0.22 °C		
		-25 °C to 120 °C	0.19 °C		
		120 ° to 410 °C	0.17 °C		
		410 °C to 1300 °C	0.28 °C		
	Type R	0 °C to 250 °C	0.63 °C		
		250 °C to 400 °C	0.38 °C		
		400 °C to 1000 °C	0.35 °C		
		1000 °C to 1767 °C	0.43 °C		
	Type S	0 °C to 250 °C	0.56 °C		
		250 °C to 1000 °C	0.39 °C		
		1000 °C to 1400 °C	0.37 °C		
		1400 °C to 1767 °C	0.45 °C		
	Type T	-250 °C to -150 °C	0.64 °C		
		-150 °C to 0 °C	0.23 °C		
		0 °C to 120 °C	0.15 °C		
		120 °C to 400 °C	0.14 °C		
	Type U	-200 °C to 0 °C	0.50 °C		
		0 °C to 600 °C	0.25 °C		
TE 13 1	Self-indicating thermometers				
	Dry Block Calibrators	-50 °C to 650 °C	0.1 °C	<i>See remark at bottom of the scope</i>	ZOE

of **Tradinco Instrumenten en Apparaten B.V.**
Tradinco Calibration Laboratory (TCL)

This annex is valid from: **31-01-2018** to **30-11-2020**

Replaces annex dated: **17-01-2018**

HCS code	Measured quantity, Instrument, Measure	Range	CMC ¹	Remarks	Location
PV 0 0	Pressure and vacuum				
PV 1 1	Absolute gas pressure	1 kPa to 3.5 kPa	$6 \text{ Pa} + 1.0 \cdot 10^{-4} \cdot p$	Gas	ZOE
		3.5 kPa to 67 kPa	$45 \cdot 10^{-6} \cdot p$ (not smaller than 2 Pa)		
		67 kPa to 7 MPa	$35 \cdot 10^{-6} \cdot p$ (not smaller than 3 Pa)	Gas	
		7 MPa to 40 MPa	$7.0 \cdot 10^{-5} \cdot p + 3 \cdot 10^{-5} \cdot p_{amb} + 6.0 \text{ Pa}$		
PV 1 2	Over atmospheric gas pressure	20 Pa to 1.4 kPa	$2.5 \cdot 10^{-4} \cdot p_e$ (not smaller than 0.15 Pa)		ZOE
		1.4 kPa to 7 MPa	$35 \cdot 10^{-6} \cdot p_e$ (not smaller than 0.35 Pa)		
		7 MPa to 40 MPa	$7.0 \cdot 10^{-5} \cdot p_e$	Oil	
PV 2 1	Absolute liquid pressure	350 kPa to 1.5 MPa	$1.1 \cdot 10^{-4} \cdot p + 6.0 \text{ Pa}$		ZOE
		1.5 MPa to 80 MPa	$7.0 \cdot 10^{-5} \cdot p + 3 \cdot 10^{-5} \cdot p_{amb} + 6.0 \text{ Pa}$		
		80 MPa to 100 MPa	$1.2 \cdot 10^{-4} \cdot p + 6.0 \text{ Pa}$		
		100 MPa to 250 MPa	$2.0 \cdot 10^{-4} \cdot p + 6.0 \text{ Pa}$		
PV 2 2	Over atmospheric liquid pressure	260 kPa to 1.5 MPa	$1.1 \cdot 10^{-4} \cdot p_e$	Oil	ZOE
		1.5 MPa to 80 MPa	$70 \cdot 10^{-6} \cdot p_e$ (not smaller than 150 Pa)		
		80 MPa to 100 MPa	$1.2 \cdot 10^{-4} \cdot p_e$		
		100 MPa to 250 MPa	$2.0 \cdot 10^{-4} \cdot p_e$		

Annex to declaration of accreditation (scope of accreditation)
Normative document: EN ISO/IEC 17025:2005
Registration number: **K 050**

of **Tradinco Instrumenten en Apparaten B.V.**
Tradinco Calibration Laboratory (TCL)

This annex is valid from: **31-01-2018** to **30-11-2020**

Replaces annex dated: **17-01-2018**

HCS code	Measured quantity, Instrument, Measure	Range	CMC ¹	Remarks	Location
PV 3 1	Under atmospheric pressure	-100 kPa to -4 kPa	$1.0 \cdot 10^{-4} \cdot p_e $	Gas <i>See remark at bottom of the scope</i>	ZOE

Remarks:

The calibrations are carried out at an ambient temperature of 20 °C (nominal).
 $p_e = p - p_{amb}$; p_e is the gauge pressure, p_{amb} is the ambient pressure.

This annex is applicable to calibrations carried out in the own laboratory.

PV 3 1: Under atmospheric pressure is measure from ambient air pressure in the negative direction, therefor -100 kPa can only be achieved when the barometric pressure is at least 101 KPa absolute.

TE 13 1: The CMC value excludes the characteristic influence of the Dry Block Calibrators.