

Annex to declaration of accreditation (scope of accreditation)
 Normative document: EN ISO/IEC 17025:2005
 Registration number: **K 063**

of **Dijkstra Advice, Research & EMC Services B.V.**
Trading as DARE!! Calibrations

This annex is valid from: **19-12-2018** to **01-11-2021**

Replaces annex dated: **07-11-2018**

Location(s) where activities are performed under accreditation

Head Office

Vijzelmolenlaan 7
 3447 GX
 Woerden
 The Netherlands

Location	Abbreviation/ location code
Vijzelmolenlaan 7 3447 GX Woerden The Netherlands	W
On-site	O

HCS code	Measured quantity, Instrument, Measure	Range	CMC ¹	Remarks	Location
LF 0 0	DC/LF ELECTRICITY				
LF 1 0	DIRECT VOLTAGE				W
	0 - 2 mV		$6 \cdot 10^{-6} \cdot U + 1.1 \mu V$	Generating	
	2 - 20 mV		$8 \cdot 10^{-6} \cdot U + 1.1 \mu V$		
	20 mV - 200 mV		$1.0 \cdot 10^{-5} \cdot U + 1.0 \mu V$		
	200 mV - 2 V		$1.1 \cdot 10^{-5} \cdot U + 1.0 \mu V$		

¹ Calibration and Measurement Capability (CMC): Demonstrated measurement uncertainty, with coverage probability of 95%, in a given measurement point or measurement range. Measurement uncertainty, *U*, is calculated according to EA-4/02 "Evaluation of the Uncertainty of Measurement in Calibration".

This annex has been approved by the Board of the
 Dutch Accreditation Council, on its behalf,

J.A.W.M. de Haas
 Director of Operations

of **Dijkstra Advice, Research & EMC Services B.V.**
Trading as DARE!! Calibrations

This annex is valid from: **19-12-2018** to **01-11-2021**

Replaces annex dated: **07-11-2018**

HCS code	Measured quantity, Instrument, Measure	Range	CMC ¹	Remarks	Location
	2 V – 20 V		$6 \cdot 10^{-6} \cdot U + 22 \mu V$		W
	20 V – 200 V		$8 \cdot 10^{-6} \cdot U + 0.2 \text{ mV}$		
	200 V – 1000 V		$9 \cdot 10^{-6} \cdot U + 2.7 \text{ mV}$		
	0– 200 mV		$3 \cdot 10^{-5} \cdot U + 0.8 \mu V$	Measuring	
	200 mV – 2 V		$7 \cdot 10^{-6} \cdot U + 2.0 \mu V$		
	2 V – 20 V		$7 \cdot 10^{-6} \cdot U + 20 \mu V$		
	20 V – 200 V		$1.2 \cdot 10^{-5} \cdot U + 0.15 \text{ mV}$		
	200 V – 1000 V		$1.1 \cdot 10^{-5} \cdot U + 1.8 \text{ mV}$		
LF 2 0	DIRECT CURRENT				W
	0 - 200 μA		$5 \cdot 10^{-3} \cdot I + 5 \text{ nA}$	Generating	
	200 μA - 2 mA		$5 \cdot 10^{-4} \cdot I + 12 \text{ nA}$		
	2 - 20 mA		$7 \cdot 10^{-5} \cdot I + 0.12 \mu A$		
	20 - 200 mA		$5 \cdot 10^{-5} \cdot I + 1.8 \mu A$		
	200 mA - 1 A		$1.2 \cdot 10^{-4} \cdot I + 30 \mu A$		
	1 – 2 A		$2.4 \cdot 10^{-4} \cdot I + 0,04 \text{ mA}$		
	0 - 200 μA		$5 \cdot 10^{-2} \cdot I + 5 \text{ nA}$	Measuring	
	200 μA - 2 mA		$5 \cdot 10^{-3} \cdot I + 0.05 \mu A$		
	2 - 20 mA		$5 \cdot 10^{-4} \cdot I + 0.5 \mu A$		
	20 - 200 mA		$1.3 \cdot 10^{-4} \cdot I + 5 \mu A$		
	200 mA - 2 A		$2.4 \cdot 10^{-4} \cdot I + 0.05 \text{ mA}$		

of **Dijkstra Advice, Research & EMC Services B.V.**
Trading as DARE!! Calibrations

This annex is valid from: **19-12-2018 to 01-11-2021**

Replaces annex dated: **07-11-2018**

HCS code	Measured quantity, Instrument, Measure	Range	CMC ¹	Remarks	Location
LF 3 0	ALTERNATING VOLTAGE				W
	1 - 2 mV	30 Hz – 3,3 kHz	$1.1 \cdot 10^{-3} \cdot U + 7 \mu V$	Generating, 2-wire	
		3,3 – 10 kHz	$1.8 \cdot 10^{-3} \cdot U + 7 \mu V$		
		10 - 33 kHz	$4 \cdot 10^{-3} \cdot U + 7 \mu V$		
		33 - 100 kHz	$1.0 \cdot 10^{-2} \cdot U + 7 \mu V$		
	2 mV - 20 mV	30 Hz - 1 kHz	$3.2 \cdot 10^{-4} \cdot U + 7 \mu V$		
		1 - 3,3 kHz	$5 \cdot 10^{-4} \cdot U + 7 \mu V$		
		3,3 – 10 kHz	$1.3 \cdot 10^{-3} \cdot U + 6 \mu V$		
		10 - 33 kHz	$3.5 \cdot 10^{-3} \cdot U + 6 \mu V$		
		33 - 100 kHz	$1.2 \cdot 10^{-2} \cdot U + 6 \mu V$		
	20 mV - 200 mV	30 - 330 Hz	$2.3 \cdot 10^{-4} \cdot U + 12 \mu V$		
		330 Hz - 1 kHz	$3.1 \cdot 10^{-4} \cdot U + 10 \mu V$		
		1 – 3,3 kHz	$6 \cdot 10^{-4} \cdot U + 6 \mu V$		
		3,3 - 10 kHz	$1.3 \cdot 10^{-3} \cdot U + 6 \mu V$		
		10 - 33 kHz	$3.6 \cdot 10^{-3} \cdot U + 6 \mu V$		
		33 - 100 kHz	$1.2 \cdot 10^{-2} \cdot U + 6 \mu V$		
	90 mV - 2V	10 - 32 Hz	$1.3 \cdot 10^{-4} \cdot U + 50 \mu V$	Generating, 4-wire	
		32 - 330 Hz	$9 \cdot 10^{-5} \cdot U + 50 \mu V$		
		330 Hz – 3,3 kHz	$6 \cdot 10^{-5} \cdot U + 30 \mu V$		
		3.3 - 33 kHz	$9 \cdot 10^{-5} \cdot U + 25 \mu V$		
		33 - 100 kHz	$1.0 \cdot 10^{-4} \cdot U + 0.22 \text{ mV}$		
		100 - 330 kHz	$1.6 \cdot 10^{-3} \cdot U + 0.8 \text{ mV}$		
		330 kHz - 1 MHz	$1.4 \cdot 10^{-2} \cdot U + 2.0 \text{ mV}$		
	2 V - 20 V	10 - 32 Hz	$1.1 \cdot 10^{-4} \cdot U + 0.6 \text{ mV}$		
		32 - 330 Hz	$7 \cdot 10^{-5} \cdot U + 0.5 \text{ mV}$		
		330 Hz - 33 kHz	$6 \cdot 10^{-5} \cdot U + 0.4 \text{ mV}$		
		33 - 100 kHz	$2.2 \cdot 10^{-4} \cdot U + 2.0 \text{ mV}$		

of **Dijkstra Advice, Research & EMC Services B.V.**
Trading as DARE!! Calibrations

This annex is valid from: **19-12-2018** to **01-11-2021**

Replaces annex dated: **07-11-2018**

HCS code	Measured quantity, Instrument, Measure	Range	CMC ¹	Remarks	Location
	2 V – 20 V	100 - 330 kHz	$1.6 \cdot 10^{-3} \cdot U + 7 \text{ mV}$		W
		330 kHz - 1 MHz	$1.1 \cdot 10^{-2} \cdot U + 12 \text{ mV}$		
	20 V – 200 V	10 - 32 Hz	$1.7 \cdot 10^{-4} \cdot U + 9 \text{ mV}$		
		32 - 330 Hz	$1.2 \cdot 10^{-4} \cdot U + 6 \text{ mV}$		
		330 Hz - 10 kHz	$7 \cdot 10^{-5} \cdot U + 5 \text{ mV}$		
		10 - 33 kHz	$8 \cdot 10^{-5} \cdot U + 6 \text{ mV}$		
		33 kHz – 100 kHz	$4 \cdot 10^{-4} \cdot U + 20 \text{ mV}$		
	200 V – 1000 V	50 - 330 Hz	$9 \cdot 10^{-4} \cdot U + 50 \text{ mV}$		
		330 Hz- 10 kHz	$7 \cdot 10^{-4} \cdot U + 40 \text{ mV}$		
		10 - 33 kHz	$9 \cdot 10^{-4} \cdot U + 50 \text{ mV}$		
	2 mV – 200 mV	20 - 40 Hz	$3 \cdot 10^{-4} \cdot U + 15 \text{ } \mu\text{V}$	Measuring	
		40 Hz - 2 kHz	$2.8 \cdot 10^{-4} \cdot U + 15 \text{ } \mu\text{V}$		
		2 - 10 kHz	$2.7 \cdot 10^{-4} \cdot U + 15 \text{ } \mu\text{V}$		
		10 - 30 kHz	$5 \cdot 10^{-4} \cdot U + 20 \text{ } \mu\text{V}$		
		30 - 100 kHz	$1.0 \cdot 10^{-3} \cdot U + 40 \text{ } \mu\text{V}$		
	200 mV – 2 V	20 - 40 Hz	$2.1 \cdot 10^{-4} \cdot U + 55 \text{ } \mu\text{V}$		
		40 - 100 Hz	$1.9 \cdot 10^{-4} \cdot U + 55 \text{ } \mu\text{V}$		
		100 - 300 Hz	$1.7 \cdot 10^{-4} \cdot U + 55 \text{ } \mu\text{V}$		
		300 - 1000 Hz	$1.6 \cdot 10^{-4} \cdot U + 40 \text{ } \mu\text{V}$		
		1 - 3 kHz	$1.8 \cdot 10^{-4} \cdot U + 40 \text{ } \mu\text{V}$		
		3 - 10 kHz	$3 \cdot 10^{-4} \cdot U + 0.05 \text{ mV}$		
		10 - 60 kHz	$6 \cdot 10^{-4} \cdot U + 0.30 \text{ mV}$		
		60 - 100 kHz	$6 \cdot 10^{-4} \cdot U + 0.30 \text{ mV}$		
		100 - 300 kHz	$4 \cdot 10^{-3} U + 2.5 \text{ mV}$		
		300 kHz - 1 MHz	$1.2 \cdot 10^{-2} \cdot U + 24 \text{ mV}$		
	2 V – 20 V	20 - 40 Hz	$2.1 \cdot 10^{-4} \cdot U + 0.6 \text{ mV}$		
		40 - 100 Hz	$1.9 \cdot 10^{-4} \cdot U + 0.6 \text{ mV}$		

of **Dijkstra Advice, Research & EMC Services B.V.**
Trading as DARE!! Calibrations

This annex is valid from: **19-12-2018** to **01-11-2021**

Replaces annex dated: **07-11-2018**

HCS code	Measured quantity, Instrument, Measure	Range	CMC ¹	Remarks	Location
	2 V – 20 V	100 Hz - 3 kHz	$1.9 \cdot 10^{-4} \cdot U + 0.6 \text{ mV}$		W
		3 - 10 kHz	$3.0 \cdot 10^{-4} \cdot U + 0.5 \text{ mV}$		
		10 - 60 kHz	$6 \cdot 10^{-4} \cdot U + 3.0 \text{ mV}$		
		60 - 300 kHz	$3.7 \cdot 10^{-3} \cdot U + 30 \text{ mV}$		
		300 kHz - 1 MHz	$1.2 \cdot 10^{-2} \cdot U + 0.24 \text{ V}$		
	20 V - 200 V	20 - 40 Hz	$2.2 \cdot 10^{-4} \cdot U + 6 \text{ mV}$		
		40 - 100 Hz	$1.9 \cdot 10^{-4} \cdot U + 7 \text{ mV}$		
		100 Hz – 300 Hz	$1.7 \cdot 10^{-4} \cdot U + 6 \text{ mV}$		
		300 Hz - 10 kHz	$1.8 \cdot 10^{-4} \cdot U + 5 \text{ mV}$		
		10 – 30 kHz	$3.0 \cdot 10^{-4} \cdot U + 6 \text{ mV}$		
		30 - 100 kHz	$7 \cdot 10^{-4} \cdot U + 30 \text{ mV}$		
	200 V - 1000 V	40 Hz - 3 kHz	$2.2 \cdot 10^{-4} \cdot U + 30 \text{ mV}$		
		3 - 10 kHz	$1.9 \cdot 10^{-4} \cdot U + 0,04 \text{ V}$		
		10 - 30 kHz	$4 \cdot 10^{-4} \cdot U + 0,08 \text{ V}$		
LF 4 0	ALTERNATING CURRENT				W
	100 - 200 μA	10 - 32 Hz	$3.0 \cdot 10^{-4} \cdot I + 13 \text{ nA}$	Generating	
		32 - 330 Hz	$1.9 \cdot 10^{-3} \cdot I + 11 \text{ nA}$		
		330 - 1000 Hz	$6 \cdot 10^{-3} \cdot I + 5 \text{ nA}$		
	200 μA - 2 mA	10 - 32 Hz	$1.6 \cdot 10^{-4} \cdot I + 0.15 \mu\text{A}$		
		32 - 330 Hz	$1.6 \cdot 10^{-4} \cdot I + 0.17 \mu\text{A}$		
		330 - 1000 Hz	$1.5 \cdot 10^{-4} \cdot I + 0.17 \mu\text{A}$		
		1 - 3.3 kHz	$2.0 \cdot 10^{-4} \cdot I + 0.20 \mu\text{A}$		
		3.3 - 5 kHz	$2.8 \cdot 10^{-4} \cdot I + 0.17 \mu\text{A}$		
	2 mA - 20 mA	10 - 32 Hz	$1.6 \cdot 10^{-4} \cdot I + 1.5 \mu\text{A}$		
		32 - 330 Hz	$1.6 \cdot 10^{-4} \cdot I + 1.7 \mu\text{A}$		
		330 - 1000 Hz	$1.4 \cdot 10^{-4} \cdot I + 1.7 \mu\text{A}$		
		1 - 3.3 kHz	$2.6 \cdot 10^{-4} \cdot I + 1.7 \mu\text{A}$		

of **Dijkstra Advice, Research & EMC Services B.V.**
Trading as DARE!! Calibrations

This annex is valid from: **19-12-2018 to 01-11-2021**

Replaces annex dated: **07-11-2018**

HCS code	Measured quantity, Instrument, Measure	Range	CMC ¹	Remarks	Location
	2 mA - 20 mA	3.3 - 5 kHz	$2.7 \cdot 10^{-4} / + 1.7 \mu\text{A}$		W
	20 mA - 200 mA	10 - 32 Hz	$1.6 \cdot 10^{-4} / + 15 \mu\text{A}$		
		32 - 330 Hz	$1.6 \cdot 10^{-4} / + 16 \mu\text{A}$		
		330 - 1000 Hz	$1.4 \cdot 10^{-4} / + 17 \mu\text{A}$		
		1 - 3.3 kHz	$2.6 \cdot 10^{-4} / + 17 \mu\text{A}$		
		3.3 - 5 kHz	$2.6 \cdot 10^{-4} / + 17 \mu\text{A}$		
	200 mA - 1 A	10 - 32 Hz	$4 \cdot 10^{-4} / + 0.15 \text{ mA}$		
		32 - 330 Hz	$6 \cdot 10^{-4} / + 0.15 \text{ mA}$		
		330 - 1000 Hz	$1.5 \cdot 10^{-3} / + 0.10 \text{ mA}$		
		1 - 3.3 kHz	$5 \cdot 10^{-3} / + 0.09 \text{ mA}$		
		3.3 - 5 kHz	$8 \cdot 10^{-3} / + 33 \mu\text{A}$		
	1 A - 2 A	10 - 32 Hz	$8 \cdot 10^{-4} / + 0.20 \text{ mA}$		
		32 - 330 Hz	$9 \cdot 10^{-4} / + 0.20 \text{ mA}$		
		330 - 1000 Hz	$1.6 \cdot 10^{-3} / + 0,20 \text{ mA}$		
		1 - 3.3 kHz	$5 \cdot 10^{-3} / + 0.09 \text{ mA}$		
		3.3 - 5 kHz	$8 \cdot 10^{-3} / + 33 \mu\text{A}$		
	100 μA - 200 μA	50 - 1000 Hz	$4 \cdot 10^{-4} / + 25 \text{ nA}$	Measuring	
		1 - 5 kHz	$6 \cdot 10^{-4} / + 0,05 \text{ nA}$		
	200 μA - 2 mA	50 - 300 Hz	$4 \cdot 10^{-4} / + 0.25 \mu\text{A}$		
		300 - 1000 Hz	$4 \cdot 10^{-4} / + 0.25 \mu\text{A}$		
		1 - 5 kHz	$6 \cdot 10^{-4} / + 0.5 \mu\text{A}$		
	2 mA - 20 mA	50 - 300 Hz	$4 \cdot 10^{-4} / + 2.5 \mu\text{A}$		
		300 - 1000 Hz	$4 \cdot 10^{-4} / + 2.5 \mu\text{A}$		
		1 - 5 kHz	$6 \cdot 10^{-4} / + 5 \mu\text{A}$		
	20 mA - 200 mA	50 - 1000 Hz	$4 \cdot 10^{-4} / + 25 \mu\text{A}$		
		1 - 5 kHz	$6 \cdot 10^{-4} / + 0,05 \text{ mA}$		
	200 mA - 2 A	50 - 1000 Hz	$1.2 \cdot 10^{-3} / + 0.5 \text{ mA}$		

of **Dijkstra Advice, Research & EMC Services B.V.**
Trading as DARE!! Calibrations

This annex is valid from: **19-12-2018** to **01-11-2021**

Replaces annex dated: **07-11-2018**

HCS code	Measured quantity, Instrument, Measure	Range	CMC ¹	Remarks	Location
	200 mA - 2 A	1 - 5 kHz	$2.5 \cdot 10^{-3} \cdot I + 1.3 \text{ mA}$		W
LF 6 0	IMPEDANCE (DC/LF)				
LF 6 2	DC resistance				W
	1 Ω		0.2 m Ω	Generating, 4-wire	
	10 Ω		0.5 m Ω		
	100 Ω		1.9 m Ω		
	1 k Ω		21 m Ω		
	10 k Ω		0.2 Ω		
	100 k Ω		2.3 Ω		
	1 M Ω		33 Ω		
	10 M Ω		0.7 k Ω		
	100 M Ω		22 k Ω		
	10 Ω		0.24 Ω	Generating, 2-wire	
	100 Ω		0.24 Ω		
	1 k Ω		0.35 Ω		
	10 k Ω		0.5 Ω		
	100 k Ω		2.6 Ω		
	1 M Ω		33 Ω		
	10 M Ω		0.7 k Ω		
	100 M Ω		22 k Ω		
	0 - 20 Ω		$2.0 \cdot 10^{-5} \cdot R + 0.12 \text{ m}\Omega$	Measuring, 4-wire	
	20 - 200 Ω		$1.5 \cdot 10^{-5} \cdot R + 0.5 \text{ m}\Omega$		
	200 - 2000 Ω		$1.2 \cdot 10^{-5} \cdot R + 2.5 \text{ m}\Omega$		
	2 - 20 k Ω		$1.2 \cdot 10^{-5} \cdot R + 25 \text{ m}\Omega$		
	20 - 200 k Ω		$2.1 \cdot 10^{-5} \cdot R + 0.4 \Omega$		

of **Dijkstra Advice, Research & EMC Services B.V.**
Trading as DARE!! Calibrations

This annex is valid from: **19-12-2018 to 01-11-2021**

Replaces annex dated: **07-11-2018**

HCS code	Measured quantity, Instrument, Measure	Range	CMC ¹	Remarks	Location
	200 kΩ - 2 MΩ		$3.1 \cdot 10^{-5} \cdot R + 10 \Omega$		W
	2 - 20 MΩ		$4.0 \cdot 10^{-5} \cdot R + 0.25 \text{ k}\Omega$	Measuring, 2-wire	
	20 - 200 MΩ		$3.5 \cdot 10^{-4} \cdot R + 15 \text{ k}\Omega$		
	200 MΩ - 2 GΩ		$3.5 \cdot 10^{-3} \cdot R + 1.1 \text{ M}\Omega$		
RF 0 0	HIGH FREQUENCY QUANTITIES				
RF 2 0	Impedance				W
	LISN Impedance	9 kHz – 30 MHz	$0.3 \Omega - 1.1 \Omega$ $1.2^\circ - 3.8^\circ$	$50 \Omega // (50 \mu\text{H} + 5 \Omega)$ and $50 \Omega // 50 \mu\text{H}$	
		100 kHz – 150 MHz	$0.5 \Omega - 0.9 \Omega$ $3.3^\circ - 8^\circ$	$50 \Omega // (5 \mu\text{H} + 1 \Omega)$ and $50 \Omega // 5 \mu\text{H}$	
	CDN Impedance	150 kHz – 300 MHz	$5 \Omega - 6 \Omega$ $2.4^\circ - 3.9^\circ$	$150 \Omega, 0^\circ$ nominal	
RF 2 1	Reflection coefficient			3)	W
	Magnitude 0 to 1.0	9 kHz – 1 MHz	$0.005 + 0.007 \cdot \Gamma + 0.005 \cdot \Gamma^2$	Nom. impedance 50 Ω at nominal -10 dBm RF power	
		1 MHz - 2 GHz	$0.005 + 0.003 \cdot \Gamma + 0.005 \cdot \Gamma^2$		
		2 GHz - 8 GHz	$0.02 + 0.004 \cdot \Gamma + 0.02 \cdot \Gamma^2$		
		8 GHz – 18 GHz	$0.03 + 0.004 \cdot \Gamma + 0.04 \cdot \Gamma^2$		
	Phase –180° to +180°	9 kHz -18 GHz	$u(\theta) = \arcsin\left(\frac{u(\Gamma)}{ \Gamma }\right)$	If the magnitude is less than its uncertainty, the phase uncertainty is 180°	

of **Dijkstra Advice, Research & EMC Services B.V.**
Trading as DARE!! Calibrations

This annex is valid from: **19-12-2018 to 01-11-2021**

Replaces annex dated: **07-11-2018**

HCS code	Measured quantity, Instrument, Measure	Range	CMC ¹	Remarks	Location
RF 2 2	Transmission Coefficient			For coaxial 50 Ω devices 1, 3)	W
	Magnitude (0 to -30 dB)	9 kHz – 10 MHz	0.04 dB	at nom. -10 dBm RF power	
		10 MHz – 1500 MHz	0.08 dB		
		1500 MHz – 18 GHz	0.12 dB		
	Magnitude (-30 to -50 dB)	9 kHz – 1500 MHz	0.08 dB	at nom. 0 dBm RF power	
		1500 MHz – 8 GHz	0.12 dB		
		8 GHz – 18 GHz	0.15 dB		
	Magnitude (-50 to -70) dB	9 kHz – 18 GHz	0.25 dB	at nom. +10 dBm * RF power	
	Magnitude (-70 to -80) dB	9 kHz – 18 GHz	0.7 dB	at nom. +10 dBm * RF power	
	Magnitude (-80 to -90)dB	30 kHz – 18 GHz	2.0 dB	at nom. +10 dBm * RF power	
	Magn. (-90 to – 100)dB	30 kHz – 18 GHz	5 dB	at nom. +10 dBm * RF power	
				* +5dBm above 8 GHz	
	Antenna Reflection coefficient				
	Magnitude 0 – 1.0	30 MHz – 700 MHz	$0.06 + 0.020 \cdot \Gamma + 0.008 \cdot \Gamma^2$	Nominal impedance 50 Ω at nominal -10 dBm RF power	
		700 MHz – 1500 MHz	$0.07 + 0.020 \cdot \Gamma + 0.013 \cdot \Gamma^2$		
		1500 MHz – 3000 MHz	$0.08 + 0.020 \cdot \Gamma + 0.013 \cdot \Gamma^2$		

of **Dijkstra Advice, Research & EMC Services B.V.**
Trading as DARE!! Calibrations

This annex is valid from: **19-12-2018** to **01-11-2021**

Replaces annex dated: **07-11-2018**

HCS code	Measured quantity, Instrument, Measure	Range	CMC ¹	Remarks	Location
RF 3 0	HIGH FREQUENCY POWER			1, 3)	W
	Calibration Factors of Power Sensors	9 kHz – 10 MHz	0.05 dB	0 dBm nominal	
		10 MHz – 3 GHz	0.06 dB	0 dBm nominal	
		3 GHz – 6 GHz	0.07 dB	0 dBm nominal	
		6 GHz – 10 GHz	0.07 – 0.09 dB	0 dBm nominal	
		10 GHz – 18 GHz	0.09 dB	0 dBm nominal	
		10 MHz – 2 GHz	0.06 dB	-30 dBm nominal	
		2 GHz – 6 GHz	0.07 dB	-30 dBm nominal	
		6 GHz – 10 GHz	0.07 – 0.10 dB	-30 dBm nominal	
		10 GHz – 18 GHz	0.10 dB	-30 dBm nominal	
	Linearity of RF power				
	0 – -10 dBm	9 kHz – 6 GHz	0.04 dB		
	0 – -20 dBm	9 kHz – 6 GHz	0.05 dB		
	0 – -30 dBm	9 kHz – 6 GHz	0.07 dB		
	0 – -40 dBm	10 MHz – 6 GHz	0.09 dB		
	0 – -50 dBm	10 MHz – 6 GHz	0.09 dB		
	0 – -60 dBm	10 MHz – 6 GHz	0.15 dB		
	Absolute power -60 to +20 dBm			2, 3)	
		9 kHz – 10 MHz	0.06 dB	0 dBm nominal	
		10 MHz – 6 GHz	0.07 dB	0 dBm nominal	
		6 GHz – 10 GHz	0.10 dB	0 dBm nominal	
		10 GHz – 18 GHz	0.12 dB	0 dBm nominal	

of **Dijkstra Advice, Research & EMC Services B.V.**
Trading as DARE!! Calibrations

This annex is valid from: **19-12-2018 to 01-11-2021**

Replaces annex dated: **07-11-2018**

HCS code	Measured quantity, Instrument, Measure	Range	CMC ¹	Remarks	Location
	Absolute power -60 to +20 dBm			2, 3)	W
		10 MHz – 6 GHz	0.07 dB	-30 dBm nominal	
		6 GHz – 10 GHz	0.10 dB	-30 dBm nominal	
		10 GHz – 18 GHz	0.14 dB	-30 dBm nominal	
	Absolute broadband power, 0 to +45 dBm				W
		9 kHz – 6 GHz	0.6 dB 1.0 dB	RadiField All other DUTs	
	Frequency response of power measuring devices			2, 3)	W
		9 kHz – 10 MHz	0.05 dB	0 dBm nominal	
		10 MHz – 3 GHz	0.06 dB	0 dBm nominal	
		3 GHz – 6 GHz	0.07 dB	0 dBm nominal	
		6 GHz – 10 GHz	0.07 – 0.10 dB	0 dBm nominal	
		10 GHz – 18 GHz	0.10 dB	0 dBm nominal	
		10 MHz – 3 GHz	0.06 dB	-30 dBm nominal	
		3 GHz – 6 GHz	0.07 dB	-30 dBm nominal	
		6 GHz – 10 GHz	0.07 – 0.10 dB	-30 dBm nominal	
		10 GHz – 18 GHz	0.10 dB	-30 dBm nominal	
	Response of CISPR receivers				W
	QP and AV	9 kHz - 1 GHz	0.3 dB		

Annex to declaration of accreditation (scope of accreditation)
 Normative document: EN ISO/IEC 17025:2005
 Registration number: **K 063**

of **Dijkstra Advice, Research & EMC Services B.V.**
Trading as DARE!! Calibrations

This annex is valid from: **19-12-2018 to 01-11-2021**

Replaces annex dated: **07-11-2018**

HCS code	Measured quantity, Instrument, Measure	Range	CMC ¹	Remarks	Location
	Bandwidth of RBW filters				W
	1 Hz to 10 MHz	9 kHz - 2.4 GHz	0.8 + 0.02·BW		
	1 Hz to 10 MHz	10 MHz - 18 GHz	0.08 + 0.02·BW		
RF 5 0	ELECTRICAL / MAGNETIC FIELD QUANTITIES /EMC				W
	Electrical Field Strength 1 – 200 V/m	9 kHz – 30 MHz	0.5 – 0.6 dB	Temcell 4)	
		30 MHz – 75 MHz	0.6 – 1.3 dB		
		75 MHz – 200 MHz	1.3 dB		
	Electrical Field Strength 1 – 100 V/m	200 MHz – 1 GHz	1.2 dB	Anechoic Chamber	
		1 GHz – 8 GHz	1.1 dB		
		8 GHz – 12 GHz	1.2 dB		
		12 GHz –15 GHz	1.2 - 1.5 dB		
		15 GHz – 18 GHz	1.5 dB		
		18 GHz – 40 GHz	2.4 dB		
	Antenna factor				W
	ANSI C63.5 CISPR 16-1-6	30 MHz – 5 GHz	0.8 dB	OATS, Standard Site method 5)	
	ANSI C63.5 CISPR 16-1-6	30 MHz – 1 GHz	0.8 dB	OATS, Reference antenna method 5)	
	Antenna symmetry - Dipole - Biconical - Hybrid	30 MHz – 1 GHz	0.25 dB	OATS ANSI C63.5 CISPR 16-1-4	

of **Dijkstra Advice, Research & EMC Services B.V.**
Trading as DARE!! Calibrations

This annex is valid from: **19-12-2018 to 01-11-2021**

Replaces annex dated: **07-11-2018**

HCS code	Measured quantity, Instrument, Measure	Range	CMC ¹	Remarks	Location
	Quasi Free Space	20 MHz – 100 MHz	0.9 dB	Free Space Environment, Three Antenna Method 5)	W
		100 MHz – 200 MHz	0.8 dB		
		200 MHz – 5 GHz	0.7 dB		
		1 GHz – 10 GHz	1.4 dB	Full Anechoic Room, Three Antenna Method 4)	
		10 GHz – 12 GHz	1.4 dB – 2.1 dB		
		12 GHz – 18 GHz	2.1 dB		
	SAE ARP 958	20 MHz – 100 MHz	0.9 dB	For military or automotive use 5)	W
		100 MHz – 200 MHz	0.8 dB		
		200 MHz – 5 GHz	0.7 dB		
	Shielding Effectiveness			According to EN50147 and Mil Std 285 5)	O
	Magnetic Field 100dB	10 kHz – 30 MHz	± 5 dB		
	Electric Field 120 – 150 dB	10 MHz – 300 MHz	± 5 dB		
	Plane wave 110 – 140 dB	30 MHz – 1 GHz	± 5 dB		
	Plane wave 110 – 140 dB	1 GHz – 18 GHz	± 6 dB		
	Normalized Site Attenuation			According to CISPR 16-1-4 using broadband antennae Horizontal and vertical polarization, distance between 3 m and 30 m 5)	O
	NSA	30 MHz – 1000 MHz	± 1.6 dB		

of **Dijkstra Advice, Research & EMC Services B.V.**
Trading as DARE!! Calibrations

This annex is valid from: **19-12-2018** to **01-11-2021**

Replaces annex dated: **07-11-2018**

HCS code	Measured quantity, Instrument, Measure	Range	CMC ¹	Remarks	Location
	Site Voltage Standing Wave Ratio			According to CISPR 16-1-4 using reciprocal method 4)	O
	S _{VSWR}	1 GHz – 18 GHz	± 2.0 dB		
	Field Uniformity			According to IEC 61000-4-3 4)	O
	Forward Power	80 MHz – 18 GHz	± 1.3 dB		
	Field Uniformity	80 MHz – 18 GHz	± 1.7 dB		
	Surge generators and coupling/decoupling networks waveform surge voltage			According to EN 61000-4-5 1.2/50 µs pulse 10/700 µs pulse	W, O
	0 V – 550 V		$6.7 V + 0.022 \cdot U$ $6.7 V + 0.025 \cdot U$	Coupling/decoupling networks for AC/DC power supply circuits only in combination with appropriate surge generator	W O
	0 V – 1.1 kV		$13.4 V + 0.022 \cdot U$ $13.4 V + 0.025 \cdot U$		W O
	0 V – 2.8 kV		$33.5 V + 0.022 \cdot U$ $33.5 V + 0.025 \cdot U$		W O
	0 V – 5.5 kV		$67 V + 0.022 \cdot U$ $67 V + 0.025 \cdot U$		W O
	Waveform surge current				W, O
	Current amplitude			1.2/50 µs pulse 10/700 µs pulse 0.5/700 µs pulse	
	0 – 15 A		$0.18 + 0.022 \cdot I$ $0.18 + 0.029 \cdot I$	Measurements at coupling/decoupling network input, output; coupling modes line to neutral, line to earth and neutral to earth	W O
	0 – 30 A		$0.36 + 0.022 \cdot I$ $0.36 + 0.029 \cdot I$		W O
	0 – 60 A		$0.72 + 0.022 \cdot I$ $0.72 + 0.029 \cdot I$		W O

of **Dijkstra Advice, Research & EMC Services B.V.**
Trading as DARE!! Calibrations

This annex is valid from: **19-12-2018** to **01-11-2021**

Replaces annex dated: **07-11-2018**

HCS code	Measured quantity, Instrument, Measure	Range	CMC ¹	Remarks	Location
	0 – 150 A		1.8 + 0.022 * I 1.8 + 0.029 * I		W O
	0 – 300 A		3.6 + 0.022 * I 3.6 + 0.029 * I		W O
	0 – 600 A		7.2 + 0.022 * I 7.2 + 0.029 * I		W O
	0 – 1500 A		18 + 0.022 * I 18 + 0.029 * I		W O
	0 – 3000 A		36 + 0.022 * I 36 + 0.029 * I		W O
	Front time. Voltage				W, O
	1.2/50 µs	0.65 – 1.75 µs	0.08 µs	4)	
	10/700 µs	5.5 – 15.5 µs	0.5 µs		
	Front time. current				W, O
	1.2/50 µs	6.4 – 9.6 µs 1.4 – 3.6 µs	0.18 µs 0.07 µs	line – line line - PE	
	10/700 µs	3.5 – 6.5µs	0.27 µs		

of **Dijkstra Advice, Research & EMC Services B.V.**
Trading as DARE!! Calibrations

This annex is valid from: **19-12-2018** to **01-11-2021**

Replaces annex dated: **07-11-2018**

HCS code	Measured quantity, Instrument, Measure	Range	CMC ¹	Remarks	Location
	Duration. Voltage			4)	W,O
	1.2/50 µs	35 µs – 65 µs	1.0 µs		
	10/700 µs	490 - 910 µs	14µs		
	Pulse duration time Current			4)	W, O
	1.2/50 µs	11 - 21 µs	0.23 µs	Current, line to line	
		14 - 36 µs	0.23 µs	Current, line to earth	
	10/700 µs	210 - 390 µs	6 µs		
	0.5/700 µs	210 - 390 µs	6 µs		
	EFT/burst generators waveform (im)pulse. voltage into 50 ohms			According to EN 61000-4-4 (July 2004)	W, O
	0 V – 150V		$2.0 V + 0.022 \cdot U$ $2.0 V + 0.025 \cdot U$	4)	W O
	0 – 300 V		$4.0 V + 0.022 \cdot U$ $4.0 V + 0.025 \cdot U$	4)	W O
	0 – 600 V		$8.0 V + 0.022 \cdot U$ $8.0 V + 0.025 \cdot U$	4)	W O
	0 – 1.5 kV		$20 V + 0.022 \cdot U$ $20 V + 0.025 \cdot U$	4)	W O
	0 – 3 kV		$37 V + 0.022 \cdot U$ $37 V + 0.025 \cdot U$	4)	W O

of **Dijkstra Advice, Research & EMC Services B.V.**
Trading as DARE!! Calibrations

This annex is valid from: **19-12-2018** to **01-11-2021**

Replaces annex dated: **07-11-2018**

HCS code	Measured quantity, Instrument, Measure	Range	CMC ¹	Remarks	Location
	EFT/burst generators Waveform pulse voltage into 1 kΩ				W, O
	0 V – 500 V		6 V + 0.045·U 6 V + 0.048·U	4)	W O
	0 V – 1 kV		12 V + 0.045·U 12 V + 0.048·U	4)	W O
	0 V – 2 kV		24 V + 0.045·U 24 V + 0.048·U	4)	W O
	0 V – 5 kV		60 V + 0.045·U 60 V + 0.048·U	4)	W O
	Rise time (10%-90%) 3 ns to 7 ns		0.3 ns	4)	W,O
	Pulse duration time (50%-50%) 30 ns to 70 ns		2.0 ns	4)	W,O
	Repetition rate			4)	W,O
	5 μs – 15 μs		0.15 μs		
	150 μs – 600 μs		2.5 μs		
	Burst duration			4)	W,O
	10 ms to 20 ms		0.5μs + 0.0005 * T		
	Burst period			4)	
	200 ms to 400 ms		500 μs		W,O

of **Dijkstra Advice, Research & EMC Services B.V.**
Trading as DARE!! Calibrations

This annex is valid from: **19-12-2018** to **01-11-2021**

Replaces annex dated: **07-11-2018**

HCS code	Measured quantity, Instrument, Measure	Range	CMC ¹	Remarks	Location
	ESD Simulators Waveform discharge current				W
	First peak current		± 7%	Networks and standard: 150pF / 330 Ω IEC61000-4-2 (2008-12)	
	Rise time		± 15%		
	Current at t1 and t2		± 7%		
	ESD Simulators Waveform discharge current				
	First peak current 0 – 10 A 0 – 20 A 0 – 50 A		0.08 + 0.032*I 0.16 + 0.032*I 0.4 + 0.032*I	Networks and standard: 150pF / 330 Ω IEC61000-4-2 (1995-01) ISO10605 (2008-07)	
	Rise time 0.5 – 1.2 ns		0.08 ns	330pF / 330 Ω ISO10605 (2008-07)	
	Current at t1 and t2 0 – 10 A 0 – 20 A 0 – 50 A		0.08 + 0.023*I 0.16 + 0.023*I 0.4 + 0.023*I		

of **Dijkstra Advice, Research & EMC Services B.V.**
Trading as DARE!! Calibrations

This annex is valid from: **19-12-2018 to 01-11-2021**

Replaces annex dated: **07-11-2018**

HCS code	Measured quantity, Instrument, Measure	Range	CMC ¹	Remarks	Location
	ESD Simulators Waveform discharge current				W
	First peak current 0 – 1.0 A 0 – 2.0 A 0 – 5.0 A		0.08 + 0.032*I 0.16 + 0.032*I 0.4 + 0.032*I	Networks and standard: 150pF / 2000 Ω 330pF / 2000 Ω ISO10605 (2008-07)	
	Rise time 0.5 – 1.2 ns		0.08 ns		
	Current at t1 and t2 0 – 1.0 A 0 – 2.0 A 0 – 5.0 A		0.008 + 0.023*I 0.02 + 0.023*I 0.04 + 0.023*I		
TF 0 0	TIME AND FREQUENCY				W
TF 2 1	Frequency				
	10 mHz - 2.7 GHz		$(8 \cdot 10^{-10} / \tau + 1.8 \cdot 10^{-10}) \cdot f$	Measuring, 10 ms ≤ τ ≤ 400s	
	10 mHz – 10 Hz		$3 \cdot 10^{-5} \cdot f$	Generating	
	10 Hz – 100 kHz		$(1 \cdot 10^{-5} / \tau) \cdot f$	Generating, 10 ms ≤ τ ≤ 400s	
	100 kHz – 2.16 GHz		$(1 \cdot 10^{-9} / \tau + 1.8 \cdot 10^{-10}) \cdot f$	Generating, 10 ms ≤ τ ≤ 400s	
TF 2 2	Time interval				
	0.5 ns – 10 μs		$(1.5 \cdot 10^{-4}) \cdot t + 15\text{ps}$	Generating	
	10 μs – 1000 s		$(3 \cdot 10^{-5}) \cdot t$		

The calibrations are carried out at an ambient temperature of (23 ± 2)°C and a relative humidity of (50 ± 10) %, with an exception for calibrations marked 4 or 5.

1. |Γ_{dut}| < 0.02
2. |Γ_{dut}| < 0.2
3. All calibrations are based on equipment using N-type connectors.
4. The calibrations are carried out at ambient conditions within (23 ± 7)°C and (50 ± 20) %.
5. The calibrations are carried out at ambient conditions within (20 ± 15)°C and (50 ± 40) %.