



ROHDE & SCHWARZ

Test and Measurement
Division

Release Notes

International Keyboards

for R&S Spectrum Analyzer/ Signal Analyzer

| | | |
|---|------------------------|---------------------|
| ➤ | R&S FSV | 1307.9002.xx |
| ➤ | R&S FSL | 1300.2502.xx |
| ➤ | R&S FSP(XP) | 1164.4391.xx |
| ➤ | R&S FSU(XP) | 1166.1660.xx |
| ➤ | R&S FSQ(XP) | 1155.5001.xx |
| ➤ | R&S FSG(XP) | 1309.0002.xx |
| ➤ | R&S FMU | 1303.3500.02 |

New Features:

- Installation of the German, Spanish, French, Italian and Portuguese keyboard support

Release Note Revision: 3

Contents

| | |
|--|---|
| History | 2 |
| General Topics | 2 |
| Installation of the international keyboard support | 2 |
| Known Problems | 4 |

History

| Date | Rel Note Rev | Changes |
|------------|--------------|--|
| 02.05.2007 | 1 | First revision |
| 21.05.2008 | 2 | Pictures and instrument R&S FSG added. |
| 22.08.2011 | 3 | Instrument R&S FSV added. |

General Topics


This update will enhance the embedded XP operating system of the following keyboard layouts:

- German Keyboards
- Spanish Keyboards
- French Keyboards
- Italian Keyboards
- Portuguese Keyboards

Installation of the international keyboard support

For the installation of the German keyboard support, proceed with the following steps:

- 1) Copy the file **FsxInternationalKeyboards.msi** onto USB stick or floppy disk.
- 2) Connect keyboard and mouse to the analyzer.
- 3) Switch the analyzer on and wait until the boot process is finished.
- 4) Terminate the analyzer firmware by pressing <ALT> <F4>.
- 5) Start Windows Explorer
- 6) Start the installation procedure with double click onto the file **FsxInternationalKeyboards.msi**.
- 7) Reboot the analyzer.
- 8) Now open the windows start menu by pressing the windows key or <CTRL> <ESC>.
- 9) Select *Settings - Control Panel - Regional and Language Options - Languages*.




10) Select *Details*.



11) Select *Add*, choose one of the available input languages and confirm with OK and the dialog will be close.

12) Now change the *Default input language* as needed and close all open dialogs with OK.



13) Reboot the analyzer.

Known Problems

For FSP/FSU/FSQ: A base system firmware 4.3x (for FSG: V4.39 SP1) or newer is required to support the international keyboard driver. In older versions some of the front panel keys will not work properly.

For FSV: A base system Firmware 1.61Sp2 or newer is required to support the international keyboard driver. In older versions some of the front panel keys will not work properly