

**TPRO-VME/TSAT-VME
VERSA MODULE EUROCARD (VME)
SYNCHRONIZABLE TIMECODE GENERATOR
WITH VMEbus INTERFACE**

User Manual

*95 Methodist Hill Drive
Rochester, NY 14623*

Phone: US +1.585.321.5800

Fax: US +1.585.321.5219



www.spectracomcorp.com

Part Number 1156-5001-0050

Manual Revision B (2.0)

16 October 2006

Copyright © 2006 Spectracom Corporation. The contents of this publication may not be reproduced in any form without the written permission of Spectracom Corporation. Printed in USA.

Specifications subject to change or improvement without notice.

Spectracom, NetClock, Ageless, TimeGuard, TimeBurst, TimeTap, LineTap, MultiTap, VersaTap, and Legally Traceable Time are Spectracom registered trademarks. All other products are identified by trademarks of their respective companies or organizations. All rights reserved.

SPECTRACOM LIMITED WARRANTY

LIMITED WARRANTY

Spectracom warrants each new product manufactured and sold by it to be free from defects in software, material, workmanship, and construction, except for batteries, fuses, or other material normally consumed in operation that may be contained therein AND AS NOTED BELOW, for five years after shipment to the original purchaser (which period is referred to as the "warranty period"). This warranty shall not apply if the product is used contrary to the instructions in its manual or is otherwise subjected to misuse, abnormal operations, accident, lightning or transient surge, repairs or modifications not performed by Spectracom.

The GPS receiver is warranted for one year from date of shipment and subject to the exceptions listed above. The power adaptor, if supplied, is warranted for one year from date of shipment and subject to the exceptions listed above.

THE ANALOG CLOCKS ARE WARRANTED FOR ONE YEAR FROM DATE OF SHIPMENT AND SUBJECT TO THE EXCEPTIONS LISTED ABOVE.

THE TIMECODE READER/GENERATORS ARE WARRANTED FOR ONE YEAR FROM DATE OF SHIPMENT AND SUBJECT TO THE EXCEPTIONS LISTED ABOVE.

The Rubidium oscillator, if supplied, is warranted for two years from date of shipment and subject to the exceptions listed above.

All other items and pieces of equipment not specified above, including the antenna unit, antenna surge suppressor and antenna pre-amplifier are warranted for 5 years, subject to the exceptions listed above.

WARRANTY CLAIMS

Spectracom's obligation under this warranty is limited to in-factory service and repair, at Spectracom's option, of the product or the component thereof, which is found to be defective. If in Spectracom's judgment the defective condition in a Spectracom product is for a cause listed above for which Spectracom is not responsible, Spectracom will make the repairs or replacement of components and charge its then current price, which buyer agrees to pay.

Spectracom shall not have any warranty obligations if the procedure for warranty claims is not followed. Users must notify Spectracom of the claim with full information as to the claimed defect. Spectracom products shall not be returned unless a return authorization number is issued by Spectracom.

Spectracom products must be returned with the description of the claimed defect and identification of the individual to be contacted if additional information is needed. Spectracom products must be returned properly packed with transportation charges prepaid.

Shipping expense: Expenses incurred for shipping Spectracom products to and from Spectracom (including international customs fees) shall be paid for by the customer, with the following exception. For customers located within the United States, any product repaired by Spectracom under a "warranty repair" will be shipped back to the customer at Spectracom's expense unless special/faster delivery is requested by customer.

Spectracom highly recommends that prior to returning equipment for service work, our technical support department be contacted to provide trouble shooting assistance while the equipment is still installed. If equipment is returned without first contacting the support department and "no problems are found" during the repair work, an evaluation fee may be charged.

EXCEPT FOR THE LIMITED WARRANTY STATED ABOVE, SPECTRACOM DISCLAIMS ALL WARRANTIES OF ANY KIND WITH REGARD TO SPECTRACOM PRODUCTS OR OTHER MATERIALS PROVIDED BY SPECTRACOM, INCLUDING WITHOUT LIMITATION ANY IMPLIED WARRANTY OR MERCHANTABILITY OR FITNESS FOR A PARTICULAR PURPOSE.

Spectracom shall have no liability or responsibility to the original customer or any other party with respect to any liability, loss, or damage caused directly or indirectly by any Spectracom product, material, or software sold or provided by Spectracom, replacement parts or units, or services provided, including but not limited to any interruption of service, excess charges resulting from malfunctions of hardware or software, loss of business or anticipatory profits resulting from the use or operation of the Spectracom product or software, whatsoever or howsoever caused. In no event shall Spectracom be liable for any direct, indirect, special or consequential damages whether the claims are grounded in contract, tort (including negligence), or strict liability.

EXTENDED WARRANTY COVERAGE

Extended warranties can be purchased for additional periods beyond the standard five-year warranty. Contact Spectracom no later than the last year of the standard five-year warranty for extended coverage.

Table of Contents

1	OVERVIEW	1-1
1.1	General Information about GPS	1-3
1.2	Your Spectracom GPS Receiver	1-4
1.3	Inventory	1-4
1.4	Inspection and Support	1-4
1.5	External Connections	1-5
1.6	GPS Antenna Connector (TSAT-VME Only)	1-5
2	SPECIFICATIONS.....	2-1
2.1.1	VMEbus Interface	2-1
2.1.2	Inputs	2-1
2.1.3	Outputs	2-2
2.1.4	LED Indicators	2-2
2.1.5	Timing	2-2
2.2	Front Panel	2-2
2.2.1	J6 Timecode Input	2-2
2.2.2	J1 20-pin Ribbon Cable Connector (TPRO-VME Only)	2-3
2.3	Rear Panel	2-3
2.3.1	P1 and P2 VMEbus Interface and User-Defined Signals	2-3
2.4	GPS Receiver/Antenna(TSAT-PMC Only)	2-5
2.4.1	Optional Cable	2-5
3	CONFIGURATION	3-1
3.1	General	3-1
3.2	VME Base Address	3-1
3.2.1	Case 1	3-1
3.2.2	Case 2	3-1
3.2.3	Case 3 (default)	3-2
3.2.4	Case 4	3-2
3.3	VMEbus Interrupt Level	3-6
3.4	Rate Outputs Configuration	3-7
3.5	VMEbus P2 Configuration (Does Not Apply to Option -32P2)	3-8
3.6	Time Tag Enable/Disable	3-10
3.7	FIFO Buffer Clear Configuration	3-11
3.8	VMEbus Address Configuration	3-12
3.8.1	A16 Base Address Configuration	3-13
3.8.2	A24 Base Address Configuration	3-13
3.8.3	A32 Base Address Configuration	3-14
3.9	Interrupt Request Level Configuration	3-15
3.10	VMEbus P2 I/O Configuration	3-15
3.11	Rate Output Configuration	3-16
3.12	External/Internal Event Configuration	3-17
3.13	FIFO Buffer Clear Configuration	3-17
4	PIN ASSIGNMENTS.....	4-1
4.1	Test Points	4-1
4.2	Input/Output Pin Assignments	4-1
5	INSTALLATION	5-1

5.1	General	5-1
5.2	External GPS Receiver/Antenna Installation	5-2
5.2.1	Routine Maintenance	5-2
5.3	Propagation Delay Adjustment	5-3
6	PROGRAMMING.....	6-1
6.1	Introduction	6-1
6.2	Register Map	6-1
6.3	Determining Address Space	6-1
6.4	Reading Time	6-2
6.5	Reading Latitude, Longitude, and Elevation (TSAT-VME Only)	6-3
6.6	Report Latitude (TSAT-VME Only)	6-3
6.7	Report Longitude (TSAT-VME Only)	6-4
6.8	Reading the Year, Gregorian Date, and Altitude (TSAT-VME Only)	6-4
6.9	Report Number of GPS Satellites Being Tracked (TSAT-VME Only)	6-5
6.10	Reading External Event Times	6-5
6.11	Interrupt Mode	6-6
6.12	Setting Propagation Delay Correction	6-7
6.13	Presetting Time	6-8
6.14	External 1PPS Time Synchronization (Option –M)	6-8
6.15	Disabling and Re-enabling Code Input	6-9
6.16	Handlers	6-9
7	PREVENTIVE MAINTENANCE	7-1
7.1	Necessary Maintenance	7-1
7.1.1	Oscillator Aging Adjustment	7-1
7.1.2	IRIG-B Output Adjustment	7-1
7.2	Optional Preventive Maintenance	7-1
7.2.1	Oscillator Aging Adjustment	7-1
8	TROUBLESHOOTING	8-1
8.1	The program has trouble talking to the TPRO-VME	8-1
8.2	Bad Data from the D32 Ports	8-1
8.3	Bad Data from the FIFO Port	8-1
8.4	Interrupt Crashes	8-2
8.5	Board Never Syncs to Input Code or 1PPS	8-2
8.6	Before You Contact Spectracom	8-2
9	OPTIONS AND ACCESSORIES.....	9-1
9.1	Common Options	9-1
9.2	Options—TPRO-VME	9-1
9.3	Options—TSAT-VME	9-1
10	DRIVER SUPPORT	10-1

List of Figures

Figure 2-1: TPRO-VME J1 Pin Identification	2-3
Figure 3-1: VME Address Jumper Locations	3-1

1 Overview

This manual provides comprehensive information on the system architecture, specifications for and operation of the Spectracom Model TPRO VME and Model TSAT-VME Synchronizable Time Code Generators with VMEbus Interface.

The TPRO VME and TSAT VME are used in VMEbus systems for time tagging, synchronizing multiple systems, and generating pulse rates at exact times and exact frequencies. The TPRO VME is basically an accurate clock whose time can be captured and read by the host VMEbus computer. The time capture can be caused by a logic pulse from the outside world (an “external” event) or can be caused by the host computer program reading the time data directly from hardware counters with two longword transfers from the TPRO VME.

The Z80 microcomputer in the TPRO VME transfers the captured external time information (i.e., hundreds of days through units of microseconds (μ S)—a total of 10 bytes) to the host by loading each of the 10 bytes sequentially from the FIFO through the TPRO VME’s VMEbus interface. The host that is testing the OUTPUT READY bit of the TPRO VME status register before the host from the FIFO output reads each byte does “handshaking”.

The longword time stamp data transfer requires no waiting by the host program. For external event time stamps, the time required for the complete transfer is from 30 to 100 μ S after the event occurs. The time fluctuates because the microcomputer may be interrupted while putting data in the FIFO. The time data is accurate to the exact microsecond at which the event occurred, and the accuracy is not affected by the transfer time. Using the on-board bus interrupter, users can interrupt the host VME processor when the first byte of data is loaded in the FIFO. In many cases, the interrupt handling (task switching) overhead exceeds the time required for the complete loading of the FIFO, so that the VME host may never actually need to wait for the microcomputer to complete FIFO loading.

Typically, the longword time stamp is used for software-initiated time measurements and the external time stamp is used for intermittent and unpredictable events triggered by the outside world.

After power-on reset, the clock starts counting automatically at tie 0 days through μ S without any commands. To maintain synchronization with the outside world, the clock can be initialized and its counting frequency adjusted (disciplined) to correct for any inherent errors due to adjustment, temperature change, or aging. The Z80 performs these tasks automatically, using a standard modulated time code signal (IRIG-B, IRIG-A, NASA36, XR3 or 2137) as a reference input or, optionally, a precision 1 PPS from a GPS system receiver (TSAT-VME). The clock time can also be set via commands from the host VMEbus computer to the TPRO VME.

The on-board clock drives the on-board LED display option and is also used to generate an IRIG-B output time code signal that can be fed to other systems. If no input reference is available, the clock counts at its natural undisciplined frequency. If the input reference disappears after being initially present, the TPRO VME continues to count from the current time, using the frequency discipline calculated while the reference was present. If the input reference reappears, the TPRO VME determines its type and phase, and then uses it for time and frequency reference. If the accumulated error during open loop operation does not exceed 200

μ S, the time count continues unperturbed by the interruption (except for a short-term change in count frequency to cancel out the error). Otherwise, an abrupt jump in time (called a “jam sync”) will occur at the beginning of the next second.

In VMEbus standard terminology, the TPRO VME is an A16 (short supervisory or short non-privileged addressing), an A24 (standard supervisory or standard non-privileged addressing), or an A32 (extended supervisory or extended non-privileged addressing) D08 (0) VME SLAVE (D16 transfers are also supported, but the high order byte is not used) with a user-configurable base address (in steps of 256 bytes) and a user-configurable interrupt level. Longword time stamp transfers are VMEbus aligned D32 longword transfers.

A status register, a command register, an internal event register, a FIFO read register, two longword time stamp registers, and several registers for the bus interrupter are offset from the configured base address. The address of a given register in user address space is generated by adding the VMEbus 16-bit short, 24-bit standard, or 32-bit extended address base (user processor specific—may be 24 or 32 bits) and the configured base address of the TPRO VME (16, 24 or 32 bits) and the offset (6 bits) for the specific register. Status bits tell the host when event data is in the FIFO, when the input code signal is present, and when the clock is known to be in time lock with the input time code or the 1PPS input.

The time code signals encode “major” time (days, hours, minutes, and seconds) using pulse width and amplitude modulation onto a sine wave carrier waveform. The time code repeats once a second (IRIG-B, NASA36, XR3, 2137) or ten times a second (IRIG-A) and includes reference marks to mark the exact beginning of a second or a tenth of a second. Each repetition is known as a frame, and the encoded time is the time at the reference mark at the beginning of the frame. The TPRO VME automatically determines the type of input time code. During each frame, the TPRO VME uses the carrier frequency (10 kHz for IRIG-A; 1 kHz for IRIG-B, NASA36, and 2137; 250 Hz for XR3) for multiple high accuracy time comparisons between the input time code and the on-board clock. Because any single comparison has jitter, multiple measurements are averaged for an accurate result, which is then used by the Z80 for frequency disciplining.

Since time information needs only about half of each frame, the remaining time is sometimes used to encode countdown information or other application-specific information. Specialized firmware options for the TPRO VME can recover or generate this information.

Depending on the actual absolute time accuracy required in the user’s application, the TPRO VME automatically corrects for the time that the time code signal requires to travel the distance between the time code source and the host computer. This “propagation delay time” is about 3.3 μ S per kilometer for radio time code transmission and about 5 μ S per kilometer for copper wire transmission. There is also a time delay on the order of 25 μ S that may be caused by small phase shifts that are due to reactances at the time code input. To correct for propagation delay, the Z80 can use a propagation delay correction setting ranging between -1000μ S (because some time sources transmit early) and $+8999 \mu$ S. The default setting is 0 μ S after the TPRO VME is reset at power-on or after the user issues a RESET command. Users can change the setting using a sequence of programmed commands to the command register on the TPRO VME.

If absolute μ S accuracy is required, the user will most likely need to calibrate for propagation delay correction by comparing the on-board clock time with a portable reference (a 1PPS GPS

pulse is good for this) when the TPRO VME is installed. The correct propagation delay correction setting is converged on rapidly by trial and error. This setting does not need to be changed unless the location of cabling of the installation is changed.

A 1PPS reference input option allows a 1-pulse-per-second TTL pulse (usually from a GPS receiver) to be used for oscillator discipline. TPRO VMEs that are configured for this option use the input code for clock time setting and disciplining if it is present, and automatically switch to the 1PPS if the input code is not present. Because only minor time (fractions of a second) can be initialized from the 1PPS, the TPRO VME needs information from the user to set the days through seconds major time. The user obtains this information from the GPS receiver (usually over a serial data link) and should then calculate the day, hour, minute and second that the next 1PPS pulse will occur. This information should be formatted into a sequence of 9 data bytes (one per digit) followed by a SET NEXT 1PPS TIME command. The user should send the 10-byte sequence to the TPRO VME command port at least once after the TPRO VME jam syncs to the 1PPS signal. The loss-of-sync status bit will be asserted after the jam sync until the NEXT 1PPS TIME SET commands are performed. If using 1PPS input, the user can simply check the loss-of-synchronization bit periodically to see if the SET NEXT 1PPS TIME command sequence needs to be performed. If the user does not set major time, the TPRO VME continue to count from the last major time (or from "0" if starting from power-on reset).

Inputs to the TPRO VME (via the 20-pin, right-angle connector, the VME P2 connector, or the front panel coaxial connectors) are modulated time code or 1PPS reference and external event pulses up to 1000 events per second.

IRIG-B time code and two TTL user-configurable rates are provided as outputs. All frequencies have a 50% duty cycle and all rates have a 1 μ S pulse width.

If P2 is a 96-pin connector (without Option –32P2), four pins of P2 Row C are available for user-configured I/O and the remaining pins of P2 Rows A and C are driven by parallel TTL time data. If P2 is a 32-pin connector (Option –32P2), there are no connections to backplane P2 Rows A and C from the TPRO VME.

1.1 General Information about GPS

NOTE: GPS applies only to the TSAT-VME board; the TPRO-VME is not equipped for GPS.

The United States government operates a set of approximately 32 satellites, collectively known as the "GPS Constellation" or "GPS Satellites." Each satellite has an internal atomic clock and transmits a signal specifying the time and satellite position. On the ground, the GPS receiver determines its position (longitude, latitude, and elevation) and the time by decoding the signals simultaneously from at least four of the GPS satellites.

The satellite orbits are circular, inclined approximately 56 degrees from the equator, orbiting the Earth once every 11 hours. There are several different orbital planes providing continuous coverage to all places on Earth. The GPS receiver uses an omni-directional antenna; the satellites move slowly across the sky (they are not at fixed locations).

Each satellite transmits a spread-spectrum signal, centered at 1575.42 MHz. When power is first applied, the GPS receiver begins searching for the satellites. It does this by searching for each satellite individually, listening for each satellite's distinct spread-spectrum hopping

sequence. This process can take a few minutes, as the receiver iteratively locates satellites, refines its position, and determines for which satellites to search.

The GPS receiver retains the last known position when the power is switched off. This results in faster satellite acquisition the next time the receiver is switched on. If the antenna has been moved more than a few miles, however, acquisition time will be slightly longer, as it must first recompute the position.

1.2 Your Spectracom GPS Receiver

Your board's GPS receiver is built into the antenna housing and communicates to the board via a serial (RS-422) interface. Power (+12V) is supplied from the board. The unit comes with a 100-foot cable. Extension cables are available in 100-foot lengths. The maximum total length is 500 feet. The connectors on the extension cables are not weatherproof; only the first 100-feet can be outdoors. The cable consists of several twisted pairs (not coaxial cable) and a foil shield.

NOTE: Spectracom recommends applying an appropriate silicon grease to the cable connection at the GPS antenna in order to protect the connection from moisture. An additional weatherproofing kit (P/N 221213) containing butyl rubber and plastic tape is also available from Andrews Corporation, US 800.255.1479.

1.3 Inventory

Before installing the board, please verify that all material ordered has been received. If there is a discrepancy, please contact Spectracom Customer Service at US 585.321.5800.

Item	Quantity	Description
1	1	Operation & Maintenance Manual
2	1	Board
3	1	Breakout Cable (micro DB-25 to several BNC connectors)
4	1	GPS Receiver/Antenna (TSAT-VME only)
5	1	100-foot Antenna Cable (TSAT-VME only)

1.4 Inspection and Support

Unpack the equipment and inspect it for damage. If any equipment has been damaged in transit, please contact Spectracom Customer Service at US 585.321.5800.

If any problems occur during installation and configuration of your Spectracom product, please contact Spectracom Technical Support at US 585.321.5823 or US 585.321.5824.

CAUTION:



Electronic equipment is sensitive to Electrostatic Discharge (ESD). Observe all ESD precautions and safeguards when handling the timecode generator.

NOTE: If equipment is returned to Spectracom, it must be shipped in its original packing material. Save all packaging material for this purpose.

1.5 External Connections

Only those functions that are actually used need to be connected. Always switch off the computer's power before connecting or disconnecting.

1.6 GPS Antenna Connector (TSAT-VME Only)

The TSAT-VME has a commercial Micro-D 15 pin plug connector that connects to the GPS antenna via the supplied cable. The supplied cable and the optional extension cable include shielding to meet EMI requirements. Use of other cables is not recommended. The optional extension cable is for indoor use only.

When power is first applied, the board sends initialization commands to the receiver/antenna. For this reason, do not disconnect and reconnect the antenna while power is applied.

2 Specifications

Table 2.1—Dimensions, Power, Fabrication & Temperature	
Dimensions	261.8mm (10.3") height x 172.2mm (6.8") depth x 20mm (0.79") width (VME 6U single width, double height)
Power	+5 VDC: $\pm 5\%$ at 3.5A max; +12 VDC: $\pm 5\%$ at 150 mA (max.); -12 VDC: $\pm 5\%$ at 100mA (max.)
PWB Fabrication	.062" thick FR4 fire-retardant glass-filled epoxy
Temperature	0° to 50° C (32° to 122° F) operating, -40° to 60° C (-40° to 140° F) storage

2.1.1 VMEbus Interface

2.1.1.1 Data Type

D08 (O) Slave

D16 Slave (even bytes n/u)

D32 Aligned Slave for longword time stamp data only

2.1.1.2 Addressing Type

A16 (short supervisory address, short non-privileged address)

A24 (standard supervisory address, standard non-privileged address)

A32 (extended supervisory address, extended non-privileged address)

2.1.2 Inputs

Table2.2—Inputs	
Connector	Configurable for 2 pins of 20-pin J2 connector, via P2 backplane C rows, front panel isolated BNC
Input Code Types	Modulated IRIG-A, IRIG-B, NASA36, 2137, XR3
Configuration	Automatic
Modulation	Amplitude modulated. Ratio of large amplitude to small Amplitude cycles between 2:1 and 4:1.
Amplitude	1.2V to 8V peak to peak.
Frequency Error	± 100 ppm maximum
Impedance	>100K Ω allows multiple units to be driven by single time code source.
Common Mode Rejection	Balanced instrumentation amplifier input withstands ± 100 V common mode voltage.
External Event	Positive going TTL pulse. The user may select either connector J1 or (by configuring P11 & P12) the VME P2 connector for external event input.

2.1.3 Outputs

Table 2.3—Outputs	
Parallel TTL Time	P2 Rows A and C. P2 Rows A and C not connected for option 32P2.
Rates	Two rates are selectable from 1PPS, 5PPS, 10PPS, 20PPS, 100PPS, 1KPPS, 10KPPS, 50KPPS and 100KPPS. The user can access these output signals on connector J1 or the VME P2 connector.
IRIG-B	1KHz amplitude modulated sine wave 3V pk/pk into 600 ohms.

2.1.4 LED Indicators

Red LED: This indicator lights when input code peak to peak amplitude <500 mV pk/pk.

Green LED: This indicator lights when time error between clock and input code (or 1PPS) is <5 μ S.

2.1.5 Timing

Table 2.4—Timing	
Resolution	1 μ S
Clock Range	1 μ S through 366 days, 23 hours, 59 minutes, 59.999999 seconds
Propagation Delay Correction Range	–1000 μ seconds through 8999 μ second Propagation Delay setting from VME host using command sequence
Synchronization Time	<20 seconds for <1 μ S error from power-on or change of propagation delay setting
Open Loop Frequency Drift	Open loop timing error shall not exceed 2 μ seconds in 10 seconds after 60 minutes with input code or 1PPS at a constant temperature.

2.2 Front Panel

2.2.1 J6 Timecode Input

This BNC jack is the differential time code input. It is isolated from ground. The center conductor is time code +, the outer conductor is time code –. It is acceptable, but not required, to connect the outer conductor to signal ground for single-ended time code signals.

The board automatically detects the polarity of this input.

2.2.2 J1 20-pin Ribbon Cable Connector (TPRO-VME Only)

The pinout is as shown below. Note that “N/C” denotes “no connection.”

Table 2.5—Pin-out of Front Panel J1 Connector for TPRO-VME			
Pin	Function	Pin	Function
1	Signal Ground	2	Time code+
3	Signal Ground	4	Time code–
5	N/C	6	N/C
7	Signal Ground	8	IRIG out
9	Signal Ground	10	Rate Output #1
11	N/C	12	Rate Output #2
13	Signal Ground	14	Time-Tag Input
15	N/C	16	N/C
17	Signal Ground	18	N/C
19	Signal Ground	20	1 MHz Output

Time code+ and time code– are described above. Rate Output #1 and #2 are described in *Chapter Three, “Rate Outputs Configuration”*. Time-Tag input is described in *Chapter Three, “Time Tag Enable/Disable”*.

Pins are located on J1 as shown in Figure 2-1. This view is looking into the J1 connector from the front panel. The component side of the board is “up” in this drawing. The indicator light is the red LED, which indicates “Lo Code Lvl”. (Illustration is not to scale.)




Figure 2-1: TPRO-VME J1 Pin Identification

2.3 Rear Panel

2.3.1 P1 and P2 VMEbus Interface and User-Defined Signals

P1 contains all of the VMEbus interfacing signals except the high-order data lines (D31–D16). P2 has the high-order data lines D31–D16. Also, the VMEbus specification states that the two outermost rows of P2 (Rows A and C) are user-defined. These pins provide direct outputs from the on-board clock, plus four jumper-selected input/outputs (*unless the board was ordered with Option –32P2*).

The direct clock time outputs (which are sometimes referred to as “zero-latency” outputs because they provide the instantaneous time) consist of 54 pins, and provide days through microseconds in binary-coded decimal (BCD) format.

Table 2.6—Direct Time (“Zero Latency”) Outputs on P2					
Pin	Output	Pin	Output	Pin	Output
C10	Day Hundreds (2)	C28	Minute Tens (1)	A15	mS Tens (2)
C11	Day Hundreds (1)	C29	Minute Ones (8)	A16	mS Tens (1)
C12	Day Tens (8)	C30	Minute Ones (4)	A17	mS Ones (8)
C13	Day Tens (4)	C31	Minute Ones (2)	A18	mS Ones (4)
C14	Day Tens (2)	C32	Minute Ones (1)	A19	mS Ones (2)
C15	Day Tens (1)	A02	Second Tens (4)	A20	mS Ones (1)
C16	Day Ones (8)	A03	Second Tens (2)	A21	uS Hundreds (8)
C17	Day Ones (4)	A04	Second Tens (1)	A22	uS Hundreds (4)
C18	Day Ones (2)	A05	Second Ones (8)	A23	uS Hundreds (2)
C19	Day Ones (1)	A06	Second Ones (4)	A24	uS Hundreds (1)
C20	Hour Tens (2)	A07	Second Ones (2)	A25	uS Tens (8)
C21	Hour Tens (1)	A08	Second Ones (1)	A26	uS Tens (4)
C22	Hour Ones (8)	A09	mS Hundreds (8)	A27	uS Tens (2)
C23	Hour Ones (4)	A10	mS Hundreds (4)	A28	uS Tens (1)
C24	Hour Ones (2)	A11	mS Hundreds (2)	A29	uS Ones (8)
C25	Hour Ones (1)	A12	mS Hundreds (1)	A30	uS Ones (4)
C26	Minute Tens (4)	A13	mS Tens (8)	A31	uS Ones (2)
C27	Minute Tens (2)	A14	mS Tens (4)	A32	uS Ones (1)

Pin C1 outputs a 1 MHz square wave to be used as a strobe for the direct time outputs. The outputs change approximately 10 nS after the strobe’s rising edge; the falling edge occurs 500 nS (± 50 nS) after the rising edge.

Also on P2 Row C are four jumper-selectable signals.

2.4 GPS Receiver/Antenna(TSAT-PMC Only)

Table 2.7—GPS Receiver/Antenna Specifications	
Description:	GPS Antenna and Receiver
Size:	5.8" (147 mm) Diameter, 3.9" (100 mm) Height
Weight:	13.4 oz (0.38 kg)
Color:	White
Mounting Threads:	1.0" ID, 14 turns/inch, straight (not tapered); will accept .75-inch galvanized water pipe for mast
Operating Frequency:	1575.42 MHz
Operating Temperature:	−30 C to +70 C
Storage Temperature:	−55 C to +100 C
Operating Humidity:	95% R.H., non-condensing
Waterproof:	Submersion to 1 meter
Altitude:	−400 m to −18,000 m
Accuracy Position Velocity Time	25 m SEP without SA 0.1 m/s without SA UTS ± 1 uS (nominal)
Power Dissipation:	2.5 Watts continuous
Acquisition Time:	<15 minutes (5–8 minutes typical)
Re-acquisition:	<2 seconds

2.4.1 Optional Cable

Table 2.8—Optional Extension Cable for (TRIM-CAB-PMC-100)	
Length	30.3 m ± 0.3 m (99.5 ft ± 1 ft)
Cable Size	9 mm (0.4 inch) O.D.
Connector Size (both ends)	34 mm X 16 mm (1.4 inch x 0.6 inch)

3 Configuration

3.1 General

The board has several configuration jumpers. The user is responsible for setting these jumpers to meet the application requirements.

To locate the jumpers, place the board component side up on a table. Position it so that the front panel is facing toward you, as shown below.

3.2 VME Base Address




Figure 3-1: VME Address Jumper Locations

Figure 3-1 shows the locations of the jumpers used to specify the board's base address in the VME memory map.

First, determine how the board is to be mapped into the VMEbus memory map. The board uses 32 (decimal) consecutive addresses, the first of which must be on a boundary of 256 addresses (A05–A07 must be zero). Determine which type of addressing mode the host will use to access the board. The board performs equally well using any of these addressing modes; choose the mode based on the host processor's requirements.

3.2.1 Case 1

A32 Extended Supervisory Data Access (Address Modifier 0D)

OR...

A32 Extended Non-privileged Data Access (Address Modifier 09)

3.2.2 Case 2

A24 Standard Supervisory Data Access (Address Modifier 3D)

OR...

A24 Standard Non-privileged Data Access (Address Modifier 39)

3.2.3 Case 3 (default)

A16 Short Supervisory Access..... (Address Modifier 2D)

OR...

A16 Short Non-privileged Access(Address Modifier 29)

3.2.4 Case 4

Permit any of the above

Distinguish length of address by address modifier field

Other address modifiers can be supported by factory customization, but those shown above work with most hosts. Note that the board does not distinguish between Supervisory Access and Non-privileged Access.

For example, if Case 1 is chosen, the board responds only if the address modifier is 0D or 09 and if A08–A15 match the settings on P4, A16–A23 match the settings on P6, and A24–A31 match the settings on P7.

If Case 2 is chosen, the board responds only if the address modifier is 3D or 39 and if A08–A15 match the settings on P4, and A16–A23 match the settings on P6. Address lines A24–A31 are ignored.

If Case 3, the factory default, is chosen, the board responds only if the address modifier is 2D or 29 and if A08–A15 match the settings on P4. Address lines A16–A31 are ignored.

If Case 4 is chosen, the board determines how many address lines to check as a function of the address modifier code. If the address modifier is 0D or 09, A08–A31 must match the settings on P4, P6, and P7. If the address modifier is 3D or 39, A08–A23 must match the settings on P4 and P6; the higher address lines will be ignored. If the address modifier is 2D or 29, A08–A15 must match the settings on P4; the higher address lines will be ignored.

First, set P16 based on one of the cases described above. P16 consists of two jumpers, which must be set as follows:

Table 3.1—P16 Jumper Set		
Case	Jumper 1	Jumper 2
1	Pin 4 to Pin 6	Pin 3 to Pin 1
2	Pin 4 to Pin 2	Pin 3 to Pin 5
3 (default)	Pin 4 to Pin 6	Pin 3 to Pin 5
4	Pin 4 to Pin 2	Pin 3 to Pin 1




Figure 3-2: P16 Default Settings

Set jumpers on P4 to select the state of address lines A15–A08. These must be set regardless of which case was chosen (see above). For each address line that needs to be logic “0”, connect the corresponding pin of P4 to P4 Pin 9. For each address line that needs to be logic “1”, leave its pin unconnected.

Table 3.2—Address Line and P4 Pin	
Address Line	P4 Pin
A08	1
A09	3
A11	4
A12	5
A13	6
A14	7
A15	8
Connect Here for “0”	9




Figure 3-3: P4 Default Settings for A08 through A15

The example shown here specifies A08, A09, and A10 must be “1” and A11–A15 must be “0”. This corresponds to address xxxx0700 (hex).

In the same manner, set jumpers on P6 to select the state of address lines A23–A16. For each address line that needs to be logic “0”, connect the corresponding pin of P6 to Pin 1 or 2. For each address line that needs to be logic “1”, leave its pin unconnected. It is necessary to set P6 only if the addressing mode is Case 1, Case 2, or Case 4.

Table 3.3—Address Line and P6 Pin	
Address Line	P6 Pin
Connect Here for “0”	1
Connect Here for “0”	2
A16	3
A17	4
A18	5
A19	6
A20	7
A21	8
A22	9
A23	10




Figure 3-4: P6 Default Settings for A16 through A23

The example shown specifies A16–A19 must be “0” and A20–A23 must be “1”. This corresponds to address `xxF0xx00` (hex).

NOTE: Pins 1 and 2 must always be connected together.

In the same manner, set the jumpers on P7 to select the state of address lines A31–A24. For each address line that needs to be logic “0”, connect the corresponding pin of P7 to Pin 1 or 2. For each address line that needs to be logic “1”, leave its pin unconnected. It is necessary to set P7 only if the addressing mode is Case 1 or Case 4.

Table 3.4—Address Line and P7 Pin	
Address Line	P7 Pin
A31	10
A30	9
A29	8
A28	7
A27	6
A26	5
A25	4
A24	3
Connect here for "0"	2
Connect here for "0"	1




Figure 3-5: P7 Default Settings for A24 through A31

The example shown here specifies A24–A28 must be “0” and A29–A31 must be “1”. This corresponds to address E0xxx00 (hex).

NOTE: Pins 1 and 2 must always be connected together.

The factory default base addresses are therefore as follows:

A16 access xxxx0700
A24 access xxF00700
A32 access E0F00700

3.3 VMEbus Interrupt Level

The board can generate an interrupt when time data has been loaded in the FIFO. This occurs when the host computer requests the current time or after a time-tag event. If desired, the user can poll the status register instead of using interrupts. The factory can also customize the board to provide periodic interrupts.

Two jumpers on P5 are used to select any interrupt IRQ1–IRQ7. Set both jumpers as shown in Table 3.5.

Table 3.5—P5 Jumper Set		
IRQ	Jumper 1	Jumper 2
1	Pin 16 to Pin 2	Pin 15 to Pin 1
2	Pin 16 to Pin 4	Pin 15 to Pin 3
3	Pin 16 to Pin 6	Pin 15 to Pin 5
4	Pin 16 to Pin 8	Pin 15 to Pin 7
5 (default)	Pin 16 to Pin 10	Pin 15 to Pin 9
6	Pin 16 to Pin 12	Pin 15 to Pin 11
7	Pin 16 to Pin 14	Pin 15 to Pin 13

If no interrupt is desired, either make no connection to P5, or leave P5 set for the factory default and program the interrupt controller to disable interrupts.




Figure 3-6: VME Interrupt P5 Jumper Location




Figure 3-7: P5 Default Settings for IRQ5

3.4 Rate Outputs Configuration

Two jumper-selected pulse rate outputs are available on the front-panel J1 connector. Each of these is a positive going pulse, nominally 1.5 μ S in duration, with the rising edge synchronized to the on-board clock. Separate output drivers are present for each output, even if the same rate is selected for both.

The outputs are TTL levels, driven by the Q output of a 74LS123 integrated circuit.




Figure 3-8: Rate Output Selection and P2 Configuration Jumper Locations

Rate Output 1 and Rate Output 2 are selected by connecting jumper wires on P13. Rate outputs and various other signals can be routed to the rear-panel P2 connector by connecting jumpers on P11 and P12, as described in the next section.

Select the output rate for Rate 1 by installing a jumper from P13 Pin 10 to one of the following locations on P13. In the same manner, select Rate 2 by installing a jumper from P13 Pin 11 to one of the following locations. It is acceptable to connect both Pin 10 and Pin 11 to the same rate, if desired. It is also acceptable to leave the jumpers unconnected if the rate output is not used.

Table 3.6—Rate Pin Connectors	
Rate	Connect to
1 Hz	P13 Pin 1
5 Hz	P13 Pin 2
10 Hz	P13 Pin 3
20 Hz	P13 Pin 4
100 Hz	P13 Pin 5
1 kHz	P13 Pin 6
10 kHz	P13 Pin 7
50 kHz	P13 Pin 8
100 kHz	P13 Pin 9

The factory default sets Rate 1 to 100 Hz, and leaves Rate 2 unconnected.




Figure 3-9: P13 Default Setting for 100 Hz Rate 1, No Rate 2

3.5 VMEbus P2 Configuration (Does Not Apply to Option -32P2)

If the board has a P2 connector with all three rows of pins installed (i.e., the board does not have option -32P2), four of these pins can be used for user-selectable purposes. The possible selections are:

- Time Code input (differential)
- 100 Hz (square wave)
- Rate Output 1
- Rate Output 2 (see paragraph 3.4)
- Time-tag input
- Counter reset output
- 1 MHz output

Any four of these can be connected to P2, Pins C6–C9, and by adding jumper wires between headers P11 and P12, as shown in Table 3.7.

Table 3.7—P2 Configuration		
P11	Pin 1	P2–C6
	Pin 2	P2–C7
	Pin 3	P2–C8
	Pin 4	P2–C9
P12	Pin 1	Time code Input +
	Pin 2	Time code Input –
	Pin 3	(RESERVED)
	Pin 4	(UNUSED)
	Pin 5	Rate Output #1 (see Table 3.6)
	Pin 6	Rate Output #2 (see Table 3.6)
	Pin 7	Time Tag input
	Pin 8	Counter Reset Output (low true)
	Pin 9	(UNUSED)
	Pin 10	(RESERVED)

The factory default is to make **no** connection to P2 Pins C6–C9. As an example, to connect the time-tag input to P2 Pin C6, add a jumper wire as shown:




Figure 3-10: Example P11, P12 Connection (Not Default)

In Figure 3-10, P13 (center row) is shown for the purpose of clarity only.

On P12, Pins 1 and 2 are the differential time code input. These two pins are also connected to the front panel J1 connector (Pins 2 and 4, respectively) and to the front panel BNC jack. **The user must not make connection to more than one of these places.**

P12 Pins 5 and 6 correspond to Rate Outputs #1 and #2, and are also connected to front panel J1 connector Pins 10 and 12, respectively. The user must be careful not to exceed the output driver capability. Also, it is not recommended that the user make connection to both the front panel pins and the rear panel P2 connector.

P12 Pin 7 is the time-tag input. This is the same as front panel J1 Pin 14. The user must not make connection to more than one of these places.

P12 Pin 8 is a low-going pulse (duration approximately 1.5 μ S). The falling edge occurs when the on-board counters are synchronized to the incoming time code (initial sync). This event, which occurs only once each time sync is established, can be used to synchronize other external devices.

3.6 Time Tag Enable/Disable

The factory default setting disables the time tag input. This is to prevent spurious events from being logged. If the time tag featured is going to be used, it is necessary to change this jumper setting as follows:

Enable Time Tag W7 to W6 (factory default)
Disable Time Tag W7 to W8 (factory default)

If time tagging is enabled, the user must ensure that no more than 200 events occur per second.

CAUTION:



If more than 200 events per second occur and W7 is connected to W6, the board may stop functioning properly. Noise can cause this when a long cable is connected to J1 with no connection to the time tag input at the far end.




Figure 3-11: Time Tag Enable/Disable Jumper Location)




Figure 3-12: Default Setting for Time Tag Disable

3.7 FIFO Buffer Clear Configuration

Normally, the FIFO needs to be cleared only at power-on reset. However, when using the time tag function, the user can configure the FIFO to be cleared when each time tag event occurs. This ensures that the data in the FIFO is always that for the most recent event. (In most applications this is not necessary, as it is possible to read and process one event before the next occurs, or several events can be stored in the FIFO before the host reads them.)

Another mode, in which the FIFO is cleared when the host accesses the interrupt controller, allows the time tag input to be used as a bus interrupt. In this application, the user is not interested in the time tag data. Time tag data is stored in the FIFO, and an interrupt is generated, when the time tag event occurs. When the host acknowledges the interrupt, the FIFO data is cleared (the host never reads the FIFO data.)




Figure 3-13: FIFO Clear Configuration Jumper Location




Figure 3-14: Default Setting for FIFO Clear on Reset

NOTE: In the preceding two figures, P9 is shown for clarity but is not related to the FIFO Clear function.

Set P14 to select the method of clearing the FIFO as shown in Table 3.8.

Table 3.8—Clear FIFO Settings	
Action	Setting
Clear FIFO	P14 Jumper
Power-on Reset	Pin 1 to Pin 2 (factory default)
Each Time Tag Event	Pin 2 to Pin 3
Interrupt Acknowledge	Pin 2 to Pin 4 ^H
^H A jumper wire, not a push-on jumper, is needed for connecting Pin 2 to Pin 4	




Figure 3-15: Other Jumper Locations (Unused)

Do not install jumpers at locations P9, P10, and P15.

NOTE: The optional display connects to P10.

3.8 VMEbus Address Configuration

The TPRO-VME base VMEbus address, VMEbus interrupt level and rate outputs, and external/internal event options are configured using wire wrap jumpers. Input and Output connections to the VME P2 connector are also selected via wire wrap jumpers. The TPRO-VME is factory configured to the following:

Base VMEbus A16 (short) address space address	0700
Base VMEbus A24 (standard) address space address	F00700
Base VMEbus A32 (extended) address space address	F0F00700
VMEbus Interrupt Level	IRG5*
External Event Enable	Enabled
FIFO Reset	Reset to empty by SYSRESET

3.8.1 A16 Base Address Configuration

The TPRO-VME base address in A16 (short) address space is configurable via wire wrap P4. Table 3.9 gives the address bits that correspond to the pins on terminal strip P4. Address bits that are “0” should be connected to GROUND (P4–9), and address that are “1” should be left unconnected.

Table 3.9—A16 Short Address Configuration	
Address Bits	Pin #
A08	P4–1
A09	P4–2
A10	P4–3
A11	P4–4
A12	P4–5
A13	P4–6
A14	P4–7
A15	P4–8

EXAMPLE

If the factory configuration address 01110000xxxxxxx (where “x” indicates offsets to I/O ports on the TPRO-VME) is to be the TPRO-VME A16 base address, then the jumpers shown in Table 3.10 must be installed.

Table 3.10— A16 Base Address Configuration Example		
Pin	TO	Pin
P4–9		P4–8, 4, 3, 2, 1

3.8.2 A24 Base Address Configuration

The TPRO-VME base address in A24 (standard) address space is configurable via wire wrap P6. Table 3.11 gives the address bits that correspond to the pins on terminal strip P6. Address bits that are “0” should be connected to GROUND (P6–2) and address bits that are “1” should be left unconnected.

Table 3.11—A24 Standard Address Configuration	
Address Bits	Pin #
A16	P6–3
A17	P6–4
A18	P6–5
A19	P6–6
A20	P6–7
A21	P6–8
A22	P6–9
A23	P6–10

EXAMPLE

The factory configuration A24 base address F00700 is configured by the jumpers shown in Table 3.10 (for the A16 portion) and in Table 3.12 (for the extra bits for A24).

Table 3.12— Sample A24 Base Address Configuration		
Pin	TO	Pin
P6-2		P6-6, 5, 4, 3

3.8.3 A32 Base Address Configuration

The TPRO-VME base address in A32 (extended) address space is configurable via wire wrap P7. Table 3.13 gives the address bits that correspond to the pins on terminal strip P7. Address bits that are “0” should be connected to GROUND (P7-2) and address bits that are “1” should be left unconnected.

Table 3.13—A32 Extended Address Configuration	
Address Bits	Pin #
A24	P7-3
A25	P7-4
A26	P7-5
A27	P7-6
A28	P7-7
A29	P7-8
A30	P7-9
A31	P7-10

EXAMPLE

The factory configuration A32 base address F0F00700 is configured by the jumpers shown in Table 3.10 (for the A16 portion), in Table 3.12 (for the extra bits for A24), and in Table 3.14 (for the extra bits for A32).

Table 3.14— Sample A32 Base Address Configuration		
Pin	TO	Pin
P7-2		P7-6, 5, 4, 3

The synchronization error is asserted until synchronization to input to $<5\ \mu\text{S}$ is verified and when the observed time base error exceeds $5\ \mu\text{S}$ for 5 successive observations spaced at 1 second intervals. If the observed time base error exceeds $200\ \mu\text{S}$, jam resynchronization is initiated. Synchronization error is also asserted when input code is inconsistent or lost (unless 1PPS is present to verify time error).

Status Bit (3-7)Not used
Offset (1B-1F)Not used

3.9 Interrupt Request Level Configuration

The TPRO-VME can be configured as an **interrupter** on the VME Priority Interrupt Bus. Any one of the seven VMEbus interrupt request signal lines IRQ1* to IRQ7* can be chosen by the user to request service from the VME Interrupt Handler. The user must configure terminal strip P5 with two jumpers (one from an odd numbered pin to P5-15, one from an even numbered pin to P5-16) for the desired interrupt level. Table 3.15 shows the configuration of P5 for the seven configurable interrupt levels.

Table 3.15—Interrupt Priority Configuration		
Interrupt Level	Even Jumper	Odd Jumper
VMEbus IRQ1*	P5-2 to P5-16	P5-1 to P5-15
VMEbus IRQ2*	P5-4 to P5-16	P5-3 to P5-15
VMEbus IRQ3*	P5-6 to P5-16	P5-5 to P5-15
VMEbus IRQ4*	P5-8 to P5-16	P5-7 to P5-15
VMEbus IRQ5*	P5-10 to P5-16	P5-9 to P5-15
VMEbus IRQ6*	P5-12 to P5-16	P5-11 to P5-15
VMEbus IRQ7*	P5-14 to P5-16	P5-13 to P5-15

EXAMPLE

For the TPRO-VME factory configuration IRQ5* on the VME Priority Interrupt Bus then P5-10 should connect to P5-16 and P5-9 connect to P5-15.

3.10 VMEbus P2 I/O Configuration

If a 96-pin VMEbus connector is installed in P2 (i.e., Option -32P2 is not present), the TPRO-VME allows four output signals to interface via CSEL1, CSEL2, CSEL3 and CSEL4 signals on the VMEP2 bus via CSEL1, CSEL2, CSEL3 and CSEL4 by configuring P11.

Table 3.16—Terminal Strip P11 Pin Assignments		
Signal Name	Pin Number	P2 Pin
CSEL1	P11-1	P2C-6
CSEL2	P11-2	P2C-7
CSEL3	P11-3	P2C-8
CSEL4	P11-4	P2C-9

Table 3.17 gives pin assignments for terminal strip P12.

Table 3.17—Terminal Strip P12 Pin Assignments	
Signal Name	Pin Number
Input Code +	P12-1
Input Code –	P12-2
100 PPS	P12-3
IRIG-B output	P12-4
SEL1	P12-5
SEL2	P12-6
EXTERNAL EVENT	P12-7
GATE*	P12-8
N/C	P12-9
1MHZFREQ	P12-10

EXAMPLE

If Input Code+ and Input Code– are distributed on VME P2 bus lines 2C:06 and 2C:07 respectively then the jumpers shown in Table 3.18 must be installed in order to use the VME P2 connector input option.

Table 3.18—VME P2 Connector Configuration Example		
Pin	TO	Pin
P12-1		P11-1
P12-2		P11-2

3.11 Rate Output Configuration

The TPRO-VME can be configured to provide any two rates provided on terminal strip P13. Table 3.19 gives the rates connected to terminal strip P13.

Table 3.19—Terminal Strip P13 Pin Assignments	
Signal Name	Pin Number
1 PPS	P13-1
5 PPS	P13-2
10 PPS	P13-3
20 PPS	P13-4
100 PPS	P13-5
1 KPPS	P13-6
10 KPPS	P13-7
50 KPPS	P13-8
100 KPPS	P13-9
RATE1	P13-10
RATE2	P13-11

EXAMPLE

To select 100 PPS as SEL1 (J1–10) and the 50 kHz PPS rate output as SEL2 (J1–12), the jumpers shown in Table 3.20 must be installed.

Table 3.20—Rate Configuration Example		
Pin	TO	Pin
P13–5		P13–10
P13–8		P13–11

3.12 External/Internal Event Configuration

The TPRO-VME is configured to enable external events at the factory (W6 connects to W7). Should the user wish to disable this external events capability, then W7 should be jumpered to W8 rather than W6.

3.13 FIFO Buffer Clear Configuration

P14 configures the FIFO buffer that has been cleared by SYSRESET, by the occurrence of an external event, or by a host access to the bus 68153 interrupter. Clearing on event can reduce the time needed by a program to empty the FIFO, especially if low-order time measurement bits are not needed. It can also eliminate the need for the user to flush the FIFO of stale data before capturing event times. Connect P14–2 to P14–1 for SYSRESET clear (factory default) or P14–2 to P14–3 for clear on event. Connect P14–2 to P14–4 to clear the FIFO whenever the BIM is accessed.

4 Pin Assignments

4.1 Test Points

Table 4.1—Test Point Assignments	
Test Points	Signal Name
TP1	6 MHz undisciplined
TP2	100 PPS
TP3	Ground

Table 4.2—Test Point Assignments	
J2 Pin	Signal
Center	CODE IN +
Shield	CODE IN –

4.2 Input/Output Pin Assignments

Table 4.3—J1 (20 pin) I/O	
J1 Pin	Signal
1	Ground
2	Time Code IN +
3	Ground
4	Time Code IN –
5	N/C
6	N/C
7	Ground
8	IRIG-B OUT
9	Ground
10	SEL RATE 1
11	N/C
12	SEL RATE 2
13	Ground
14	External Event IN
15	N/C
16	N/C
17	Ground
18	N/C
19	Ground
20	1 MHz Disciplined

Table 4.4—J3 (SMB) IRIG-B Output J5 (BNC)

J3 Pin	Signal
Center	IRIG-B OUT +
Shield	GROUND

Table 4.5—P8 GP STAR I/O

P8 Pin	Signal
01	Ground
03	+5V
05	+5V
07	+12V
09	Ground
11	+12V
13	–12V
15	TTL Serial IN
02	TTL Serial IN
04	TTL Serial OUT
06	Rest*
08	TTL Serial OUT
10	1PPS TTL IN
12	Ground
14	Ground

Table 4.6—P2 Parallel Outputs			
P2 Row A		P2 Row C	
Pin Number	Signal	Pin Number	Signal
01	N/C	01	!DATA VALID
02	40 seconds	02	Reserved
03	20 seconds	03	Reserved
04	10 Seconds	04	Reserved
05	8 seconds	05	Reserved
06	4 seconds	06	CSEL1
07	2 seconds	07	CSEL2
08	1 second	08	CSEL3
09	800 mS	09	CSEL4
10	400 mS	10	200 days
11	200 mS	11	100 days
12	100 mS	12	80 days
13	80 mS	13	40 days
14	40 mS	14	20 days
15	20 mS	15	10 days
16	10 mS	16	8 days
17	8 mS	17	4 days
18	4 mS	18	2 days
19	2 mS	19	1 day
20	1 mS	20	20 hours
21	800 μ S	21	10 hours
22	400 μ S	22	8 hours
23	200 μ S	23	4 hours
24	100 μ S	24	2 hours
25	80 μ S	25	1 hour
26	40 μ S	26	40 minutes
27	20 μ S	27	20 minutes
28	10 μ S	28	10 minutes
29	8 μ S	29	8 minutes
30	4 μ S	30	4 minutes
31	2 μ S	31	2 minutes
32	1 μ S	32	1 minute

5 Installation

5.1 General

The TPRO-VME's inputs, outputs, and terminal strips should be connected and configured for the user requirements before the board is installed. Front panel BNC cutouts allow users to install front panel BNC connectors for code input and IRIG-B output.

The TPRO-VME module should be inserted or removed from the VMEbus backplane only when the VME system power is off.

There are up to 3 inputs from the external world: an *input time code* (if used), a *TTL external event pulse* (if used), and a *1PPS TTL time pulse* (if used).

- **Time code input** to the board is connected at J1 (+side to Pin 2, –side to Pin 4), to J2 (if installed), a front panel BNC wired to J2 (if installed) or the VME P2 connector. To use P2 see Chapter Three. (If input time code is not provided, the board provides “local” timing starting at 000 days, 00 hours, 00 minutes, 00 seconds).
- **TTL external event pulses** are connected at J1 or the VME P2 connector. To use P2 see Chapter Three. The external event pulses may come from external user equipment, or pulse outputs from the TPRO-VME at J1 may be looped back and used as external event inputs with a simple jumper connection.
- If the **1PPS option** is used, a front panel BNC will be factory installed for 1PPS TTL (rising edge) input.

Custom factory configurations have a manual addendum that details non-standard I/O configuration.

The user should refer to Chapter Three of this manual in order to be certain that jumpers on configuration headers P4, P5, P6 and P7 are installed correctly. These configurations control the TPRO-VME Base Address and Interrupt Request Level. If “on-time” rates are required as outputs, then terminal strip P13 must be configured.

First-time users are advised to configure the board with the address and interrupt level settings used for the demonstration program, and then to run the demonstration program to gain confidence in correct operation. The board can then be configured for the users requirement.

It should be noted that although the VMEbus is an asynchronous bus, operations within the MC68153 Bus Interrupter Module portion of the TPRO-VME are synchronous. If the TPRO-VME will be used in an interrupt driven mode, a VMEbus system clock (SYSCLK) signal must be provided on Pin A10 of the VMEbus P1 connector. Systems that meet the C.1 VMEbus specification will always have SYSCLK present.

After configuring the wire wrap jumpers, connecting the inputs and outputs, and powering the VME system down, the TPRO-VME board installs in a dual VME (6U) backplane slot. If the

TPRO-VME is to be used in an interrupt driven mode, there should be no empty connector slots in the interrupt acknowledge signal path between the INTERRUPT HANDLER and the TPRO-VME.

5.2 External GPS Receiver/Antenna Installation

The GPS receiver and antenna are housed in a single enclosure. This unit must be mounted outdoors with an unobstructed view of the sky. Best timing accuracy is obtained when the unit has a clear view of the entire sky (i.e., the antenna must be mounted horizontally, and the mast should not be more than 30 degrees off of vertical).

For the most reliable operation, the antenna module should be mounted with a full horizon view and the 100-foot serial/power cable and coaxial cable run to the VMEbus system location. Any excess cable length can be coiled without affecting correct operation. The adapter cable on the TPRO-VME is equipped with three connectors to connect the power, serial data, and 1PPS signals to the mast module.

Once the mast module and cabling have been installed, and the TPRO-VME has been installed and connected to the mast module, the system can be powered up again. The TPRO-VME ignores the GPS time until time lock is achieved. Then, the TPRO-VME synchronizes to the GPS time as indicated by the green “GOOD SYNC” LED on the TPRO-VME.

Connection to the unit is made with the supplied 12-conductor cable. This cable brings power to the unit and conveys the serial (RS-422) communication and one-pulse-per-second from the unit to the computer board. Note that the cable uses no RF signals. **The user is responsible for weatherproofing the cable connection.** Suitable weatherproofing compounds, (e.g., “marine putty”) are available at hardware stores.

The connector on the antenna end of the cable will pass through a ¾-inch or greater hole. The connector on the computer end of the cable is larger; it will pass through a 1¾-inch or greater hole.

There are no special precautions regarding lightning. The unit will not survive a direct or close-proximity lightning strike. Installation of lightning arrestors is neither practical nor possible. The unit is connected to ground through the cable (and thus through the computer). No additional grounding is possible. The brass object in the center of the mounting hole is not an electrical contact. It is acceptable, but not required, that a metal mast make contact with this point.

5.2.1 Routine Maintenance

Inspect the connection to the unit at least once each year. It is not necessary to de-mate this connection, just verify that the weatherproofing is still intact. Replace the weatherproofing if necessary.

NOTE: The GPS receivers included with your product are burned in at the Spectracom facility. Because this burn in is, of course, carried out outdoors, the units may appear soiled. Do not be concerned; the units are not used. This is just weathering—a normal part of the burn-in process, which helps to insure that you receive the most reliable product possible

5.3 Propagation Delay Adjustment

The user can command the TPRO-VME to correct for the period necessary for the time code signal to travel the distance between the time code source and the host computer, based on the actual absolute time accuracy that the user's application requires.

This "propagation delay time" is approximately 3.3 μS per kilometer for radio time code transmission, and approximately 5 μS per kilometer for copper wire transmission. Also, small phase shifts that are due to reactances at the time code input may cause a time delay on the order of 25 μS . To correct for propagation delay, the Z80 can use a propagation delay correction setting ranging from -1000 (because some time sources transmit early) to +8999 μS . The default setting is 0 μS after the board is reset at power-on, or after a RESET command from the user. Users can change the setting using a sequence of programmed commands to the command register on the TPRO-VME. If the 1PPS input is used for time reference, propagation delay settings are neither needed nor used.

If absolute μS accuracy is required, the user may need to calibrate for propagation delay correction by comparing the on-board clock time with a portable reference (a 1PPS GPS pulse is good for this) when the TPRO-VME is installed. The correct propagation delay correction setting is converged on rapidly by trial and error. This setting does not need to be changed unless the location or cabling of the installation is changed.

Determining the correct propagation delay setting often needs the help of a special user program that experiments with a variety of propagation delay settings while zeroing in on the correct setting. The normal user program should have the ability to use the correct setting once it is determined.

6 Programming

6.1 Introduction

In addition to the programming information relating to time data given in this manual, there are command codes for triggering the TPRO-VME to add latitude, longitude and elevation data in the FIFO buffer.

6.2 Register Map

Table 6.1—Addressing Base: A6 to A31 Configurable by Wire Wrap	
Offset 1	D08/D16 MC68153 control register 0
Offset 3	D08/D16 MC68153 control register 1
Offset 5	D08/D16 MC68153 control register 2
Offset 7	D08/D16 MC68153 control register 3
Offset 9	D08/D16 MC68153 vector register 0
Offset B	D08/D16 MC68153 vector register 1
Offset D	D08/D16 MC68153 vector register 3
Offset F	D08/D16 MC68153 vector register 4
Offset 11	D08/D16 MC68153 Command register 0
Offset 13	D08/D16 MC68153 Internal event register 1
Offset 14	D08/D16 MC68153 D32 longword high order time read and freeze release
Offset 15	D08/D16 FIFO Data read
Offset 17	Not used 1
Offset 18	D32 longword time freeze ad low order data
Offset 19	D08/D16 Status Register
Status Bit 0	"0" FIFO not empty, "1" FIFO empty
Status Bit 1	"0" input code amplitude <500mV pk/pk, "1" Input code amplitude> 500mV pk/pk
Status Bit 2	"0" Synchronization error, "1" Good synchronization

The synchronization error is asserted until synchronization to input to <5 μ S is verified by the Z80 or when the observed time base error exceeds 5 μ S for five successive observations spaced at one-second intervals.

If the observed time base error exceeds 200 μ seconds, jam resynchronizations are initiated. Synchronization error is also asserted whenever input code is inconsistent or lost (unless 1PPS is present to verify time error).

Status Bit (3–7) Not used
 Offset (1B–1F) Not used

6.3 Determining Address Space

The user's VMEbus processor defines address windows for mapping user addresses into VMEbus addresses. These mappings vary from manufacturer and even from model to model for a given manufacturer. There will usually be a different window for A32 (extended) VMEbus addressing, A24 (standard) VMEbus programming and A16 (short) VMEbus addressing. To

access the various registers used in programming the TPRO-VME, the user must decide what address space they are going to use (short space is recommended for compatibility with the AITG-VME). The address used in the user program to access a given register will be the sum of the address window base, plus the configured base address of the TPRO-VME, plus the register offset.

6.4 Reading Time

Time can be measured instantaneously by host programs by two sequential D32 longword reads from the Base+18 and Base+14 addresses. The first read (to Base+18) freezes the time in hardware latches and also transfers the low order (seconds through microseconds) time bits. The second read (to Base+14) transfers the high order (days through minutes) time bits and re-enables the freeze capability. Bits marked n/u may be “1” or “0”.

Table 6.2—Low Order D32 Register (Base+18)

Bit	Bit	Bit	Bit	Bit	Bit	Bit	Bit	Bit	Bit	Bit	Bit	Bit	Bit	Bit	Bit
31	30	29	28	27	26	25	24	23	22	21	20	19	18	17	16
n/u	40	20	10	8	4	2	1	800	400	200	100	80	40	20	10
Seconds								Milliseconds							
Bit	Bit	Bit	Bit	Bit	Bit	Bit	Bit	Bit	Bit	Bit	Bit	Bit	Bit	Bit	Bit
15	14	13	12	11	10	9	8	7	6	5	4	3	2	1	0
8	4	2	1	800	400	200	100	80	40	20	10	8	4	2	1
Milliseconds				Microseconds											

Table 6.3—High Order D32 Register (Base+14)

Bit	Bit	Bit	Bit	Bit	Bit	Bit	Bit	Bit	Bit	Bit	Bit	Bit	Bit	Bit	Bit
31	30	29	28	27	26	25	24	23	22	21	20	19	18	17	16
n/u	n/u	n/u	n/u	n/u	n/u	200	100	80	40	20	10	8	4	2	1
—						Days									
Bit	Bit	Bit	Bit	Bit	Bit	Bit	Bit	Bit	Bit	Bit	Bit	Bit	Bit	Bit	Bit
15	14	13	12	11	10	9	8	7	6	5	4	3	2	1	0
n/u	n/u	20	11	8	4	2	1	n/u	40	20	10	8	4	2	1
—		Hours							Minutes						

6.5 Reading Latitude, Longitude, and Elevation (TSAT-VME Only)

Readings of latitude, longitude or elevation can be caused by the host VME program writing the appropriate one byte command to the TSAT-VME command port at the base address +11 (hex). The Z80 microcomputer in the TSAT-VME then transfers the 10 bytes of latitude or longitude or elevation information sequentially into the time data FIFO on the TSAT-VME board according to the formats shown in the tables following. The host program reads the 10 bytes of information (format shown below) sequentially from the FIFO through the TSAT-VMEbus interface. Handshaking is done by the host testing the OUTPUT READY bit (bit 0) of the TSAT-VME status register (offset 19 hex) for “0” before each byte is read from the FIFO output. Note that **S** latitudes and **W** longitudes are biased by +180 degrees to distinguish them from **N** latitudes and **E** longitudes. The TSAT-VME automatically updates the position data with data received from the mast module every second. The OUTPUT READY status bit will be asserted (0) whenever the FIFO is not empty. The time required for the complete transfer is up to 100 μ S after the command. It fluctuates because the Z80 microcomputer inside the TSAT-VME may be interrupted while putting data in the FIFO.

Table 6.4—Extra TSAT-VME Commands for Position Reports	
Command Byte	Command
5F (hex)	Latitude Report
5E (hex)	Longitude Report
5D (hex)	Elevation Report
5C (hex)	#SAT Report

6.6 Report Latitude (TSAT-VME Only)

The GPS receiver's computation for latitude is put into the FIFO in response to the 0x005f command. The units are degrees, minutes, and fractional minutes. Latitudes north of the equator are reported as 0–90 degrees, latitudes south of the equator are reported as 180+ degrees South—i.e., if the reported value is greater than or equal to 180, subtract 180 degrees and change the sign to negative (south). It is important to convert minutes to decimal degrees before comparing to 180 degrees. The response is in the format shown here:

Table 6.5—10-byte FIFO Latitude Data Format		
Word	Bits 7-4	Bits 3-0
0 (first)	5	F
1	5	F
2	0	10^2 degrees
3	10^1 degrees	10^0 degrees
4	10^1 minutes	10^0 minutes
5	10^{-1} minutes	10^{-2} minutes
6	10^{-3} minutes	0
7	0	0
8	0	0
9 (last)	0	0

6.7 Report Longitude (TSAT-VME Only)

The GPS receiver's computation for longitude is put into the FIFO in response to the 0x005e command. The units are degrees, minutes, and fractions of minutes. The response is in the following format:

Table 6.6—10-byte FIFO Longitude Data Format		
Word	Bits 7-4	Bits 3-0
0 (first)	5	E
1	5	E
2	0	10^2 degrees
3	10^1 degrees	10^0 degrees
4	10^1 minutes	10^0 minutes
5	10^{-1} minutes	10^{-2} minutes
6	10^{-3} minutes	0
7	0	0
8	0	0
9 (last)	0	0

NOTE: If the reported value is less than 180 degrees, the longitude is East (for Europe, Africa, Asia, etc.). If the reported value is greater than or equal to 180 degrees, subtract 180 degrees to obtain west longitude (for USA, Canada, Central & South America, etc.). For example, the reported value for Los Angeles, CA, USA, will be approximately 298 degrees. Since this is greater than 180, subtract 180 to get the correct value, i.e., 118 degrees West. Be sure to convert minutes to decimal degrees before comparing to 180 degrees.

6.8 Reading the Year, Gregorian Date, and Altitude (TSAT-VME Only)

The GPS receiver's computations for the altitude and Gregorian date are reported in the FIFO in response to the 0x005d command. Altitude is in units of meters relative to mean sea level.

Table 6.7—10-byte FIFO Elevation Data Format		
Word	Bits 7-4	Bits 3-0
0 (first)	5	D
1	5	D
2	10^1 day	10^0 day
3	10^5 meters	10^4 meters
4	10^3 meters	10^2 meters
5	10^1 meters	10^0 meters
6	10^{-1} meters	0
7	10^1 year	10^0 year
8	10^3 year	10^2 year
9 (last)	10^1 month	10^0 month

The board increments the Gregorian date approximately one second after time rolls over from 23:59:59 to 00:00:00 (HH:MM:SS). This occurs on every date in every year. Consequently, the Gregorian date should not be used during minutes 59 and 00—this will avoid rollover problems.

6.9 Report Number of GPS Satellites Being Tracked (TSAT-VME Only)

The GPS receiver determines automatically which satellites to track. The number (quantity) of satellites being tracked is reported in the FIFO in response to command (5Ch). The response is of the form shown here:

Table 6.8—10-byte FIFO #SAT Data Format		
Word	Bits 7–4	Bits 3 – 0
0 (first)	5	C
1	5	C
2	Reserved	Reserved
3	Reserved	Reserved
4	Reserved	Reserved
5	Reserved	Reserved
6	Reserved	Reserved
7	0	# of Satellites
8	Reserved	Reserved
9 (last)	Reserved	Reserved

NOTE: The number of satellites tracked is for diagnostic purposes only. Valid time and position require that at least four satellites be tracked. The In-Sync flag (status bit 2) must be used to actually determine whether the board is synchronized and providing valid position.

There is no difference in timing accuracy when 5, 6, 7, or 8 satellites are being tracked. It is normal for the number to fluctuate between 5 and 8, given that the GPS receiver picks different satellites to optimize the timing.

It is normal for the number of satellites to be between 0 and 3 during acquisition and re-acquisition.

6.10 Reading External Event Times

A logic pulse can cause time capture from the outside world (an “external” event). The Z80 microcomputer in the TPRO-VME then transfers the 10 bytes of captured time information sequentially into the time data FIFO on the TPRO-VME board.

It is possible to simulate an external event (e.g., for purpose of testing interrupt-handling software) by writing to the INTERNAL EVENT register (offset 13). Custom versions of the TPRO-VME (as detailed in manual addendum supplied with customized versions) may also allow the external event pulse to latch time in the longword time freeze registers.

The host program reads captured time information (11s of days through units of microseconds—a total of 10 bytes) sequentially from the FIFO through the TPRO-VME VMEbus interface. The host that is testing the OUTPUT READY bit (Bit 0) does handshaking of the TPRO-VME status register (offset 19) for “0” before each byte is read from the FIFO output. Complete transfer requires a time period of from 30 to 100 μ S after the event occurs; it fluctuates because the microcomputer may be interrupted while putting data in the FIFO. The time data is accurate to the exact microsecond the event occurred, and the accuracy is not affected by the transfer time. The user should be sure to empty the FIFO of stale data when their program is initialized. Do this by continuously reading the FIFO until the OUTPUT READY status bit is “1” for at least 100 μ S, which could take about 512 reads if the FIFO is full.

Table 6.9—10-byte FIFO Event Data Format		
Word	Bits 7-4	Bits 3-0
0 (first)	0	0
1	0	0
2	0	Hundreds of days
3	Tens of days	Units of days
4	Tens of hours	Units of hours
5	Tens of minutes	Units minutes
6	Tens of seconds	Units of seconds
7	Hundreds of milliseconds	Tens of milliseconds
8	Units of milliseconds	Hundreds of microseconds
9 (last)	Tens of microseconds	Units of microseconds

6.11 Interrupt Mode

Using the on-board bus interrupter, users can interrupt the host VME processor when the first byte of data is loaded in the FIFO. In many cases, the interrupt handling (task switching) overhead exceeds the time required for the complete loading of the FIFO, so that the VME host may never actually need to wait for the microcomputer to complete loading the FIFO. The INT3 input to the MC68153 bus interrupter is connected to the FIFO output ready. If the TPRO-VME is used to generate interrupts, then MC68153 Control Register 3 (offset 7) and MC68153 Vector Register 3 must be programmed correctly by the users program. The interrupt level selected should match the level configured with P4 and P5. Bit assignments for MC68153 Control Register 3 are shown in Table 6.10. *(To use interrupts, the user should consult documentation for the MC68153 Bus Interrupter Module, and read and understand the programming information.)*

Typically, the INT3 interrupts are disabled after the first interrupt from a time stamp while the bytes are read from the FIFO. If not, a new interrupt is generated for each of the remaining 9 bytes. Remember that the interrupt occurs when the first byte of the 10 bytes is written to the FIFO. The programmer should check the FIFO OUTPUT READY for “0” (asserted) before reading each of the following 9 bytes. To disable their inputs, the INT0 and INT1 inputs to the MC68153 are tied to a pull-up resistor. INT2 is disabled in software unless otherwise instructed in a manual for the custom version. See the interrupt driven programming example. Remaining MC68153 INT inputs are not connected. The user may connect them to the on-board pulse rates to generate periodic interrupts without event processing overhead.

Table 6.10—MC68153 Interrupt Control Register							
Bit	Bit	Bit	Bit	Bit	Bit	Bit	Bit
7	6	5	4	3	2	1	0
N/U	N/U	N/U	IRE	N/U	L2	L1	L0
X	X	X	W	X	W	W	W

The least-significant, 3-bit field of the interrupt control register determines the level at which an interrupt is generated.

Table 6.11—Control Register Interrupt Level			
L2	L1	L0	IRQ Level
0	0	0	Disabled
0	0	1	IRQ1*
0	1	0	IRQ2*
0	1	1	IRQ3*
1	0	0	IRQ4*
1	0	1	IRQ5*
1	1	0	IRQ6*
1	1	1	IRQ7*

The IRE field (Bit 4) must be set (high level) to enable the bus interrupt request associated with the control register. If this bit is cleared a VMEbus interrupt will not be generated.

Vector Register 3 (offset F) is an eight-bit register within the MC68153 that the user must program with the interrupt vector number associated with the TPRO-VME interrupt service routine.

This appendix of this manual contains two demonstration programs. The first program illustrates the TPRO-VME operating in the non-interrupt mode using the longword time reads. The second program shows the TPRO-VME operating in the interrupt mode, using 100PPS output signal from J1 as a source of external event pulses.

6.12 Setting Propagation Delay Correction

The user may set the propagation delay to compensate for propagation delay between the time code source and the TPRO-VME location. The default setting is “0”. The propagation delay is programmed by sending a wake-up byte to the TPRO-VME command offset, then sending one command byte for each propagation delay digit to the toe command offset, and then sending a “go” command to restart with the new propagation delay. After power-up, allow nine seconds to pass before setting the propagation delay. Remember to delay at least 100 μ S between sending each command byte. Table 6.12 shows the command sequence for setting the propagation delay.

Table 6.12—Propagation Delay Programming Protocol

Byte	High Nibble	Low Nibble
0	F	0 (wake-up command)
1	3	BCD for 10 ³ μ S digit
2	2	BCD for 10 ² μ S digit
3	1	BCD for 10 ¹ μ S digit
4	0	BCD for 10 ⁰ μ S digit
5	E	0 (“go” command)

For example, to set a propagation delay correction value of 1234 μ S, the sequence “F0 31 22 13 04 E0” is sent. For negative propagation delay settings, add 10000 μ S to the desired value. For example, to set –500 μ S use a setting of 9500 μ S.

6.13 Presetting Time

For applications in which code input is not used, the user can preset the TPRO-VME time to a predetermined value. If the user does not preset a time, a default preset of 0 days through seconds is used. If code input is connected to the TPRO-VME, it will override the preset time unless the DISABLE RESYNC command has been sent to the TPRO-VME. As each digit is set, the corresponding digit is set in the TPRO-VME’s internal time-set buffer. When the GO command (E0) is received, the time set buffer is copied to the clock. At least 100 μ S should be allowed between sending each time set command byte. The command bytes should be sent to the command register (offset hex 11).

Table 6.13—Time Initialization Programming Protocol

Byte	High Nibble	Low Nibble
0	F	0 (ZERO time set buffer commands)
1	5	BCD for 100s of days digit
2	6	BCD for 10s of days digit
3	7	BCD for 1s of days digit
4	8	BCD for 10s of hours digit
5	9	BCD for 1s of hours digit
6	A	BCD for 10s of minutes digit
7	B	BCD for 1s of minutes digit
8	C	BCD for 10s of seconds digit
9	D	BCD for 1s of seconds digit
0	E	0 (GO command)

For example, to set time to “123 days 01 hours 23 minutes 45 seconds”, the sequence “F0 51 62 73 80 91 A2 B3 C4 E0” is sent.

6.14 External 1PPS Time Synchronization (Option –M)

A 1PPS reference input option allows a 1 pulse-per-second TTL pulse (usually from a GPS receiver) to be used for oscillator disciplining. TPRO-VMEs that are configured for this option

use the code input for clock time set and disciplining, if it is preset, and automatically switch to the 1PPS if the input code is not preset.

Because only minor time (fractions of a second) can be initialized from the 1PPS input, the TPRO-VME needs information from the user to set the days through seconds major time. The user obtains this information from the GPS receiver (typically over a serial data link), and should then calculate the day, hour, minute and second that the next 1PPS pulse will occur. This information should be formatted into a sequence of 9 data bytes (one per digit) followed by a SET NEXT 1PPS TIME command. The user should send the 10-byte sequence to the TPRO-VME command port at least once after the TPRO-VME jam syncs to the 1PPS signal. The loss-of-sync status will be asserted after the jam sync until the NEXT 1PPS TIME SET commands are performed. If using 1PPS input, the user can simply check the loss-of-synchronization bit periodically to see if the SET NEXT 1PPS TIME command sequence needs to be performed. If the user does not set major time, the TPRO-VME continues to count from the last major time (or “0” if starting from power-on reset). At least 100 μ S should be allowed between sending each byte in the sequence.

Table 6.14—1PPS Time Initialization Programming Protocol		
Byte	High Nibble	Low Nibble
0	5	BCD for 100s of days digit
1	6	BCD for 10s of days digit
2	7	BCD for 1s of days digit
3	8	BCD for 10s hours digit
4	9	BCD for 1s of hours digit
5	A	BCD for 10s of minutes digit
6	B	BCD for 1s of minutes digit
7	C	BCD for 10s of seconds digit
8	D	BCD for 1s of seconds digit
9	4	C (SET NEXT 1PPS TIME command)

For example, to set next 1PPS time to “123 days 01 hours 23 minutes 45 seconds”, the sequence “51 62 73 80 91 A2 B3 C4 D5 4C” is sent.

6.15 Disabling and Re-enabling Code Input

Table 6.15 gives the two commands used to tell the TPRO-VME to ignore code—i.e., not to synchronize to it even if there is a sync error—or to enable re-synchronizing the on-board clock to the input code (the default).

Table 6.15—Sync Control and Reset	
Command Byte	Command
4D	Enable Re-sync
4E	Disable Re-sync

6.16 Handlers

Handlers and drivers are available for TPRO-VME for SunOS, other Unix variants, and other operating systems.

7 Preventive Maintenance

7.1 Necessary Maintenance

7.1.1 Oscillator Aging Adjustment

The oscillator aging adjustment for the TPRO-VME corrects for the effects of aging on the natural crystal oscillator frequency to insure that the undisciplined frequency of the 1MHz oscillator is 1.000000 MHz \pm 10 Hz.

This preventive maintenance should be performed once every 2 years. A digital frequency counter with 1PPM or better accuracy and 1Hz or better resolution (be sure that the counter is calibrated), and a trimmer capacitor adjustment tool (non-metallic screwdriver) are required. For units with a custom crystal oscillator, consult the oscillator data sheet for adjustment method.

Put the TPRO-VME on a card extender and connect a calibrated frequency counter to the 1 MHz signal at P1 Pin 20; use P1 Pin 19 for ground. Before powering up the system, disconnect the time code input and/or external 1PPS input to the board so that the oscillator is undisciplined. Then, power the system up and wait at least five minutes for the on-board crystal oven temperature to stabilize. Then adjust C9 for the frequency specified above. Power down the system, leaving time code and 1PPS disconnected, and re-install the TPRO-VME without the extender. Because some systems may have extremely high airflow, which affects the crystal temperature, power up the system and recheck the frequency after five minutes.

If the frequency with the board installed is outside the tolerance given above, it may be necessary to bias the frequency setting when extended to reach the desired tolerance when installed.

When frequency adjustment is completed, re-attach the time code input or 1PPS for normal operation.

7.1.2 IRIG-B Output Adjustment

With the TPRO-VME at approximate operating temperature, adjust R108 for zero crossover discontinuity at the transition from large to small amplitude cycles. This adjustment ensures that the sine wave phase is in proper lock to the amplitude modulation signal.

The remaining R104, R112 and R113 adjustments are seldom needed in the field. Output amplitude adjustment is made with R104. R112 adjusts waveform distortion (triangle vs. sine), and R113 adjusts wave waveform symmetry.

7.2 Optional Preventive Maintenance

7.2.1 Oscillator Aging Adjustment

The oscillator aging adjustment for the TPRO/TSAT-VME corrects for the effects of aging on the natural crystal oscillator frequency to insure that the undisciplined frequency of the 10 MHz oscillator is 10.000000 MHz \pm 10 Hz.

This preventive maintenance, although optional, should be performed once every 2 years.

7.2.1.1 Necessary Equipment

1. A digital frequency counter with 1 PPM or better accuracy and 1 Hz or better resolution (*make certain* that the counter is calibrated).
2. A trimmer capacitor adjustment tool (non-metallic screwdriver). If you have a custom crystal oscillator in your unit, consult the oscillator data sheet for adjustment method.

CAUTION:

Electronic equipment is sensitive to Electrostatic Discharge (ESD). Observe all ESD precautions and safeguards prior to and during calibration.

7.2.1.2 Procedure

1. Power down the system.
2. Connect a calibrated frequency counter to the 1 MHz signal at JP9 Pin 4. (Use JP10 Pin 3 for ground.)
3. Before powering up the system, disconnect the time code input and/or the external 1PPS input to the board so that the oscillator is undisciplined.
4. Power up the system and wait at least five minutes for the on-board crystal oven temperature to stabilize. Adjust the oscillator's on-board trimmer for 1MHz \pm 1Hz.
5. Power down the system and reconnect any disconnected inputs.

8 Troubleshooting

8.1 The program has trouble talking to the TPRO-VME

Try to run versions of the Spectracom demo programs (modified as little as possible) to reduce the possibility that what appears to be a hardware problem is really a software problem. If your program crashes due to a bus error when trying to access the TPRO-VME the problem may be:

- The configured board address does not agree with what your program uses
- You are using the wrong window address for the address mode (A16, 24, A32) being used
- You are performing D32 longword transfers to addresses other than Base+14 and Base+18

8.2 Bad Data from the D32 Ports

If your program does not crash, but receives “crazy” data (e.g., illegal BCD, etc.) from the D32 ports your problem may be:

- You are using an address that maps into memory or into a different VMEbus device.
- You are using a base address that maps into both the TPRO-VME and another VMEbus device.
- You are accessing high order bits at Base+14 before freezing & reading low order bits at Base+18.
- You forgot that unused bits may be “0” or “1”
- You are reading data (especially if there are many zeros) before the TPRO-VME has synchronized to the input code. Remember that there is an approximate 20-second delay from power on before the TPRO-VME jam syncs to the input code.
- You issued a TIME SET command sequence with invalid values.

8.3 Bad Data from the FIFO Port

If your program does not crash, but has invalid data (e.g., illegal BCD, etc.) from the FIFO your problem may be:

- Your code does not check the FIFO OUTPUT READY bit in the STATUS REGISTER for “0” before reading each byte from the FIFO. If your data contains “FF” or seems to slip (hours show up where you expected minutes, etc.) this is very likely the reason.
- You are reading data (especially if there are many zeros) before the TPRO-VME has synchronized to the input code. Remember that there is about a 20-second delay from power on before the TPRO-VME jam syncs to the input code.
- You performed a TIME SET command sequence with invalid values.
- You are using an address that maps into memory or into a different VMEbus device.

- You are using a base address that maps into both the TPRO-VME and another VMEbus device.

8.4 Interrupt Crashes

- Check that the configured IRQ level matches the level programmed into bus interrupter.
- Check that there are no empty slots in the IACKIN/IACKOUT daisy chain between the INTERRUPT CONTROLLER (usually in the CPU board) and the slot occupied by the TPRO-VME.
- Check that the host interrupt vector is correctly initialized
- Check that you are disabling INT3 while reading the FIFO

8.5 Board Never Syncs to Input Code or 1PPS

If the “loss of sync” status bit is “0” (indicating an error), loss of sync should be asserted in many cases. Its assertion does not mean that there is a fault in the board.

- If the loss-of-signal LED (red) is on, the TPRO-VME does not see enough signal amplitude. Check signal amplitude and connections.
- Check that both the “+” (signal) and “–” side of the input code are connected. Remember that the input is differential for common mode noise rejection.
- If using 1PPS, is there a good 1PPS? Has a SET NEXT 1PPS TIME command sequence been performed?
- Is the time code carrier frequency stable to ± 100 PPM? Does it make periodic large ($> 5\mu\text{S}$) time jumps? It is very likely that tape playback will have high frequency error unless a calibrated servo track has been used to accurately control speed.

8.6 Before You Contact Spectracom

- Try to exhibit the problem in as reduced (fewest boards in the system) a configuration as you can.
- Run Spectracom examples, modified as little as possible, to ascertain that your problem is not a programming error.
- Have the board in front of you when you call. If this is not possible, have the serial number (marked on the board) and firmware version (marked on the EPROM U44) ready.
- Be prepared to answer questions about your host CPU and its address windows—especially if the product had been working when using a different CPU.
- Be prepared to interact with the board using a debugger running on your host. Be familiar with its command and syntax.

9 Options and Accessories

9.1 Common Options

9.1.1.1 -D

9-digit LED Display (days through seconds) front panel .3" digit height

9.1.1.2 -32P2

Connector option—eliminates 54-bit output on VME P2 connector

9.2 Options—TPRO-VME

9.2.1.1 -MX5

Sync to 1 PPS Output

9.2.1.2 -MJ5

1 PPS input on J5 connector

9.3 Options—TSAT-VME

9.3.1.1 TRIM-CAB-D-D-100

This cable acts as an extension cord for a board that is using the Trimble GPS Receiver. It consists of a 100' cable with DB-15 connectors (one male, one female) on the ends. It connects to a board on one end, and to the standard TRIM-CAB-STD cable on the other end. It does not connect directly to the Trimble GPS Receiver.

9.3.1.2 GPS Optic Isolator

The GPS Optic Isolator system combines a GPS receiver/antenna, a fibre optic transmitter, a fibre optic receiver, and a standard Spectracom GPS timing board. The satellite information that is received via the GPS antenna is passed to a fibre optic transmitter via an extension cable. The fibre optic transmitter converts the signal and feeds it to a fibre optic receiver, which then converts the data back and sends it to a standard GPS timing board that can be controlled via a graphical user interface on standard PC. All of this is possible while the GPS Receiver and the actual timing board are up to 500 meters away.

10Driver Support

Please contact your sales representative for information about Spectracom's bus-level timing board driver support for Windows, Linux, VxWorks, and a variety of other platforms. You may also visit our website at www.spectracomcorp.com to download datasheets and manuals.

REVISION HISTORY

[illegible]

Spectracom Corporation

95 Methodist Hill Drive

Rochester, NY 14623

www.spectracomcorp.com

Phone: US +1.585.321.5800

Fax: US +1.585.321.5219