User Manual

Model 9100 (single-channel) Model 9200 (dual-channel)

HIGH VOLTAGE WIDEBAND AMPLIFIER

Publication No. 140803

Tabor Electronics Ltd.

Tabor Electronics Ltd.P.O. Box 404, Tel Hanan Israel 20302
Tel: +972-4-821-3393, FAX: +972-4-821-3388

PUBLICATION DATE: August 14, 2003

Copyright 2003 by Tabor Electronics Ltd. Printed in Israel. All rights reserved. This book or parts thereof may not be reproduced in any form without written permission of the publisher.

WARRANTY STATEMENT

Products sold by Tabor Electronics Ltd. are warranted to be free from defects in workmanship or materials. Tabor Electronics Ltd. will, at its option, either repair or replace any hardware products which prove to be defective during the warranty period. You are a valued customer. Our mission is to make any necessary repairs in a reliable and timely manner.

Duration of Warranty

The warranty period for this Tabor Electronics Ltd. hardware is three years, except software and firmware products designed for use with Tabor Electronics Ltd. Hardware is warranted not to fail to execute its programming instructions due to defect in materials or workmanship for a period of ninety (90) days from the date of delivery to the initial end user.

Return of Product

Authorization is required from Tabor Electronics before you send us your product for service or calibration. Call your nearest Tabor Electronics support facility. A list is located on the last page of this manual. If you are unsure where to call, contact Tabor Electronics Ltd. Tel Hanan, Israel at 972-4-821-3393 or via fax at 972-4-821-3388. We can be reached at: tabor@netvision.net.il.

Limitation of Warranty

Tabor Electronics Ltd. shall be released from all obligations under this warranty in the event repairs or modifications are made by persons other than authorized Tabor Electronics service personnel or without the written consent of Tabor Electronics.

Tabor Electronics Ltd. expressly disclaims any liability to its customers, dealers and representatives and to users of its product, and to any other person or persons, for special or consequential damages of any kind and from any cause whatsoever arising out of or in any way connected with the manufacture, sale, handling, repair, maintenance, replacement or use of said products.

Representations and warranties made by any person including dealers and representatives of Tabor Electronics Ltd., which are inconsistent or in conflict with the terms of this warranty (including but not limited to the limitations of the liability of Tabor Electronics Ltd. as set forth above), shall not be binding upon Tabor Electronics Ltd. unless reduced to writing and approved by an officer of Tabor Electronics Ltd.

Except as stated above, Tabor Electronics Ltd. makes no warranty, express or implied (either in fact or by operation of law), statutory or otherwise; and except to the extent stated above, Tabor Electronics Ltd. shall have no liability under any warranty, express or implied (either in fact or by operation of law), statutory or otherwise.

PROPRIETARY NOTICE

This document and the technical data herein disclosed, are proprietary to Tabor Electronics, and shall not, without express written permission of Tabor Electronics, be used, in whole or in part to solicit quotations from a competitive source or used for manufacture by anyone other than Tabor Electronics. The information herein has been developed at private expense, and may only be used for operation and maintenance reference purposes or for purposes of engineering evaluation and incorporation into technical specifications and other documents, which specify procurement of products from Tabor Electronics.

FOR YOUR SAFETY

Before undertaking any troubleshooting, maintenance or exploratory procedure, read carefully the **WARNINGS** and **CAUTION** notices.





This equipment contains voltage hazardous to human life and safety, and is capable of inflicting personal injury.



If this instrument is to be powered from the AC line (mains) through an autotransformer, ensure the common connector is connected to the neutral (earth pole) of the power supply.



Before operating the unit, ensure the conductor (green wire) is connected to the ground (earth) conductor of the power outlet. Do not use a two-conductor extension cord or a three-prong/two-prong adapter. This will defeat the protective feature of the third conductor in the power cord.



Maintenance and calibration procedures sometimes call for operation of the unit with power applied and protective covers removed. Read the procedures and heed warnings to avoid "live" circuit points.

Before operating this instrument:

- 1. Ensure the proper fuse is in place for the power source to operate.
- 2. Ensure all other devices connected to or in proximity to this instrument are properly grounded or connected to the protective third-wire earth ground.

If the instrument:

- fails to operate satisfactorily
- shows visible damage
- has been stored under unfavorable conditions
- has sustained stress

Do not operate until, performance is checked by qualified personnel.

Table of Contents

Chapter 1 - PORTRAYAL	
What's in This Chapter	1-1
Introduction	1-1
Conventions Used in this Manual	1-3
9200 Feature Highlights	1-4
Functional Description	1-4
Options	1-4
Specifications	1-5
Front Panel Connectors	1-5
Input	1-5
Output	1-5
Grounding Considerations	1-6
Operating Instructions	1-6
Chapter 2 - INSTALLATION	
Installation Overview	2-1
Unpacking and Initial Inspection	2-1
Safety Precautions	2-1
Operating Environment	2-2
Power Requirements	2-2
Grounding Requirements	2-3
Floating the Input/Output Grounds	2-3
Calibration	2-5
Abnormal Conditions	2-5
Long Term Storage or Repackaging for Shipment	2-5
Preparation for Use	2-6
Bench Installation	2-6
Rack Mounting	2-6
Appendix A - SPECIFICATIONS	A-1

This page intentionally left blank

What's In This Chapter

This chapter contains general and functional description of the Models 9100 and 9200 – high voltage wideband amplifier series. It also describes the front panel connectors and operational modes and provides description of all features available with the instrument.

Introduction

The Model 9100 is shown in Figure 1-1 and the 9200 is shown in Figure 1-2. Both instruments are built on the same platform and share specifications and features except the 9100 is a single channel version and while the 9200 has two channel.



This manual is common to both the 9100 and 9200. Description is given for the Model 9200 only. Ignore references to the second channel if you use model 9100.

The 9200 is a 2U, $\frac{1}{2}$ rack size, bench-top power amplifier. The amplifier is used for general-purpose signal amplification purpose. Offering unprecedented signal purity, the 9200 amplifies signals from dc to over 500 KHz. The unit has a fixed gain of x15 however; one may order the same amplifier with custom gain without jeopardizing signal purity and amplifier performance.

The 9200 is intended to operate as an amplifying buffer for signals such as available from waveform, function, or pulse generators. Most of these generators can produce signals that are limited to 20Vp-p into high impedance; The 9200 can be used to convert these voltages to levels as high as 300Vp-p.

The amplifier has a drive capability of over 100mA from a 0.1Ω source. While the output can drive small capacitive or inductive loads, for full high speed potential it is recommended that the load characteristics should be mainly resistive. Load capacitance and inductance up to $0.01\mu F$ and 0.5 mH, respectively, can be tolerated without performance deterioration.

The 9200 is supplied with floating input and output connectors. Therefore, the signal can float from ground level up to $\pm 250 \text{Vdc}$. The only limiting factor is that both input and output grounds must connect to the same level. This capability is extremely important in applications where the amplifying device should reside on the same ground level as its source. The floating capability can be added or removed using a simple, user-accessible, jumper connection.

The amplifier case was designed to stack on top or below other Tabor products. It also can be mounted next to a Tabor generator in a standard 19" rack. The waveform-amplifier combination is a perfect fit for almost every high-voltage, wide bandwidth application.



Figure 1-1. Model 9100 – Single channel High Voltage Amplifier



Figure 1-2. Model 9200 – Dual Channel high Voltage Amplifier

Conventions Used in this Manual

The following conventions may appear in this manual:



A Note contains information relating to the use of this product



A Caution contains information that should be followed to avoid personal damage to the instrument or the equipment connected to it.



A Warning alerts you to a potential hazard. Failure to adhere to the statement in a WARNING message could result in personal injury.

9200 Feature Highlights

- High voltage output to 300Vp-p (±150V)
- Output current to 100 mA per channel (150 mA in 9100)
- Wideband range to over 200 KHz
- Slew rate to 200V/μs
- Low distortion
- Low cost
- Custom configuration of:

Gain Signal Ground

Functional Description

Detailed functional description of the features, operation and options available with the 9200 is given in the following paragraphs. The high voltage amplifiers can be ordered with different configurations such as output impedance, gain, etc. therefore read the following description carefully and make sure your amplifier is configured correctly for your application before you start operating the amplifier.

Options

Model 9200 must be ordered from the factory already configured for your application. There will be no schematics, nor instructions how to modify the amplifier for other configurations as any configuration change may affect the amplifier performance.

Below, you'll find a list of optional configurations for the wideband amplifier:

Gain – specifies gain magnitude of the input signal. Factory default setting is 15 however, any custom gain from 10 to 20 can be specified. Note that some characteristics of the output may change for gain setting above 15.

Circuit Ground – determines if the amplifier ground is floated or tied to case ground. Note that this is the only option that is user accessible and may be changed by adding or removing two jumper connection.

As explained above, all options must be specified at the time of your purchase and the 9200 is supplied fully configured.

Specifications

Instrument specifications are listed in Appendix A. These specifications are the performance standards or limits against which the instrument is tested. Specifications apply under the following conditions: output terminated into matching impedance, after 30 minutes of warm up time, and within a temperature range of 20°C to 30°C. Specifications outside this range are degraded by 0.1% per °C.

Front Panel Connectors

The 9200 has 4 BNC connectors on its front panel, designated as INPUT and OUTPUT. These connectors are described below.

Input

The input connector accepts signals within the range of DC to over 500 KHz and amplifies them by a fixed gain. Input impedance is $1M\Omega$

The amplifier input can not tolerate high voltage therefore, before applying the cable to the input connector, make sure your signal does not exceed input rating, as specified in Appendix A of this manual.

Output

The output connector outputs amplified signals. The output connector may generate inverted signals, depending on the amplifier configuration. Gain at this output is fixed at 15 (or another gain factor that was specified at the time of your purchase). Output source impedance is 0.1Ω .



WARNING

Applying the output signal on inductive or capacitive loads may damage the amplifier.

Grounding Considerations

Understanding how to connect your ground path could be critical to preserving the integrity of your output signal. If you are using a single-ended output then it will probably be safe for you to connect the circuit ground to case ground. However, in applications requiring floated ground connection, it is imperative that the amplifier ground be made floating as well. In this case, refer to Chapter 2 and Figure 2-1 for instructions how to float the circuit ground. Always bear in mind the following warning:



WARNING

Input and output grounds are tied together and therefore, it is absolutely forbidden to connect the output ground to a different level than the input ground. Failure to adhere to this limitation may damage the TE3322 and the surrounding equipment connected to its I/O connectors.

Also note:



Normal amplifier operation is recommended with circuit ground connected to case ground (float disabled), thus minimizing susceptibility to system noise. Enable float configuration for floated application only.

Operating Instructions

Being a passive device, there are no controls, nor computer programming required to operate the 9200. The following procedure is recommended for proper operation of the high voltage power amplifier:


- 1. Make sure your amplifier is configured for input and output impedance, gain and grounding
- 2. Follow the installation instructions given in Chapter 2 of this manual

- 3. Connect the output terminal to your load
- 4. Connect the input terminal to your source
- 5. Turn on power to your 9200



WARNING

There is no switch control to tunr amplification on and off and therefore, the amplifier is active immediately after you power it up. Always make sure your load is protected from inadvertent power up conditions before you turn on your 9200.



Chapter 2 INSTALLATION

Installation Overview

This chapter contains information and instructions necessary to prepare the 9200 for operation. Details are provided for initial inspection, grounding requirements, repackaging instructions for storage or shipment and installation information.

Unpacking and Initial Inspection

Unpacking and handling of the generator requires normal precautions and procedures applicable to handling of sensitive electronic equipment. The contents of all shipping containers should be checked for included accessories and certified against the packing slip to determine that the shipment is complete.

Safety Precautions

The following safety precautions should be observed before using this product and associated computer. Although some instruments and accessories would normally be used with non-hazardous voltages, there are situations where hazardous conditions may be present.

This product is intended for use by qualified persons who recognize shock hazards and are familiar with the safety precautions required to avoid possible injury. The following sections contain information and cautions that must be observed to keep the 9200 operating in a correct and safe condition.



CAUTION

For maximum safety, do not touch the product, test cables, or any other instrument parts while power is applied to the circuit under test. ALWAYS remove power from the entire test system before connecting cables or jumpers. Do not touch any object that could provide a current path to the common side of the circuit under test or power line (earth) ground. Always keep your hands dry while handling the instrument.

Operating Environment

The 9200 is intended for indoor use and should be operated in a clean, dry environment with an ambient temperature within the range of 0 °C to 40 °C.



WARNING

The 9200 must not be operated in explosive, dusty, or wet atmospheres. Avoid installation of the module close to strong magnetic fields.

The design of the 9200 has been verified to conform to EN 61010-1 safety standard per the following limits: Installation (Overvoltage) Category I (Measuring terminals) Pollution Degree 2

Pollution Degree 2 refers to an operating environment where normally only dry non-conductive pollution occurs. Occasionally a temporary conductivity caused by condensation must be expected

Power Requirements

The 9200 operates from one of the following nominal ac sources: 100V, 115V, or 230V ac. Voltage selection is done using the rearpanel power line selector switch. The instrument operates over the power mains frequency range of 47 to 63 Hz. Always verify that the operating power mains voltage is the same as that specified on the rear panel voltage selector switch.

The 9200 is supplied with the correct power line setting. If this setting needs to be change, use a flat-head screwdriver to set the Line Selector switch on the rear panel to the required position.

The instrument is not intended for operation from two phases of a multi-phase ac system or across the legs of a single-phase, three-wire ac power system. Crest factor (ratio of peak voltage to rms.) should be typically within the range of 1.3 to 1.6 at 10% of the nominal rms mains voltage.



WARNING

DO not connect the line cord to the 9200 before you verify the correct power line setting. Failure to switch the instrument to match the operating line voltage will damage the instrument and may void the warranty.

Grounding Requirements

To conform to the applicable safety and EMC requirements, ensure that the 9200 instrument panel and the PC chassis is "earth" grounded.



CAUTION

The outer shells of the front panel terminals (Input, Output) can float from case ground. Refer to Figures 2-1, 2-2 and the instructions in this manual to disconnect/connect the circuit ground from/to case ground.

Floating the Input/Output Grounds

The 9200 ground circuit is designed so it can float from case ground. The only limitation is that the input and output grounds must reside on the same ground level.

Looking at the front panel, you will notice that the BNC connectors are housed in plastic material and therefore are isolated from the front panel metal. The amplifier circuit is also floated from the backplan power supply ground through isolating transformer and therefore the amplifier ground circuit can be configured to float from case ground in applications requiring isolation from circuit to case grounds.



WARNING

Input and output grounds are tied together and therefore, it is absolutely forbidden to connect the output ground to a different level than the input ground. Failure to adhere to this limitation may damage the amplifier and the surrounding equipment connected to its I/O connectors.

The 9200 can be ordered from the factory already configured with the required ground setting however, ground can be configured differently for various applications using simple jumper connections. Use the following procedure to enable/disable floated ground configuration.

- 1. Locate LK1 at the left side of the board, as shown in Figure 2-1
- 2. Locate LK2 at the right side of the board, as shown in Figure 2-2
- 3. Remove both shorting jumpers if you want to float the input/output grounds
- 4. Leave the jumpers on LK1/LK2 if you want to connect circuit ground to case ground



Normal amplifier operation is recommended with circuit ground connected to case ground (float disabled), thus minimizing susceptibility to system noise. Enable float configuration for floated application only.




Figure 2-1. Float Enable/Disable Jumper LK1



Figure 2-2. Float Enable/Disable Jumper LK2

Calibration

The recommended calibration interval is three years. Calibration should be performed by qualified personnel only.

Abnormal Conditions

Operate the 9200 only as intended by the manufacturer. If you suspect the product has been impaired, remove the power cord and secure against any unintended operation. The 9200 protection is likely to be impaired if, for example, the instrument fails to perform the intended operation or shows visible damage.



WARNING

Any use of the 9200 in a manner not specified by the manufacturer may impair the protection provided by the instrument

Long Term Storage or Repackaging For Shipment

If the instrument is to be stored for a long period of time or shipped immediately, proceed as directed below. If you have any questions, contact your local Tabor representative or the Tabor Customer Service Department.

- 1. Repack the instrument using the wrappings, packing material and accessories originally shipped with the unit. If the original container is not available, purchase replacement materials.
- 2. Be sure the carton is well sealed with strong tape or metal straps.
- 3. Mark the carton with the model and serial number. If it is to be shipped, show sending and return address on two sides of the box.

4.



NOTE

If the instrument is to be shipped to Tabor for calibration or repair, attach a tag to the instrument identifying the owner. Note the problem, symptoms, and service or repair desired. Record the model and serial number of the instrument. Show the returned authorization order number (RMA) as well as the date and method of shipment. ALWAYS OBTAIN A RETURN AUTHORIZATION NUMBER FROM THE FACTORY BEFORE SHIPPING THE INSTRUMENT TO TABOR.

Preparation For Use

Preparation for use includes removing the instrument from the box, the bag and installing the 9200 either on the beck or in a 19" rack.

Bench Installation

The 9200 is fully solid state and dissipates only a small amount of power. No special cooling is required. However, the instrument should not be operated where the ambient temperature exceeds 50 °C, when the relative humidity exceeds 80% or condensation appears anywhere on the instrument. Avoid operating the instrument close to strong magnetic fields, which may be found near high power equipment such as motors, pumps, solenoids, or high power cables. Use care when rack mounting to locate the instrument away from sources of excessive heat or magnetic fields. Always leave 4 cm (1.5 inches) of ventilation space on all sides of the instrument.

Rack Mounting

The 9200 can be rack mounted inside a standard 19" rack. It can be mounted using one of two configurations: 1) single, or 2) side-by-side. Tabor offers rack-mounting ears for both options. Consult the factory for the appropriate part number.

Using the single rack mounting option, the 9200 is supplied with a blank panel that covers the empty side. The case can be mounted either on the left or the right of the rack with the blank panel covering the empty space.

Side-by-side option is available only with other Tabor products. In this case, the two boxes are latched in the middle and side ears connect the assembly to the rack. Figure 2-3 shows how the 9200 can be mounted next to a dual-channel arb – Model 8026.




Figure 2-3. Dual Rack Mounting Option

Appendix A

SPECIFICATIONS

Channels

9100 1 9200 2

Input Characteristics

Connector BNC

Impedance $1M\Omega$, DC coupled

Damage Level 50Vp-p

Frequency Range DC to 500KHz

Output Characteristics

General

Connector BNC

Impedance 0.1Ω , DC coupled Protection Short-circuit, 10 seconds

Gain x15, fixed (can be ordered from factory with different gain setting)

Polarity Output normal

Amplitude 0 to 300Vp-p (± 150 V)

Square Wave Characteristics

Transition Time <1.5μs
Aberrations <15%

Sine Wave Characteristics

Small Signal

Bandwidth 1MHz, at 20Vp-p (-3dB)

Large Signal

Bandwidth 500KHz, at 300Vp-p (-3dB)

Accuracy $\pm (2\% \text{ of full-scale amplitude range} + 25\text{mV})$, Square wave at 1KHz

THD <0.1%, 10Hz to 10KHz

<1.2%, 10KHz to 200KHz

General

Physical Size Single-slot, 2U high, half rack size metal case

Power Requirements 100V/115V/230V, 47-63Hz, <60W

Signal Ground Floated to the same level as the source, 250V dc maximum

EMC Certification CE marked

Reliability MTBF per MIL-HDBK-217E, 25 °C, Ground Benign Safety Designed to meet IEC EN61010-1, UL 3111-1


Workmanship Std. Conform to IPC-A-610D

Environmental

Operating Temperature 0 °C - 50 °C, RH 80% (non-condensing)

Storage Temperature -30 °C - 80 °C

Ordering Information



^(*) Custom gain from x10 to x20 can be ordered however, bandwidth cannot be maintained. Consult the factory before ordering gain above 15.