

**SA2600 Spectrum Analyzer  
Option EP1 Software Upgrade  
Instructions**

[www.tektronix.com](http://www.tektronix.com)



075-0985-00

**Tektronix**

Copyright © Tektronix. All rights reserved. Licensed software products are owned by Tektronix or its subsidiaries or suppliers, and are protected by national copyright laws and international treaty provisions.

Tektronix products are covered by U.S. and foreign patents, issued and pending. Information in this publication supersedes that in all previously published material. Specifications and price change privileges reserved.

TEKTRONIX and TEK are registered trademarks of Tektronix, Inc.

DPX is a registered trademark of Tektronix, Inc.

## **Contacting Tektronix**

Tektronix, Inc.  
14200 SW Karl Braun Drive  
P.O. Box 500  
Beaverton, OR 97077  
USA

For product information, sales, service, and technical support:

- In North America, call 1-800-833-9200.
- Worldwide, visit [www.tektronix.com](http://www.tektronix.com) to find contacts in your area.

---

# SA2600 Option EP1 Software Upgrade Kit

This document explains how to upgrade the SA2600 Spectrum Analyzer to support the Option EP1 DPX Enhancement feature.

## Instruments

**SA2600:** all

## Minimum Tools and Equipment List

<b>Required tools and equipment</b>	<b>Part number</b>
Microsoft Windows-based PC with Web access to download a 4.5 MB file	NA
Tektronix SA2600 Spectrum Analyzer instrument	SA2600
10 MB or greater USB flash drive or CompactFlash memory card with PCMCIA adapter	NA

## Kit Parts List

<b>Quantity</b>	<b>Part number</b>	<b>Description</b>
1 ea.	075-0985-XX	KIT INSTRUCTION; SOFTWARE OPTION UPGRADE, SA2600 OPTION EP1
1 ea.		UPGRADE LABEL SHEET (NEW OPTION KEY)

## Install the Option Software

1. Open your preferred browser and go to **www.tektronix.com/software**.
2. Enter **SA2600** in the **Search by keyword** field and click **Go**.
3. Click the **SA2600 Software Upgrade Archive File** link. If there are multiple versions listed, select the link whose description has the most recent release date or the highest version number.
4. On the **Software Upgrade Archive File** page, click **Download File** and save the file to your PC.
5. On the PC, unzip the ZIP file. The ZIP archive file contains a readme file and an ARQ file.
6. Open the readme file and follow the instructions on how to copy the ARQ file to the memory device and install the software upgrade on your instrument.
7. When the software upgrade installation is complete, remove and discard the existing Software Option Key label (that contains the software key number/letter sequence) from the back of the instrument (on the back of the battery compartment).

---

**NOTE.** Do not remove the name/serial number/barcode label that is on the bottom of the instrument.

---

8. Remove the new Software Option Key label from the label sheet and place it on the back of the instrument, where the original label was positioned (on the back of the battery compartment).
9. Discard the label sheet.

## Verify the Option Software

1. Start the SA2600 application.
  2. Tap **Tools > Upgrade Software Options**.
  3. Verify that **Option EP1** is listed and the check box is checked. If Option EP1 is not listed, repeat the upgrade procedure. If Option EP1 is not listed after the second upgrade attempt, contact a Tektronix Customer support representative.
  4. Tap **Cancel** to close the dialog box.
- End of document •