

**Read This First**

**Tektronix**

**P6330/P7330  
Differential Probes  
Accessory Kit Changes**

**001-1337-00**

**[www.tektronix.com](http://www.tektronix.com)**



**001133700**

# Read This First

Occasionally, design changes must be made to improve the reliability of our products. This sometimes requires changes to the product or accessories that are described in the Instruction Manual.


## Accessory Kit Changes

Due to design changes, probe tips are no longer replaceable. If your probe tips need replacement, contact Tektronix for service. The following accessories that are described in the manual are no longer included with your accessory kit:

- Replaceable Probe Tip, 2 each, Tektronix P/N 016-1883-01
- Square Pin Adapter, 4 each, Tektronix P/N 016-1884-00
- Y-Lead Adapter, 2 each, Tektronix P/N 196-3468-00
- TwinFoot Adapter, 4 each, Tektronix P/N 016-1785-00

Use the adapter below (included in the accessory kit), to complete connections where the square pin adapter is used in the manual.

- Adapter, 1 each, Tektronix part number 196-3473-00



### Alternate connection method

If you want to use the square pin and Y-lead adapters as specified in the manual procedures, you can obtain prototype samples by contacting Tektronix at:

[www.quality.comments@tek.com](mailto:www.quality.comments@tek.com), or 1-800-833-9200, ext. 7-1399

Include your name, shipping address, phone number or e-mail address, and product type and serial number with all correspondence.