

NOTE REGARDING FACTORY CALIBRATION PROCEDURES

AND TEST SPECIFICATIONS

Factory Calibration Procedures and Test Specifications are intended for use at the factory as a general guide for calibrators and quality control men. Most of the tolerances listed in these sheets are closer than advertised specifications. This is done purposely in order to insure that the instrument will meet or exceed advertised specifications when it reaches the customer.

These calibration procedures and test specifications should be used, therefore, as a guide only.

Some of the test equipment referred to in the calibration procedures is not available commercially; the Tektronix field engineer will be glad to suggest alternate approaches.

FACTORY CALIBRATION PROCEDURE

CONTENTS:

General
Circuit Specifications
Calibration
Standard Checks
Special Test Equipment

GENERAL:

This isn't a field recalibration procedure as is the procedure in your instruction manual. This is a guide in calibrating brand-new instruments, just assembled instruments that have never been turned on before. Therefore it calls out many procedures and adjustments that are rarely required for subsequent recalibration.

Even though we wrote this procedure primarily for our own factory test department, it's valuable to others also if used with some caution:

1. **Special test equipment**, if mentioned, is not available from Tektronix unless it's listed also in our current catalog. This special equipment is used in our test department to speed calibration. Usually you can either duplicate its function with standard equipment in your facility, devise alternate approaches, or build the special test equipment yourself.
2. **Factory circuit specifications** are not guaranteed unless they also appear as catalog or instruction manual specifications. Factory circuit specs usually are tighter than advertised specs. This helps insure the instrument will meet or exceed advertised specs after shipment and during subsequent field recalibrations over several years of use. Your instrument may not meet factory circuit specs but should meet catalog or instruction manual specs.
3. **Presetting controls**, if mentioned, usually is unnecessary. This is helpful for "first-time" calibration only. If internal controls are preset, you'll have to perform a 100% recalibration. So don't preset controls unless you're certain a "start-from-scratch" policy is the best.
4. **Quality control men steps**. Factory calibration procedures are for our test department calibrators who first calibrate the instrument. Quality control men then check the initial calibration and perform additional fine points such as trimming resistor leads, installing shields, etc. In some cases a factory calibration procedure instructs the calibrator not to perform these fine points. You'll ordinarily have to include these fine points in your calibration.

In this procedure, all front panel controls are in capital letters (SENSITIVITY) and internal adjustments are capitalized only (Gain Adj).

Publication:
061-112
March 1962

For all serial numbers.

CA



Copyright 1962 by Tektronix, Inc., P. O. Box 500,
Beaverton, Oregon. Printed in the United States of
America. All rights reserved. Contents of this pub-
lication may not be reproduced without permission
of the copyright owner.

FACTORY CIRCUIT SPECIFICATIONS

Factory circuit specifications are not guaranteed unless they also appear as catalog or instruction manual specifications. Factory circuit specs usually are tighter than advertised specs. This helps insure the instrument will meet or exceed advertised specs after shipment and during subsequent field recalibrations over several years of use. Your instrument may not meet factory circuit specs but should meet catalog or instruction manual specs.

The numbers listed beside the specifications are the factory calibration procedure steps where the check or adjustment is made.

BALANCE:

- 7a. Output V4383: electrical center ± 2 cm.
- 8c. DC BAL: final adjust within $\pm 90^\circ$ of mr.

GAS AND MICROPHONICS:

- 9a. Gas: 2 mm max.
- 9b. Microphonics: no ringing type.

VERTICAL POSITION:

- 10b. Traces within 3 cm of each other and each within ± 2 cm of electrical center.
- 10b. Vert Pos Range R4376: final adjust within $\pm 90^\circ$ of mr.
- 11a. Trace shift between NORMAL and INVERTED: ± 1 cm max; slow drift: ± 2 mm max.
- 11c. Trace shift between A or B ONLY and ADDED ALGEBRAICALLY: ± 2 cm max.

OUTPUT DC LEVEL:

- 12a. 67.5 v ± 2.5 v.

DUAL TRACE:

- 13a. Alternate: must alternate on all sweep rates between .1 sec and .1 μ sec/cm.

Chopped:

- 14a. frequency: 100 kc $\pm 20\%$.
- 14a. symmetry: ratio 3 to 2 max.
- 14a. flat top: distortion .5 mm max.

GAIN:

GAIN ADJ:

- 15b. range: $\pm 10\%$.
- 15b. final adjust within $\pm 90^\circ$ of mr.
- 17a. ADDED accuracy: $\pm 2\%$.
- 17b. Common mode rejection: 20 to 1.
- 26b. dc to 24 mc.
- 18a. VOLTS/CM accuracy: $\pm 2\%$.
- 18a. VARIABLE range: 2.5 to 1.

FREQUENCY RESPONSE:

- 25b. DC: dc to 24 mc (-3 db point).
- AC: 2 cps (-3 db point) to 24 mc (-3 db point).

4. RESISTANCE CHECKS:

a. Check resistance to ground:

use	amphenol connector pin	resistance (approx)
output	1	8.5 k
gnd	2	0
output	3	8.5 k
not used	4 to 7	inf
alt trig	8	0
-150 v	9	12 k
+100 v	10	2 k
+225 v	11	6.3 k
+350 v	12	inf
6.3 vac	13	inf
6.3 vac	14	inf
+75 v fil	15	65 Ω
alt trig	16	*320 k

*inf below sn 34790

5. PLUG-IN SCOPE PRESETS:

- a. Sweep: 1 millisecc/cm.
Trigger: +int, ac, stability cw.
- b. Determine plug-in scope's vertical amplifier electrical center by either:
1. Inserting a test load unit, pushing PRESS TO SHORT INPUT button and noting trace vertical position, or
 2. Shorting amphenol connector pins 1 and 3 together and noting trace vertical position.

6. SETUP:

- a. Use a plug-in extension (013-019) and plug CA Unit into plug-in scope. Turn power on.

7. OUTPUT BALANCE:

- a. Connect a jumper between CA Unit output of V4383 12AT7 pins 2 and 7. Select V4383 so trace falls within ± 2 cm of plug-in scope's electrical center (step 5b). Remove jumper.

8. PRELIMINARY DC BALANCE:

- a. Move A trace to plug-in scope's electrical center with A VERTICAL POSITION. Rotate A VARIABLE volts/cm back and forth while adjusting A DC BAL for no trace shift.
- b. Repeat for B.
- c. Both A and B DC BAL adjustments must be within $\pm 90^\circ$ of mr after adjustment. Return VARIABLES cw.

9. GAS AND MICROPHONICS:

- a. Ground A input. Switch A AC-DC back and forth between AC and DC. View trace shift (gas): 2 mm max. Return AC-DC to DC.
- b. Rap lightly on CA Unit front panel. View microphonics: no ringing type. Leave input grounded.
- c. Ground B and repeat for B.

10. VERTICAL POSITION RANGE:

- a. Set both A and B VERTICAL POSITION controls to mr. Adjust Vert Pos Range R4376 so both traces are equidistant above and below plug-in scope's electrical center.
- b. Traces must be within 3 cm of each other and each trace must be within ± 2 cm of scope's electrical center. Vert Pos Range R4376 must be within $\pm 90^\circ$ of mr after adjustment.

11. POLARITY:

- a. Move A trace to scope's electrical center with A VERTICAL POSITION. Switch POLARITY back and forth between NORMAL and INVERTED. View trace shift: ± 1 cm max. View slow drift after initial shift (gas): ± 2 mm max.
- b. Repeat for B.
- c. Return both A and B POLARITY controls to NORMAL. Move both traces to scope's electrical center with VERTICAL POSITION controls. Change MODE to ADDED ALGEBRAICALLY. View trace shift: ± 2 cm max. Return MODE to ALTERNATE.

12. OUTPUT DC LEVEL:

a. Keep both traces to electrical center. Measure voltage between amphenol connector pin 1 and gnd. Read +65 to +70 v. Measure between pin 3 and gnd. Read +65 to +70 v.

13. ALTERNATE MODE:

a. Change VERTICAL POSITION controls so traces are about 2 cm apart. Change scope to .1 sec/cm. View display alternate between A and B traces. Check that display alternates between A and B traces from .1 sec to .1 μ sec/cm.

14. CHOPPED MODE:

a. Keep traces about 2 cm apart. Change MODE to CHOPPED. Change scope to 2 μ sec/cm, triggered. View one cycle of chopped waveform: 4.25 to 6.25 cm duration; waveform top duration to waveform bottom duration ratio (or vice versa) 3 to 2 max; .5 mm max flat-top distortion.

b. Set scope crt cathode selector (scope rear) to dual trace chopped blanking. View rise and fall portions of chopped waveform are blanked. Change crt cathode selector back to normal. View rise and fall portions unblanked.

15. GAIN:

a. Change scope to 1 millisecc/cm. Change MODE to A ONLY. Keep VOLTS/CM at .05 and VARIABLE cw. Apply .1v from standard calibrator (special) to A input.

b. Keep GAIN ADJ cw; view 2.2 cm deflection, min. Change GAIN ADJ ccw; view 1.8 cm deflection, max. Adjust GAIN ADJ for exactly 2 cm deflection. GAIN ADJ must be within $\pm 90^\circ$ of mr after adjustment.

c. Change MODE to B ONLY, apply calibrator to B input, and repeat for B.

16. AC-DC:

a. Keep MODE to B ONLY and keep calibrator applied to B input. Move bottom of calibrator waveform to center graticule line with B VERTICAL POSITION. Change B AC-DC to AC. View waveform shifts so it's approximately centered about center graticule line. Return B AC-DC to DC.

b. Change MODE to A ONLY, apply calibrator to A input, and repeat for A.

17. ADDED ALGEBRAICALLY:

a. Apply .1v calibrator to *both* A and B inputs. Change MODE to ADDED ALGEBRAICALLY. Keep both POLARITY controls to NORMAL. View 4 cm deflection, ± 8 mm.

b. Change A (or B) POLARITY to INVERTED. Change calibrator to 1v. View 1cm deflection, max. If over 1cm, vary VERTICAL POSITION controls. If still over 1cm, reverse POLARITY controls and again vary VERTICAL POSITION controls, if necessary. One method must result in 1 cm or less deflection.

18. VOLTS/CM:

a. Change MODE to A ONLY and both POLARITY controls to NORMAL. Keep calibrator applied to both inputs. Check as follows:

VOLTS/ CM	calibrator	deflection VARIABLE	
		cw	ccw
.1	.05 v	*2 cm	.8 cm max
.2	.1	2 +2%	.8
.5	.2	2.5	1
1	.5	2	.8
2	1	2	.8
5	2	2.5	1
10	5	2	.8
20	10	2	.8

*Adjusted, step 15

b. Change MODE to B ONLY and repeat for B. Remove calibrator.

19. A INPUT CAPACITY:

a. Remove plug-in extension and reinsert CA Unit, less extension, into scope. Change MODE to A ONLY and A VOLTS/CM to .05.

b. Connect a 52 Ω termination (011-001) to 105 output. Connect a 52 Ω cable (012-001) to 52 Ω termination and a 20 pf input capacitance standardizer (011-022) to other end of cable. Connect 20 pf standardizer to CA Unit, A input.

c. Set 105 for about 3.5 cm of 1 kc. Adjust C3322, input capacitor, for best square wave.

20. A VOLTS/CM COMPENSATION:

a. Keep 3.5 cm 1 kc 105 deflection and adjust for best square wave as follows:

VOLTS/CM	spike	adjust level
.05		C3322 (adj, step 19)
.1	C3311C	C3311B
.2	12C	12B
.5	13C	13B
1	14C	14B
2	15C	15B
5	16C	16B
10	17C	17B
20	18C	18B

21. B INPUT CAPACITY:

a. Change MODE to B ONLY and B VOLTS/CM to .05. Change 105 (via 20pf standardizer) to B input, keep 3.5 cm of 1kc. Adjust C4322, input capacitor, for best square wave.

22. B VOLTS/CM COMPENSATION:

a. Keep 3.5 cm 1 kc 105 deflection and adjust for best square wave as follows:

VOLTS/CM	spike	adjust level
.05		C4322 (adj, step 21)
.1	C4311C	C4311B
.2	12C	12B
.5	13C	13B
1	14C	14B
2	15C	15B
5	16C	16B
10	17C	17B
20	18C	18B

b. Remove 105.

23. HIGH FREQUENCY:

- a. Connect a $52\ \Omega$ cable to 107 output, connect a $52\ \Omega$ termination to $52\ \Omega$ cable and connect termination to CA Unit A input.
- b. Change MODE to A ONLY and A VOLTS/CM to .1. Set 107 for about 3 cm of approx 500 kc. Adjust L3364, L4374, L3362, L3372, L3334 and L3354 for best square wave with no overshoot.
- c. Change A VOLTS/CM to .05 and check for no square-wave overshoot or roll-off.
- d. Change MODE to B ONLY and B VOLTS/CM to .1. Change 107 to B input. Adjust L4362, L4372, L4334 and L4354 for best square wave with no overshoot.
- e. Change B VOLTS/CM to .05 and check for no square-wave overshoot or roll-off.
- f. NOTE: L3364 and L4374 affect both A and B channels. Adjust for best square wave on both channels.
- g. Remove 107.

24. FINAL DC BALANCE:

- a. Change MODE to ALTERNATE. Move A trace to scope's electrical center with A VERTICAL POSITION. Rotate A VARIABLE volts/cm back and forth while adjusting A DC BAL for no trace shift.
- b. Repeat for B.

25. FREQUENCY RESPONSE:

- a. Change scope to .1 millisecc/cm. Change MODE to A ONLY. Attach a $52\ \Omega$ termination to 190 output and apply exactly 3 cm of 50 kc to A input.
- b. Increase 190 frequency to 24 mc. View 2.1 cm deflection, min.
- c. Change MODE to B ONLY and repeat for B.
- d. Remove 190.

**26. HIGH FREQUENCY COMMON MODE
REJECTION:**

- a. Attach a dual connector (003-035) to CA Unit A and B inputs. Keep 52Ω termination on 190 output and connect 190 to dual connector. Set both VOLTS/CM to .5. Set 190 for 2 cm of 50 kc.
- b. Change both VOLTS/CM to .05, MODE to ADDED ALGEBRAICALLY and A POLARITY to INVERTED. View deflection: 1 cm max. If over 1 cm, change A POLARITY to NORMAL and B POLARITY to INVERTED. One method must result in 1 cm or less deflection.
- c. Change both AC-DC controls to AC. View deflection: 1 cm max.

27. THE END.

