CERTIFICATE OF CALIBRATION

Issued By Transmille Ltd.

Date of Issue 09 December 2008

_____0





Transmille Ltd. Unit 4, Select Business Centre Lodge Road Staplehurst, Kent. TN12 0QW. TEL 01580 890700 FAX 01580 890711

Approved Signatory

EXAMPLE CERTIFICATE

Certificate Number EXAMPLE

Customer:

Date Received : 11 August 2008

Instrument :	System ID : Description : Manufacturer : Model Number : Serial Number : Procedure Version :	EXAMPLE Picoamp Source Adapter Transmille EA013 EXAMPLE 3.00/N
	Procedure Version :	3.00/N

Environmental Conditions

Temperature :20°C +/- 1°CRelative Humidity : 50% +/- 20%

Mains Voltage : 240V +/- 12V Mains Frequency : 50Hz +/- 1Hz

Comments

Instrument was allowed to stabilise for at least 12 hours before calibration.

Calibration Information

The instrument was calibrated against laboratory standards whose values are traceable to recognised National Standards. The uncertainty limits quoted refer to the measured values only, with no account being taken of the instruments ability to maintain its calibration.

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor k=2, providing a level of confidence of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.

Calibrated By : EXAMPLE

Date of Calibration : EXAMPLE

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to recognised national standards, and to the units of measurement realised at the National Physical Laboratory or other recognised national standards laboratories. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

CERTIFICATE OF CALIBRATION

UKAS Accredited Calibration Laboratory No. 0324 AFTER ADJUSTMENT RESULTS

Certificate Number EXAMPLE

Page 2 of 2 Pages

Test Title		Reading	Uncertainties
	Applied Value	Reading	Uncertainties
100uA Range			
100uA Range	100.000uA	100.003uA	1nA
100uA Range	0.000uA	-0.040uA	1nA
100uA Range	-100.000uA	-100.085uA	1nA
10uA Range			
10uA Range	10.000 OuA	10.000 8uA	0.5nA
10uA Range	0.000 OuA	-0.0046uA	0.5nA
10uA Range	-10.0000uA	-10.0099uA	0.5nA
Aut Dawns			
1uA Range	1.000 00uA	1.000 28uA	0.5~4
1uA Range	0.000 00uA	-0.00048uA	0.5nA 0.5nA
1uA Range	-1.00000uA	-1.00125uA	0.5nA
1uA Range	-1.00000uA	-1.001250A	0.5HA
100nA Range			
100nA Range	100.000nA	100.008nA	0.2nA
100nA Range	0.000nA	-0.054nA	0.2nA
100nA Range	-100.000nA	-100.133nA	0.2nA
roon, ritange	100.0001/1	100.1001/1	0.21/7
10nA Range			
10nA Range	10.000nA	9.994nA	0.2nA
10nA Range	0.000nA	-0.009nA	0.2nA
10nA Range	-10.000nA	-10.016nA	0.2nA
5			
Linearity - 100uA Range			
100uA Range	90.000uA	89.993uA	1nA
100uA Range	80.000uA	79.989uA	1nA
100uA Range	70.000uA	69.987uA	1nA
100uA Range	60.000uA	59.981uA	1nA
100uA Range	50.000uA	49.976uA	1nA
100uA Range	40.000uA	39.973uA	1nA
100uA Range	30.000uA	29.969uA	1nA
100uA Range	20.000uA	19.962uA	1nA
100uA Range	10.000uA	9.959uA	1nA
100uA Range	-10.000uA	-10.050uA	1nA
100uA Range	-20.000uA	-20.056uA	1nA
100uA Range	-30.000uA	-30.060uA	1nA
100uA Range	-40.000uA	-40.065uA	1nA
100uA Range	-50.000uA	-50.069uA	1nA
100uA Range	-60.000uA	-60.076uA	1nA
100uA Range	-70.000uA	-70.079uA	1nA
100uA Range	-80.000uA	-80.082uA	1nA
100uA Range	-90.000uA	-90.086uA	1nA

End of results