



Precision Impedance Analyzers 6500B Series

Product Specification

Issue B

UK

Wayne Kerr Electronics
Vinnetrow Business Park
Vinnetrow Road
Chichester
West Sussex PO20 1QH
Tel: +44 (0)1243 792200
Fax: +44 (0)1243 792201
email: sales@wayne-kerr.co.uk
email: service@wayne-kerr.co.uk
www.waynekerrtest.com

© Wayne Kerr Electronics 2008

The copyright in this work is vested in Wayne Kerr Electronics and this document is issued for the purpose only for which it is supplied. No licence is implied for the use of any patented feature. It must not be reproduced in whole or in part, or used for tendering or manufacturing purposes except under an agreement or with the consent in writing of and then only on the condition that this notice is included in any such reproduction. Information furnished is believed to be accurate but no liability in respect of any use of it is accepted by Wayne Kerr Electronics.

CONTENTS

1	SPECIFICATION.....	1-1
1.1	Measurement Parameters	1-1
1.2	AC Functions	1-1
1.2.1	<i>Equivalent Circuit</i>	1-1
1.3	Test Conditions	1-1
1.3.1	<i>AC Drive</i>	1-2
1.3.2	<i>Internal DC Bias /DI (Optional)</i>	1-2
1.3.3	<i>Measurement Speeds</i>	1-2
1.3.4	<i>Measurement Accuracy</i>	1-3
1.3.5	<i>Resistance / Reactance (R / X)</i>	1-3
1.3.6	<i>Conductance / Susceptance (G / B)</i>	1-3
1.3.7	<i>Capacitance (C)</i>	1-3
1.3.8	<i>Inductance (L)</i>	1-3
1.3.9	<i>Dissipation Factor (D)</i>	1-3
1.3.10	<i>Quality Factor (Q)</i>	1-3
1.4	Measurement Ranges	1-3
1.4.1	<i>Hardware Ranges</i>	1-4
1.5	Measurement Connections.....	1-4
1.6	Analysis Mode (Graphical Sweep).....	1-4
1.7	Meter Mode.....	1-4
1.8	Setup Data.....	1-4
1.9	General.....	1-4
1.9.1	<i>Power Supply</i>	1-4
1.9.2	<i>Display</i>	1-5
1.9.3	<i>Printer Output</i>	1-5
1.9.4	<i>Remote Control (Optional)</i>	1-5
1.9.5	<i>Remote Trigger</i>	1-5
1.9.6	<i>Universal Serial Bus (USB)</i>	1-5
1.9.7	<i>VGA External Monitor</i>	1-5
1.9.8	<i>Local Area Network (LAN)</i>	1-5
1.9.9	<i>Mouse</i>	1-5
1.9.10	<i>Keyboard interface</i>	1-5
1.9.11	<i>Mechanical</i>	1-5
1.10	Environmental Conditions	1-6
1.10.1	<i>Temperature Range</i>	1-6
1.10.2	<i>Relative Humidity</i>	1-6
1.10.3	<i>Altitude</i>	1-6
1.10.4	<i>Installation Category</i>	1-6
1.10.5	<i>Pollution Degree</i>	1-6
1.10.6	<i>Safety</i>	1-6
1.10.7	<i>EMC</i>	1-6
2	THEORY REFERENCE.....	2-1
2.1	Abbreviations.....	2-1
2.2	Formulae	2-1

2.3	Series/Parallel Conversions.....	2-2
2.4	Polar Derivations.....	2-2



1 SPECIFICATION

Wayne Kerr Electronics Limited reserves the right to change specification without notice.

1.1 Measurement Parameters

Any of the following parameters can be measured and displayed.

1.2 AC Functions

Capacitance (C), Inductance (L), Resistance (R), Reactance (X), Conductance (G), Susceptance (B), Dissipation Factor (D), Quality Factor (Q), Impedance (Z), Admittance (Y) and Phase Angle (θ).

1.2.1 Equivalent Circuit

Series or Parallel.

1.3 Test Conditions

6505B

20Hz to 5MHz

6510B

20Hz to 10MHz

6515B

20Hz to 15MHz

6520B

20Hz to 20MHz

6530B

20Hz to 30MHz

6550B

20Hz to 50MHz

65120B

20Hz to 120MHz

For all models: Accuracy of set frequency: $\pm 0.005\%$

Frequency step size: $< 1\text{mHz}$



1.3.1 AC Drive

Drive Level (AC Measurements)

Open Circuit Voltage	Short Circuit Current
10mV to 1V rms (< 50 MHz)	200 μ A to 20mA rms (< 50 MHz)
10mV to 0.5V rms (> 50 MHz)	200 μ A to 10mA rms (> 50 MHz)

Signal source impedance: 50 Ω nominal

1.3.2 Internal DC Bias /D1 (Optional)

Current

DC bias 0 to +100mA.

Voltage

DC bias 0 to +40V

1.3.3 Measurement Speeds

Four selectable speeds for all measurement functions. Selecting slower measurement speed increases reading resolution and reduces measurement noise.

Maximum speed.

Fast speed.

Medium speed.

Slow speed.

A custom speed may be set which allows for user defined measurement speeds to be set to enhance noise performance.

Custom speed 1-256

When set to the maximum speed the typical 6500 measurement times are as follows:-

	Meter Mode	Analysis Mode
1kHz	250ms	190ms
10kHz	70ms	70ms
100kHz	70ms	70ms
>1MHz	60ms	60ms



1.3.4 Measurement Accuracy

The accuracy statements given apply when the instrument is used under the following measurement conditions.

AC Drive Level: 1V/20mA

Speed: SLOW

Fixture: 1J1011

Instrument fully trimmed

Temperature range: $23 \pm 5^\circ\text{C}$

30 minute warm up period.

1.3.5 Resistance / Reactance (R / X)

$\pm 0.05\%$ *

1.3.6 Conductance / Susceptance (G / B)

$\pm 0.05\%$ *

1.3.7 Capacitance (C)

$\pm 0.05\%$ *

1.3.8 Inductance (L)

$\pm 0.05\%$ *

1.3.9 Dissipation Factor (D)

$\pm 0.0005 (1+D^2)^*$

1.3.10 Quality Factor (Q)

$\pm 0.05\% (Q+1/Q)^*$

***Varies with frequency, drive level and measured impedance.**

1.4 Measurement Ranges

R, Z, X 0.01m Ω to >2G Ω

G, Y, B 0.01nS to >2kS

L 0.1nH to >2kH

C 1fF to >1F

D 0.00001 to >1000

Q 0.00001 to >1000



1.4.1 Hardware Ranges

The impedance of the Device Under Test and the measurement frequency determine the hardware range used. Auto ranging is available which sets the most appropriate range for a measurement.

Range	Impedance Ω	Frequency Range
1	< 5	Full range
2	<50	Full range
3	>50	Full range
4	>500	Full range
5	>5000	Up to 1MHz
6	>50000	Up to 100kHz
7	>500000	Up to 10kHz

1.5 Measurement Connections

4 front panel BNC connectors permit 3- and 4-terminal connections with the screens at ground potential.

Terminals withstand connection of charged capacitor up to 500V.

1.6 Analysis Mode (Graphical Sweep)

Any two measurement parameters may be selected and measurements made while varying frequency, AC drive level or dc bias (option).

1.7 Meter Mode

Provides a standard LCR meter interface presenting numerical results of single or repetitive measurements. All instrument measurement parameters may be set prior to making measurements.

1.8 Setup Data

Up to 20 instrument setups may be locally stored for each mode.

1.9 General

1.9.1 Power Supply

Input Voltage 90V – 264V AC, auto-ranging

Frequency 47 – 63Hz

VA rating 150VA max

Input fuse rating 2.5AT



1.9.2 Display

8.4" high contrast colour VGA (640 x 480 pixels) TFT module with CPL back lighting.

Touch screen interface.

Visible area 170 x 130mm.

1.9.3 Printer Output

HP-PCL compatible graphics printing

Ethernet direct print

Centronics/parallel printer port, Epson compatible text/ticket printing

1.9.4 Remote Control (Optional)

Designed to GPIB IEEE-488.2 and SCPI 1992.0.

1.9.5 Remote Trigger

Rear panel BNC with internal pull-up, operates on logic low or contact closure.

1.9.6 Universal Serial Bus (USB)

Two Universal Serial Bus Interfaces

USB 1.0 compliant

1.9.7 VGA External Monitor

15-way D-type connector to drive an external monitor in addition to the instrument display.

1.9.8 Local Area Network (LAN)

10/100-BASET Ethernet controller.

RJ45 connector

1.9.9 Mouse

Standard USB or PS/2 mouse port. Touch screen remains enabled when the mouse is connected.

1.9.10 Keyboard interface

Standard USB or PS/2 keyboard port. Instrument front panel remains active with keyboard plugged in

1.9.11 Mechanical

Height 190mm (7.5")

Width 440mm (17.37")



Depth 525mm (20.7")
Weight 13.8kg (30.4 lbs)

1.10 Environmental Conditions

This equipment is intended for indoor use only in a non-explosive and non-corrosive atmosphere.

1.10.1 Temperature Range

Storage: -20°C to +60°C.
Operating: 0°C to 40°C.
Normal accuracy: 18°C to 28°C.

1.10.2 Relative Humidity

Up to 80% non-condensing.

1.10.3 Altitude

Up to 2000m.

1.10.4 Installation Category

II in accordance with IEC664.

1.10.5 Pollution Degree

2 (mainly non-conductive)

1.10.6 Safety

Complies with the requirements of EN61010-1.

1.10.7 EMC

Complies with EN61326 for emissions and immunity.



2 THEORY REFERENCE

2.1 Abbreviations

B	Susceptance (= 1/X)	R	Resistance
C	Capacitance	X	Reactance
D	Dissipation factor (tan δ)	Y	Admittance (= 1/Z)
E	Voltage	Z	Impedance
G	Conductance (= 1/R)	ω	2π x frequency
I	Current		
L	Inductance		Subscript s (s) = series
Q	Quality (magnification) factor		Subscript p (p) = parallel

2.2 Formulae

$$Z = \frac{E}{I} \quad (\text{all terms complex})$$

$$Y = \frac{I}{E} = \frac{1}{Z}$$

$$Z_s = R + jX = R + j\omega L = R - \frac{j}{\omega C}$$

$$|Z_s| = \sqrt{(R^2 + X^2)}$$

$$|Z_p| = \frac{RX}{\sqrt{(R^2 + X^2)}}$$

$$Y_p = G + jB = G + j\omega C = G - \frac{j}{\omega L}$$

$$|Y_p| = \sqrt{(G^2 + B^2)}$$

$$|Y_s| = \frac{GB}{\sqrt{(G^2 + B^2)}}$$

$$\text{where} \quad X_L = \omega L \quad X_C = \frac{1}{\omega C} \quad B_C = \omega C \quad B_L = \frac{1}{\omega L}$$

$$Q = \frac{\omega L_s}{R_s} = \frac{1}{\omega C_s R_s} \quad (\text{series R, L, C values})$$

$$Q = \frac{R_p}{\omega L_p} = \omega C_p R_p \quad (\text{parallel R, L, C values})$$



$$D = \frac{G_p}{\omega C_p} = \omega L_p G_p \quad (\text{parallel G, L, C values})$$

$$D = \frac{R_s}{\omega L_s} = \omega C_s R_s \quad (\text{series R, L, C values})$$

Note : The value $Q = \frac{1}{D}$ is constant regardless of series/parallel convention

2.3 Series/Parallel Conversions

$$R_s = \frac{R_p}{(1+Q^2)}$$

$$R_p = R_s(1+Q^2)$$

$$C_s = C_p(1+D^2)$$

$$C_p = \frac{C_s}{(1+D^2)}$$

$$L_s = \frac{L_p}{\left(1+\frac{1}{Q^2}\right)}$$

$$L_p = L_s\left(1+\frac{1}{Q^2}\right)$$

Conversions using the above formulae will be valid only at the test frequency.

2.4 Polar Derivations

$$R_s = |Z| \cos\theta$$

$$G_p = |Y| \cos\theta$$

$$X_s = |Z| \sin\theta$$

$$B_p = |Y| \sin\theta$$

Note that, by convention, +ve angle indicates an inductive impedance or capacitive admittance.

If capacitance is measured as inductance, the L value will be -ve.

If inductance is measured as capacitance, the C value will be -ve.

$D = \tan \delta$ where $\delta = (90 - \theta)^\circ$ admittance measurement.

$Q = \frac{1}{\tan \delta}$ where $\delta = (90 - \theta)^\circ$ impedance measurement.