

## Correction to Sampling a Biomarker of the Human Immunodeficiency Virus across a Synthetic Nanopore

[*ACS Nano* 2013, 7,

DOI: 10.1021/nn400125c]. David J. Niedzwiecki,

Raghuvaran Iyer, Philip N. Borer, and Liviu Movileanu\*

In the Experimental Section of our article, the last term in eq 3 should read  $1/d$  not  $1/d^2$ . The correct form of the full equation is

$$I = V([\mu_{\text{Na}^+} + \mu_{\text{Cl}^-}]n_{\text{NaCl}}e) \left( \frac{4h_{\text{eff}}}{\pi d^2} + \frac{1}{d} \right)^{-1}$$

This erratum does not affect any of the experimental results, discussions, or conclusions reported in the paper. The authors sincerely apologize for this unintended oversight.

Published online March 29, 2013

10.1021/nn401345w