

Full-length article

Development of a complex scintillation proximity assay for high-throughput screening of PPAR γ modulators¹

Bin WU, Jie GAO, Ming-wei WANG²

The National Center for Drug Screening, Shanghai Institute of Materia Medica, Shanghai Institutes for Biological Sciences, Chinese Academy of Sciences; Graduate School of Chinese Academy of Sciences, Shanghai 201203, China

Key words

peroxisome proliferator-activated receptor gamma; retinoid X receptor alpha; scintillation proximity assay; high-throughput screening

¹ Project supported by grants from the Ministry of Science and Technology of China (2002AA-2Z343A), Chinese Academy of Sciences (KSCX1-SW-11-2) and Shanghai Municipality Science and Technology Development Fund (03dz19224).

² Correspondence to Dr Ming-wei WANG.

Phn 86-21-5080-0598.

Fax 86-21-5080-0721.

E-mail mwwang@siniwest.com

Received 2004-09-06

Accepted 2004-11-26

doi: 10.1111/j.1745-7254.2005.00040.x

Abstract

Aim: To develop a complex high-throughput screening (HTS) assay based on scintillation proximity assay (SPA) technology for identification of novel peroxisome proliferator-activated receptor gamma (PPAR γ) modulators. **Methods:** Full-length PPAR γ and retinoid X receptor alpha (RXR α), biotinylated PPAR response element (PPRE), [³H]BRL49653 and streptavidin-coated FlashPlate or microbead were used to develop an HTS assay based on SPA technology. This ‘ABCDE’ method was validated against conventional hydroxyapatite (HA) assay and applied to large-scale screening of 16 000 synthetic compounds and natural product extracts. **Results:** (1) IC₅₀ values of positive control compounds (BRL49653 and troglitazone) obtained from the ‘ABCDE’ method and HA assay were comparable and consistent with those reported elsewhere; (2) Approximately 178 compounds, showing more than 70% competitive inhibition on BRL49653 binding to PPAR γ , were identified initially by the ‘ABCDE’ method (microbead); (3) Secondary screening using FlashPlate and cross-reactivity studies with RAR α , β , γ and RXR α , β , γ confirmed that 12 compounds possessed specific PPAR γ binding properties including 2 with IC₅₀ values less than 0.5 μ mol/L and novel chemical structures. **Conclusions:** The ‘ABCDE’ method using either FlashPlate or microbead, is a highly efficient, automatable, and robust tool to screen potential PPAR γ modulators in HTS setting. Its application may be expanded to other nuclear receptors that form heterodimers upon activation.

Introduction

Nuclear receptors (NRs) are a superfamily of ligand activated transcription factors that modulate specific gene expression. To date, more than 100 NRs have been identified including class I (ligand-dependent), class II (ligand-independent), and orphan receptors. A common feature of NRs is that they all contain a DNA binding domain that interacts with respective target genes to exert physiological functions^[1]. Peroxisome proliferator-activated receptors (PPARs) with three isoforms (α , β , and γ) regulate gene transcription in response to small, lipophilic ligands^[2-5]. PPAR α is present in the liver, kidney, and heart, PPAR β (also known as PPAR δ) is expressed ubiquitously, and PPAR γ is mainly found in the adipose tissue and muscle. Upon ligand binding, PPARs release relevant co-repressors and form heterodimers

with retinoid X receptors (RXRs)^[6]. The heterodimers bind to peroxisome proliferator response elements (PPREs)^[7,8] and recruit co-activators to initiate transcription of target genes. It is known that PPAR γ is activated by fatty acids and prostaglandin J2 derivatives, although the identities of its physiologically relevant activators are not certain^[9,10].

Because PPAR γ activation can cause insulin sensitization, its synthetic agonists have been used in the treatment of type 2 diabetes^[11,12]. Recently discovered liabilities of such therapy, namely, weight gain and edema, led to regulatory concerns on the long-term administration of drugs acting through PPAR γ ^[13,14]. The elimination of such adverse effects may depend on the discovery of novel compounds with improved tissue selectivity while retaining insulin-sensitizing property.

Conventional methods to study and characterize NRs include non-homogeneous hydroxyapatite (HA) and gel shift assays which require a laborious separation procedure and thus, are not suitable for high-throughput screening (HTS). Scintillation proximity assay (SPA)^[15] technology, however, provides a homogeneous screening approach that does not involve post-reaction liquid handling steps and is well-suited to automation and HTS. In the SPA system, an isotope (eg, ^3H) is brought very close to a scintillant-impregnated microbead or FlashPlate by binding to its surface. Because the emitted β particles or auger electrons can only travel short distances in the bulk solution, the microbead or FlashPlate preferentially captures electrons from the bound radiolabeled ligand. Therefore, the amount of light emitted from the scintillant in the microbead or FlashPlate is directly proportional to the amount of bound radiolabeled ligand (Figure 1). Several SPA-based NR competitive binding assays were developed and applied to HTS using biotinylated receptor ligand binding domains (LBDs; ‘ABC’ method)^[16,17]. In this paper, we describe a more complex SPA-based assay system which includes the full-length PPAR γ and RXR α , biotinylated PPRE, [^3H]BRL49653 and streptavidin-coated FlashPlate or microbead in a homogeneous setting. This ‘ABCDE’ approach was fully validated and applied to HTS of a sizable compound library. A series of structurally diversified ‘hits’ were found, and subsequent characterization led to the discovery of two novel PPAR γ binders with sub-micromolar potency and high specificity.




Figure 1. Principle of SPA: PPAR γ forms a heterodimer with RXR α upon activation by BRL49653. The heterodimer binds to biotinylated PPRE via streptavidin-coated microbead. Because the distance between [^3H]BRL49653 and scintillant is short enough, the latter can be stimulated by [^3H] to release $h\nu$. In the case of FlashPlate, biotinylated PPRE, streptavidin and scintillant were immobilized on the surface of each well.

Materials and methods

Reagents Potassium chloride, sodium phosphate

monobasic anhydrous, and magnesium chloride hexahydrate were purchased from Shanghai Chemical Co, Ltd. Edetic acid was purchased from Sigma-Aldrich (USA). BRL49653 and troglitazone were purchased from Cayman Chemical Co (USA). 3-[(3-Cholamidopropyl)dimethylammonio]-1-propane sulfonate (CHAPS) was purchased from Boehringer Mannheim GmbH (Germany). Dithiothreitol (DTT) was purchased from BioBasic Inc (Canada) and hydroxyapatite was obtained from Bio-Rad Laboratories (USA). Aprotinin and leupeptin were purchased from Merck KGaA (Germany). [^3H]BRL49653 (53 Ci/mmol) was obtained from American Radiolabeled Chemicals, Inc (USA), FlashPlate and flat-bottom IsoplateTM was obtained from PerkinElmer, Inc (USA), and streptavidin-coated microbead was obtained from Amersham Biosciences UK Ltd (England). The plasmids of human NRs used in this study were from Dr Shen X of Shanghai Institute of Materia Medica, Chinese Academy of Sciences and Dr Chen SJ of Shanghai Institute of Hematology. Full-length PPAR γ , RAR α , β , γ , and RXR α , β , γ were produced with a baculovirus expression system using IPLB-Sf-21 cells^[18]. The stock solutions for PPAR γ , RAR α , β , γ and RXR α , β , γ extract proteins were at 7, 12–15, and 7–13 g/L, respectively. Double-strand 5'-biotinylated-PPRE (CCTT-TGACCTATTGAACTATTACCT) was synthesized by Shanghai Sangon Biological Engineering Technology & Service Co, Ltd.

HA assay The assay buffer consists of 10% glycerol (v/v), NaH_2PO_4 25 mmol/L, MgCl_2 0.5 mmol/L, DTT 1 mmol/L, edetic acid 1 mmol/L, CHAPS 5 mmol/L, aprotinin 2 mg/L and leupeptin 100 $\mu\text{mol/L}$. PPAR γ 1 μL extract protein (70 mg/L) was loaded into each well of IsoplateTM containing the assay buffer, followed by [^3H]BRL49653 (1.2 μL , 10 nmol/L) and various concentrations of BRL49653 or troglitazone (2.5 μL), to give a final volume of 100 μL per well. The plates were sealed and incubated overnight at 4 $^\circ\text{C}$. HA (25%, v/v) 25 mL was added to each well the next morning and the plates were gently agitated twice for 5 min each. Following centrifugation at 1200 $\times g$ for 3 min, the supernatant was decanted and 100 μL assay buffer was added to each well. This washing procedure was repeated twice before adding 150 μL scintillation liquid (PerkinElmer), the plates were gently agitated to resuspend HA and counting was measured by a MicroBeta counter (PerkinElmer).

FlashPlate based SPA assay Biotinylated-PPRE (4 μL from a stock solution of 10 g/L) was mixed with the above assay buffer (20 mL) containing fish sperm DNA (Sangon; 20 μL from a 10 g/L stock solution), loaded to streptavidin-coated FlashPlate (200 μL /well) and incubated overnight at 4 $^\circ\text{C}$. It was then washed three times with the assay buffer,

200 μ L reaction solution containing 14 μ g PPAR γ extract protein (70 mg/L), 0.94 mg RXR α extract protein (4.7 mg/L), 10 nmol/L [3 H]BRL49653 and various concentrations of BRL49653 or troglitazone were added to each well. Following incubation at 4 $^{\circ}$ C for 4 h, the plates were counted by the MicroBeta counter. For validation purpose, various concentrations of PPAR γ and RXR α extract proteins, as well as different reaction time lengths, were studied to determine an optimal assay condition.

Microbead based SPA assay Biotinylated-PPRE (2 μ L from a stock solution of 10 g/L) was mixed with the assay buffer (10 mL) containing fish sperm DNA (Sangon; 10 μ L from a 10 g/L stock solution) and 4 mg streptavidin-coated microbead in a conical polypropylene centrifuge tube (Corning Inc, USA) and incubated overnight at 4 $^{\circ}$ C. The mixture was centrifuged for 10 min at 1500 \times g. The supernatant was then removed and washed three times with the 10-mL assay buffer. Reaction solution 10 mL containing 700 μ g PPAR γ extract protein (70 mg/L), 47 μ g RXR α extract protein (4.7 mg/L), 10 nmol/L [3 H]BRL49653 and various concentrations of BRL49653 or troglitazone were distributed to each well of a IsoplateTM (100 μ L/well) and incubated at 4 $^{\circ}$ C for 4 h before counting by the MicroBeta counter. For validation purposes, various amounts of microbead were used to determine an optimal assay condition.

HTS studies The compound library used for screening consists of 16 000 pure synthetic compounds and extracts of natural products. A 10-compound pool per well mix was applied to the primary screening (microbead based SPA assay), with an average concentration of 7 μ mol/L for each compound dissolved in 100% Me₂SO solution. This matrix system maximizes the advantage of HTS and allows duplicate screening of each compound^[19]. In each 96-well IsoplateTM, 16 wells were used as positive control (BRL-49653) and samples showing greater than 70% inhibition were considered 'hits'. Positive compounds were re-screened with FlashPlate based SPA assay and confirmed 'hits' studied for their binding cross-reactivities with RAR α , β , γ and RXR α , β , γ using respective HA assay.

Results

Assay validation In the present study, we first assessed the kinetics of the signal strength generated by FlashPlate assay. Time-course experiment suggested that the equilibrium reached after 3.5 h of incubation at 4 $^{\circ}$ C and prolongation of the reaction time did not improve the assay efficiency (Figure 2A). Various concentrations of PPAR γ and RXR α were used to establish the optimal assay condition. In the

absence of unlabeled BRL49653, a maximum signal was detected with a combination of 140 mg/L (1:50 of the stock solution) PPAR γ extract protein and 7 mg/L (1:1000 of the stock solution) RXR α (Figure 2B). The fully optimized assay possessed a signal to background ratio of 5. Thus, optimal receptor concentrations were determined for PPAR γ (70 mg/L; 1:100) and RXR α (4.7 mg/L; 1:1500), respectively. Under this assay condition, IC₅₀ values for the two PPAR γ agonists, BRL49653 and troglitazone were measured (Figure 2C), and found to be comparable to those calculated from the HA assay (Figure 2D). Since the principle of microbead-based SPA assay is similar to that of FlashPlate, identical receptor concentrations were used with a reduced assay volume (100 μ L/well). When different concentrations of microbead were used, a saturation reached between 2 and 4 g/L (Figure 3A) with a signal-to-background ratio equal to 5. The IC₅₀ values for BRL49653 and troglitazone determined by this assay were within the range described earlier (Figure 3B).

Assay parameters In order to apply the microbead-based SPA assay to HTS, both non-specific binding (NSB; using 22.5 mmol/L BRL49653) and maximum binding (MB; using 0 mmol/L BRL49653) were studied. Coefficient of variation (CV) values were 8.5% for NSB and 6.2% for MB, respectively (Figure 3C). The Z' factor, which estimates the suitability to HTS^[20], was calculated to be 0.71.

High-throughput screening Of the 16 000 samples initially screened, 178 'hits' (1.11%) showing greater than 70% competitive inhibition on BRL49653 binding to PPAR γ were discovered (all synthetic compounds; Figure 4A). Secondary (single compound per well) screening confirmed that 24 of the above 'hits' displayed consistent inhibitory effects with IC₅₀ values between 0.2 and 28.5 μ mol/L. Cross-reactivity studies with RAR α , β , γ and RXR α , β , γ revealed that 12 of these compounds possess specific PPAR γ binding properties including 2 with IC₅₀ values less than 0.5 μ mol/L (Table 1). In this HTS campaign, the signal-to-noise ratio (10-to-15-fold), CV (5%–8%) (Figure 4B) and Z' factor (0.66–0.75) are of high quality nature.

Discussion

Three receptor binding assays were employed and compared side-by-side in the present study to measure specific binding properties of two known PPAR γ agonists, namely, BRL49653 and troglitazone. Conventional HA assay is a non-homogeneous method widely used to assess competitive interaction between a testing agent and receptor in the presence of radiolabeled ligand. For PPAR γ , such interaction involves additional components such as RXR α and




Figure 2. (A) Time-course study of FlashPlate assay. Equilibrium of PPAR γ , RXR α , [3 H]BRL49653, unlabeled BRL49653, PPRE and streptavidin-coated FlashPlate was reached at 4 h. (B) Effects of different receptor concentrations on FlashPlate assay signal strength (the stock solution concentration is 7 g/L for both PPAR γ and RXR α). (C) Dose-response curves of BRL49653 and troglitazone measured by FlashPlate assay from which IC $_{50}$ values were calculated. (D) Dose-response curves of BRL49653 and troglitazone measured by HA assay. Data shown are duplicate samples with mean \pm SEM from at least two experiments.

PPRE. Therefore, the ‘ABCDE’ method utilizing either FlashPlate or microbead based on SPA technology was developed and validated to include both RXR α and PPRE in the assay system. The final readout is fully dependent upon specific binding of biotinylated PPRE to streptavidin-coated FlashPlate or microbead as addition of free biotin was able to block this interaction completely^[16]. Although the IC $_{50}$ values of BRL49653 and troglitazone generated by these three approaches were very similar and comparable to those reported previously with a relatively simple ‘ABC’ SPA method (BRL49653: $K_d=26$ nmol/L, IC $_{50}=36$ nmol/L; troglitazone: $K_d=310$ nmol/L, IC $_{50}=320$ nmol/L)^[16], the ‘ABCDE’ model is obviously superior in terms of physiologic mimicry, easy to use, robustness and efficiency, largely due to its inclusiveness and homogenous nature.

We found that both FlashPlate and microbead-based SPA methods could be readily adapted to automated HTS. However, the cost of FlashPlate, including assay volume (200

μ L), associated reagent consumption, and compound depletion, *etc.*, is approximately 5 times greater than that of microbead (100 μ L). This may constitute a major concern when implementing a large HTS campaign. All the key assay parameters, such as CV and Z’ factor, obtained from our microbead-based SPA validation experiments, indicate that it is well suited to HTS^[20]. Indeed, when employed in HTS of potential PPAR γ modulators, this assay system demonstrated a consistently high quality in terms of the signal-to-noise ratio, CV and Z’ factor in all the 40 pooled compound matrix plates. One implication of this is that microbead-based SPA technology may be expanded to other NRs that form heterodimers upon activation.

Following the initial screening of 16 000 samples, a ‘hit’ rate of 1.11% was achieved, of which, only 13.5% of the ‘hits’ could be confirmed by secondary screening. They were further tested for cross-reactivities with both RARs and RXRs (defined as IC $_{50}$ less than 70 μ mol/L), and 12 compounds




Figure 3. (A) Effects of different microbead concentrations on SPA assay signal strength. Saturation reached between 2 and 4 g/L. (B) Dose-response curves of BRL49653 and troglitazone measured by microbead based SPA assay from which IC₅₀ values were calculated. Data shown are duplicate samples with mean±SEM from at least two experiments. (C) Intra-plate variation of microbead based SPA assay as demonstrated by MB and NSB.

showed high specificity for PPAR γ with IC₅₀ values ranging from 0.2 to 28.5 μ mol/L. The IC₅₀ for one thiazolidinedione-like compound is 0.2 μ mol/L, better than several currently marketed PPAR γ agonists. The other compound with an IC₅₀

Table 1. Receptor binding specificity and cross-activity of the confirmed ‘hits’.

Compound	IC ₅₀ / μ mol·L ⁻¹						
	PPAR γ	RAR α	RAR β	RAR γ	RXR α	RXR β	RXR γ
Thiazolidinedione class							
SH00013951	14.4	NB	NB	NB	NB	NB	NB
SH00012671	0.2	NB	NB	NB	NB	NB	NB
Imide class							
SH00017229	6.7	NB	NB	NB	NB	NB	NB
SH00013478	10.5	NB	NB	NB	NB	NB	NB
SH00012632	0.49	NB	NB	NB	NB	NB	NB
Other classes							
SH00017545	5.2	NB	NB	NB	NB	NB	NB
SH00010267	23.3	NB	NB	NB	NB	NB	NB
SH00017475	15.2	NB	NB	NB	NB	NB	NB
SH00015594	6.7	NB	NB	NB	NB	NB	NB
SH00012095	16.7	NB	NB	NB	NB	NB	NB
SH00013003	28.5	NB	NB	NB	NB	NB	NB
SH00010675	15.1	NB	NB	NB	NB	NB	NB

NB: No binding activity was detected at the concentration (70 μ mol/L) tested. Secondary ‘hits’ showing binding activities of various degrees with RARs or RXRs were discarded.

of 0.49 μ mol/L belongs to the imide class. Both of them possess novel chemical structures (data not shown). If their activities could be demonstrated by cell-based functional assays, bioassay-guided structure modification and optimization may lead to the discovery of some entirely new PPAR γ modulators.

Acknowledgement

We thank Dr Dale E MAIS, Xin XIE, Na LI, and Chenghe JIN for their valuable discussions.

References

- Mangelsdorf DJ, Thummel C, Beato M, Herrlich P, Schutz G, Umesono K, *et al*. The nuclear receptor superfamily: the second decade. *Cell* 1995; 83: 835–9.
- Schmidt A, Endo N, Rutledge SJ, Vogel R, Shinar D, Rodan GA. Identification of a new member of the steroid hormone receptor superfamily that is activated by a peroxisome proliferators and fatty acids. *Mol Endocrinol* 1992; 6: 1634–41.
- Sher T, Yi HF, McBride OW, Gonzalez FJ. cDNA cloning, chromosomal mapping and functional characterization of the human peroxisome proliferator activated receptor. *Biochemistry* 1993; 32: 5598–604.
- Desvergne B, Wahli W. Peroxisome proliferator-activated receptors: nuclear control of metabolism. *Endocr Rev* 1999; 20: 649–88.
- Wise H. Multiple signalling options for prostacyclin. *Acta Pharmacol*




Figure 4. (A) High-throughput screening of 16 000 compounds using microbead-based SPA assay. Results are expressed as percentage inhibition of $[^3\text{H}]\text{BRL49653}$ binding to PPAR γ . (B) Signal-to-noise (S/N) and coefficient of variation (CV) of the 10-compound pool per well matrix assay system.

Sin 2003; 24: 625–30.

- 6 Kliewer SA, Umesono K, Noonan DJ, Heyman RA, Evans RM. Convergence of 9-*cis* retinoic acid and peroxisome proliferator signalling pathways through heterodimer formation of their receptors. *Nature* 1992; 358: 771–4.
- 7 Ijpenberg A, Jeannin E, Wahli W, Desvergne B. Polarity and specific sequence requirements of PPAR-RXR heterodimer binding to DNA: a functional analysis of the malic enzyme gene PPRE. *J Biol Chem* 1997; 272: 20108–17.
- 8 Juge-Aubry C, Pernin A, Favez T, Burger AG, Wahli W, Meier CA, *et al.* DNA binding properties of peroxisome proliferator-activated receptor subtypes on various natural peroxisome proliferator response elements: importance of the 5' flanking region. *J Biol Chem* 1997; 272: 25252–9.
- 9 Yu K, Bayona W, Kallen CB, Harding HP, Ravera CP, McMahon G, *et al.* Differential activation of peroxisome proliferator-activated receptors by eicosanoids. *J Biol Chem* 1995; 270: 23975–83.
- 10 Kliewer SA, Lenhard JM, Willson TM, Patel I, Morris DC, Lehmann JM. A prostaglandin J2 metabolite binds peroxisome proliferator-activated receptor gamma and promotes adipocyte differentiation. *Cell* 1995; 83: 813–9.
- 11 O'Moore-Sullivan TM, Prins JB. Thiazolidinediones and type 2 diabetes: new drugs for an old disease. *Med J Aust* 2002; 176: 381–6.
- 12 Yang L, An HX, Deng XL, Chen LL, Li ZY. Rosiglitazone reverses insulin secretion altered by chronic exposure to free fatty acid via IRS-2-associated phosphatidylinositol 3-kinase pathway. *Acta Pharmacol Sin* 2003; 24: 429–34.
- 13 Camp HS. Thiazolidinediones in diabetes: current status and future outlook. *Curr Opin Investig Drugs* 2003; 4: 406–11.
- 14 Guan Y, Zhang Y, Davis L, Breyer MD. Expression of peroxisome proliferator-activated receptors in urinary tract of rabbits and humans. *Am J Physiol* 1997; 273: F1013–22.
- 15 Hart HE, Greenwald EB. Scintillation proximity assay of antigen-antibody kinetics: concise communication. *J Nucl Med* 1979; 20: 1062–5.
- 16 Nichols JS, Parks DJ, Consler TG, Blanchard SG. Development of a scintillation proximity assay for peroxisome proliferator-activated receptor gamma ligand binding domain. *Anal Biochem* 1998; 257: 112–9.
- 17 Henke BR, Consler TG, Go N, Hale RL, Hohman DR, Jones SA, *et al.* A new series of estrogen receptor modulators that display selectivity for estrogen receptor beta. *J Med Chem* 2002; 45: 5492–505.
- 18 Gearing KL, Göttlicher M, Teboul M, Widmark E, Gustafsson J. Interaction of the peroxisome proliferator-activated receptor and retinoid X receptor. *Proc Natl Acad Sci USA* 1993; 90: 1440–4.
- 19 Qian J, Voorbach MJ, Huth JR, Coen ML, Zhang HC, Ng SC, *et al.* Discovery of novel inhibitors of Bcl-xL using multiple high-throughput screening platforms. *Anal Biochem* 2004; 328: 131–8.
- 20 Zhang JH, Chung TD, Oldenburg KR. A simple statistical parameter for use in evaluation and validation of high-throughput screening assays. *J Biomol Screen* 1999; 4: 67–73.