

A simple and cheap method of screening glass microelectrodes

I. ENGBERG, J.A. FLATMAN & J.D.C. LAMBERT (introduced by G.N. Woodruff)

Institute of Physiology, University of Aarhus, 8000 Aarhus C, Denmark

The recording of fast intracellular potential changes during micro-iontophoretic studies is complicated by the attenuation of high frequency responses by the distributed capacitance between the recording electrode and the multibarrelled iontophoretic electrode assembly. This attenuation is particularly troublesome in coaxial electrodes of the type described by Sonnhof (1973). The problem can be overcome by the use of screened recording electrodes—the screen being driven from the input amplifier at unitary gain. Provided that the input amplifier has good HF characteristics, this method will also allow accurate measurement of cell membrane resistance by current pulses through the recording electrode (cf. Eide, 1968). An efficient method to screen electrodes was developed by Eide, Engberg & Sonnhof (see Sonnhof, 1973), but this necessitates access to expensive high vacuum equipment for 'sputtering' a thin layer of gold onto the micropipettes. Further the electrodes must be filled with electrolyte *after* this goldplating when their interior is no longer visible. To avoid these complications we have used the following procedure.

Single, filled, electrodes were coated to within 1 mm of the tip with graphite from an aerosol can (Graphit 33, Kontakt Chemie, Rastatt, W. Germany). The electrode was held in a clamp so

that the tip dipped into 99% ethanol and then sprayed so that an even layer of graphite was deposited on the glass. The coated electrode was then inserted into the central canal of the iontophoretic assembly, where it was surrounded with a small amount of saline. The graphite screen or the saline was connected to a unitary gain output of the amplifier output via a 100 nF capacitor.

The present technique has proved to be quite comparable to the gold screening in conjunction with coaxial multibarrel electrodes (for illustration see Figure 1 of the communication by Engberg, Flatman & Lambert, 1975, this meeting). Graphite screening can also be used to neutralize the capacitance of single electrodes in other types of recording, if the screen is insulated by a thin layer of lacquer covering the part that enters the tissue or conducting fluid.

We acknowledge the technical assistance of Mr A Jensen and a travelling grant from the Wellcome Trust to J.D.C.L.

References

- EIDE, E. (1968). Input amplifier for intracellular potential and conductance measurements. *Acta physiol. scand.*, **73**, 1A-2A.
- ENGBERG, I., FLATMAN, J.A. & LAMBERT, J.D.C. (1975). DL-Homocysteate induced motoneurone depolarization with membrane conductance decrease. *Communication at this meeting.*
- SONNHOF, U. (1973). A multibarrelled coaxial electrode for iontophoresis and intracellular recording with a gold shield of the central pipette for capacitance neutralization. *Pflügers Arch.*, **341** 351-358.

Combined EEG recording and intraventricular administration of drugs in the conscious rat

D. GOFF, A.A. MILLER, R.E. SMITH, S.J. SMITH & P.L. WHEATLEY

Pharmacology Laboratory, Wellcome Research Laboratories, Beckenham, Kent BR3 3BS

This technique, a modification of that described previously by Hill & Miller (1973) for use in the rabbit, allows the combination of EEG recording

and intraventricular administration of drugs in the conscious rat.

EEG activity is recorded from a 4-pin miniature electrical connector (ITT Cannon, Basingstoke) linked to stainless steel screws inserted in the skull. The small dimensions of the connector (1.5 mm x 6 mm x 9 mm high) and especially the narrow width are of considerable advantage for use on the rat skull.

The intraventricular cannula was modified from that described by Hayden, Johnson & Maickel (1966). It was prepared from 3 mm diameter Perspex rod, the small diameter being of advantage for combination with the electrical connector.