## **DEMONSTRATIONS**

### A microcomputer-based signal acquisition system suitable for use in the pharmacological and physiological laboratory

C.T. KIRKPATRICK

(introduced by C.N. SCHOLFIELD)

Physiology Department, The Queen's University of Belfast

A digital signal acquisition package has been developed for use in conjunction with the Commodore PET-2001 microcomputer. It features a multichannel analog-to-digital converter (National Semiconductors ADC 0817) and a single-chip digital-to-analog converter (Ferranti ZN 425 or equivalent). These devices are interfaced to the memory bus of the microcomputer using MOS 6522 Versatile Interface Adaptors, which also contain useful timing registers.

Although data handling can be performed using BASIC programs (Baillie & Smith, 1979), this can be slow, and greater efficiency and speed can be achieved using machine code subroutines for data capture and display. A block of memory is reserved as a data buffer, and a simple subroutine written in 6502 machine code controls the sequential digitization and storage of 256 (or 512 or 1024 etc) 8-bit data values in the buffer. The contents of this array of 'captured' data can then be sequentially accessed for display via an analog-to-digical converter and oscilloscope or

pen-recorder, or can be stored on mass storage media such as magnetic tape or disc.

Facilities also exist for the internal generation of a wide variety of mathematical functions which can be displayed on the oscilloscope or written to a pen recorder.

The system offers many of the facilities available on transient recorders, e.g. capture of high-speed transients and their subsequent play-back at lower speed, but also allows the programmer to perform more sophisticated analysis such as signal averaging, measurement of maxima and minima or duration of waveforms, and even procedures such as spectral analysis could be performed.

The package could readily be adapted for use on other microcomputer systems. Current applications include the recording of transients in electrophysiology and the acquisition and analysis of data from respiratory experiments.

The approximate cost of external components is £50.

#### Reference

BAILLIE, P. & SMITH, I.C.H. (1979). Real-time computing in the teaching laboratory—an analysis of heart-rate transients. J. Physiol. proceedings (In press).

The following demonstrations were given at the Aston meeting of the British Pharmacological Society, 4th-6th April, 1979

# The binding of [3H]-noradrenaline and [14C]-propranolol to synaptosomal fractions of rat brain homogenates

#### **B.A. HEMSWORTH & J.A. STREET**

Pharmacological Research Laboratories, Department of Pharmacy, The University of Aston in Birmingham, Birmingham B4 7ET

Propranolol has been shown to inhibit both uptake and tyramine-induced release of noradrenaline in the periphery (Foo, Jowett & Stafford, 1968). More recently, it has been suggested that propranolol itself may be taken up into peripheral adrenergic neurones, stored, and released along with noradrenaline (Daniell, Walle, Gaffney & Webb, 1978), although there is conflicting evidence on this point (Lewis, 1977). Penetration of propranolol into the brain after systemic administration has been demonstrated by a number of workers (Laverty & Taylor, 1968; Hayes & Cooper, 1971; Street, Hemsworth, Roach & Day, 1979), but the action of this drug upon central adrenergic neurones has not been studied in detail. The present work was undertaken to investigate the possible inhibitory effects of propranolol upon noradrenaline uptake into synaptosomes prepared from rat brain homogenates, and further, to investigate