

# MALDI-induced fragmentation of leucine enkephalin, nitro-Tyr-leucine enkephalin, and d<sub>5</sub>-Phe-nitro-Tyr-leucine enkephalin

Xianquan Zhan<sup>a,b,\*</sup>, Dominic M. Desiderio<sup>a,b,c,d</sup>

<sup>a</sup> Charles B. Stout Neuroscience Mass Spectrometry Laboratory, The University of Tennessee Health Science Center, 847 Monroe Avenue, Room 117, Memphis, TN 38163, USA

<sup>b</sup> Department of Neurology, The University of Tennessee Health Science Center, Memphis, TN, USA

<sup>c</sup> Department of Molecular Sciences, The University of Tennessee Health Science Center, Memphis, TN, USA

<sup>d</sup> University of Tennessee Cancer Institute, The University of Tennessee Health Science Center, Memphis, TN, USA

## ARTICLE INFO

### Article history:

Received 13 May 2008

Received in revised form 17 July 2008

Accepted 25 August 2008

Available online 2 September 2008

### Keywords:

Nitrotyrosine-containing peptide

MALDI-induced fragmentation

Nitro-Tyr-leucine enkephalin

d<sub>5</sub>-Phe-Nitro-Tyr-leucine enkephalin

Sensitivity

## ABSTRACT

The long-term objective of this study is to use MALDI MS and MS/MS to study the fragmentation pattern of *in vitro* nitrotyrosine-containing peptides in order to assist the interpretation of MS-identification of endogenous nitroproteins in human tissues and fluids. The short-term objective is to study synthetic leucine enkephalin, nitro-Tyr-leucine enkephalin, and d<sub>5</sub>-Phe-nitro-Tyr-leucine enkephalin with a vacuum matrix-assisted laser desorption/ionization linear ion-trap mass spectrometer (vMALDI-LTQ). The results demonstrated the UV laser-induced photochemical decomposition of the nitro group. Although photochemical decomposition decreased the ion intensity and complicated the MS spectrum, the recognition of that unique decomposition pattern unambiguously identified a nitrotyrosine. The a<sub>4</sub>- and b<sub>4</sub>-ions were the most intense fragment ions found in the MS/MS spectra for those three synthetic peptides. Compared to the unmodified peptides, more collision energy optimized the fragmentation of the nitropeptide, increased the intensity of the a<sub>4</sub>-ion, and decreased the intensity of the b<sub>4</sub>-ion. Optimized laser fluence maximized the fragmentation of the nitropeptide. MS<sup>3</sup> analysis confirmed the MS<sup>2</sup>-derived amino acid sequence, but required much more sample. To detect a nitropeptide, the sensitivity of vMALDI-LTQ is 1 fmol for MS detection and 10 fmol for MS<sup>2</sup> detection; the S/N ratio was ca. 50:1 in those studies. Those data are important for an analysis of low-abundance endogenous nitroproteins, where preferential enrichment of nitroproteins and optimized mass spectrometry parameters are used.

© 2008 Elsevier B.V. All rights reserved.

## 1. Introduction

Reactive oxygen species and reactive nitrogen species (ROS and RNS, respectively)-mediated oxidative/nitrative stress play important roles in cellular, physiological, and pathological processes [1,2]. Specific amino acid residues in proteins are sensitive tar-

gets that can be modified by ROS/RNS. Tyrosine nitration in a protein is an important redox-related modification that alters the activity of that protein. Endogenous nitrotyrosine-containing proteins and nitrotyrosine-sites have been discovered in a human pituitary [3,4], a pituitary adenoma [5], a diabetic rat retina [6], a SOD2-knockdown mouse eye-cup [7], sickle cell disease [8], spinal cords of a mouse model of familial amyotrophic lateral sclerosis [9], aging rat skeletal muscle [10,11], aging rat heart [12], mouse brain [13], human gliomas [14], and a septic patient's rectus abdominis muscle [15]. Most of the nitrotyrosine-sites studied to date are located within a specific functional and structural domains or motifs of a protein. For example, tyrosine nitration occurred within the enzyme-substrate activity region of sphingosine-1-phosphate lyase 1 and within the receptor-ligand region of leukocyte immunoglobulin-like receptor A4 in a human pituitary adenoma [5]. The tyrosine nitration within critical regions (enzyme-substrate; receptor-ligand) of a protein could decrease interactions and interfere with protein functions because the nitro (-NO<sub>2</sub>) group is a bulky electron-withdrawing group that shifts the pK of a Tyr OH group from ca. 10 to ca. 7 and decreases the electron

**Abbreviations:** CHCA, α-cyano-4-hydroxycinnamic acid; CID, collision-induced dissociation; LE1, leucine enkephalin; LE2, nitro-Tyr-leucine enkephalin; LE3, d<sub>5</sub>-Phe-nitro-Tyr-leucine enkephalin; LTQ, linear ion-trap mass spectrometer; vMALDI, vacuum matrix-assisted laser desorption/ionization; MS, mass spectrometry; MS<sup>2</sup>, MS/MS, tandem mass spectrometry; MS<sup>3</sup>, MS/MS/MS; MW, molecular weight; m/z, mass-to-charge ratio; NCE, normalized collision energy; S/N, signal-to-noise ratio; RNS, reactive nitrogen species; ROS, reactive oxygen species; TFA, trifluoroacetic acid.

\* Corresponding author at: Charles B. Stout Neuroscience Mass Spectrometry Laboratory, Department of Neurology, The University of Tennessee Health Science Center, 847 Monroe Avenue, Room 108, Memphis, TN 38163, USA.  
Tel.: +1 901 448 2976; fax: +1 901 448 7842.

E-mail address: [xzhan@utmem.edu](mailto:xzhan@utmem.edu) (X. Zhan).

density of the phenolic ring of a nitrotyrosine residue to diminish protein binding. Tyrosine nitration occurs within a tyrosine phosphorylation motif ([R or K]-x2(3)-[D or E]-x3(2)-[Y]) such as insulin-responsive glucose transporter type 4 and protein tyrosine phosphatase  $\eta$  in the diabetic rat retina [6]; and tyrosine nitration competes with the same tyrosine site with phosphorylation.

Mass spectrometry-based proteomics effectively detects and identifies endogenous nitroprotein and nitrotyrosine-sites. Proteomics is limited by the low-abundance of the endogenous nitrotyrosine-sites, which have been estimated to be 1 in  $10^6$  tyrosines [16], and the complicated mass spectrometric behavior of a nitrotyrosine-containing peptide. Several protocols to study nitrotyrosine-containing peptides or proteins were developed to be used before mass spectrometry analysis, including (a) the chemical reduction of nitrotyrosine to aminotyrosine [17], (b) derivatization of nitrotyrosine into free sulfhydryl groups followed by enrichment of sulfhydryl-containing peptides with thiopropyl sepharose beads [18], (c) dansyl chloride labeling of the nitration sites in combination with a precursor ion scan and an MS<sup>3</sup> analysis [19], (d) 2D-PAGE fractionation in combination with nitrotyrosine Western blot analysis [3,4,20], and (e) nitrotyrosine immunoaffinity enrichment of endogenous nitroproteins [5–7]. Protocols (a), (b) and (c) have succeeded only with an *in vitro* model peptide or protein, and in an *in vitro* nitrated proteome [17–19], but not with any *in vivo* studies. Protocols (d) and (e) have succeeded in the identification of endogenous nitrotyrosine-sites [3–7]. Nevertheless, a nitrotyrosine-containing peptide is still low in concentration relative to the tryptic peptides that are derived from a nitroprotein. Most nitroproteins have been identified with only one or two nitropeptides; that low number results in a low protein-coverage [3–15]. Therefore, a high-quality MS/MS-based amino acid sequence is mandatory. However, the complicated mass spectrometric behavior of a nitropeptide makes it difficult to obtain high-quality spectra and to interpret an MS spectrum. Studies have shown that MALDI-induced photochemical decompositions of a nitro group occur in *in vitro* nitroproteins and synthetic nitropeptides [21,22], and in endogenous nitroproteins [4]. This present study shows that MALDI-induced photochemical decompositions decrease the precursor-ion intensity of a nitropeptide, and complicate the MS spectrum. Electrospray spectra do not contain those decompositions [21,22].

This study describes the use of a vMALDI linear ion-trap mass spectrometer to study the fragmentation details of three synthetic peptides – leucine enkephalin, nitro-Tyr-leucine enkephalin, and d<sub>5</sub>-Phe-nitro-Tyr-leucine enkephalin – to explore the fragmentation behavior of nitrotyrosine-containing peptides during vMALDI-MS and MS/MS with the goal to assist our interpretation of endogenous low-abundance nitropeptides that are derived from human tissues and fluids. The accumulation of hundreds of MS/MS spectra of a nitropeptide significantly optimized the signal-to-noise ratio (S/N) of that spectrum. The spectrum of the <sup>2</sup>H (d)-modified nitro-Tyr-leucine enkephalin confirmed the interpretation of the vMALDI MS and MS/MS spectra of nitro-Tyr-leucine enkephalin.

## 2. Materials and methods

### 2.1. Synthetic peptides

Leucine enkephalin acetate hydrate (LE1—Table 1) was purchased (Sigma L9133-10 MG). Nitro-Tyr-leucine enkephalin (LE2) and d<sub>5</sub>-Phe-nitro-Tyr-leucine enkephalin (LE3) were synthesized by Dr. Peter Schiller (Clinical Research Institute of Montreal, Canada). Each sample was lyophilized. Table 1 contains the code, amino acid sequence, and theoretical accurate mass of those three peptides.

### 2.2. Preparation of peptide samples

A separate stock solution (5 pmol/ $\mu$ l) of each peptide was prepared by dissolving 220  $\mu$ g LE1 (396 pmol) in 79  $\mu$ l 0.1% trifluoroacetic acid (TFA) with 2% acetonitrile; 350  $\mu$ g LE2 (490 pmol) in 98  $\mu$ l 0.1% TFA with 2% acetonitrile; and 505  $\mu$ g LE3 (702 pmol) in 140  $\mu$ l 0.1% TFA with 2% acetonitrile. Each solution was stored (4 °C). An aliquot of a peptide stock solution (5 pmol) was diluted with a solution to yield a final concentration of 50% acetonitrile and 0.1% TFA. A series of dilutions was produced: 5000, 1000, 500, 100, 50, 10, 5, and 1 fmol/ $\mu$ l. The  $\alpha$ -cyano-4-hydroxycinnamic acid (CHCA) matrix solution (5 mg/ml) was prepared by dissolving 5 mg CHCA in 1 ml of a solution to yield a final concentration of 50% acetonitrile and 0.1% TFA. Each diluted peptide solution (4  $\mu$ l) was mixed with 4  $\mu$ l of the CHCA matrix solution. The peptide-CHCA matrix solution (2  $\mu$ l) was loaded onto a vMALDI 96-well plate (1 pmol/spot), and was dried in ambient air. Each sample was analyzed in triplicate (three separate MALDI spots).

### 2.3. vMALDI-LTQ mass spectrometer

Each peptide was analyzed with a vacuum matrix-assisted laser desorption/ionization-linear ion-trap mass spectrometer (vMALDI-LTQ) (ThermoFisher Scientific, San Jose, CA, USA) in the tune-page operation mode to obtain the MS and MS/MS spectra. For the vMALDI source, the crystal-positioning system (CPS) was enabled, auto spectrum filter (ASF) disabled, and the number of scans of step 10 for the disabled ASF. The automatic gain control (AGC) was enabled to allow the vMALDI software to automatically adjust the number of laser shots to optimize the quality of the vMALDI spectrum. The number of scans/file was 30. For an MS scan, mass range ( $m/z$  150–700), normal scan-rate, full scan, polarity positive, and 5 microscans of each experiment were used. For an MS/MS scan, normal mass range, normal scan-rate, polarity negative, full scan, isolation width 1.0 Th, activation Q-value 0.25, activation time 30 ms, and 5 microscans of each experiment were used. The laser beam was attenuated to provide a series of fluences (5, 10, 15, 20, 25, 30, 35, 40, and 45% of the unattenuated beam; arbitrary units). To analyze the effect of laser fluence on the product-ion intensity, the same series of laser fluence was used, and the normalized collision energy (NCE) was fixed at 35 arbitrary units. To analyze the effect of collision energy on the product-ion intensity, the laser fluence was fixed at 30 arbitrary units, and a series of NCE (10, 20, 30, 40, 50, 60, and 70 arbitrary units) was used. For MS<sup>2</sup>, a laser fluence 30 units, NCE 35 units, isolation width 1.0 Th, activation-Q 0.25, and activation time 30 ms were used. For MS<sup>3</sup>, the laser fluence 30 units, NCE 35 units, isolation width 1.0 Th, activation-Q 0.25, and activation time 30 ms were used.

The MS and MS/MS data were managed with the Qual Browser that is a part of the Xcalibur software package that is a part of the vMALDI-LTQ system. The spectrum (MS; MS<sup>2</sup>; MS<sup>3</sup>) of each scan ( $n=30$  scans) in each file was processed and accumulated ( $n=30$ ) to obtain a synthetic spectrum. The  $m/z$  value and peak intensity in the synthetic spectrum were copied into a Microsoft Excel program for data analysis and graph construction. Each experiment was performed in triplicate.

## 3. Results and discussion


### 3.1. MS analysis of LE1, LE2, and LE3

The MALDI MS spectrum of each peptide (1 pmol) was obtained at a laser fluence of 30 units. Fig. 1 contains the MS spectrum of LE1

**Table 1**

The amino acid sequence and theoretical mass of the synthesized peptides

Code	Sequences		Accurate mass (Da)
LE1	Tyr-Gly-Gly-Phe-Leu	Y-G-G-F-L	555.1818
LE2	(3-NO <sub>2</sub> )Tyr-Gly-Gly-Phe-Leu	(3-NO <sub>2</sub> )Y-G-G-F-L	600.0909
LE3	(3-NO <sub>2</sub> )Tyr-Gly-Gly-(d <sub>5</sub> )Phe-Leu	(3-NO <sub>2</sub> )Y-G-G-(d <sub>5</sub> )F-L	605.1818

**Fig. 1.** MALDI MS spectra of LE1 (top), LE2 (middle), and LE3 (bottom). nY = nitro-Tyr. F(d<sub>5</sub>) = Phe residue with five <sup>2</sup>H (d) atoms.

(top; [M+H]<sup>+</sup> at *m/z* 556.2), LE2 (middle; [M+H]<sup>+</sup> at *m/z* 601.1), and LE3 (bottom; [M+H]<sup>+</sup> at *m/z* 606.2). The results clearly showed that the [M+H]<sup>+</sup> ion of LE2 shifted 45 mass units (due to the presence of the nitro group) relative to LE1, and that the [M+H]<sup>+</sup> of LE3 shifted 5

mass units (d<sub>5</sub>) relative to LE2 and 50 mass units (nitro group + d<sub>5</sub>) relative to LE1. Each peptide also has a sodium-adduct ion: for LE1 (*m/z* 578.2), LE2 (*m/z* 623.2), and LE3 (*m/z* 628.2); the mass of each sodium-adduct ion also shifted appropriately.




Fig. 2. MS<sup>2</sup> spectra of LE1 (top), LE2 (middle), and LE3 (bottom). nY = nitro-Tyr. F(d<sub>5</sub>) = Phe residue with five <sup>2</sup>H (d) atoms.




Fig. 3. MS<sup>3</sup> spectra ( $b_4$  ion) of LE1 (top), LE2 (middle), and LE3 (bottom). nY = nitro-Tyr. F<sub>(d5)</sub> = Phe residue with five <sup>2</sup>H (d) atoms.

Compared to LE1, a unique decomposition pattern of ions ( $[M+H]^+ - 16$  and  $[M+H]^+ - 32$ ) was observed in the LE2 and LE3 spectra; that pattern is due to the photodecomposition of the nitro group ( $-NO_2$ ) to lose one and two oxygen atoms [4,21,22]. A sim-

ilar decomposition pattern for the loss of one/two oxygen atoms also occurred for the sodium adduct of LE2 and LE3. A product at  $[M+H]^+ - 30$  was also observed in the LE2 and LE3 spectra, which could result from the reduction of the nitro group ( $-NO_2$ ) to an

amino group ( $-\text{NH}_2$ ) [22]. Furthermore, the base-peak intensity of the  $[\text{M}+\text{H}]^+$  ion of LE1 ( $\text{NL}=1.01\text{E}5$ ) was higher than that of LE2 ( $\text{NL}=3.25\text{E}4$ ) and LE3 ( $\text{NL}=9.09\text{E}4$ ) (see discussion below; Fig. 4A).

### 3.2. $\text{MS}^2$ analysis of LE1, LE2, and LE3

Each  $[\text{M}+\text{H}]^+$  ion of LE1, LE2, and LE3 was selected as the precursor ion for MS/MS ( $\text{MS}^2$ ) analysis (Fig. 2). The amino acid sequence of each peptide is listed in the upper-left of each spectrum, and each product ion is underlined. A similar product-ion pattern was found in those three MS/MS spectra;  $y_2, y_3; b_2, b_3, b_4, b_5$ ; and  $a_4, a_5$  were detected for each peptide. The  $b_2$  ( $m/z$  221.4),  $b_3$  ( $m/z$  278.3),  $b_4$  ( $m/z$  425.1), and  $b_5$  ( $m/z$  538.1) ions of LE2 shifted 45 mass units relative to the corresponding  $b_2, b_3, b_4$ , and  $b_5$  ions of LE1. Also, the product ions  $y_2$  ( $m/z$  279.4) and  $y_3$  ( $m/z$  336.2) in LE2 are the same as LE1. This result confirmed that the nitro group ( $-\text{NO}_2$ ) was on  $Y_1$  in LE2.

The  $y_2$  ( $m/z$  284.4) and  $y_3$  ( $m/z$  341.2) ions of LE3 shifted 5 mass units relative to the corresponding  $y_2$  and  $y_3$  ions of LE1 and LE2. Also, the  $b_2$  and  $b_3$  ions of LE3 are the same as LE2, but the  $b_4$  and  $b_5$  ions of LE3 shifted 5 mass units relative to LE2 and shifted 50 mass units relative to LE1. Those results clearly confirmed the  $d_5\text{-F}_4$  in LE3.

Moreover, the abundant  $a_4$  and  $a_5$  ions in each spectrum (Fig. 2) further confirmed the presence of the  $b_4$  and  $b_5$  ions because an  $a$ -ion equals the loss of CO from a  $b$ -ion. Compared to the  $b_4$  and  $b_5$  ion in the  $\text{MS}^2$  spectra of LE1, LE2, and LE3, the same mass shift occurred in the  $a_4$  and  $a_5$  ions of LE1, LE2, and LE3. That result further confirmed the nitro group ( $-\text{NO}_2$ ) on  $Y_1$  in LE2 and LE3, and the  $d_5\text{-F}_4$  in LE3. Furthermore, the base-peak intensity of the product-ion spectrum of LE2 ( $\text{NL}=1.22\text{E}3$ ) and LE3 ( $\text{NL}=5.59\text{E}2$ ) was lower than that of LE1 ( $\text{NL}=3.36\text{E}3$ ) (see discussion below; Fig. 4B).

### 3.3. $\text{MS}^3$ analysis of LE1, LE2, and LE3

The  $b_4$  ion was the second most intense ion in the  $\text{MS}^2$  spectra of LE1, LE2, and LE3 (Fig. 2).  $\text{MS}^3$  analysis was performed on the  $b_4$  product ion of LE1, LE2, and LE3. Fig. 3 contains the  $\text{MS}^3$  spectra of the  $b_4$  product ion of LE1 (top), LE2 (middle), and LE3 (bottom). The amino acid sequence is listed in the upper-left corner of each spectrum. The  $\text{MS}^3$  ions ( $y, b$ , and  $a$ ) are labeled in each spectrum. From those three  $\text{MS}^3$  spectra, the range of base-peak intensity ( $5.28\text{E}1\text{--}9.81\text{E}1$ ) was significantly lower than the  $\text{MS}^2$  base-peak intensity range ( $5.59\text{E}2\text{--}3.36\text{E}3$ ). No  $y$ -ions were detected; only  $y_3\text{-H}_2\text{O}$  was detected. Ions at  $b_2, b_3, a_4$ , and  $b_4\text{-H}_2\text{O}$  were detected in each  $\text{MS}^3$  spectrum. Compared to the  $\text{MS}^2$  product ions (Fig. 2), the  $\text{MS}^3$  product ions demonstrated the same mass shift among LE1, LE2, and LE3.

The  $\text{MS}^3$   $b_2, b_3, b_4\text{-H}_2\text{O}$ , and  $a_4$  ions of LE2 shifted 45 mass units relative to LE1. The  $\text{MS}^3$   $b_2$  and  $b_3$  ions of LE3 were the same as LE2, but the  $b_4\text{-H}_2\text{O}$  and  $a_4$  ions of LE3 shifted 5 mass units relative to LE2 and 50 mass units relative to LE1. Also, the  $\text{MS}^3$   $y_3\text{-H}_2\text{O}$  ion was the same between LE1 and LE2, but the  $y_3\text{-H}_2\text{O}$  ion in LE3 shifted 5 mass units relative to LE1 and LE2. Those  $\text{MS}^3$  data are consistent with the  $\text{MS}^2$  data.

### 3.4. Effect of laser fluence on the fragmentation of LE1, LE2, and LE3

An appropriate laser fluence optimizes the ion signal of a precursor ion and the fragmentation sensitivity of the selected product-ion  $b_4$  of LE1, LE2, and LE3. The effect of laser fluence on the intensity of a precursor-ion and product-ion was measured. The  $b_4$  product ion (Fig. 2) was selected as the index for peptide fragmentation. The normalized collision energy (NCE) was fixed at 35




Fig. 4. The effect of laser fluence on the fragmentation of nitropeptides. (A) Relationship between laser fluence and the precursor-ion intensity ( $n=3$ ). (B) Relationship between laser fluence and the product-ion  $b_4$  intensity ( $n=3$ ).


units for the  $\text{MS}^2$  laser-fluence analysis. The data in Fig. 4A show the effect of laser fluence on the precursor-ion intensity of LE1, LE2, and LE3. Each precursor-ion intensity maximized at a laser fluence of 30 arbitrary units. However, the precursor-ion intensities of LE2 and LE3 were, respectively, 3.5-fold and 4.9-fold lower than LE1 due to the photochemical decompositions described above, because the presence of a nitro group in LE2 and LE3 is the only difference with LE1, and because data described above showed ion losses ( $-16, -32$ ) that derive only from a nitro group. That photochemical decomposition pattern is consistent with previous studies [4,21,22].

Fig. 4B shows the effect of laser fluence on the intensity of the  $b_4$  product-ion at NCE 35 units. The LE1 and LE2  $b_4$  ions maximized at a laser fluence of 40 units, and the LE3  $b_4$  ion at 45 units. Furthermore, the LE2 and LE3  $b_4$  ion intensities were, respectively, 7.1-fold and 11-fold lower than the LE1  $b_4$  ion; that difference is consistent with the result that the precursor-ion intensity of LE2 and LE3 was much lower than LE1 (Fig. 4A).

A laser fluence of 30 units was used for subsequent MS and  $\text{MS}^2$  experiments.

### 3.5. Effect of collision energy on the fragmentation of LE1, LE2, and LE3


The fragmentation of a precursor ion occurs within the collision-induced dissociation (CID) zone of a mass spectrometer. Collision energy significantly affects fragmentation efficiency [23–25]. Here,



**Fig. 5.** The effect of collision energy on the fragmentation of nitropeptides. (A) Relationship between collision energy and the product-ion intensity ( $n = 3$ ). (B) Relationship between collision energy and the product-ion  $b_4$  and  $a_4$  intensities ( $n = 3$ ).

we investigated the effect of collision energy on the fragmentation of a nitropeptide, and analyzed whether a nitro group affected the fragmentation at a constant laser fluence of 30 units. Fig. 5A shows the effect of NCE on the intensity of the  $MS^2$   $b_2$ ,  $b_3$ , and  $b_4$  prod-

uct ions. The top three panels (from left to right) show NCE versus intensity of those three product-ions for LE1; the middle three panels for LE2; and the bottom three panels for LE3. The LE1  $b_2$  ion had the (first peak) highest intensity at NCE 30, LE2 at 40, and LE3 at 40



**Fig. 6.** The vMALDI-LTQ sensitivity in the detection of nitropeptides. (A) Relationship between peptide amount and the precursor-ion intensity ( $n=3$ ). (B) Relationship between peptide amount and the product-ion  $b_4$  intensities ( $n=3$ ).

units. The LE1  $b_3$  ion maximized at NCE 30, LE2 at 40, and LE3 at 40 units. The LE1  $b_4$  ion maximized at NCE 20, LE2 at 20, and LE3 at 30 units. These results clearly showed that the generation of the  $b_2$  and  $b_3$  ions from LE2 and LE3, respectively, required a higher NCE than LE1 because of the presence of the nitro group in LE2 and LE3; and that the generation of the  $b_4$  ion of LE3 required a higher NCE than LE1 and LE2 because of the  $d_5$ -F<sub>4</sub> residue in LE3. However, the generation of the  $b_4$  ion from LE1, LE2, and LE3 required a lower NCE than the  $b_2$  and  $b_3$  ions. Perhaps the distance of the F-L bond from the  $\text{NO}_2$ -Y<sub>1</sub> residue affected the NCE-versus-intensity relationship. Nevertheless, a higher collision energy is required to fragment a nitropeptide.

Thus, a potential and significant problem is raised that, when an endogenous nitroprotein is MS/MS-identified from a complex proteome, a variable NCE cannot be set for each separate nitropeptide relative to its unmodified species; that limitation could affect the optimized fragmentation of nitropeptide relative to its unmodified species.


Clearly, for an MS/MS identification of an amino acid sequence, the complete  $b$ - and  $y$ -ion series are optimal. However, for the vMALDI-LTQ experiments on LE1, LE2, and LE3, the  $a_4$  and  $a_5$  ions are present in the MS/MS spectrum (Fig. 2). The  $[b\text{-CO}]$  ion is an  $a$ -ion. Therefore, the effect of NCE on the relative intensity of  $a$ - and  $b$ -ions was investigated (Fig. 5B). For the LE1  $b_4$  and  $a_4$  ions, when  $\text{NCE} < 20$  units, the  $b_4$  ion intensity is slightly higher than the  $a_4$  ion intensity, but when  $\text{NCE} > 20$  units, the  $b_4$  ion intensity was lower than the  $a_4$  ion intensity. For the LE2  $b_4$  and  $a_4$  ions, when  $\text{NCE} < 32$  units, the  $b_4$  ion intensity was higher than the  $a_4$  ion intensity, but when  $\text{NCE} > 32$  units, the  $a_4$  ion intensity was higher than the  $b_4$  ion intensity. For the LE3  $b_4$  and  $a_4$  ions, when  $\text{NCE} < 35$  units, the  $b_4$  ion

intensity was higher than the  $a_4$  ion intensity, but when  $\text{NCE} > 35$  units, the  $a_4$  ion intensity was higher than the  $b_4$  ion intensity. Those results clearly showed that a higher NCE produced more  $a$ -ions and fewer  $b$ -ions. That result is reasonable, and reflects for the  $a_4$  ion, the energetically favorable loss of a molecule ( $\text{CO}$ ;  $\Delta H_f = -26$  kcal) from the  $b_4$ -ions.

### 3.6. The detection sensitivity of a precursor-ion and product-ion of LE1, LE2, and LE3

Sensitivity is the key issue for the MS-detection of endogenous low-abundance nitroproteins (and for their future quantification). The optimized experimental conditions described above were used to determine the sensitivity of vMALDI-LTQ to detect the synthetic nitropeptides LE2 and LE3. A series of peptide dilutions (5000, 1000, 500, 100, 50, 5, and 1 fmol) was analyzed. Data in Fig. 6A show the sensitivity to detect the precursor ion of LE1, LE2, and LE3. The precursor ion of each peptide at 1 fmol can be effectively detected, and an excellent MS spectrum was obtained. When the peptide amount was  $>500$  fmol, no linear response relationship existed between the precursor-ion intensity and peptide amount. Fig. 6B shows the sensitivity to detect the  $b_4$  ion of LE1, LE2, and LE3. The  $b_4$  ion can be detected at 1 fmol for LE1, and at 10 fmol for LE2 and LE3; and a good MS<sup>2</sup> spectrum was obtained (Fig. 7). The S/N of those ions in Fig. 7a–c are ca. 40–50:1. When the peptide amount was  $>500$  fmol, no linear response relationship existed between the  $b_4$  ion intensity and peptide amount. Therefore, the sensitivity to detect the  $b_4$  ion of LE2 and LE3 is lower than for LE1; that difference is consistent with the result that photochemical decomposition decreased the precursor-ion intensity of LE2 and LE3 relative to LE1.





**Fig. 7.** The MS<sup>2</sup> sensitivity to detect LE1 (1 fmol) (top), LE2 (10 fmol) (middle), and LE3 (10 fmol) (bottom). The S/N was ca. 50:1.

#### 4. Conclusions

Tyrosine nitration is an important redox-mediated protein post-translational modification, is associated with a wide-range of diseases, and is involved in physiological and pathological processes. A variety of protein chemical approaches have been used to detect endogenous nitroproteins and nitropeptides [3,5,6,16,26–28]. However, mass spectrometry-based proteomics is an effective, sensitive, and specific method to detect and characterize the sites of nitrotyrosines in endogenous proteins. The MS method has been documented [3–15], including studies of human pituitary adenoma, diabetic retina, and lung disease. However, usually only one, or a few, nitropeptides were identified for each endogenous nitroprotein, and a low protein-coverage was found. Therefore, a high-sensitivity mass spectrometer and high-quality MS/MS spectra are needed.

This present study used vMALDI-LTQ to study the fragmentation of a synthetic peptide and two synthetic nitropeptides in order to assist in the interpretation of MS spectra of endogenous nitropeptides. We analyzed synthetic leucine enkephalin (LE1), nitrotyrosine-containing leucine enkephalin (LE2), and nitrotyrosine-containing d<sub>5</sub>-F<sub>4</sub> leucinine enkephalin (LE3). First, the results showed that UV laser-induced photochemical decomposition of the nitro group (loss of 16, 32 Da) decreased the precursor-ion intensity, and complicated the MS spectrum. However, in turn, the recognition of the unique photochemical decomposition pattern significantly assisted in the identification of a nitropeptide and the location of the nitrotyrosine site. Second, the optimal fragmentation of the two nitropeptides (LE2, LE3) required more collision energy compared to the unmodified peptide LE1. However, more collision energy increased the intensity of a-ions and decreased the intensity of b-ions. Third, MS<sup>3</sup> analysis confirmed the MS<sup>2</sup>-derived amino acid sequence, although an MS<sup>3</sup> analysis requires a higher amount of peptides relative to MS<sup>2</sup>. Therefore, MS<sup>3</sup> analysis might not be suitable for routine analysis of endogenous low-abundance nitroproteins. Only when a target is determined can MS<sup>3</sup> be used for confirmation. Fourth, the optimization of collision energy could depend on the position of nitrotyrosine and length of the peptide, whereas optimization of the laser fluence could rely on the nature of matrix and the ratio of analyte to matrix because laser beam mainly excites the production of the precursor ion ([M+H]<sup>+</sup>) and might not contribute to the fragmentation of the C–N bond of nitropeptide. Fifth, vMALDI-LTQ MS could detect (with a S/N of ~50:1) 1 fmol of a nitropeptide, and MS<sup>2</sup> 10 fmol nitropeptide and 1 fmol unmodified peptide, because the photochemical decomposition of the nitro group decreased the sensitivity of fragmentation of nitropeptides. Finally, for vMALDI-LTQ, the laser fluence should be evaluated periodically with a standard peptide so that an effective laser energy is used because a laser degrades with time.

#### Acknowledgements

The authors acknowledge the financial support from Chiesi Pharmaceuticals (Parma, Italy) and NIH (RR016679 to DMD). We thank Dr. Peter Schiller for the synthetic peptides.

#### References

- [1] I. Dalle-Donne, A. Scaloni, D. Giustarini, E. Cavarra, G. Tell, G. Lungarella, R. Colombo, R. Rossi, A. Milzani, *Mass Spectrom. Rev.* 24 (2005) 55.
- [2] I. Dalle-Donne, A. Scaloni, D.A. Butterfield (Eds.), *Redox Proteomics: From Protein Modifications to Cellular Dysfunction and Diseases*, John Wiley & Sons, Inc., Hoboken, New Jersey, 2006.
- [3] X. Zhan, D.M. Desiderio, *Biochem. Biophys. Res. Commun.* 325 (2004) 1180.
- [4] X. Zhan, D.M. Desiderio, *Int. J. Mass Spectrom.* 259 (2007) 96.
- [5] X. Zhan, D.M. Desiderio, *Anal. Biochem.* 354 (2006) 279.
- [6] X. Zhan, Y. Du, J.S. Crabb, X. Gu, T.S. Kern, J.W. Crabb, *Mol. Cell. Proteomics* 7 (2008) 864.
- [7] V. Justilien, J.J. Pang, K. Renganathan, X. Zhan, J.W. Crabb, S.R. Kim, J.R. Sparrow, W.W. Hauswirth, A.S. Lewin, *Invest. Ophthalmol. Vis. Sci.* 48 (2007) 4407.
- [8] M. Aslan, T.M. Ryan, T.M. Townes, L. Coward, M.C. Kirk, S. Barnes, C.B. Alexander, S.S. Rosenfeld, B.A. Freeman, *J. Biol. Chem.* 278 (2003) 4194.
- [9] F. Casoni, M. Basso, T. Massignan, E. Gianazza, C. Cheroni, M. Salmona, C. Bendotti, V. Bonetto, *J. Biol. Chem.* 280 (2005) 16295.
- [10] J. Kanski, S.J. Hong, C. Schoneich, *J. Biol. Chem.* 280 (2005) 24261.
- [11] V.S. Sharov, N.A. Galeva, J. Kanski, T.D. Williams, C. Schoneich, *Exp. Gerontol.* 41 (2006) 407.
- [12] J. Kanski, A. Behring, J. Pelling, C. Schoneich, *Am. J. Physiol. Heart Circ. Physiol.* 288 (2005) H371.
- [13] C.A. Sacksteder, W.J. Qian, T.V. Knyushko, H.W. Wang, M.H. Chin, G. Lacan, W.P. Melega, D.G. Camp II, R.D. Smith, D.J. Smith, T.C. Squier, D.J. Bigelow, *Biochemistry* 45 (2006) 8009.
- [14] G. Fiorce, C.D. Cristo, G. Monti, A. Amoresano, L. Columbano, P. Pucci, F.A. Cioffi, A.D. Cosmo, A. Palumbo, M. d'Ischia, *Neurosci. Lett.* 394 (2006) 57.
- [15] S. Lanone, P. Manivet, J. Callebort, J.M. Launay, D. Payen, M. Aubier, J. Boczkowski, A. Mebazaa, *Biochem. J.* 366 (2002) 399.
- [16] M.K. Shigenaga, H.H. Lee, B.C. Blount, S. Christen, E.T. Shigeno, H. Yip, B.N. Ames, *Proc. Natl. Acad. Sci. U.S.A.* 94 (1997) 3211.
- [17] B. Ghesquiere, M. Goethals, J. Van Damme, A. Staes, E. Timmerman, J. Vandekerckhove, K. Gevaert, *Rapid Commun. Mass Spectrom.* 20 (2006) 2885.
- [18] Q. Zhang, W.J. Qian, T.V. Knyushko, T.R. Claus, S.O. Purvine, R.J. Moore, C.A. Sacksteder, M.H. Chin, D.J. Smith, D.G. Camp 2nd, D.J. Bigelow, R.D. Smith, *J. Proteome Res.* 6 (2007) 2257.
- [19] A. Amoresano, G. Chiappetta, P. Pucci, M. D'Ischia, G. Marino, *Anal. Chem.* 79 (2007) 2109.
- [20] Y.K. Butt, S.C. Lo, *Methods Enzymol.* 440 (2008) 17.
- [21] A.S. Petersson, H. Steen, D.E. Kalume, K. Caidahl, P. Roepstorff, *J. Mass Spectrom.* 36 (2001) 616.
- [22] A. Sarver, N.K. Scheffler, M.D. Shetlar, B.W. Gibson, *J. Am. Soc. Mass Spectrom.* 12 (2001) 439.
- [23] O.V. Borisov, M.B. Goshe, T.P. Conrads, V.S. Rakov, T.D. Veenstra, R.D. Smith, *Anal. Chem.* 74 (2002) 2284.
- [24] H. Yin, A. Chacon, N.A. Porter, D.S. Masterson, *J. Am. Soc. Mass Spectrom.* 18 (2007) 807.
- [25] H. Lioe, R.A. O'Hair, *Anal. Bioanal. Chem.* 389 (2007) 1429.
- [26] K.S. Aulak, M. Miyagi, L. Yan, K.A. West, D. Massillon, J.W. Crabb, D.J. Stuehr, *Proc. Natl. Acad. Sci. U.S.A.* 98 (2001) 12056.
- [27] J. Khan, D.M. Brennan, N. Bradley, B. Gao, R. Bruckdorfer, M. Jacobs, *Biochem. J.* 330 (1998) 895.
- [28] I.Y. Haddad, G. Pataki, P. Hu, C. Galliani, J.S. Beckman, S. Matalon, *J. Clin. Invest.* 94 (1994) 2407.