Note

The Density of 1,1-Dichloro-l-Fluoroethane (HCFC 141b)

A. T. Sousa,¹ P. S. Fialho,² and C. A. Nieto de Castro¹

Receil,ed September 9, 1993

In this Note we present the density of HCFC 141b, measured between 293.15 and 300.15 K, with an mechanical oscillator densimeter, with an uncertainty of 0.007%. The results are compared with the densities estimated by the reduced hard-sphere-DeSantis equation of state and with the experimental data obtained by several authors.

KEY WORDS: density; l,l-dichloro-I-fluoroethane; equation of state; HCFC 14lb.

The density of HCFC 141b was measured between 293.15 and 300.15 K (ITS90) with a mechanical oscillator densimeter (Anton Paar type DMA 55) at atmospheric pressure with an uncertainty of 0.04 kg \cdot m⁻³. or 0.007%. For this work the densimeter was calibrated with toluene and 2,2,4-trimethylpentane, between 293.15 and 323.15 K. Details of the calibrations and of the working equation for this apparatus are described in Refs. 1 and 2. The density of 2,2,4-trimethylpentane was obtained from the IUPAC recommendation [3].

The data obtained are compared with the densities estimated with the reduced hard-sphere-DeSantis equation of state (RHSDS), developed by Fialho [4], and with the experimental data obtained by different authors [5-7]. Table I shows the densities obtained in this work corrected for nominal temperatures, the predictions of the RHSDS, and the results obtained by other authors [6, 7]. The results in this work for temperatures

^t Departamento de Química, Faculdade de Ciências, Universidade de Lisboa, Campo Grande, Bloco CI, 1700 Lisboa, Portugal.

² Departamento de Quimica, Universidade de Évora, Largo dos Colegiais 2, 7000 Évora, Portugal.

| T(K) | ρ (kg = m ⁻³) | | | |
|--------|--------------------------------|-------------------|---------------------|--------|
| | This work | RHSDS | Ref. 6 ^a | Ref. 7 |
| 273.15 | 1283.54 | 1289.21" | 1281.2 | |
| 283.15 | 1263.48 | 1271.01 | 1262.5 | |
| 288.15 | 1253.56 | 1261.78 | | |
| 293.15 | 1243.68 | 1252.49 | 1243.0 | |
| 298.15 | 1233.86 | 1243.18 | 1233.2 | 1232 |
| 300.15 | 1229.94 | 1239.32 | | |
| 303.15 | 1224.07 | 1233.71 | 1223.2 | |
| 313.15 | | 1214.59" | 1203.7 | |
| 323.15 | | 1195.08" | 1183.7 | |
| 333.15 | | 1175.08" | 1162.4 | |
| 343.15 | | 1154.45° | 1141.4 | |
| 353.15 | | 1133.05" | 1118.8 | |

Table I. Density of HCFC 141b at 0.1 MPa

"Density of the liquid at the saturation vapor pressure.

below 293.15 K and higher than 300.15 K were obtained by extrapolation of the experimental dependence of the period of vibration of the oscillator in the densimeter cell [2]. It can be seen that the agreement between our experimental data and those of Kumagai Takahashi [6] is well within the claimed uncertainties. Although their data were obtained at the saturation

Fig. I. Deviations of the density of HCFC 141b from the values predicted by the RHSDS. Deviations = $(\rho_{exp} - \rho_{RHSDS})/\rho_{RHSDS} \times 100$.

vapor pressure, which is very close to 0.1 M Pa, this comparison does not introduce an error greater than the uncertainty of the experimental data obtained with the pycnometer method. The result of Sukornick [7] is 0.15% lower than our data.

To calculate the densities with the RHSDS at the same nominal temperatures and pressure as in the work by Kumagai and Takahashi [6], it is necessary to know the saturation vapor pressure. These values were obtained from the work of Maezewa et al. [5].

Figure I shows the deviations between the available experimental data and the data predicted by the RHSDS. The maximum deviation between our data and the RHSDS is -0.8% . The figure also shows that this EOS can predict the data obtained by Maezawa et al. [5] for the compressed liquid within -1.7% and the data of Kumagai Takahashi [6] for the liquid at the saturation vapor pressure within a maximum deviation of -1.3% . These results are very encouraging because, as described in Ref. 4, the critical parameters of HCFC 141b necessary to apply the theory were estimated $\lceil 5 \rceil$.

ACKNOWLEDGMENT

The authors would like to thank SOLVAY FLUOR UND DERIVATE, Hannover, for supplying the HCFC 141b, with an estimated purity of 99.9 %.

REFERENCES

- 1. A. T. Sousa, C. A. Nieto de Castro, and J. A. Lourenço, *High Temp.-High Press.* 21:149 (1989).
- 2. A. T. Sousa, Tese de Doutoramento (Universidade de Lisboa, Lisboa, 1994).
- 3. K. N. Marsh, *Recommended Reference Materials for the Realization of Physicochemical Properties* (IUPAC, Blackwell Scientific, Oxford, 1988).
- 4. P. S. Fialho, Tese de Doutoramento (Universidade de Lisboa, Lisboa, 1993).
- 5. Y. Maezawa, H. Sato, and K. Watanabe, *J. Chem. Eng. Data* 36:151 (1991).
- 6. A. Kumagai and S. Takahashi, *hit. J. Thermophys.* 14:339 (1993).
- 7. B. Sukornick, *Int. J. Thermophys.* **10:553** (1989).