



ELSEVIER

Journal of Chromatography A, 868 (2000) 321

JOURNAL OF
CHROMATOGRAPHY A

www.elsevier.com/locate/chroma

Erratum

Erratum to: "Optimization of conductivity of low-molecular-mass anions in capillary zone electrophoresis" [J. Chromatogr. A, 850 (1999) 355–362][☆]

Martin U. Katzmayer, Christian W. Klampfl*, Wolfgang Buchberger

Department of Analytical Chemistry, Johannes-Kepler-University Linz, Altenbergersatrasse 69, A-4040 Linz, Austria

Received 2 November 1999

Keywords: Conductivity detection; Detection, electrophoresis; Background electrolyte composition; Organic acids; Inorganic anions

Page 361, Table 2 should read as follows:

Table 2

Peak areas obtained for a 10 mg/l formate standard solution as a function of the electrolyte counter-ion mobility

Counter-ion	$\mu(\text{counter-ion})$ ($\text{cm}^2 \text{ kV}^{-1} \text{ s}^{-1}$)	Peak area (arbitrary units)
K ⁺	0.693	146 518
Na ⁺	0.474	143 382
Li ⁺	0.363	132 369
TEA ⁺	0.297	123 032
Tris ⁺	0.259	120 203
Arg ⁺	0.229	117 865

[☆]PII of original article: S0021-9673(99)00168-5

*Corresponding author. +43-732-2468-722; +61-732-2468-679.

E-mail address: christian.kampfl@jk.uni-linz.ac.at (C.W. Klampfl)