





www.elsevier.com/locate/jmr

Journal of Magnetic Resonance 182 (2006) 178

Corrigendum

Corrigendum to "Radiation damping in micro-coil NMR probes" [J. Magn. Reson. 179 (2006) 294–298]

V.V. Krishnan

Department of Molecular, Cell and Developmental Biology, University of California, Santa Cruz, CA 95064, USA Department of Applied Science and Center for Comparative Medicine, University of California, Davis, CA 95616, USA

Available online 10 July 2006

In the discussion section, to determine which factor (η , filling factor, or Q_c , quality factor) contributes to the radiation damping effects, the estimate of ($Q_c^{\text{5mm}}/Q_c^{\text{cap}}$) must be corrected to ($Q_c^{\text{cap}}/Q_c^{\text{5mm}}$) = 19.7. This suggests that in both the Cap-NMR probe and cryogenically cooled probe, the increase in radiation damping is due to an increase in the quality factor. I apologize for this error, which does not change the results or any other conclusions of the paper.

DOI of original article: 10.1016/j.jmr.2005.12.011.

E-mail address: vvkrish@gmail.com.