

Erratum

J. H. Beijnen *et al.*, Gas chromatographic analysis, with electron capture detection, of the antifungal drug fluconazole in microvolumes of human plasma. *J. Pharm. Biomed. Anal.* **9**, 1173–1175 (1991).

The publishers regret that errors occurred in Table 1 of this paper. The corrected table is given below.

Table 1

Equations of calibration lines, accuracy and precision for the bio-analysis of fluconazole

Conc. range (ng ml ⁻¹)	Equation		<i>r</i> ²	<i>n</i>
10–100	$y^1 = 0.423 (\pm 0.063)$	+ 0.0124 (± 0.0008) <i>x</i>	0.9898	8
100–1000	$y^2 = 0.0032 (\pm 0.0031)$	+ 0.00024 (± 0.000003) <i>x</i>	0.9991	8

Theoretical conc. (ng ml ⁻¹)	Measured conc. (ng ml ⁻¹)	Accuracy (%)	RSD*	<i>n</i>
49.4	45.2	91.5	9.2%	4
494.0	489.3	99.1	4.8%	4

*y*¹ is the peak height ratio between fluconazole and the external standard medazepam, *y*² is the peak area ratio between fluconazole and medazepam and *x* is the fluconazole concentration in ng ml⁻¹.

* RSD: relative standard deviation.