

This article was downloaded by:

On: 25 January 2011

Access details: *Access Details: Free Access*

Publisher *Taylor & Francis*

Informa Ltd Registered in England and Wales Registered Number: 1072954 Registered office: Mortimer House, 37-41 Mortimer Street, London W1T 3JH, UK



Journal of Macromolecular Science, Part A

Publication details, including instructions for authors and subscription information:

<http://www.informaworld.com/smpp/title~content=t713597274>

Nomenclature in Thermal Analysis

R. C. Mackenzie^a

^a The Macaulay Institute for Soil Research, Aberdeen, Scotland

To cite this Article Mackenzie, R. C.(1970) 'Nomenclature in Thermal Analysis', Journal of Macromolecular Science, Part A, 4: 4, 1015 – 1020

To link to this Article: DOI: 10.1080/00222337008060995

URL: <http://dx.doi.org/10.1080/00222337008060995>

PLEASE SCROLL DOWN FOR ARTICLE

Full terms and conditions of use: <http://www.informaworld.com/terms-and-conditions-of-access.pdf>

This article may be used for research, teaching and private study purposes. Any substantial or systematic reproduction, re-distribution, re-selling, loan or sub-licensing, systematic supply or distribution in any form to anyone is expressly forbidden.

The publisher does not give any warranty express or implied or make any representation that the contents will be complete or accurate or up to date. The accuracy of any instructions, formulae and drug doses should be independently verified with primary sources. The publisher shall not be liable for any loss, actions, claims, proceedings, demand or costs or damages whatsoever or howsoever caused arising directly or indirectly in connection with or arising out of the use of this material.

LETTER TO THE EDITOR

Nomenclature in Thermal Analysis

Nomenclature in thermal analysis is neither uniform nor consistent and can at times be confusing. Because of this a committee consisting of Dr. R. C. Mackenzie (Chairman), Mr. C. J. Keatch (Secretary), Dr. J. P. Redfern, and Dr. A. A. Hodgson was appointed at the First International Conference on Thermal Analysis to explore this field. A report, arrived at after consultation with experts in all major English-speaking countries and in other countries interested in this aspect, was approved in principle at the Second International Conference on Thermal Analysis and the Council of the International Confederation for Thermal Analysis (ICTA) have directed that it be published.

This first report is offered as a definitive document of ICTA, the recommendations in which ought to be adhered to in all publications in the English language. It is appreciated that this is only a beginning and that many aspects still require attention; furthermore, new developments in the science may lead to minor revision. Such matters will be the subject of later reports.

Since linguistic considerations render difficult universal application of terms and it may be that names unacceptable in one language are normal usage in another, sub-committees are at present considering the position regarding the French, German, Japanese, and Russian languages. The decisions of these sub-committees will be published later and developments in the field of nomenclature will be reported from time to time in the *ICTA Newsletter*.

I. GENERAL RECOMMENDATIONS

(a) *Thermal analysis* and not "thermography" should be the acceptable name in English, since the latter has at least two other meanings in this language, the major one being medical [*Sci. Progr.* (London), 55, 167 (1967)]. The adjective should then be *thermoanalytical* (cf. physical chemistry and physicochemical): the term "thermoanalysis" is not supported (on the same logical basis).

(b) *Differential* should be the adjectival form of *difference*; *derivative* should be used for the first derivative (mathematical) of any curve.

(c) The term "analysis" should be avoided as far as possible since the methods considered do not comprise analysis as generally understood chemically: terms such as *differential thermal analysis* are too widely accepted, however, to be changed.

(d) The term *curve* is preferred to "thermogram" for the following reasons:

1. "Thermogram" is used for the results obtained by the medical technique of thermography—see (a).
2. If applied to certain curves (e.g., thermogravimetric curves), "thermogram" would not be consistent with the dictionary definition.
3. For clarity there would have to be frequent use of terms such as differential thermogram and thermogravimetric thermogram, which are not only cumbersome but also confusing.

(e) In multiple techniques, *simultaneous* should be used for the application of two or more techniques to the same sample at the same time: *combined* would then indicate the use of separate samples for each technique.

(f) *Thermal decomposition* and similar terms are being further considered by the Committee.

II. TERMINOLOGY

Acceptable names and abbreviations, together with names which were for various reasons rejected, are listed in Table 1. The Committee are in accord with the suggestion, made during discussion of the report, that the limited number of abbreviations considered permissible should be adopted internationally, irrespective of language.

The committee do not wish to pronounce on nomenclature in borderline techniques (such as thermometric titrimetry or calorimetry) which are, to its knowledge, being considered by other bodies. Consideration of techniques not yet extensively employed has been deferred.

III. DEFINITIONS AND CONVENTIONS

A. General

Thermal analysis. A general term covering a group of related techniques

Table 1. Recommended Terminology

Acceptable name	Acceptable abbreviation ^a	Rejected name(s)
A. General		
Thermal analysis		Thermography Thermoanalysis
B. Methods associated with weight change		
1. Static		
Isobaric weight-change determination		
Isothermal weight-change determination		Isothermal thermo- gravimetric analysis
2. Dynamic		
Thermogravimetry	TG	Thermogravimetric analysis Dynamic thermo- gravimetric analysis
Derivative thermogravimetry	DTG	Differential thermo- gravimetry Differential thermo- gravimetric analysis Derivative thermo- gravimetric analysis
C. Methods associated with energy change		
Heating curves ^b		Thermal analysis
Heating-rate curves ^b		Derivative thermal analysis
Inverse heating-rate curves ^b		
Differential thermal analysis	DTA	Dynamic differential calorimetry
Derivative differential thermal analysis		
Differential scanning calorimetry	DSC	
D. Methods associated with evolved volatiles		
Evolved gas detection	EGD	Effluent gas detection
Evolved gas analysis ^c	EGA	Effluent gas analysis Thermovaporimetric analysis
E. Methods associated with dimensional change		
Dilatometry		
Derivative dilatometry		
Differential dilatometry		

(continued)

Table 1. Recommended Terminology (continued)

Acceptable name	Acceptable abbreviation ^a	Rejected name(s)
F. Multiple techniques Simultaneous TG and DTA, etc.		DATA (differential and thermogravimetric analysis) Derivatography Derivatographic analysis

^aAbbreviations should be in capital letters without full-stops, and should be kept to the minimum to avoid confusion.

^bWhen determinations are performed during the cooling cycle these become Cooling curves, Cooling-rate curves, and Inverse cooling-rate curves, respectively.

^cThe method of analysis should be clearly stated and abbreviations such as MTA (mass-spectrometric thermal analysis) and MDTA (mass spectrometry and differential thermal analysis) avoided.

whereby the dependence of the parameters of any physical property of a substance on temperature is measured.

B. Methods Associated with Weight Change

1. *Static. Isobaric weight-change determination.* A technique of obtaining a record of the equilibrium weight of a substance as a function of temperature (T) at a constant partial pressure of the volatile product or products.

The record is the isobaric weight-change curve; it is normal to plot weight on the ordinate with weight decreasing downwards and T on the abscissa increasing from left to right.

Isothermal weight-change determination. A technique of obtaining a record of the dependence of the weight of a substance on time (t) at constant temperature.

The record is the isothermal weight-change curve; it is normal to plot weight on the ordinate with weight decreasing downwards and t on the abscissa increasing from left to right.

2. *Dynamic. Thermogravimetry (TG).* A technique whereby the weight of a substance, in an environment heated or cooled at a controlled rate, is recorded as a function of time or temperature.

The record is the thermogravimetric of TG curve; the weight should be plotted on the ordinate with weight decreasing downwards and t or T on the abscissa increasing from left to right.

Derivative thermogravimetry (DTG). A technique yielding the first derivative of the thermogravimetric curve with respect to either time or temperature.

The curve is the derivative thermogravimetric or DTG curve; the derivative should be plotted on the ordinate with weight losses downwards and t or T on the abscissa increasing from left to right.

C. Methods Associated with Energy Change

Heating curves. These are records of the temperature of a substance against time, in an environment heated at a controlled rate.

T should be plotted on the ordinate increasing upwards and t on the abscissa increasing from left to right.

Heating-rate curves. These are records of the first derivative of the heating curve with respect to time (i.e., dT/dt) plotted against time or temperature.

The function dT/dt should be plotted on the ordinate and t or T on the abscissa increasing from left to right.

Inverse heating-rate curves. These are records of the first derivative of the heating curve with respect to temperature (i.e., dt/dT) plotted against either time or temperature.

The function dt/dT should be plotted on the ordinate and t or T on the abscissa increasing from left to right.

Differential thermal analysis (DTA). A technique of recording the difference in temperature between a substance and a reference material against either time or temperature as the two specimens are subjected to identical temperature regimes in an environment heated or cooled at a controlled rate.

The record is the differential thermal or DTA curve; the temperature difference (ΔT) should be plotted on the ordinate with endothermic reactions downwards and t or T on the abscissa increasing from left to right.

Derivative differential thermal analysis. A technique yielding the first derivative of the differential thermal curve with respect to either time or temperature.

The record is the derivative differential thermal or derivative DTA curve; the derivative should be plotted on the ordinate and t or T on the abscissa increasing from left to right.

Differential scanning calorimetry (DSC). A technique of recording the energy necessary to establish zero temperature difference between a substance and a reference material against either time or temperature as the two specimens are subjected to identical temperature regimes in an environment heated or cooled at a controlled rate.

The record is the DSC curve; it represents the amount of heat applied per unit time as ordinate against either t or T as abscissa.

D. Methods Associated with Evolved Volatiles

Evolved gas detection (EGD). This term covers any technique of detecting whether or not a volatile product is formed during thermal analysis.

Evolved gas analysis (EGA). A technique of determining the nature and/or amount of volatile product or products formed during thermal analysis.

E. Methods Associated with Dimensional Change

Dilatometry. A technique whereby changes in dimension(s) of a substance are measured as a function of temperature.

The record is the dilatometric curve.

Derivative dilatometry; differential dilatometry. These terms carry the connotations given in I(b) above.

F. Multiple Techniques

This term covers simultaneous DTA and TG, etc., and definitions follow from the above.

ACKNOWLEDGMENTS

The Committee express their thanks to the Society for Analytical Chemistry and the Thermal Analysis Group of that Society for assistance rendered and also to thermal analysts in many countries for detailed comments at various stages of the program. Reprinted by permission from Pergamon Press.

R. C. Mackenzie
Chairman, ICTA Nomenclature Committee

*The Macaulay Institute for Soil Research
Craigiebuckler
Aberdeen, Scotland*

*Accepted by editor January 19, 1970
Received for publication February 20, 1970*