

Computer-Based Instruments

NI 4551/4552 User Manual

Dynamic Signal Acquisition Instruments for PCI

Internet Support

E-mail: support@natinst.com

FTP Site: <ftp.natinst.com>

Web Address: <http://www.natinst.com>

Bulletin Board Support

BBS United States: 512 794 5422

BBS United Kingdom: 01635 551422

BBS France: 01 48 65 15 59

Fax-on-Demand Support

512 418 1111

Telephone Support (USA)

Tel: 512 795 8248

Fax: 512 794 5678

International Offices

Australia 03 9879 5166, Austria 0662 45 79 90 0, Belgium 02 757 00 20, Brazil 011 288 3336,
Canada (Ontario) 905 785 0085, Canada (Québec) 514 694 8521, Denmark 45 76 26 00, Finland 09 725 725 11,
France 01 48 14 24 24, Germany 089 741 31 30, Hong Kong 2645 3186, Israel 03 6120092, Italy 02 413091,
Japan 03 5472 2970, Korea 02 596 7456, Mexico 5 520 2635, Netherlands 0348 433466, Norway 32 84 84 00,
Singapore 2265886, Spain 91 640 0085, Sweden 08 730 49 70, Switzerland 056 200 51 51, Taiwan 02 377 1200,
United Kingdom 01635 523545

National Instruments Corporate Headquarters

6504 Bridge Point Parkway Austin, Texas 78730-5039 USA Tel: 512 794 0100

Important Information

Warranty

The NI 4551 and NI 4552 are warranted against defects in materials and workmanship for a period of one year from the date of shipment, as evidenced by receipts or other documentation. National Instruments will, at its option, repair or replace equipment that proves to be defective during the warranty period. This warranty includes parts and labor.

The media on which you receive National Instruments software are warranted not to fail to execute programming instructions, due to defects in materials and workmanship, for a period of 90 days from date of shipment, as evidenced by receipts or other documentation. National Instruments will, at its option, repair or replace software media that do not execute programming instructions if National Instruments receives notice of such defects during the warranty period. National Instruments does not warrant that the operation of the software shall be uninterrupted or error free.

A Return Material Authorization (RMA) number must be obtained from the factory and clearly marked on the outside of the package before any equipment will be accepted for warranty work. National Instruments will pay the shipping costs of returning to the owner parts which are covered by warranty.

National Instruments believes that the information in this manual is accurate. The document has been carefully reviewed for technical accuracy. In the event that technical or typographical errors exist, National Instruments reserves the right to make changes to subsequent editions of this document without prior notice to holders of this edition. The reader should consult National Instruments if errors are suspected. In no event shall National Instruments be liable for any damages arising out of or related to this document or the information contained in it.

EXCEPT AS SPECIFIED HEREIN, NATIONAL INSTRUMENTS MAKES NO WARRANTIES, EXPRESS OR IMPLIED, AND SPECIFICALLY DISCLAIMS ANY WARRANTY OF MERCHANTABILITY OR FITNESS FOR A PARTICULAR PURPOSE. CUSTOMER'S RIGHT TO RECOVER DAMAGES CAUSED BY FAULT OR NEGLIGENCE ON THE PART OF NATIONAL INSTRUMENTS SHALL BE LIMITED TO THE AMOUNT THEREOF PAID BY THE CUSTOMER. NATIONAL INSTRUMENTS WILL NOT BE LIABLE FOR DAMAGES RESULTING FROM LOSS OF DATA, PROFITS, USE OF PRODUCTS, OR INCIDENTAL OR CONSEQUENTIAL DAMAGES, EVEN IF ADVISED OF THE POSSIBILITY THEREOF. This limitation of the liability of National Instruments will apply regardless of the form of action, whether in contract or tort, including negligence. Any action against National Instruments must be brought within one year after the cause of action accrues. National Instruments shall not be liable for any delay in performance due to causes beyond its reasonable control. The warranty provided herein does not cover damages, defects, malfunctions, or service failures caused by owner's failure to follow the National Instruments installation, operation, or maintenance instructions; owner's modification of the product; owner's abuse, misuse, or negligent acts; and power failure or surges, fire, flood, accident, actions of third parties, or other events outside reasonable control.

Copyright

Under the copyright laws, this publication may not be reproduced or transmitted in any form, electronic or mechanical, including photocopying, recording, storing in an information retrieval system, or translating, in whole or in part, without the prior written consent of National Instruments Corporation.

Trademarks

ComponentWorks™, CVI™, LabVIEW™, Measure™, NI-DAQ™, and VirtualBench™ are trademarks of National Instruments Corporation.

Product and company names listed are trademarks or trade names of their respective companies.

WARNING REGARDING MEDICAL AND CLINICAL USE OF NATIONAL INSTRUMENTS PRODUCTS

National Instruments products are not designed with components and testing intended to ensure a level of reliability suitable for use in treatment and diagnosis of humans. Applications of National Instruments products involving medical or clinical treatment can create a potential for accidental injury caused by product failure, or by errors on the part of the user or application designer. Any use or application of National Instruments products for or involving medical or clinical treatment must be performed by properly trained and qualified medical personnel, and all traditional medical safeguards, equipment, and procedures that are appropriate in the particular situation to prevent serious injury or death should always continue to be used when National Instruments products are being used. National Instruments products are NOT intended to be a substitute for any form of established process, procedure, or equipment used to monitor or safeguard human health and safety in medical or clinical treatment.

Contents

About This Manual

Organization of This Manual	ix
Conventions Used in This Manual.....	x
National Instruments Documentation	xi
Related Documentation.....	xi
Customer Communication	xi

Chapter 1

Introduction

What You Need to Get Started	1-2
Unpacking	1-2
Software Programming Choices	1-3
National Instruments Application Software	1-3
Optional Equipment	1-4
Custom Cabling	1-5

Chapter 2

Installation and Configuration

Software Installation	2-1
Hardware Installation.....	2-1
Instrument Configuration.....	2-2

Chapter 3

Hardware Overview

Analog Input	3-4
Input Mode	3-4
Input Coupling.....	3-4
Input Polarity and Input Range.....	3-4
Considerations for Selecting Input Ranges	3-5
Analog Output.....	3-6
Output Mode.....	3-6
Output Polarity and Output Range	3-6
Trigger	3-7
RTSI Triggers	3-11
Digital I/O	3-11
Timing Signal Routing.....	3-11

Selecting Sample/Update Clock Frequency	3-12
NI 4551/4552 LEDs	3-13

Chapter 4 Signal Connections

I/O Connectors.....	4-1
Analog I/O Connector Signal Descriptions	4-2
Digital I/O Connector Signal Descriptions	4-5
Analog Input Signal Connections.....	4-6
Types of Signal Sources	4-8
Floating Signal Sources	4-8
Ground-Referenced Signal Sources	4-8
Analog Output Signal Connections	4-8
Analog Power Connections	4-10
Digital I/O Signal Connections	4-10
Digital Power Connections.....	4-12
Field Wiring Considerations.....	4-12

Chapter 5 Calibration

Loading Calibration Constants	5-1
Self-Calibration	5-2
External Calibration.....	5-2
Traceable Recalibration.....	5-3

Chapter 6 Theory of Analog Operation

Analog Input Circuitry	6-1
Input Coupling	6-2
Calibration.....	6-2
Antialias Filtering	6-2
The ADC	6-7
Noise	6-8
Analog Output Circuitry.....	6-9
Anti-Image Filtering	6-9
The DAC	6-11
Calibration.....	6-12
Mute Feature	6-12

Appendix A Specifications

Appendix B Pin Connections

Appendix C Customer Communication

Glossary

Index

Figures

Figure 3-1.	Digital Function Block Diagram	3-2
Figure 3-2.	Analog Function Block Diagram	3-3
Figure 3-3.	Below-Low-Level Triggering Mode	3-9
Figure 3-4.	Above-High-Level Triggering Mode	3-9
Figure 3-5.	Inside-Region Triggering Mode	3-10
Figure 3-6.	High-Hysteresis Triggering Mode	3-10
Figure 3-7.	Low-Hysteresis Triggering Mode	3-10
Figure 4-1.	Analog Pin Connections	4-2
Figure 4-2.	Digital Pin Connections	4-5
Figure 4-3.	Analog Input Stage	4-7
Figure 4-4.	Analog Output Channel Block Diagram	4-9
Figure 4-5.	Digital I/O Connections	4-11
Figure 6-1.	Input Frequency Response	6-4
Figure 6-2.	Input Frequency Response Near the Cutoff	6-5
Figure 6-3.	Alias Rejection at the Oversample Rate	6-6
Figure 6-4.	Comparison of a Clipped Signal to a Proper Signal	6-7
Figure 6-5.	Signal Spectra in the DAC	6-11
Figure A-1.	Idle Channel Noise (Typical)	A-4
Figure B-1.	68-Pin Digital Connector for Any Digital Accessory	B-2

Tables

Table 3-1.	Actual Range and Measurement Precision of Input	3-5
Table 3-2.	Actual Range and Measurement Precision of Output	3-7
Table 4-1.	Analog I/O Connector Pin Assignment	4-3
Table 4-2.	Analog I/O Signal Summary	4-4
Table 4-3.	Digital I/O Connector Pin Assignment	4-6
Table 4-4.	Digital I/O Signal Summary	4-6

About This Manual

This manual describes the electrical and mechanical aspects of the NI 4551 and NI 4552 instruments and contains information concerning their operation. Unless otherwise noted, the text applies to both instruments.

The NI 4551 and NI 4552 are high-performance, high-accuracy analog input/output (I/O) instruments for the PCI bus. These instruments also support digital I/O (DIO) functions, counter/timer functions, and external trigger functions.

Organization of This Manual

The *NI 4551/4552 User Manual* is organized as follows:

- Chapter 1, *Introduction*, describes the NI 4551 and NI 4552 instruments, lists what you need to get started, explains how to unpack your instruments, and describes the optional software and optional equipment.
- Chapter 2, *Installation and Configuration*, explains how to install and configure your NI 4551/4552 instrument.
- Chapter 3, *Hardware Overview*, presents an overview of the hardware functions on your NI 4551/4552 instrument.
- Chapter 4, *Signal Connections*, describes how to make input and output connections to your NI 4551/4552 instrument via the analog I/O and digital I/O connectors of the instrument.
- Chapter 5, *Calibration*, discusses the calibration procedures for your NI 4551/4552 instrument.
- Chapter 6, *Theory of Analog Operation*, contains a functional overview and explains the operation of each analog functional unit making up the NI 4551/4552.
- Appendix A, *Specifications*, lists the specifications of the NI 4551/4552.
- Appendix B, *Pin Connections*, describes the pin connections on the optional 68-pin digital accessories for the NI 4551/4552 instruments.
- Appendix C, *Customer Communication*, contains forms you can use to request help from National Instruments or to comment on our products and manuals.

- The *Glossary* contains an alphabetical list and description of terms used in this manual, including abbreviations, acronyms, metric prefixes, mnemonics, and symbols.
- The *Index* contains an alphabetical list of key terms and topics in this manual, including the page where you can find each one.

Conventions Used in This Manual

The following conventions are used in this manual:

<>

Angle brackets enclose the name of a key on the keyboard—for example, <shift>. Angle brackets containing numbers separated by an ellipsis represent a range of values associated with a bit or signal name—for example, DBIO<3..0>.

◆

The ◆ symbol indicates that the text following it applies only to a specific product, a specific operating system, or a specific software version.

*

An asterisk following a signal name denotes an active low signal.



This icon to the left of bold italicized text denotes a note, which alerts you to important information.



This icon to the left of bold italicized text denotes a caution, which advises you of precautions to take to avoid injury, data loss, or a system crash.

bold italic

Bold italic text denotes an activity objective, note, caution, or warning.

DSA

DSA refers to dynamic signal acquisition.

italic

Italic text denotes variables, emphasis, a cross reference, or an introduction to a key concept. This font also denotes text from which you supply the appropriate word or value, as in NI-DAQ 6.x.

SE

SE refers to single ended and is equivalent to RSE (referenced single ended).

National Instruments Documentation

The *NI 4551/4552 User Manual* is one piece of the documentation set for your system. You could have any of several types of manuals depending on the hardware and software in your system. Use the manuals you have as follows:

- Hardware documentation—This manual presents information about using your instrument, such as modes of operation and high-level features.
- Software documentation—You may have both application software and NI-DAQ software documentation. National Instruments application software includes ComponentWorks, LabVIEW, LabWindows/CVI, Measure, and VirtualBench. After you set up your hardware system, use either your application software documentation or the NI-DAQ documentation to help you write your application. If you have a large, complicated system, it is worthwhile to look through the software documentation before you configure your hardware.
- Accessory installation guides or manuals—If you are using accessory products, read the terminal block and cable assembly installation guides. They explain how to physically connect the relevant pieces of the system. Consult these guides when you are making your connections.

Related Documentation

The following documents contain information you may find helpful:

- *BNC-2140 User Manual*
- National Instruments Application Note 025, *Field Wiring and Noise Considerations for Analog Signals*
- *PCI Local Bus Specification* Revision 2.0

Customer Communication

National Instruments wants to receive your comments on our products and manuals. We are interested in the applications you develop with our products, and we want to help if you have problems with them. To make it easy for you to contact us, this manual contains comment and configuration forms for you to complete. These forms are in Appendix C, [Customer Communication](#), at the end of this manual.

Introduction

This chapter describes the NI 4551 and NI 4552 instruments, lists what you need to get started, explains how to unpack your instruments, and describes the optional software and optional equipment.

Thank you for buying the NI 4551/4552 dynamic signal analyzer instrument for PCI. The NI 4551/4552 are high-performance, high-accuracy analog I/O instruments for the PCI bus. These instruments are members of the PCI-DSA instrument family and are specifically designed for demanding dynamic signal acquisition applications. The NI 4551 has two channels of 16-bit simultaneously sampled input at 204.8 kS/s and two channels of 16-bit simultaneously updated output at 51.2 kS/s. The NI 4552 has four channels of 16-bit simultaneously sampled analog input at 204.8 kS/s. Information on analog output applies only to the NI 4551, but information on analog input applies to both the NI 4551 and the NI 4552.

Both the analog input and the analog output circuitry have oversampling delta-sigma modulating converters. Delta-sigma converters are inherently linear, provide built-in brick-wall antialiasing/imaging filters, and have specifications that exceed other conventional technology for this application with regard to THD, SNR, and amplitude flatness. You can use these high-quality specifications and features to acquire or generate signals with high accuracy and fidelity without introducing noise or out-of-band aliases.

Applications include audio signal processing and analysis, acoustics and speech research, sonar, audio frequency test and measurement, vibration and modal analysis, or any application requiring high-fidelity signal acquisition of signals with a bandwidth up to 95 kHz or signal generation with a bandwidth up to 23 kHz.

What You Need to Get Started

To set up and use your NI 4551 or NI 4552, you will need the following:

- One of the following instruments:
 - NI 4551
 - NI 4552
- NI 4551/4552 User Manual*
- Dynamic Signal Analyzer Software CD
- You may have one or more of the following software packages and documentation:
 - LabVIEW for Windows
 - LabWindows/CVI for Windows
 - NI-DAQ for PC Compatibles
 - VirtualBench-DSA
 - ComponentWorks
 - Measure
- Your computer
- SHC68-C68-A1 analog cable
- BNC-2140 accessory

Unpacking

Your NI 4551/4552 is shipped in an antistatic plastic package to prevent electrostatic damage to the instrument. Electrostatic discharge can damage components on the instrument. To avoid such damage in handling the instrument, take the following precautions:

- Ground yourself via a grounding strap or by holding a grounded object.
- Touch the plastic package to a metal part of your computer chassis before removing the instrument from the package.

- Remove the instrument from the package and inspect the instrument for loose components or any other sign of damage. Notify National Instruments if the instrument appears damaged in any way. *Do not* install a damaged instrument into your computer.
- *Never* touch the exposed pins of connectors.

Software Programming Choices

You have several options to choose from to program and use your National Instruments computer-based instrument. You can use National Instruments application software or the NI-DAQ driver software.

The NI 4551/4552 can operate in two distinct modes that have different programming requirements. When you operate the instrument in the instrument mode, you must do all programming through the instrument driver. When you operate the instrument in the NI-DAQ compatible mode (default state), you must do all programming through NI-DAQ.

National Instruments Application Software

LabVIEW and LabWindows/CVI are innovative program development software packages for data acquisition and control applications. LabVIEW uses graphical programming, whereas LabWindows/CVI enhances traditional programming languages. Both packages include extensive libraries for data acquisition, instrument control, data analysis, and graphical data presentation.

LabVIEW features interactive graphics, a state-of-the-art user interface, and a powerful graphical programming language. The LabVIEW Data Acquisition VI Library, a series of VIs for using LabVIEW with National Instruments computer-based instrument hardware, is included with LabVIEW. The LabVIEW Data Acquisition VI Library is functionally equivalent to the NI-DAQ software.

LabWindows/CVI features interactive graphics, a state-of-the-art user interface, and uses the ANSI C programming language. The LabWindows/CVI Data Acquisition, a series of functions for using LabWindows/CVI with National Instruments computer-based instruments hardware, is included with the NI-DAQ software kit. The LabWindows/CVI Data Acquisition library is functionally equivalent to the NI-DAQ software.

VirtualBench is a suite of VIs for using your DAQ products just as you use standalone instruments, but you benefit from processing, display, and

storage capabilities of PCs. VirtualBench instruments load and save waveform data to disk in the same forms used in popular spreadsheet programs and word processors. A report generation capability complements the raw data storage by adding timestamps, measurements, user name, and comments.

The complete VirtualBench suite contains VirtualBench-Scope, VirtualBench-DSA, VirtualBench-FG, VirtualBench-Arb, VirtualBench-AODC, VirtualBench-DIO, VirtualBench-DMM, and VirtualBench-Logger. Your NI 4551/4552 comes with VirtualBench-DSA. VirtualBench-DSA is a turnkey application you can use to make measurements as you would with a stand-alone dynamic analyzer.

ComponentWorks contains tools for data acquisition and instrument control built on NI-DAQ driver software. ComponentWorks provides a higher-level programming interface for building virtual instruments with Visual Basic, Visual C++, Borland Delphi, and Microsoft Internet Explorer. With ComponentWorks, you can use all of the configuration tools, resource management utilities, and interactive control utilities included in NI-DAQ.

Measure is a data acquisition and instrument control add-in for Microsoft Excel. With Measure, you can acquire data directly from plug-in DAQ boards, GPIB instruments, or serial (RS-232) devices. Measure has easy-to-use dialog boxes for configuring your measurements. Your data is placed directly into Excel worksheet cells, from which you can perform your analysis and report generation using the full power and flexibility of Excel.

Optional Equipment

National Instruments offers a variety of products to use with your NI 4551/4552 series instruments, including cables and connector blocks as follows:

- SHC50-68 digital cable
- Shielded and DIN rail-mountable 68-pin connector blocks
- RTSI cables

Custom Cabling

National Instruments offers cables of different lengths and the BNC-2140 DSA accessory to connect your analog I/O to the NI 4551/4552. National Instruments recommends you do not develop your own cabling solution due to the difficulty of working with the high-density connector and the need to maintain high signal integrity. However, if your application requires that you develop your own cable, use the following guidelines:

- Use shielded twisted-pair wires for each differential analog input or output channel pair. Since the signals are differential, using this type of wire yields the best results.
- When connecting the cable shields, be sure to connect the analog input grounds to the AIGND pins and the analog output grounds to the AOGND pins. For a connector pin assignment, refer to Table 4-1.
- To create your own accessories, you can use an AMP 68-pin right-angle PWB receptacle header, part number 787254-1.
- Recommended manufacturer part numbers for the 68-pin mating connector for the cable assembly are as follows:
 - AMP 68-position straight cable plug, part number 787131-3
 - AMP 68-position backshell with jackscrews, part number 787191-1

National Instruments also offers cables of different lengths and accessories to connect your digital I/O signals to the NI 4551/4552. To develop your own cable, remember that the digital I/O mating connector is a 50-position receptacle. For a connector pin assignment, refer to Table 4-3. Recommended manufacturer part numbers for this mating connector are as follows:

- 50-position straight cable plug, part number 787131-1
- 50-position backshell with jackscrews, part number 787233-1

Refer to Appendix B, *Pin Connections*, for pin assignments of digital accessories and cables.

Installation and Configuration

This chapter explains how to install and configure your NI 4551/4552 instrument.

Software Installation



Note

Install your software before you install your NI 4551/4552 instrument.

If you are using NI-DAQ, refer to your NI-DAQ release notes. Find the installation section for your operating system and follow the instructions given there. If you are using LabVIEW, LabWindows/CVI, or other National Instruments application software, refer to the appropriate release notes. After you have installed your application software, refer to your NI-DAQ release notes and follow the instructions given there for your operating system and application software package.

Hardware Installation

You can install the NI 4551/4552 instrument in any available 5 V PCI expansion slot in your computer. However, to achieve the best noise performance, leave as much room as possible between the NI 4551/4552 instrument and other devices and hardware. The following are general installation instructions, but consult your computer user manual or technical reference manual for specific instructions and warnings.

1. Write down the instrument serial number in the *NI 4551/4552 Hardware and Software Configuration Form* in Appendix C, *Customer Communication*, of this manual.
2. Turn off and unplug your computer.
3. Remove the top cover or access port to the I/O channel.
4. Remove the expansion slot cover on the back panel of the computer.
5. Insert the NI 4551/4552 instrument into a 5 V PCI slot. It should fit snugly, but *do not force* the instrument into place.
6. Screw the mounting bracket of the NI 4551/4552 instrument to the back panel rail of the computer.

7. Check the installation.
8. Replace the cover.
9. Plug in and turn on your computer.

The NI 4551/4552 instrument is now installed. You are now ready to configure your software.

Instrument Configuration

The NI 4551/4552 instruments are completely software configurable. However, you must perform two types of configuration—bus-related and data acquisition-related.

The NI 4551/4552 instruments are fully compatible with the industry standard *PCI Local Bus Specification* Revision 2.0. The PCI system automatically performs all bus-related configurations and requires no interaction from you. Bus-related configuration includes setting the instrument base memory address and interrupt channel.

Data acquisition-related configuration includes such settings as analog input polarity and range, analog input mode, and others. You can modify these settings through National Instruments application-level software, such as ComponentWorks, LabVIEW, LabWindows/CVI, and VirtualBench, or driver-level software such as NI-DAQ.

Hardware Overview

This chapter presents an overview of the hardware functions on your NI 4551/4552 instrument. Figure 3-1 shows a block diagram of the digital functions. Figure 3-2 shows a block diagram of the analog functions. The two function blocks connect through the analog mezzanine bus.




Figure 3-1. Digital Function Block Diagram




Figure 3-2. Analog Function Block Diagram

Analog Input

The analog input section of each NI 4551/4552 instrument is software configurable. You can select different analog input configurations through application software. The following sections describe in detail each of the analog input categories.

Input Mode

The NI 4551/4552 instruments use differential (DIFF) inputs. You can configure the input as a referenced single-ended (SE) channel using the BNC-2140 DSA accessory. For more information, please refer to the *BNC-2140 User Manual*. In DIFF mode, one line connects to the positive input of the channel, and the other connects to the negative input of the same channel. You can connect the differential input to SE or DIFF signals, either floating or ground-referenced. However, grounding the negative input from floating sources may improve the measurement quality by removing the common-mode noise.

Input Coupling

The NI 4551/4552 has a software-programmable switch that determines whether a capacitor is placed in the signal path. If the switch is set for DC, the capacitor is bypassed and any DC offset present in the source signal is passed to the ADC. If the source has a significant amount of unwanted offset (bias voltage), you must set the switch for AC coupling to place the capacitor in the signal path and take full advantage of the input signal range.

Input Polarity and Input Range

The NI 4551/4552 instruments operate in bipolar mode. Bipolar input means that the input voltage range is between $-V_{\text{ref}}/2$ and $+V_{\text{ref}}/2$. The NI 4551/4552 has a bipolar input range of 20 V (± 10 V) for a gain of 1.0 (0 dB).

You can program the range settings on a per channel basis so that you can configure each input channel uniquely. The software-programmable gain on these instruments increases their overall flexibility by matching the input signal ranges to those that the ADC can accommodate. With the proper gain setting, you can use the full resolution of the ADC to measure the input signal. Table 3-1 shows the overall input range and precision according to the input range configuration and gain used.

Table 3-1. Actual Range and Measurement Precision of Input

Linear Gain	Gain	Input Range	Precision [†]
0.1	-20 dB	±42.4 V ^{††}	3.0518 mV ^{††}
0.316	-10 dB	±31.6 V	965.05 μV
1.0	0 dB	±10.0 V	305.18 μV
3.16	10 dB	±3.16 V	96.505 μV
10	20 dB	±1.00 V	30.518 μV
31.6	30 dB	±0.316 V	9.6505 μV
100	40 dB	±0.100 V	3.0518 μV
316	50 dB	±31.6 mV	965.05 nV
1000	60 dB	±10.0 mV	305.18 nV

[†] The value of 1 LSB of the 16-bit ADC; that is, the voltage increment corresponding to a change of one count in the ADC 16-bit count.

^{††} The actual input range is by design ±100 V; however, the instrument is not tested or certified to operate in this range.

See Appendix A, *Specifications*, for absolute maximum ratings.

All data read from the ADC is interpreted as two's complement format. In two's complement mode, digital data values read from the analog input channel are either positive or negative.

Considerations for Selecting Input Ranges

The input range you select depends on the expected range of the incoming signal. A large input range can accommodate a large signal variation but reduces the voltage resolution. Choosing a smaller input range improves the voltage resolution but can result in the input signal going out of range. For best results, match the input range as closely as possible to the expected range of the input signal.



Caution *If you exceed the rated input voltages, you can damage the computer and the connected equipment.*

If you do not choose the input range appropriately, an input signal can be clipped and can introduce large errors that are easily identified in the frequency spectrum. The NI 4551/4552 is equipped with overrange-detection circuits in both the analog and digital sections of each input channel. These circuits determine if an input signal has exceeded the

selected input voltage. Chapter 6, *Theory of Analog Operation*, provides a more in-depth explanation of how overranges can occur.

Analog Output

The analog output section of the NI 4551 instrument is software configurable. You can select different analog output configurations through application software designed to control the NI 4551. The following sections describe in detail each of the analog output categories. The NI 4551 instrument has two channels of analog output voltage at the I/O connector.

Output Mode

The NI 4551 instrument uses DIFF outputs. You can configure the outputs as an SE channel using the BNC-2140 DSA accessory. For more information, please refer to the *BNC-2140 User Manual*. In DIFF mode, one line connects to the positive input of the channel, and the other connects to the negative input of that same channel. You can connect the differential output to either SE or DIFF loads, either floating or ground-referenced. However, grounding the negative output is recommended when driving floating single-ended loads.

Output Polarity and Output Range

The NI 4551 instrument operates in bipolar mode. Bipolar output means that the output voltage range is between $-V_{\text{ref}}/2$ and $+V_{\text{ref}}/2$. The NI 4551 has a bipolar output range of 20 V (± 10 V) for an attenuation of 1.0 (0 dB).

You can program the range settings on a per channel basis so that you can configure each output channel uniquely. The software-programmable attenuation on these instruments increases their overall flexibility by matching the output signal ranges to your application. Table 3-2 shows the overall output range and precision according to the attenuation used.

Table 3-2. Actual Range and Measurement Precision of Output

Attenuation Linear	Attenuation	Range	Precision ¹
1.0	0 dB	±10.0 V	305.18 µV
10	20 dB	±1.00 V	30.158 µV
100	40 dB	±0.100 V	3.0518 µV
∞	∞ dB	0 V	0 V

¹ The value of 1 LSB of the 16-bit DAC; that is, the voltage increment corresponding to a change of one count in the DAC 16-bit count.

See Appendix A, *Specifications*, for absolute maximum ratings.



Note

The instrument powers up in a mode with the outputs disabled AND infinitely (∞) attenuated. Although these functions appear similar, they are quite distinct and are implemented to protect your external equipment from startup transients.

When the DACs no longer have data written to them, they automatically retransmit the last data point they received. If you expect the data to return to 0 V or any other voltage level, you MUST append the data to make it do so.

All data written to the DACs are interpreted as two's complement format. In two's complement mode, data values written to the analog output channel are either positive or negative.

Trigger

In addition to supporting internal software triggering and external digital triggering to initiate a data acquisition sequence, the NI 4551/4552 also supports analog level triggering. You can configure the trigger circuit to monitor any one of the analog input channels to generate the level trigger. Choosing an input channel as the level trigger channel does not influence the input channel capabilities. The level trigger circuit compares the full 16 bits of the programmed trigger level with the digitized 16-bit sample. The trigger-level range is identical to the analog input voltage range. The

trigger-level resolution is the same as the precision for a given input range. Refer to Table 3-1 for more information about input range and precision.

The trigger circuit generates an internal digital trigger based on the input signal and the user-defined trigger levels. Any of the timing sections of the timing input/output (TIO) ASIC can use this level trigger, including the analog input, analog output, RTSI, and general-purpose counter/timer sections. For example, you can configure the analog input section to acquire a given number of samples after the analog input signal crosses a specific threshold. As another example, you can configure the analog output section to generate an output waveform whenever the analog input signal crosses a specific threshold.

Due to the nature of delta-sigma converters, the triggering circuits operate on the digital output of the converter. Since the trigger is generated at the output of the converter, triggers can occur only when a sample is actually generated. Placing the triggering circuits on the digital side of the converter does not affect most measurements unless an analog output is generated based on the input trigger. In this case, you must be aware of the inherent delays of the finite impulse response (FIR) filters internal to the delta-sigma converters and you must account for the delays. The delay through the input converter is 42 sample periods, while the delay through the output converter is 34.6 ± 0.5 sample periods.

During repetitive sampling of a waveform, you may observe jitter due to the uncertainty of where a trigger level falls compared to the actual digitized data. Although this trigger jitter is never greater than one sample period, it can seem quite bad when the sample rate is only twice the bandwidth of interest. This jitter has no effect on the processing of the data, and you can decrease this jitter by oversampling.

There are five analog level triggering modes available, as shown in Figures 3-3 through 3-7. You can set **lowValue** and **highValue** independently in the software.

In below-low-level triggering mode, shown in Figure 3-3, the trigger is generated when the signal value is less than **lowValue**. **HighValue** is unused.




Figure 3-3. Below-Low-Level Triggering Mode

In above-high-level triggering mode, the trigger is generated when the signal value is greater than **highValue**. **LowValue** is unused.




Figure 3-4. Above-High-Level Triggering Mode

In inside-region triggering mode, the trigger is generated when the signal value is between the **lowValue** and the **highValue**.




Figure 3-5. Inside-Region Triggering Mode

In high-hysteresis triggering mode, the trigger is generated when the signal value is greater than **highValue**, with the hysteresis specified by **lowValue**.




Figure 3-6. High-Hysteresis Triggering Mode

In low-hysteresis triggering mode, the trigger is generated when the signal value is less than **lowValue**, with the hysteresis specified by **highValue**.




Figure 3-7. Low-Hysteresis Triggering Mode

You can use digital triggering through the RTSI bus and the external digital 50-pin connector. EXT_TRIG is the pin dedicated to external digital triggering. You can use the digital trigger to start an acquisition, a waveform generation, or to synchronize the start of a simultaneous acquisition and waveform generation.

You can trigger the NI 4551/4552 instruments from any other PCI-DSA or National Instruments device that has the RTSI bus feature. You can connect the devices through the RTSI bus cable. An external digital trigger can also trigger multiple devices simultaneously by distributing that trigger through the RTSI bus. You can also select the polarity of the external digital trigger.

RTSI Triggers

The seven RTSI trigger lines on the RTSI bus provide a flexible interconnection scheme for any NI 4551/4552 instrument sharing the RTSI bus. These bidirectional lines can drive the digital trigger onto the RTSI bus and can receive this signal.

Digital I/O

The NI 4551/4552 instruments contain 32 lines of digital I/O for general-purpose use through the 50-pin connector. You can individually software-configure each line for either input or output.



Note

At system power-on and reset, the hardware sets the DIO lines to high impedance. Table 4-4 shows that there is a 100 k Ω pull-down resistor. These pull-down resistors set the DIO pin to a logic low when the output is in a high-impedance state. Take careful consideration of the power-on state of the system to prevent any damage to external equipment.

Timing Signal Routing

The TIO ASIC provides a flexible interface for connecting timing signals to other instruments or to external circuitry. Your NI 4551/4552 instrument uses the RTSI bus to connect timing signals between instruments, and uses the DIO pins on the I/O connector to connect the instrument to external circuitry. These connections enable the NI 4551/4552 instrument to both control and be controlled by other devices and circuits.

Selecting Sample/Update Clock Frequency

The two analog input channels of the NI 4551 and the four inputs of the NI 4552 are simultaneously sampled at any software-programmable rate from 5.0 kS/s to 204.8 kS/s in 190.7 μ S/s increments (worst case). The instruments use direct digital synthesis (DDS) technology so that you can choose the correct sample rate required for your application. All the input channels acquire data at the same rate. One input channel *cannot* acquire data at a different rate than another input channel.

The two analog output channels of the NI 4551 are updated simultaneously at any software programmable rate from 1.25 kS/s to 51.2 kS/s in 47.684 μ S/s increments (worst case). The input sample rate and output update rate on the NI 4551 are synchronized and derived from the same DDS clock. The input and output clocks may differ from each other by a factor of 2 (1, 2, 4, 8, ..., 128) while still maintaining their synchronization as long as the lower bounds for update and sample rate are maintained. All the output channels update data at the same rate. One output channel *cannot* update data at a different rate than another output channel.

**Note**

You cannot generate the sample rate and update rate directly using external clock signals. You can only generate sample and update rates using the DDS clock circuitry.

The DDS clock signal and the synchronization start signal (digital trigger) are transmitted to other PCI-DSA instruments via the RTSI bus. The NI 4551/4552 can also receive these signals to synchronize the acquisition or waveform generation with other devices. In a multidevice system, a master device would drive the clock and synchronization signal to other slave or receiving devices.

Selecting a sample rate that is less than two times the frequency of a band of interest seems to indicate that the board is functioning improperly. By undersampling the signal, you could receive what appears to be a DC signal. This situation is due to the sharp antialiasing filters that remove frequency components above the sampling frequency. If you have a situation where this occurs, increase the sample rate until it meets the requirements of the Shannon Sampling Theorem. For more information on the filters and aliasing, refer to Chapter 6, *Theory of Analog Operation*.

Unlike other converter technologies, delta-sigma converters must be run continuously and at a minimum clock rate. To operate within guaranteed specifications, the ADCs should operate at a minimum sample rate of 5.0 kS/s and the DACs should operate at a minimum update rate of

1.25 kS/s. This minimum rate is required to keep the internal circuitry of the converters running within specifications. You are responsible for selecting sample and update rates that fall within the specified limits. Failure to do so could greatly affect the specifications.

NI 4551/4552 LEDs

The NI 4551/4552 has a green LED that flashes at power up. This LED indicates that the onboard processor is running in the NI-DAQ-compatible mode and is ready to acquire or generate data. In this mode, the instrument acts as a standard DAQ device and you can program it using LabVIEW, LabWindows/CVI, or any other supported National Instruments application software package. The green LED may flash at a different rate when the NI 4551/4552 is accessed by VirtualBench-DSA or by the instrument driver software.

The NI 4551/4552 has four red LEDs that National Instruments currently uses for internal debugging purposes. The state of these red LEDs has no significance to your application.

Signal Connections

This chapter describes how to make input and output connections to your NI 4551/4552 instrument via the analog I/O and digital I/O connectors of the instrument.

The analog I/O connector for the NI 4551/4552 connects to the BNC-2140 DSA accessory through the SHC68-C68-A1 shielded cable. You can access the analog I/O of the NI 4551/4552 using standard BNC connectors on the BNC-2140. You can connect the analog I/O to the shielded cable through a single 68-pin connector.

The digital I/O connector for the NI 4551/4552 has 50 pins that you can connect to generic 68-pin terminal blocks through the SHC50-68 shielded cable. You can connect the digital I/O signals to the shielded cable through a single 50-pin connector.

I/O Connectors

Table 4-1 describes the pin assignments for the 68-pin analog I/O connector. Table 4-3 describes the 50-pin digital connector on the NI 4551/4552 instruments. A signal description follows the connector pinouts.

**Caution**

Connections that exceed any of the maximum ratings of input or output signals on the NI 4551/4552 instruments can damage the NI 4551/4552 instrument, the computer, and accessories. Maximum input ratings for each signal are given in the Protection column of Table 4-2 and 4-4. National Instruments is not liable for any damages resulting from such signal connections.

Analog I/O Connector Signal Descriptions

Figure 4-1 shows the analog pin connections for the NI 4551/4552.

-ACH0	1	35	+ACH0
AIGND [†]	2	36	AIGND
-ACH1	3	37	+ACH1
AIGND [†]	4	38	AIGND
-ACH2 ¹	5	39	+ACH2 ¹
AIGND [†]	6	40	AIGND
-ACH3 ¹	7	41	+ACH3 ¹
AIGND [†]	8	42	AIGND
NC	9	43	NC
AIGND [†]	10	44	AIGND [†]
NC	11	45	NC
AIGND [†]	12	46	AIGND [†]
NC	13	47	NC
AIGND [†]	14	48	AIGND [†]
NC	15	49	NC
AIGND [†]	16	50	AIGND [†]
NC	17	51	NC
AIGND [†]	18	52	AIGND [†]
NC	19	53	NC
AIGND [†]	20	54	AIGND [†]
NC	21	55	NC
AIGND [†]	22	56	AIGND [†]
NC	23	57	NC
AIGND [†]	24	58	AIGND [†]
-DAC0OUT ²	25	59	+DAC0OUT ²
AOGND [†]	26	60	AOGND
-DAC1OUT ²	27	61	+DAC1OUT ²
AOGND [†]	28	62	AOGND
NC	29	63	NC
AOGND [†]	30	64	AOGND [†]
NC	31	65	NC
AOGND [†]	32	66	AOGND [†]
+5 V	33	67	+5 V
DGND	34	68	DGND

¹Not available on NI 4551
²Not available on NI 4552
[†]These AIGND and AOGND pins are not connected in the SHC68-C68-A1 cable.

Figure 4-1. Analog Pin Connections

Table 4-1. Analog I/O Connector Pin Assignment

Signal Name	Reference	Direction	Description
+ACH<0..3>	AIGND	Input	+Analog Input Channel 0 through 3—The NI 4551 uses +ACH<0..1> and the NI 4552 uses +ACH<0..3>.
-ACH<0..3>	AIGND	Input	-Analog Input Channel 0 through 3—The NI 4551 uses -ACH<0..1> and the NI 4552 uses -ACH<0..3>.
AIGND	—	—	Analog Input Ground—These pins are the reference point for single-ended measurements in SE configuration and the bias current return point for differential measurements. All three ground references—AIGND, AOGND, and DGND—are connected together on your NI 4551/4552 instrument, but each serves a separate purpose.
+DAC0OUT	-DAC0OUT	Output	+Analog Output Channel 0—This pin supplies the analog non-inverting output channel 0. This pin is available only on the NI 4551.
-DAC0OUT	+DAC0OUT	Output	-Analog Output Channel 0—This pin supplies the analog inverting output channel 0. This pin is available only on the NI 4551.
+DAC1OUT	-DAC1OUT	Output	+Analog Output Channel 1—This pin supplies the analog non-inverting output channel 1. This pin is only available on the NI 4551.
-DAC1OUT	+DAC1OUT	Output	-Analog Output Channel 1—This pin supplies the analog inverting output channel 1. This pin is only available on the NI 4551.
AOGND	—	—	Analog Output Ground—The analog output voltages are ultimately referenced to this node. All three ground references—AIGND, AOGND, and DGND—are connected together on your NI 4551/4552 instrument, but each serves a separate purpose.
+5 V	DGND	Output	+5 VDC Source—These pins are fused for up to 0.5 A and supply power to the DSA signal conditioning accessories. The fuse is self resetting.
DGND	—	—	Digital Ground—This pin supplies the reference for the +5 VDC supply. All three ground references—AIGND, AOGND, and DGND—are connected together on your NI 4551/4552 instrument, but each serves a separate purpose.

Table 4-2. Analog I/O Signal Summary

Signal Name	Signal Type and Direction	Impedance	Protection	Source	Sink	Rise Time (ns)	Bias
+ACH<0..3>	AI	1 M Ω in parallel with 50 pF to AIGND	± 42.4 V (on) [†] ± 42.4 V (off) [†]	—	—	—	± 100 pA
-ACH<0..3>	AI	1 M Ω in parallel with 50 pF to AIGND	± 42.4 V (on) [†] ± 42.4 V (off) [†]	—	—	—	± 100 pA
AIGND	AI	—	—	—	—	—	—
+DAC0OUT	AO	22 Ω to -DAC0OUT, 4.55 k Ω to AOGND	Short-circuit to -DAC0OUT, ground	16.7 mA at 10 V	—	—	—
-DAC0OUT	AO	22 Ω to +DAC0OUT, 4.55 k Ω to AOGND	Short-circuit to +DAC0OUT, ground	16.7 mA at 10 V	—	—	—
+DAC1OUT	AO	22 Ω to -DAC1OUT, 4.55 k Ω to AOGND	Short-circuit to -DAC1OUT, ground	16.7 mA at 10 V	—	—	—
-DAC1OUT	AO	22 Ω to +DAC1OUT, 4.55 k Ω to AOGND	Short-circuit to +DAC1OUT, ground	16.7 mA at 10 V	—	—	—
AOGND	AO	—	—	—	—	—	—
DGND	DIO	—	—	—	—	—	—
+5 V	DO	0.7 Ω	Short-circuit to ground	0.5A	—	—	—
AI = Analog Input DIO = Digital Input/Output AO = Analog Output DO = Digital Output [†] ± 400 V/ ± 400 V guaranteed by design, but not tested or certified to operate beyond ± 42.4 V							

Digital I/O Connector Signal Descriptions

Figure 4-2 shows the digital pin connections for the NI 4551/4552.

DIO0	1	26	DGND
DIO2	2	27	DIO1
DIO3	3	28	DGND
DIO5	4	29	DIO4
DIO6	5	30	DGND
DIO8	6	31	DIO7
DIO9	7	32	DGND
DIO11	8	33	DIO10
DIO12	9	34	DGND
DIO14	10	35	DIO13
DIO15	11	36	DGND
DIO22	12	37	DIO21
DIO20	13	38	DGND
DIO17	14	39	DIO19
DIO18	15	40	DGND
+5 V	16	41	DIO16
+5 V	17	42	DGND
+5 V	18	43	+5 V
DIO23	19	44	DGND
DIO25	20	45	DIO24
DIO26	21	46	DGND
DIO28	22	47	DIO27
DIO29	23	48	DGND
DIO31	24	49	DIO30
EXT_TRIG	25	50	DGND

Figure 4-2. Digital Pin Connections

Refer to Appendix B, [Pin Connections](#), for the digital pin connections of the 68-pin connector.

Table 4-3. Digital I/O Connector Pin Assignment

Signal Name	Reference	Direction	Description
DIO<0..31>	DGND	Input or Output	Digital I/O channels 0 through 31
DGND	—	—	Digital Ground—This pin supplies the reference for the digital signals at the I/O connector as well as the +5 VDC supply.
+5 V	DGND	Output	+5 VDC Source—These pins are fused for up to 1 A of +5 V supply. The fuse is self-resetting.
EXT_TRIG	DGND	Input or Output	External trigger—This pin is used as the trigger to start an acquisition and waveform generation.

Table 4-4. Digital I/O Signal Summary

Signal Name	Signal Type and Direction	Impedance	Protection (Volts)	Source	Sink	Bias
DIO<0..31>	DIO	—	Vcc +0.5	3.5 mA at 2.4 V	10 mA at 0.45 V	100 k Ω pd
DGND	DIO	—	—	—	—	—
+5 V	DO	0.15 Ω	Short-circuit to ground	1A	—	—
EXT_TRIG	DIO	—	Vcc +0.5	3.5 mA at 2.4 V	10 mA at 0.45 V	100 k Ω pu
DIO = digital input/output DO = digital output		pu = pullup pd = pulldown				
Note The tolerance on the 100 k Ω resistors is very large. Actual value may range between 20 and 100 k Ω						

Analog Input Signal Connections

The analog input signals for the NI 4551/4552 instruments are +ACH<0..3>, -ACH<0..3>, and AIGND. The \pm ACH<0..1> signals are tied to the two analog input channels of your NI 4551, and \pm ACH<0..3> are tied to the four analog input channels of your NI 4552 instrument.



Caution *Exceeding the differential and common-mode input ranges distorts your input signals.*

AIGND is an analog input common signal that connects directly to the ground system on the NI 4551/4552 instruments. You can use this signal for a general analog ground tie point to your NI 4551/4552 instrument if necessary, but connecting AIGND to other earth-connected grounds is not recommended. AIGND is not directly available if you are using a BNC-2140 accessory.

Figure 4-3 shows a diagram of your NI 4551/4552 instrument analog input stage.




Figure 4-3. Analog Input Stage

The analog input stage applies gain and common-mode voltage rejection and presents high input impedance to the analog input signals connected to your NI 4551/4552 instrument. Signals are routed directly to the positive and negative inputs of the analog input stage on the instrument. The analog input stage converts two input signals to a signal that is the difference between the two input signals multiplied by the gain setting of the amplifier. The amplifier output voltage is referenced to the ground for the instrument. Your NI 4551/4552 instrument ADC measures this output voltage when it performs A/D conversions.

Connection of analog input signals to your NI 4551/4552 instrument depends on the configuration of the input signal sources. For most signals, you use a DIFF configuration and connect the signal to $+ACHx$ (where x is the NI 4551/4552 channel) and the signal ground (or signal minus, as appropriate) to $-ACHx$. However, if a signal has a high output impedance (greater than $1\text{ k}\Omega$) and is floating, you may find it useful to use an SE configuration and tether the signal minus to AIGND to reduce common-mode interference. You can make the DIFF and SE connections through the BNC-2140 accessory.

Types of Signal Sources

When configuring the input channels and making signal connections, first determine whether the signal sources are floating or ground-referenced. The following sections describe these two types of signals.

Floating Signal Sources

A floating signal source does not connect in any way to the building ground system but instead has an isolated ground-reference point. Some examples of floating signal sources are outputs of transformers, thermocouples, battery-powered devices, optical isolator outputs, and isolation amplifiers. An instrument or device that has an isolated output is a floating signal source.

Ground-Referenced Signal Sources

A ground-referenced signal source connects in some way to the building system ground and is, therefore, already connected to a common ground point with respect to the NI 4551/4552 instrument, assuming that you plug the computer into the same power system. Nonisolated outputs of instruments and devices that plug into the building power system fall into this category.

The difference in ground potential between two instruments connected to the same building power system is typically between 1–100 mV but can be much higher if power distribution circuits are not properly connected. For this reason, National Instruments does not recommend connecting AIGND to the source signal ground system, since the difference between the grounds can induce currents in the NI 4551/4552 ground system.

Analog Output Signal Connections

The analog output signals for the NI 4551 instrument are +DAC0OUT, –DAC0OUT, +DAC1OUT, –DAC1OUT, and AOGND. +DAC0OUT and –DAC0OUT are the plus and minus voltage output signals for analog output channel 0. +DAC1OUT and –DAC1OUT are the plus and minus voltage output signal for analog, output channel 1.

AOGND is a ground-reference signal for both analog output channels. It is connected directly to the ground system on the NI 4551 instrument. You can use this signal for a general analog ground tie point to your NI 4551 instrument if necessary, but connecting AOGND to other earth-connected

grounds is not recommended. AOGND is not directly available if you are using the BNC-2140 accessory.

The NI 4551 has two analog output channels, either of which is illustrated in Figure 4-4.




Figure 4-4. Analog Output Channel Block Diagram

The analog output stage is differential and balanced. Each output signal consists of a plus connection, a minus connection, and a ground (AOGND) connection. The actual output signal is the *difference* between the plus and minus connections. The pair is *balanced*, meaning that, if the impedance from each of the pair to AOGND is the same (or infinite), then the voltage at the plus and minus terminals is equal but opposite, so that their difference is the desired signal and their sum (or average) is zero. If impedance from each of the pair to AOGND is not the same, the connection is unbalanced, but the difference between the plus and minus terminals is still equal to the desired signal. If the minus side is grounded, the plus voltage is equal to the signal. Conversely, if the plus side is grounded, the minus voltage is equal to the negative of the signal. In all cases, the difference is equal to the signal.

Connection of analog output signals from your NI 4551 instrument depends on the configuration of the devices receiving the signals. For most signals, you use a DIFF configuration and simply connect +DAC x OUT (where x is the NI 4551 channel) to the signal and -DAC x OUT to the signal ground (or signal minus, as appropriate). When driving some floating devices, however, you may sometimes find it helpful to use the SE configuration and connect the floating ground system of the instrument to AOGND to reduce common-mode noise coupled from an interfering source to the

instrument. You can make DIFF and SE connections through the BNC-2140 accessory.

Analog Power Connections

Two pins on the analog I/O connector supply +5 V from the computer power supply via a self-resetting fuse. The fuse resets automatically within a few seconds after the overcurrent condition is removed. These pins are referenced to DGND and you can use them to power external analog accessories like the BNC-2140. The following is the power rating for the fuse:

- Power rating +4.65 to +5.25 VDC at 0.5 A



Caution *Do not under any circumstances connect these +5 V power pins directly to analog ground, digital ground, or to any other voltage source on the NI 4551/4552 or any other device. Doing so can damage the NI 4551/4552 and the computer. National Instruments is not liable for damages resulting from such a connection.*

Digital I/O Signal Connections

The digital I/O signals are DIO<0..31> and DGND. DIO<0..31> are the signals making up the DIO port. DGND is the ground-reference signal for the DIO port. You can program all lines individually to be inputs or outputs.

Figure 4-5 shows signal connections for three typical digital I/O applications. Eight of the 32 DIO signals are shown.




Figure 4-5. Digital I/O Connections

Figure 4-5 shows DIO<0..3> configured for digital input and DIO<4..7> configured for digital output. Digital input applications include receiving TTL signals and sensing external device states such as the state of the switch shown in Figure 4-5. Digital output applications include sending TTL signals and driving external devices such as the LED shown in Figure 4-5.

Digital Power Connections

Four pins on the digital I/O connector supply +5 V from the computer power supply via a self-resetting fuse. The fuse will reset automatically within a few seconds after the overcurrent condition is removed. These pins are referenced to DGND and you can use them to power external digital circuitry. The following is the power rating for the fuse:

- Power rating +4.65 to +5.25 VDC at 1 A

Take careful consideration of the current drawn from the DIO connector, Although the fuse allows a 1 A current draw, system specifications can limit the current draw to less than 1 A.



Caution *Do not under any circumstances connect these +5 V power pins directly to analog ground, digital ground, or to any other voltage source on the NI 4551/4552 or any other device. Doing so can damage the NI 4551/4552 and the computer. National Instruments is not liable for damages resulting from such a connection.*

Field Wiring Considerations

Environmental noise can seriously influence the accuracy of measurements made with your NI 4551/4552 if you do not take proper care when running signal wires between signal sources and the instrument. The following recommendations apply mainly to analog input signal routing to the instrument, although they also apply to signal routing in general.

Minimize noise pickup and maximize measurement accuracy by taking the following precautions:

- Use differential analog input connections to reject common-mode noise.
- Use individually shielded, twisted-pair wires to connect analog input signals to the instrument. With this type of wire, the signals attached to the +ACH_x and –ACH_x inputs are twisted together and then covered with a shield. You then connect this shield only at one point to the signal source ground. This kind of connection is required for signals traveling through areas with large magnetic fields or high electromagnetic interference.
- Route signals to the instrument carefully. Keep cabling away from noise sources. The most common noise source in a PCI DAQ system is the video monitor. Separate the monitor from the analog signals as much as possible.

The following recommendations apply for all signal connections to digital signal routing from your NI 4551/4552:

- The digital output signal integrity is greatly influenced by the length of the cable being driven. Minimize cable lengths and use schmitt-trigger devices to deglitch signals. Further conditioning may be required to create a clean signal.
- Always try to couple a ground with a signal to minimize noise pickup and radiation.

The following recommendations apply for all signal connections to your NI 4551/4552:

- Separate NI 4551/4552 signal lines from high-current or high-voltage lines. These lines can induce currents in or voltages on the NI 4551/4552 instrument signal lines if they run in parallel paths at a close distance. To reduce the magnetic coupling between lines, separate them by a reasonable distance if they run in parallel, or run the lines at right angles to each other.
- Do not run signal lines through conduits that also contain power lines.
- Protect signal lines from magnetic fields caused by electric motors, welding equipment, breakers, or transformers by running them through special metal conduits.

For more information, refer to the application note, *Field Wiring and Noise Consideration for Analog Signals*, available from National Instruments.

Calibration

This chapter discusses the calibration procedures for your NI 4551/4552 instrument. Your NI 4551/4552 is shipped with a calibration certificate. The traceability information is stored in National Instruments corporate databases and is not actually shown on your certificate. The certificate contains a unique tracking number linking your instrument to the database. You can get a detailed calibration report from National Instruments for an additional charge.

If you are using NI-DAQ, that software includes calibration functions for performing all of the steps in the calibration process. Calibration refers to the process of minimizing measurement and output voltage errors by making small circuit adjustments. On the NI 4551/4552 instruments, these adjustments take the form of writing values to onboard calibration DACs (CalDACs). Some form of instrument calibration is required for all but the most forgiving applications. If you do not calibrate your instrument, your signals and measurements could have very large offset and gain errors. The four levels of calibration available are described in this chapter. The first level is the fastest, easiest, and least accurate, whereas the last level is the slowest, most difficult, and most accurate.

Loading Calibration Constants

Your NI 4551/4552 instrument is factory calibrated before shipment at approximately 25° C to the levels indicated in Appendix A, [Specifications](#). The associated calibration constants—the values that were written to the CalDACs to achieve calibration in the factory—are stored in the onboard nonvolatile memory (EEPROM). Because the CalDACs have no memory capability, they do not retain calibration information when the instrument is unpowered. Loading calibration constants refers to the process of loading the CalDACs with the values stored in the EEPROM. NI-DAQ software determines when this is necessary and does it automatically.

The EEPROM contains a user-modifiable calibration area in addition to the permanent factory calibration area. This means that you can load the CalDACs with values either from the original factory calibration or from a calibration that you subsequently performed. This method of calibration is

not very accurate because it does not take into account the fact that the instrument measurement and output voltage errors can vary with time and temperature. It is better to self-calibrate when you install the instrument in your environment.

Self-Calibration

Your NI 4551/4552 can measure and correct for almost all of its calibration-related errors without any external signal connections. With National Instruments software you can self-calibrate generally in less than a minute. This is the preferred method of assuring accuracy in your application. Initiate self-calibration to minimize the effects of any offset and gain drifts, particularly those due to warmup.

Your NI 4551/4552 instrument has an onboard calibration reference to ensure the accuracy of self-calibration. Its specifications are listed in Appendix A, *Specifications*. The reference voltage is measured at the factory and stored in the EEPROM for subsequent self-calibrations.

Immediately after self-calibration, the only significant residual calibration error could be gain error due to time or temperature drift of the onboard voltage reference. This error is addressed by external calibration, which is discussed in the following section, *External Calibration*. If you are interested primarily in relative measurements, you can ignore a small amount of gain error, and self-calibration should be sufficient.

To calibrate your NI 4551/4552 device while it is connected to a BNC-2140 accessory, set each input channel to SE and connect each channel + terminal to a channel – terminal through a BNC shunt. You can also calibrate your NI 4551/4552 device by removing the external cable connected to the BNC-2140 accessory.

External Calibration

The onboard calibration reference voltage is stable enough for most applications, but if you are using your instrument at an extreme temperature or if the onboard reference has not been measured for a year or more, you should externally calibrate your instrument.

In external calibration you calibrate your instrument with a known external reference rather than relying on the onboard reference. Redetermining the value of the onboard reference is part of this process and you can save the results in the EEPROM, so you should not have to perform an external

calibration very often. To externally calibrate your instrument call the NI-DAQ calibration function. You must be sure to use a very accurate external DC reference. This reference should be several times more accurate than the instrument itself. For example, to calibrate the NI 4551/4552, the external reference should have a DC accuracy better than ± 115 ppm (± 0.001 dB).

Traceable Recalibration

Traceable recalibration is divided into three different areas—factory, onsite and third party precision instruments typically require this type of recalibration every year.

If you require factory recalibration, send your NI 4551/4552 back to National Instruments. The instrument will be sent back to you with a new calibration certificate. You can request a detailed report for an additional fee. Please check with National Instruments for additional information such as cost and delivery times.

If your company has a metrology laboratory, you can recalibrate the NI 4551/4552 at your location (onsite). You can also send out your NI 4551/4552 for recalibration by a third party. Please contact National Instruments for approved third-party calibration service providers.

Theory of Analog Operation

This chapter contains a functional overview and explains the operation of each analog functional unit making up the NI 4551/4552. See Figure 3-2 for a general block diagram of the NI 4551/4552 analog functions.

Analog Input Circuitry

The NI 4551 has two identical analog input channels. The NI 4552 has four identical analog input channels. An analog input channel is illustrated in Figure 4-3.

These input channels have 16-bit resolution and are simultaneously sampled at software-programmable rates from 5 to 204.8 kS/s in 190.7 μ S/s increments. This flexibility in sample rates makes the instrument well suited for a wide variety of applications, including audio and vibration analysis.

The differential analog inputs have AC/DC coupling. You can use a programmable gain amplifier stage on the inputs to select gains from -20 to 60 dB in 10 dB steps. The input stage has differential connections, allowing quiet measurement of either single-ended or differential signals.

The analog inputs have both analog and real-time digital filters implemented in hardware to prevent aliasing. Input signals first pass through lowpass analog filters to attenuate signals with frequency components beyond the range of the ADCs. Then digital antialiasing filters automatically adjust their cutoff frequency to remove frequency components above half the programmed sampling rate. Because of this advanced analog input design, you do not have to add any filters to prevent aliasing. These filters, however cause a delay of 42 conversion periods between the input analog data and the digitized data.

The 90 dB dynamic range of the NI 4551/4552 instruments is the result of low noise and distortion and makes possible high-accuracy measurements. The instruments have excellent amplitude flatness of ± 0.1 dB, and have a maximum total harmonic distortion (THD) specification of -92 dB at 1 kHz and a worst case THD of -80 dB at higher frequencies.

State-of-the-art, 128-times oversampling, delta-sigma modulating ADCs achieve the low noise and low distortion of the NI 4551/4552. Because these ADCs sample at 128 times the specified sampling rate with 1-bit resolution, they produce nearly perfect linearity. Extremely flat, linear-phase, lowpass digital filters then remove the quantization noise from outside the band of interest, divide the sample rate by 128, and increase the resolution to 16 bits. Using the delta-sigma modulating ADCs, the NI 4551/4552 are immune to the DNL distortion associated with conventional data acquisition devices.

Input Coupling

The NI 4551/4552 has a software-programmable switch to individually configure each input channel for AC or DC coupling. If the switch is set for DC, the capacitor is bypassed, and any DC offset present in the source signal passes to the ADC. The DC configuration is preferred because it places one less component in the signal path and thus has higher fidelity. The DC configuration is recommended if the signal source has only small amounts of offset voltage (less than ± 100 mV), or if the DC content of the acquired signal is important.

If the source has a significant amount of unwanted offset (or bias voltage), you must set the switch for AC coupling to take full advantage of the input signal range. Using AC coupling results in a drop in the low-frequency response of the analog input. The -3 dB cutoff frequency is approximately 3.4 Hz, but the -0.01 dB cutoff frequency, for instance, is considerably higher at approximately 70.5 Hz. The input coupling switch can connect the input circuitry to ground instead of to the signal source. This connection is usually made during offset calibration, which is described in Chapter 5, [Calibration](#).

Calibration

The NI 4551/4552 analog inputs have calibration adjustments. Onboard CalDACs remove the offset and gain errors for each channel. For complete calibration instructions, refer to Chapter 5, [Calibration](#).

Antialias Filtering

A sampling system (such as an ADC) can represent only signals of limited bandwidth. Specifically according to the Shannon Sampling Theorem, a sampling rate of F_s can only represent signals with a maximum frequency of $F_s/2$. This maximum frequency is known as the *Nyquist frequency*. If a signal is input to the sampling system with frequency components that exceed the Nyquist frequency, the sampler cannot distinguish these parts of

the signal from some signals with frequency components less than the Nyquist frequency.

For example, suppose an ADC is sampling at 1,000 S/s. If a 400 Hz sine wave is input, the resulting samples accurately represent a 400 Hz sine wave. However, if a 600 Hz sine wave is input, the resulting samples again appear to represent a 400 Hz sine wave because this signal exceeds the Nyquist frequency (500 Hz) by 100 Hz. In fact, any sine wave with a frequency greater than 500 Hz that is input is represented incorrectly as a signal between 0 and 500 Hz. The apparent frequency of this sine wave is the absolute value of the difference between the frequency of the input signal and the closest integer multiple of 1,000 Hz (the sampling rate). Therefore, if a 2,325 Hz sine wave is input, its apparent frequency is calculated as follows:

$$2,325 - (2)(1,000) = 325 \text{ Hz}$$

If a 3,975 Hz sine wave is input, its apparent frequency is calculated as follows:

$$(4)(1,000) - 3,975 = 25 \text{ Hz}$$

The process by which the sampler modulates these higher frequency signals back into the 0 through 500 Hz baseband is called *aliasing*.

If the signal in the previous example is not a sine wave, the signal can have many components (harmonics) that lie above the Nyquist frequency. If present, these harmonics are erroneously aliased back into the baseband and added to the parts of the signal that are sampled accurately, producing a distorted sampled data set. Input to the sampler only those signals that can be accurately represented. All frequency components of such signals lie below the Nyquist frequency. To make sure that only those signals go into the sampler, a lowpass filter is applied to signals before they reach the sampler. The NI 4551/4552 has complete antialiasing filters.

The NI 4551/4552 includes two stages of antialias filtering in each input channel lowpass filter. This filter has a cutoff frequency of about 4 MHz and a rejection of greater than 40 dB at 20 MHz. Because its cutoff frequency is significantly higher than the data sample rate, the analog filter has an extremely flat frequency response in the bandwidth of interest, and it has very little phase error.

The analog filter precedes the analog sampler, which operates at 128 times the selected sample rate (26.2144 MS/s in the case of a 204.8 kS/s sample rate) and is actually a 1-bit ADC. The 1-bit, 128 times oversampled data

that the analog sampler produces is passed on to a digital antialiasing filter that is built into the ADC chip. This filter also has extremely flat frequency response and no phase error, but its roll-off near the cutoff frequency (about 0.493 times the sample rate) is extremely sharp, and the rejection above 0.536 times the sample rate is greater than 85 dB. The output stage of the digital filter resamples the higher frequency data stream at the output data rate, producing 16-bit digital samples.

With the NI 4551/4552 filters, you have the complete antialiasing protection needed to sample signals accurately. The digital filter in each channel passes only those signal components with frequencies that lie below the Nyquist frequency or within one Nyquist bandwidth of multiples of 128 times the sample rate. The analog filter in each channel rejects possible aliases (mostly noise) from signals that lie near these multiples. Figures 6-1 and 6-2 show the frequency response of the NI 4551/4552 input circuitry.




Figure 6-1. Input Frequency Response




Figure 6-2. Input Frequency Response Near the Cutoff

Because the ADC samples at 128 times the data rate, frequency components above 64 times the data rate can alias. The digital filter rejects most of the frequency range over which aliasing can occur. However, the filter can do nothing about components that lie close to 128 times the data rate, 256 times the data rate, and so on, because it cannot distinguish these components from components in the baseband (0 Hz to the Nyquist frequency). If, for instance, the sample rate is 200 kS/s and a signal component lies within 100 kHz of 25.6 MHz (128×200 kHz), this signal is aliased into the passband region of the digital filter and is not attenuated. The purpose of the analog filter is to remove these higher frequency components near multiples of the oversampling rate before they get to the sampler and the digital filter.

While the frequency response of the digital filter scales in proportion to the sample rate, the frequency response of the analog filter remains fixed. The filter response is optimized to produce good high-frequency alias rejection while having a flat in-band frequency response. Because this filter is third order, its roll-off is rather slow. This means that, although the filter has good alias rejection for high sample rates, it does not reject as well at lower sample rates. The alias rejection near 128 times the sample rate versus the

sample rate is illustrated in Figure 6-3. For frequencies not near multiples of the oversampling rate, the rejection is better than 85 dB.




Figure 6-3. Alias Rejection at the Oversample Rate

There is a form of aliasing that no filter can prevent. When a waveform exceeds the voltage range of the ADC, it is said to be *clipped* or *overranged*. When clipping occurs, the ADC assumes the closest value in its digital range to the actual value of the signal, which is always either $-32,768$ or $+32,767$. Clipping nearly always results in an abrupt change in the slope of the signal and causes the corrupted digital data to have high-frequency energy. This energy is spread throughout the frequency spectrum, and because the clipping happens *after* the antialiasing filters, the energy is aliased back into the baseband. The remedy for this problem is simple: do not allow the signal to exceed the nominal input range. Figure 6-4 shows the spectra of $10.5 V_{\text{rms}}$ and $10.0 V_{\text{rms}}$, 3.0 kHz sine waves digitized at 48 kS/s. The signal-to-THD plus noise ratio is 35 dB for the clipped waveform and 92 dB for the properly ranged waveform. Notice that aliases of all the harmonics due to clipping appear in Figure 6-4.




Figure 6-4. Comparison of a Clipped Signal to a Proper Signal

An overrange can occur on the analog signal as well as on the digitized signal. Furthermore, an analog overrange can occur independently from a digital overrange and vice-versa. For example, a piezoelectric accelerometer may have a resonant frequency that, when stimulated, can produce an overrange in the analog signal, but because the delta-sigma technology of the ADC uses very sharp antialiasing filters, the overrange is not passed into the digitized signal. Conversely, a sharp transient on the analog input may not overrange, but due to the step response of those same delta-sigma antialiasing filters, the digitized data may be clipped.

The ADC

The NI 4551/4552 ADCs use a method of A/D conversion known as delta-sigma modulation. If the data rate is 204.8 kS/s, each ADC actually samples its input signal at 26.2144 MS/s (128 times the data rate) and produces 1-bit samples that are applied to the digital filter. This filter then expands the data to 16 bits, rejects signal components greater than 102.4 kHz (the Nyquist frequency), and resamples the data at the more conventional rate of 204.8 kS/s.

Although a 1-bit quantizer introduces a large amount of quantization error to the signal, the 1-bit, 26 MS/s from the ADC carry all the information used to produce 16-bit samples at 204.8 kS/s. The delta-sigma ADC achieves this conversion from high speed to high resolution by adding a large amount of random noise to the signal so that the resulting quantization noise, although large, is restricted to frequencies above 102.4 kHz. This noise is not correlated with the input signal and is almost completely rejected by the digital filter.

The resulting output of the filter is a band-limited signal with a dynamic range of over 90 dB. One of the advantages of a delta-sigma ADC is that it uses a 1-bit DAC as an internal reference, whereas most 16-bit ADCs use 16-bit resistor network DACs or capacitor-network DACs. As a result, the delta-sigma ADC is free from the kind of differential nonlinearity (DNL) that is inherent in most high-resolution ADCs. This lack of DNL is especially beneficial when the ADC is converting low-level signals, in which noise and distortion are directly affected by converter DNL.

Noise

The NI 4551/4552 analog inputs typically have a dynamic range of more than 90 dB. The dynamic range of a circuit is the ratio of the magnitudes of the largest signal the circuit can carry and the residual noise in the absence of a signal. In a 16-bit system, the largest signal is taken to be a full-scale sine wave that peaks at the codes +32,767 and -32,768. Such a sine wave has an rms magnitude of $32,768/1.414 = 23,170.475$ least significant bits (LSBs).

A grounded channel of the NI 4551/4552 has a noise level of about 0.65 LSB rms; this amount fluctuates. The ratio of $23,170.475 / 0.65$ is about 35647, or 91.0 dB—the dynamic range. Several factors can degrade the noise performance of the inputs.

First, noise can be picked up from nearby electronics. The NI 4551/4552 works best when it is kept as far away as possible from other plug-in devices, power supplies, disk drives, and computer monitors. Cabling is also critical. Make sure to use well shielded coaxial or balanced cables for all connections, and route the cables away from sources of interference such as computer monitors, switching power supplies, and fluorescent lights.

Finally, choose the sample rate carefully. Take advantage of the antialias filtering that removes signals beyond the band of interest. Computer monitor noise, for example, typically occurs at frequencies between 15 and 50 kHz. If the signal of interest is restricted to below 10 kHz, for example, the antialias filters reject the monitor noise outside the frequency band of interest. The frequency response inside the band of interest is not influenced if the sample rate were between approximately 21.6 and 28 kS/s.

Analog Output Circuitry

- ◆ NI 4551 only

The NI 4551 has two analog output channels, either of which is illustrated in Figure 4-4.

A common application for the analog output is to stimulate a system under test while measuring the response with the analog inputs. The input and output sample clocks are synchronized and derived from the same DDS clock. The input and output clocks can differ from each other by a factor of 2 (1, 2, 4, 8, ... 128) while still maintaining their synchronization. Output conversions occur simultaneously at software-programmable rates from 1.25 to 51.2 kS/s, in increments of 47.684 μ S/s.

The analog output circuitry uses eight-times oversampling interpolators with 64-times oversampling delta-sigma modulators to generate high-quality signals. The output channel has a range up to ± 10 V ($7.07 V_{\text{rms}}$) and can be driven as SE or DIFF. The analog output also has an attenuation stage so you can choose attenuation of 0, -20 , or -40 dB.

Because of the delta-sigma modulating DAC, the instrument is immune to DNL distortion. The analog output stage generates signals with extremely low noise and low distortion. Because the instrument has a 93 dB dynamic range, it is possible to generate low-noise waveforms. The instrument also has excellent amplitude flatness of ± 0.2 dB within the frequency range of DC to 23 kHz and has a total harmonic distortion (THD) of -95 dB at 1 kHz. With these specifications, you are assured of the quality and integrity of the output signals generated.

Anti-Image Filtering

A sampled signal repeats itself throughout the frequency spectrum. These repetitions begin above one-half the sample rate (F_s) and, at least in theory, continue up through the spectrum to infinity, as shown in Figure 6-5a. Because the sample data actually represents only the frequency components below one-half the sample rate (the baseband), it is better to filter out all these extra images of the signal. The NI 4551 accomplishes this filtering in two stages.

First, the data is digitally resampled at eight times the original sample rate. Then, a linear-phase digital filter removes almost all energy above one-half the original sample rate and sends the data at the eight-times rate to the DAC, as shown in Figure 6-5b. Some further (inherent) filtering occurs at

the DAC because the data is digitally sampled and held at eight times the sample rate. This filtering has a $\sin x/x$ response, yielding nulls at multiples of eight times the sample rate, as shown in Figure 6-5c. Still, images remain and they must be filtered out. Each output channel of the NI 4551 has discrete-time (switched-capacitor) and continuous-time analog filters that remove the high-frequency images, as shown in Figure 6-5d.




Figure 6-5. Signal Spectra in the DAC

The DAC

The 64-times oversampling delta-sigma DACs on the NI 4551 work in the same way as delta-sigma ADCs, only in reverse. The digital data *first*

passes through a digital lowpass filter and *then* goes to the delta-sigma modulator.

In the ADC the delta-sigma modulator is analog circuitry that converts high-resolution *analog* signals to high-rate, 1-bit digital data, whereas in the DAC the delta-sigma modulator is digital circuitry that converts high-resolution *digital* data to high-rate, 1-bit digital data. As in the ADC, the modulator frequency shapes the quantization noise so that almost all of its energy is above the signal frequency (refer to [The ADC](#), earlier in this chapter).

The digital 1-bit data is then sent directly to a simple 1-bit DAC. This DAC can have only one of two analog values and, therefore, is inherently perfectly linear. The output of the DAC, however, has a large amount of quantization noise at higher frequencies and, as described in the section, [Anti-Image Filtering](#), some images still remain near multiples of eight times the sample rate.

Two analog filters eliminate the quantization noise and the images. The first is a fifth-order, switched-capacitor filter in which the cutoff frequency scales with the sample frequency and is approximately 0.52 times the sample frequency. This filter has a four-pole Butterworth response and an extra pole at about 1.04 times the sample frequency.

The second filter is a continuous-time, second-order Butterworth filter in which the cutoff frequency (at 80 kHz) does not scale with the sample frequency. This filter mainly removes high-frequency images from the 64-times oversampled switched-capacitor filter. These filters cause a delay between the input digital data and the output analog data of 34.6 ± 0.5 sample periods.

Calibration

The NI 4551 analog outputs have calibration adjustments. Onboard CalDACs remove the offset and gain errors for each channel. For complete calibration instructions, refer to Chapter 5, [Calibration](#).

Mute Feature

The two-channel DAC chip on the NI 4551 goes into *mute* mode if the chip receives at least 4,096 consecutive zero values on both channels at once. In mute mode, the outputs clamp to ground and the noise floor drops from about 92 dB below full-scale to about 120 dB below full-scale. Upon receiving any nonzero data, the DAC instantly reverts to normal mode. Mute mode quiets the background noise to extremely low levels when no

waveforms are being generated. Mute mode has a slightly different offset from the normal offset when zeros are being sent. As a result, the DAC has one offset for the first 4,096 zero samples and another offset in mute mode for as long as zeros are sent. This difference is usually less than 1 mV.

Specifications

This appendix lists the specifications of the NI 4551/4552. These specifications are typical at 25° C unless otherwise noted. The system must be allowed to warmup for 15 minutes to achieve the rated accuracy.



Note *Be sure to keep the cover on your computer to maintain forced air cooling.*

Analog Input

Channel Characteristics

- Number of channels 2 (NI 4551) or 4 (NI 4552),
simultaneously sampled
- Input configuration..... True differential
- Resolution 16 bits
- Type of ADC..... Delta-sigma, 128 times
oversampling
- Sample rates 5 kS/s to 204.8 kS/s in increments
of 190.735 μ S/s
- Frequency accuracy..... ± 100 ppm
- Input signal ranges Software-selectable

Gain		Full-Scale Range (Peak)
Linear	Log	
0.1	-20 dB	± 42.4 V
0.316	-10 dB	± 31.6 V
1	0 dB	± 10.0
3.16	+10 dB	± 3.16 V

Gain		Full-Scale Range (Peak)
Linear	Log	
10	+20 dB	±1.00 V
31.6	+30 dB	±0.316 V
100	+40 dB	±0.100 V
316	+50 dB	±0.0316 V
1000	+60 dB	±0.0100 V

FIFO buffer size.....256 samples

Data transfersDMA, programmed I/O, interrupt

Transfer Characteristics

INL (relative accuracy).....±2 LSB

DNL±0.5 LSB typ, ±1 LSB max, no missing codes

Offset (residual DC)

Gain	Max Offset
-20 dB	±30 mV
-10 dB	±10 mV
0 dB	±3 mV
+10 dB	±1 mV
+20 dB	±300 µV
+30, +40, +50, +60 dB	±100 µV

Gain (amplitude accuracy).....±0.1 dB, $f_{in} = 1$ kHz

Amplifier Characteristics

Input impedance	1 M Ω in parallel with 50 pF (+ and - each to AIGND)
Frequency response	
Gain	
0, +10, +20, +30, +40 dB	± 0.1 dB, 0 through 95 kHz, 204.8 kS/s, DC coupling
-20, -10, +50, +60 dB	± 1 dB, 0 – 95 kHz, ± 0.1 dB, 0 – 20 kHz
-3 dB bandwidth	0.493 f_s
Input coupling	AC or DC, software-selectable
AC -3 dB cutoff frequency	3.4 Hz
Common-mode range	
Gain ≥ 0 dB	Both + and - should remain within ± 12 V of AIGND
Gain < 0 dB	Both + and - should remain within ± 42.4 V of AIGND
Overvoltage protection	± 42.4 V, powered on or off (± 400 V guaranteed by design, but not tested or certified to operate beyond ± 42.4 V)
Inputs protected	ACH0, ACH1, ACH2, ACH3
Common-mode rejection ratio	
($f_{in} < 1$ kHz)	90 dB, Gain ≥ 0 dB 60 dB, Gain < 0 dB




Figure A-1. Idle Channel Noise (Typical)

Input noise spectral density8 nV/√Hz (achievable only at Gain = +50 dB or +60 dB)

Dynamic Characteristics

Alias-free bandwidthDC to 0.464 f_s

Alias rejection80 dB, $0.536 f_s < f_{in} < 63.464 f_s$

Spurious-free dynamic range95 dB

THD-80 dB; -90 dB for $f_{in} < 20$ kHz or signal $< 1 V_{rms}$

IMD-100 dB (CCIF 14 kHz + 15 kHz)

Crosstalk (channel separation).....-100 dB, DC to 100 kHz

Phase linearity	$\pm 1^\circ$, Gain ≥ 0 dB, $\pm 2^\circ$, Gain < 0 dB
Interchannel phase.....	$\pm 1^\circ$, Gain ≥ 0 dB, $\pm 2^\circ$, Gain < 0 dB (same configuration all input channels)
Interchannel gain mismatch	± 0.1 dB, for all gains (same configuration for all input channels)
Signal delay	42 sample periods, any sample rate (time from when signal enters analog input to when digital data is available)

Onboard Calibration Reference

DC level	5.000 V ± 2.5 mV
Temperature coefficient	± 5 ppm/ $^\circ$ C max
Long-term stability.....	± 15 ppm/ $\sqrt{1,000}$ h

Analog Output

- ◆ NI 4551 only

Channel Characteristics

Number of channels	2 simultaneously updated
Output configuration	Balanced differential
Resolution	16 bits
Type of DAC.....	Delta-sigma, 64-times oversampling
Sample rates	1.25 – 51.2 kS/s in increments of 47.684 μ S/s
Frequency accuracy.....	± 100 ppm

Output signal range, software-selectable

Attenuation		Full-Scale Range
Linear	Log	
1	0 dB	±10.0 V
10	20 dB	±1.00 V
100	40 dB	±0.100 V

FIFO buffer size.....256 samples

Data transfersDMA, programmed I/O, Interrupt

Transfer Characteristics

Offset (residual DC)±5 mV max, any gain

Gain (amplitude accuracy).....±0.1 dB, $f_{out} = 1$ kHz

Voltage Output Characteristics

Output impedance22 Ω between + and –
DACxOUT, 4.55 kΩ to AOGND

Frequency response±0.2 dB, 0 to 23 kHz, 51.2 kS/s

-3 dB bandwidth0.492 f_s

Output couplingDC

Short-circuit protection Yes (+ and – may be shorted together indefinitely)

Outputs protected.....±DAC0OUT, ±DAC1OUT

Idle channel noise–91 dB f_s , DC to 23 kHz
measurement bandwidth

Dynamic Characteristics

Image-free bandwidth	DC to $0.450 f_s$
Image rejection.....	90 dB, $0.550 f_s < f_{out} < 63.450 f_s$
Spurious-free dynamic range	90 dB, DC to 100 kHz
THD	-80 dB; -90 dB for $f_{out} < 5$ kHz or signal $< 1 V_{rms}$
IMD.....	-90 dB (CCIF 14 kHz + 15 kHz)
Crosstalk (channel separation)	-80 dB, DC to 23 kHz
Phase linearity	$\pm 1^\circ$
Interchannel phase.....	$\pm 1^\circ$ (same configuration both output channels)
Interchannel gain mismatch	± 0.1 dB, for all attenuations (same configuration both output channels)
Signal delay	34.6 ± 0.5 sample periods, any sample rate (time from when digital data is expressed to when analog signal appears at output terminals)

Digital I/O

Number of channels	32 input/output
Compatibility	TTL/CMOS

Digital logic levels

Level	Min	Max
Input low voltage	0.0 V	0.8 V
Input high voltage	2.0 V	5.0 V
Input low current ($V_{in} = 0\text{ V}$)	—	-10 μA
Input high current ($V_{in} = 5\text{ V}$)	—	10 μA
Output low voltage ($I_{OL} = 10\text{ mA}$)	—	0.45 V
Output high voltage ($I_{OH} = 3.5\text{ mA}$)	2.4 V	—

Power-on stateInput (High-Z)

Data transfersProgrammed I/O

Triggers

Analog Trigger

Source

NI 4551ACH<0..1>

NI 4552ACH<0..3>

Level \pm full-scale

SlopePositive or negative
(software selectable)

Resolution16 bits

HysteresisProgrammable

Digital Trigger

Compatibility	TTL
Response	Rising or falling edge
Pulse width.....	10 ns min

Bus Interface

Type	PCI Master/Slave
------------	------------------

Power Requirement

Power (NI 4551).....	+5 V, 2.8 A idle, 3.8 A active typical ¹ +12 V, 11 mA typical (not including momentary relay switching) -12 V, 40 mA typical +3.3 V, unused
Power (NI 4552).....	+5 V, 3.3 A idle, 4.3 A active typical ¹ +12 V, 150 mA typical (not including momentary relay switching) -12 V, unused +3.3 V, unused
Available power	+4.65 to +5.25 VDC at 0.5 A (analog I/O connector)
Available power	+4.65 to +5.25 VDC at 1.0 A (digital I/O connector)

¹ Power consumption can vary based on instrument personality.

Physical

Dimensions	10.65 by 31.19 by 1.84 cm (4.19 by 12.28 by 0.73 in.) not including connectors
Digital I/O connector	50-pin VHDIC female type
Analog I/O connector	68-pin VHDIC female type

Environment

Operating temperature	0° C to +40° C
Storage temperature range	-25° C to +85° C
Relative humidity	10% to 95%, no condensation

Calibration

Calibration interval	1 year
----------------------------	--------

Pin Connections

This appendix describes the pin connections on the optional 68-pin digital accessories for the NI 4551 and NI 4552 instruments.

DIO0	1	35	DGND
DIO1	2	36	DGND
DIO2	3	37	DGND
DIO3	4	38	DGND
DIO4	5	39	DGND
DIO5	6	40	DGND
DIO6	7	41	DGND
DIO7	8	42	DGND
DIO8	9	43	DGND
DIO9	10	44	DGND
DIO10	11	45	DGND
DIO11	12	46	DGND
DIO12	13	47	DGND
DIO13	14	48	DGND
DIO14	15	49	DGND
DIO15	16	50	DGND
DIO16	17	51	DGND
DIO17	18	52	DGND
DIO18	19	53	DGND
DIO19	20	54	DGND
DIO20	21	55	DGND
DIO21	22	56	DGND
DIO22	23	57	DGND
+5 V	24	58	+5 V
DIO23	25	59	DGND
DIO24	26	60	DGND
DIO25	27	61	DGND
DIO26	28	62	DGND
DIO27	29	63	DGND
DIO28	30	64	DGND
DIO29	31	65	DGND
DIO30	32	66	DGND
DIO31	33	67	DGND
EXT_TRIG	34	68	DGND

Figure B-1. 68-Pin Digital Connector for Any Digital Accessory



Customer Communication

For your convenience, this appendix contains forms to help you gather the information necessary to help us solve your technical problems and a form you can use to comment on the product documentation. When you contact us, we need the information on the Technical Support Form and the configuration form, if your manual contains one, about your system configuration to answer your questions as quickly as possible.

National Instruments has technical assistance through electronic, fax, and telephone systems to quickly provide the information you need. Our electronic services include a bulletin board service, an FTP site, a fax-on-demand system, and e-mail support. If you have a hardware or software problem, first try the electronic support systems. If the information available on these systems does not answer your questions, we offer fax and telephone support through our technical support centers, which are staffed by applications engineers.

Electronic Services

Bulletin Board Support

National Instruments has BBS and FTP sites dedicated for 24-hour support with a collection of files and documents to answer most common customer questions. From these sites, you can also download the latest instrument drivers, updates, and example programs. For recorded instructions on how to use the bulletin board and FTP services and for BBS automated information, call 512 795 6990. You can access these services at:

United States: 512 794 5422

Up to 14,400 baud, 8 data bits, 1 stop bit, no parity

United Kingdom: 01635 551422

Up to 9,600 baud, 8 data bits, 1 stop bit, no parity

France: 01 48 65 15 59

Up to 9,600 baud, 8 data bits, 1 stop bit, no parity

FTP Support

To access our FTP site, log on to our Internet host, `ftp.natinst.com`, as anonymous and use your Internet address, such as `joesmith@anywhere.com`, as your password. The support files and documents are located in the `/support` directories.

Fax-on-Demand Support

Fax-on-Demand is a 24-hour information retrieval system containing a library of documents on a wide range of technical information. You can access Fax-on-Demand from a touch-tone telephone at 512 418 1111.

E-Mail Support (Currently USA Only)

You can submit technical support questions to the applications engineering team through e-mail at the Internet address listed below. Remember to include your name, address, and phone number so we can contact you with solutions and suggestions.

support@natinst.com

Telephone and Fax Support

National Instruments has branch offices all over the world. Use the list below to find the technical support number for your country. If there is no National Instruments office in your country, contact the source from which you purchased your software to obtain support.

Country	Telephone	Fax
Australia	03 9879 5166	03 9879 6277
Austria	0662 45 79 90 0	0662 45 79 90 19
Belgium	02 757 00 20	02 757 03 11
Brazil	011 288 3336	011 288 8528
Canada (Ontario)	905 785 0085	905 785 0086
Canada (Québec)	514 694 8521	514 694 4399
Denmark	45 76 26 00	45 76 26 02
Finland	09 725 725 11	09 725 725 55
France	01 48 14 24 24	01 48 14 24 14
Germany	089 741 31 30	089 714 60 35
Hong Kong	2645 3186	2686 8505
Israel	03 6120092	03 6120095
Italy	02 413091	02 41309215
Japan	03 5472 2970	03 5472 2977
Korea	02 596 7456	02 596 7455
Mexico	5 520 2635	5 520 3282
Netherlands	0348 433466	0348 430673
Norway	32 84 84 00	32 84 86 00
Singapore	2265886	2265887
Spain	91 640 0085	91 640 0533
Sweden	08 730 49 70	08 730 43 70
Switzerland	056 200 51 51	056 200 51 55
Taiwan	02 377 1200	02 737 4644
United Kingdom	01635 523545	01635 523154
United States	512 795 8248	512 794 5678

Technical Support Form

Photocopy this form and update it each time you make changes to your software or hardware, and use the completed copy of this form as a reference for your current configuration. Completing this form accurately before contacting National Instruments for technical support helps our applications engineers answer your questions more efficiently.

If you are using any National Instruments hardware or software products related to this problem, include the configuration forms from their user manuals. Include additional pages if necessary.

Name _____

Company _____

Address _____

Fax (____) _____ Phone (____) _____

Computer brand _____ Model _____ Processor _____

Operating system (include version number) _____

Clock speed _____ MHz RAM _____ MB Display adapter _____

Mouse ____ yes ____ no Other adapters installed _____

Hard disk capacity _____ MB Brand _____

Instruments used _____

National Instruments hardware product model _____ Revision _____

Configuration _____

National Instruments software product _____ Version _____

Configuration _____

The problem is: _____

List any error messages: _____

The following steps reproduce the problem: _____

NI 4551/4552 Hardware and Software Configuration Form

Record the settings and revisions of your hardware and software on the line to the right of each item. Complete a new copy of this form each time you revise your software or hardware configuration, and use this form as a reference for your current configuration. Completing this form accurately before contacting National Instruments for technical support helps our applications engineers answer your questions more efficiently.

National Instruments Products

NI 4551/4552 instrument _____

NI 4551/4552 instrument serial number _____

Base memory address of the NI 4551/4552 instrument _____

Programming choice and version (NI-DAQ, LabVIEW, or other) _____

Other boards in system _____

Base I/O address of other boards _____

DMA channels of other boards _____

Interrupt level of other boards _____

Other Products

Computer make and model _____

Microprocessor _____

Clock frequency or speed _____

Type of video board installed _____

Operating system version _____

Operating system mode _____

Programming language _____

Programming language version _____

Other boards in system _____

Base I/O address of other boards _____

DMA channels of other boards _____

Interrupt level of other boards _____

Documentation Comment Form

National Instruments encourages you to comment on the documentation supplied with our products. This information helps us provide quality products to meet your needs.

Title: *NI 4551/4552 User Manual*

Edition Date: April 1998

Part Number: 321934A-01

Please comment on the completeness, clarity, and organization of the manual.

If you find errors in the manual, please record the page numbers and describe the errors.

Thank you for your help.

Name _____

Title _____

Company _____

Address _____

E-Mail Address _____

Phone (____) _____ Fax (____) _____

Mail to: Technical Publications
National Instruments Corporation
6504 Bridge Point Parkway
Austin, Texas 78730-5039

Fax to: Technical Publications
National Instruments Corporation
512 794 5678

Glossary

Prefix	Meanings	Value
p-	pico	10^{-12}
n-	nano-	10^{-9}
μ -	micro-	10^{-6}
m-	milli-	10^{-3}
k-	kilo-	10^3
M-	mega-	10^6
G-	giga-	10^9
t-	tera-	10^{12}

Numbers/Symbols

%	percent
+	positive of, or plus
-	negative of, or minus
/	per
°	degree
Ω	ohm
+5 V	+5 VDC source signal

A

A	amperes
AC	alternating current
AC coupled	allowing the transmission of AC signals while blocking DC signals

ACH	analog input channel signal
A/D	analog-to-digital
ADC	analog-to-digital converter—an electronic device, often an integrated circuit, that converts an analog voltage to a digital number
ADC resolution	the size of the discrete steps in the ADC's input-to-output transfer function; therefore, the smallest voltage difference an ADC can discriminate with a single measurement
AIGND	analog input ground signal
alias	a false lower frequency component that appears in sampled data acquired at too low a sampling rate
amplification	a type of signal conditioning that improves accuracy in the resulting digitized signal and reduces noise
amplitude flatness	a measure of how close to constant the gain of a circuit remains over a range of frequencies
antialiasing filter	a low-pass filter preceding an ADC (usually a brickwall filter) that rejects signal energy above the Nyquist frequency (1/2 the sample rate) of the ADC so that the ADC does not mistake out-of-band signals for in-band signals.
anti-imaging filter	a low-pass filter after a DAC (usually a brickwall filter) that rejects signal energy above the Nyquist frequency (1/2 the sample rate) of the DAC in order to suppress out-of-band images of the in-band signal created by the D/A conversion process.
AOGND	analog output ground signal
ASIC	Application-Specific Integrated Circuit—a proprietary semiconductor component designed and manufactured to perform a set of specific functions for a specific customer
asynchronous	(1) hardware—a property of an event that occurs at an arbitrary time, without synchronization to a reference clock (2) software—a property of a function that begins an operation and returns prior to the completion or termination of the operation

attenuate	to decrease the amplitude of a signal
attenuation ratio	the factor by which a signal's amplitude is decreased
B	
b	bit—one binary digit, either 0 or 1
B	byte—eight related bits of data, an eight-bit binary number. Also used to denote the amount of memory required to store one byte of data
bandwidth	the range of frequencies present in a signal, or the range of frequencies to which a measuring device can respond
base address	a memory address that serves as the starting address for programmable registers. All other addresses are located by adding to the base address.
binary	a number system with a base of 2
bipolar	a signal range that includes both positive and negative values (for example, -5 V to $+5\text{ V}$)
BNC	a type of coaxial signal connector
brickwall filter	a low-pass filter having a very flat passband, a very sudden, sharp transition region, and high rejection in the stopband.
buffer	temporary storage for acquired or generated data (software)
burst-mode	a high-speed data transfer in which the address of the data is sent followed by back-to-back data words while a physical signal is asserted
bus	the group of conductors that interconnect individual circuitry in a computer. Typically, a bus is the expansion vehicle to which I/O or other devices are connected. Examples of PC buses are the ISA and PCI bus.
bus master	a type of a plug-in board or controller with the ability to read and write devices on the computer bus

C

C	Celsius
CalDAC	calibration DAC
channel	pin or wire lead to which you apply or from which you read the analog or digital signal. Analog signals can be single-ended or differential. For digital signals, you group channels to form ports. Ports usually consist of either four or eight digital channels.
circuit trigger	a condition for starting or stopping clocks
clip	clipping occurs when an input signal exceeds the input range of the amplifier
clock	hardware component that controls timing for reading from or writing to groups
CMOS	complementary metal-oxide semiconductor
CMRR	common-mode rejection ratio—a measure of an instrument’s ability to reject interference from a common-mode signal, usually expressed in decibels
code width	the smallest detectable change in an input voltage of a DAQ device
common-mode range	the input range over which a circuit can handle a common-mode signal
common-mode signal	the mathematical average voltage, relative to the computer’s ground, of the signals from a differential input
common-mode voltage	any voltage present at the instrumentation amplifier inputs with respect to amplifier ground
compensation range	the range of a parameter for which compensating adjustment can be made
conditional retrieval	a method of triggering in which you simulate an analog trigger using software. Also called software triggering.
conversion device	device that transforms a signal from one form to another. For example, analog-to-digital converters (ADCs) for analog input, digital-to-analog converters (DACs) for analog output, digital input or output ports, and counter/timers are conversion devices.

conversion time	the time required, in an analog input or output system, from the moment a channel is interrogated (such as with a read instruction) to the moment that accurate data is available
counter/timer	a circuit that counts external pulses or clock pulses (timing)
coupling	the manner in which a signal is connected from one location to another
crosstalk	an unwanted signal on one channel due to an input on a different channel
current drive capability	the amount of current a digital or analog output channel is capable of sourcing or sinking while still operating within voltage range specifications
current sinking	the ability of a DAQ board to dissipate current for analog or digital output signals
current sourcing	the ability of a DAQ board to supply current for analog or digital output signals

D

D/A	digital-to-analog
DAC	digital-to-analog converter—an electronic device, often an integrated circuit, that converts a digital number into a corresponding analog voltage or current
DAC _x OUT	analog channel <i>x</i> output signal
daisy-chain	a method of propagating signals along a bus, in which the devices are prioritized on the basis of their position on the bus
DAQ	data acquisition—(1) collecting and measuring electrical signals from sensors, transducers, and test probes or fixtures and inputting them to a computer for processing; (2) collecting and measuring the same kinds of electrical signals with A/D and/or DIO boards plugged into a computer, and possibly generating control signals with D/A and/or DIO boards in the same computer
dB	decibel—the unit for expressing a logarithmic measure of the ratio of two signal levels: $\text{dB}=20\log_{10}(V_1/V_2)$, for signals in volts
DC	direct current

DC coupled	allowing the transmission of both AC and DC signals
DDS	direct digital synthesis
default setting	a default parameter value recorded in the driver. In many cases, the default input of a control is a certain value (often 0) that means <i>use the current default setting</i> . For example, the default input for a parameter may be <i>do not change current setting</i> , and the default setting may be <i>no AMUX-64T boards</i> . If you do change the value of such a parameter, the new value becomes the new setting. You can set default settings for some parameters in the configuration utility or manually using switches located on the device.
delta-sigma modulating ADC	a high-accuracy circuit that samples at a higher rate and lower resolution than is needed and (by means of feedback loops) pushes the quantization noise above the frequency range of interest. This out-of-band noise is typically removed by digital filters.
device	a plug-in data acquisition board, card, or pad that can contain multiple channels and conversion devices. Plug-in boards, PCMCIA cards, and devices such as the DAQPad-1200, which connects to your computer parallel port, are all examples of DAQ devices. SCXI modules are distinct from devices, with the exception of the SCXI-1200, which is a hybrid.
DGND	digital ground signal
DIFF	differential mode
differential input	an analog input consisting of two terminals, both of which are isolated from computer ground, whose difference is measured
differential measurement system	a way you can configure your device to read signals, in which you do not need to connect either input to a fixed reference, such as the earth or a building ground
digital port	<i>See</i> port.
digital trigger	a TTL level signal having two discrete levels—a high and a low level
DIO	digital input/output
DMA	direct memory access—a method by which data can be transferred to/from computer memory from/to a device or memory on the bus while the processor does something else. DMA is the fastest method of transferring data to/from computer memory.

DNL	differential nonlinearity—a measure in least significant bit of the worst-case deviation of code widths from their ideal value of 1 LSB
down counter	performing frequency division on an internal signal
drivers	software that controls a specific hardware device such as a DAQ board or a GPIB interface board
dynamic range	the ratio of the largest signal level a circuit can handle to the smallest signal level it can handle (usually taken to be the noise level), normally expressed in decibels

E

EEPROM	electrically erasable programmable read-only memory—ROM that can be erased with an electrical signal and reprogrammed
EMC	electromechanical compliance
encoder	a device that converts linear or rotary displacement into digital or pulse signals. The most popular type of encoder is the optical encoder, which uses a rotating disk with alternating opaque areas, a light source, and a photodetector.
EPROM	erasable programmable read-only memory—ROM that can be erased (usually by ultraviolet light exposure) and reprogrammed
event	the condition or state of an analog or digital signal
expansion ROM	an onboard EEPROM that may contain device-specific initialization and system boot functionality
EXT_TRIG	external digital trigger
external trigger	a voltage pulse from an external source that triggers an event such as A/D conversion

F

false triggering	triggering that occurs at an unintended time
FIFO	first-in first-out memory buffer—the first data stored is the first data sent to the acceptor. FIFOs are often used on DAQ devices to temporarily store incoming or outgoing data until that data can be retrieved or output. For example, an analog input FIFO stores the results of A/D conversions until the data can be retrieved into system memory, a process that requires the servicing of interrupts and often the programming of the DMA controller. This process can take several milliseconds in some cases. During this time, data accumulates in the FIFO for future retrieval. With a larger FIFO, longer latencies can be tolerated. In the case of analog output, a FIFO permits faster update rates, because the waveform data can be stored on the FIFO ahead of time. This again reduces the effect of latencies associated with getting the data from system memory to the DAQ device.
filtering	a type of signal conditioning that allows you to attenuate unwanted portions of the signal you are trying to measure
FIR	finite impulse response—a non recursive digital filter with linear phase
flash ADC	an ADC whose output code is determined in a single step by a bank of comparators and encoding logic
floating signal sources	signal sources with voltage signals that are not connected to an absolute reference or system ground. Also called nonreferenced signal sources. Some common example of floating signal sources are batteries, transformers, or thermocouples.
F_s	sample rate
ft	feet

G

gain	the factor by which a signal is amplified, sometimes expressed in decibels
gain accuracy	a measure of deviation of the gain of an amplifier from the ideal gain

GND	ground
grounded measurement system	<i>See</i> SE.
H	
h	hour
half-power bandwidth	the frequency range over which a circuit maintains a level of at least -3 dB with respect to the nominal level
handshaked digital I/O	a type of digital acquisition/generation where a device or module accepts or transfers data after a digital pulse has been received. Also called latched digital I/O.
hardware	the physical components of a computer system, such as the circuit boards, plug-in boards, chassis, enclosures, peripherals, and cables
hardware triggering	a form of triggering where you set the start time of an acquisition and gather data at a known position in time relative to a trigger signal
hysteresis	the lag between making a change and the effect of the change
Hz	hertz—cycles per second. Specifically refers to the repetition frequency of a waveform.
I	
IC	integrated circuit
IMD	intermodulation distortion—the ratio, in decibels, of the total rms signal level of harmonic sum and difference distortion products, to the overall rms signal level. The test signal is two sine waves added together according to the following standards: SMPTE—A 60 Hz sine wave and a 7 kHz sine wave added in a 4:1 amplitude ratio. DIN—A 250 Hz sine wave and an 8 kHz sine wave added in a 4:1 amplitude ratio. CCIF—A 14 kHz sine wave and a 15 kHz sine wave added in a 1:1 amplitude ratio.
in.	inches

INL	integral nonlinearity—a measure in LSB of the worst-case deviation from the ideal A/D or D/A transfer characteristic of the analog I/O circuitry
input bias current	the current that flows into the inputs of a circuit
input impedance	the measured resistance and capacitance between the input terminals of a circuit
input offset current	the difference in the input bias currents of the two inputs of an instrumentation amplifier
instrument driver	a set of high-level software functions that controls a specific GPIB, VXI, or RS-232 programmable instrument or a specific plug-in DAQ board. Instrument drivers are available in several forms, ranging from a function callable language to a virtual instrument (VI) in LabVIEW.
instrumentation amplifier	a circuit whose output voltage with respect to ground is proportional to the difference between the voltages at its two inputs
integrating ADC	an ADC whose output code represents the average value of the input voltage over a given time interval
interrupt	a computer signal indicating that the CPU should suspend its current task to service a designated activity
interrupt level	the relative priority at which a device can interrupt
I/O	input/output—the transfer of data to/from a computer system involving communications channels, operator interface devices, and/or data acquisition and control interfaces
I_{OH}	current, output high
I_{OL}	current, output low
IRQ	interrupt request
isolation	a type of signal conditioning in which you isolate the transducer signals from the computer for safety purposes. This protects you and your computer from large voltage spikes and makes sure the measurements from the DAQ device are not affected by differences in ground potentials.
isolation voltage	the voltage that an isolated circuit can normally withstand, usually specified from input to input and/or from any input to the amplifier output, or to the computer bus

K

k	kilo—the standard metric prefix for 1,000, or 10^3 , used with units of measure such as volts, hertz, and meters
K	kilo—the prefix for 1,024, or 2^{10} , used with B in quantifying data or computer memory
kbytes/s	a unit for data transfer that means 1,024 bytes/s
kS	1,000 samples
Kword	1,024 words of memory

L

LabVIEW	laboratory virtual instrument engineering workbench
latched digital I/O	a type of digital acquisition/generation where a device or module accepts or transfers data after a digital pulse has been received. Also called handshaked digital I/O.
library	a file containing compiled object modules, each comprised of one of more functions, that can be linked to other object modules that make use of these functions. NIDAQMSC.LIB is a library that contains NI-DAQ functions. The NI-DAQ function set is broken down into object modules so that only the object modules that are relevant to your application are linked in, while those object modules that are not relevant are not linked.
linearity	the adherence of device response to the equation $R = KS$, where R = response, S = stimulus, and K = a constant
linearization	a type of signal conditioning in which software linearizes the voltage levels from transducers, so the voltages can be scaled to measure physical phenomena
low frequency corner	in an AC-coupled circuit, the frequency below which signals are attenuated by at least 3 dB
LSB	least significant bit

M

m	meters
M	(1) Mega, the standard metric prefix for 1 million or 10^6 , when used with units of measure such as volts and hertz; (2) mega, the prefix for 1,048,576, or 2^{20} , when used with B to quantify data or computer memory
Mbytes/s	a unit for data transfer that means 1,048,576 bytes/s
memory buffer	<i>See</i> buffer.
MITE	MXI Interface to Everything—a custom ASIC designed by National Instruments that implements the PCI bus interface. The MITE supports bus mastering for high-speed data transfers over the PCI bus.
MS	million samples
MSB	most significant bit
MTBF	mean time between failure
MTTR	mean time to repair—predicts downtime and how long it takes to fix a product
N	
NC	normally closed, or not connected
NI-DAQ	National Instruments driver software for DAQ hardware
NIST	National Institute of Standards and Technology
noise	an undesirable electrical signal—Noise comes from external sources such as the AC power line, motors, generators, transformers, fluorescent lights, soldering irons, CRT displays, computers, electrical storms, welders, radio transmitters, and internal sources such as semiconductors, resistors, and capacitors. Noise corrupts signals you are trying to send or receive.
nonlatched digital I/O	a type of digital acquisition/generation where LabVIEW updates the digital lines or port states immediately or returns the digital value of an input line. Also called immediate digital I/O or non-handshaking.

nonreferenced signal sources signal sources with voltage signals that are not connected to an absolute reference or system ground. Also called floating signal sources. Some common example of nonreferenced signal sources are batteries, transformers, or thermocouples.

NRSE nonreferenced single-ended mode—all measurements are made with respect to a common (NRSE) measurement system reference, but the voltage at this reference can vary with respect to the measurement system ground

Nyquist Frequency one-half of F_s

0

onboard channels channels provided by the plug-in data acquisition board

operating system base-level software that controls a computer, runs programs, interacts with users, and communicates with installed hardware or peripheral devices

optical isolation the technique of using an optoelectric transmitter and receiver to transfer data without electrical continuity, to eliminate high-potential differences and transients

output settling time the amount of time required for the analog output voltage to reach its final value within specified limits

output slew rate the maximum rate of change of analog output voltage from one level to another

P

passband the range of frequencies which a device can properly propagate or measure

pattern generation a type of handshaked (latched) digital I/O in which internal counters generate the handshaked signal, which in turn initiates a digital transfer. Because counters output digital pulses at a constant rate, this means you can generate and retrieve patterns at a constant rate because the handshaked signal is produced at a constant rate.

PCI	Peripheral Component Interconnect—a high-performance expansion bus architecture originally developed by Intel to replace ISA and EISA. It is achieving widespread acceptance as a standard for PCs and work-stations; it offers a theoretical maximum transfer rate of 132 Mbytes/s.
peak to peak	a measure of signal amplitude; the difference between the highest and lowest excursions of the signal
PFI	programmable function input
Plug and Play devices	devices that do not require DIP switches or jumpers to configure resources on the devices—also called switchless devices
port	(1) a communications connection on a computer or a remote controller (2) a digital port, consisting of four or eight lines of digital input and/or output
posttriggering	the technique used on a DAQ board to acquire a programmed number of samples after trigger conditions are met
potentiometer	an electrical device the resistance of which can be manually adjusted; used for manual adjustment of electrical circuits and as a transducer for linear or rotary position
ppm	parts per million
pretriggering	the technique used on a DAQ board to keep a continuous buffer filled with data, so that when the trigger conditions are met, the sample includes the data leading up to the trigger condition
propagation	the transmission of a signal through a computer system
propagation delay	the amount of time required for a signal to pass through a circuit
pts	points
pulse trains	multiple pulses
pulsed output	a form of counter signal generation by which a pulse is outputted when a counter reaches a certain value

Q

quantization error	the inherent uncertainty in digitizing an analog value due to the finite resolution of the conversion process
quantizer	a device that maps a variable from a continuous distribution to a discrete distribution

R

real time	a property of an event or system in which data is processed as it is acquired instead of being accumulated and processed at a later time
relative accuracy	a measure in LSB of the linearity of an ADC. It includes all non-linearity and quantization errors. It does not include offset and gain errors of the circuitry feeding the ADC.
resolution	the smallest signal increment that can be detected by a measurement system. Resolution can be expressed in bits, in proportions, or in percent of full scale. For example, a system has 12-bit resolution, one part in 4,096 resolution, and 0.0244% of full scale.
resource locking	a technique whereby a device is signaled not to use its local memory while the memory is in use from the bus
retry	an acknowledge by a destination that signifies that the cycle did not complete and should be repeated
ribbon cable	a flat cable in which the conductors are side by side
rise time	the difference in time between the 10% and 90% points of a system's step response
rms	root mean square—the square root of the average value of the square of the instantaneous signal amplitude; a measure of signal amplitude
ROM	read-only memory
RSE	<i>see</i> SE.
RTSI bus	real-time system integration bus—the National Instruments timing bus that connects DAQ boards directly, by means of connectors on top of the boards, for precise synchronization of functions

S

s	seconds
S	samples
sample counter	the clock that counts the output of the channel clock, in other words, the number of samples taken. On boards with simultaneous sampling, this counter counts the output of the scan clock and hence the number of scans.
SE	single-ended—a term used to describe an analog input that is measured with respect to a common ground
self-calibrating	a property of a DSA board that has an extremely stable onboard reference and calibrates its own A/D and D/A circuits without manual adjustments by the user
sensor	a device that responds to a physical stimulus (heat, light, sound, pressure, motion, flow, and so on), and produces a corresponding electrical signal
settling time	the amount of time required for a voltage to reach its final value within specified limits
Shannon Sampling Theorem	a law of sampling theory stating that if a continuous bandwidth-limited signal contains no frequency components higher than half the frequency at which it is sampled, then the original signal can be recovered without distortion
S/H	sample-and-hold—a circuit that acquires and stores an analog voltage on a capacitor for a short period of time
signal conditioning	the manipulation of signals to prepare them for digitizing
SNR	signal-to-noise ratio—the ratio of the overall rms signal level to the rms noise level, expressed in decibels
software trigger	a programmed event that triggers an event such as data acquisition
software triggering	a method of triggering in which you simulate an analog trigger using software. Also called conditional retrieval.
source impedance	a parameter of signal sources that reflects current-driving ability of voltage sources (lower is better) and the voltage-driving ability of current sources (higher is better)

SS	simultaneous sampling—a property of a system in which each input or output channel is digitized or updated at the same instant
S/s	samples per second—used to express the rate at which a DAQ board samples an analog signal
STC	system timing controller
switchless device	devices that do not require dip switches or jumpers to configure resources on the devices—also called Plug and Play devices
synchronous	(1) hardware—a property of an event that is synchronized to a reference clock (2) software—a property of a function that begins an operation and returns only when the operation is complete
system noise	a measure of the amount of noise seen by an analog circuit or an ADC when the analog inputs are grounded
system RAM	RAM installed on a personal computer and used by the operating system, as contrasted with onboard RAM
T	
TC	terminal count—the highest value of a counter
T/H	track-and-hold—a circuit that tracks an analog voltage and holds the value on command
THD	total harmonic distortion—the ratio of the total rms signal due to harmonic distortion to the overall rms signal, in decibel or a percentage
THD+N	signal-to-THD plus noise—the ratio in decibels of the overall rms signal to the rms signal of harmonic distortion plus noise introduced
throughput rate	the data, measured in bytes/s, for a given continuous operation, calculated to include software overhead.

TIO	timing input/output—the TIO ASIC is a timing and triggering controller. It includes four general-purpose counter/ timers used for applications such as event counting, period and frequency measurement, and pulse train generation. The counters are a full 32 bits wide and include interfacing options for quadrature encoders and high-frequency signals. Additionally, the TIO has a robust trigger routing and condition mechanism for connecting RTSI bus and board specific trigger and timing signals. The TIO also provides advanced DIO capabilities for time stamping multiple I/O lines and controlling digital output lines.
transducer	<i>See</i> sensor
transducer excitation	a type of signal conditioning that uses external voltages and currents to excite the circuitry of a signal conditioning system into measuring physical phenomena
transfer rate	the rate, measured in bytes/s, at which data is moved from source to destination after software initialization and set up operations; the maximum rate at which the hardware can operate
trigger	any event that causes or starts some form of data capture
TTL	transistor-transistor logic
U	
unipolar	a signal range that is always positive (for example, 0 to +10 V)
update	the output equivalent of a scan. One or more analog or digital output samples. Typically, the number of output samples in an update is equal to the number of channels in the output group. For example, one pulse from the update clock produces one update which sends one new sample to every analog output channel in the group.
update rate	the number of output updates per second
V	
V	volts
V _{DC}	volts direct current

VI	virtual instrument—(1) a combination of hardware and/or software elements, typically used with a PC, that has the functionality of a classic stand-alone instrument (2) a LabVIEW software module (VI), which consists of a front panel user interface and a block diagram program
V_{IH}	volts, input high
V_{IL}	volts, input low
V_{in}	volts in
V_{OH}	volts, output high
V_{OL}	volts, output low
V_{ref}	reference voltage

W

waveform	multiple voltage readings taken at a specific sampling rate
word	the standard number of bits that a processor or memory manipulates at one time. Microprocessors typically use 8-, 16-, or 32-bit words.
working voltage	the highest voltage that should be applied to a product in normal use, normally well under the breakdown voltage for safety margin.

Z

zero-overhead looping	the ability of a high-performance processor to repeat instructions without requiring time to branch to the beginning of the instructions
zero-wait-state memory	memory fast enough that the processor does not have to wait during any reads and writes to the memory

Index

Numbers

- +5 V signal
 - analog I/O pin assignments (table), 4-3
 - analog I/O signal summary (table), 4-4
 - analog power connections, 4-10
 - digital I/O pin assignments (table), 4-6
 - digital I/O signal summary (table), 4-6
 - digital power connections, 4-12
 - self-resetting fuse, 4-10, 4-12

A

- AC input coupling, 3-4
- +ACH<0..3> signal
 - analog I/O pin assignments (table), 4-3
 - analog I/O signal connections, 4-7
 - analog I/O signal summary, 4-4
- ACH<0..3> signal
 - analog I/O pin assignments (table), 4-3
 - analog I/O signal connections, 4-7
 - analog I/O signal summary, 4-4
- ADC, 6-7 to 6-8
- AIGND signal
 - analog input signal connections, 4-6 to 4-7
 - analog I/O pin assignments (table), 4-3
 - analog I/O signal summary, 4-4
- amplifier characteristic specifications, A-3 to A-4
- analog function (block diagram), 3-3
- analog input, 3-4 to 3-6
 - input coupling, 3-4
 - input mode, 3-4
 - input polarity and range, 3-4 to 3-5
 - input range selection considerations, 3-5 to 3-6

- signal connections
 - analog input stage (figure), 4-7
 - description, 4-6 to 4-7
- specifications, A-1 to A-5
 - amplifier characteristics, A-3 to A-4
 - channel characteristics, A-1 to A-2
 - dynamic characteristics, A-4 to A-5
 - transfer characteristics, A-2
- analog input circuitry, 6-1 to 6-8
 - ADC, 6-7 to 6-8
 - antialias filtering, 6-3 to 6-7
 - calibration, 6-2
 - input coupling, 6-2
 - noise, 6-8
- analog I/O connector signal descriptions
 - pin assignments (table), 4-3
 - pin connections (figure), 4-2
 - signal summary (table), 4-4
- analog operation theory, 6-1 to 6-13
 - analog input circuitry, 6-1 to 6-8
 - ADC, 6-7 to 6-8
 - antialias filtering, 6-3 to 6-7
 - calibration, 6-2
 - input coupling, 6-2
 - noise, 6-8
 - analog output circuitry, 6-10 to 6-13
 - anti-image filtering, 6-10 to 6-11
 - calibration, 6-12
 - DAC, 6-11 to 6-12
 - mute feature, 6-12 to 6-13
- analog output, 3-6 to 3-7
 - output mode, 3-6
 - output polarity and range, 3-6 to 3-7

- signal connections
 - analog output channel
 - block diagram, 4-9
 - description, 4-8 to 4-10
- specifications, A-5 to A-7
 - channel characteristics, A-5 to A-6
 - dynamic characteristics, A-7
 - transfer characteristics, A-6
 - voltage output, A-6
- analog output circuitry, 6-10 to 6-13
 - anti-image filtering, 6-10 to 6-11
 - calibration, 6-12
 - DAC, 6-11 to 6-12
 - mute feature, 6-12 to 6-13
- analog power connections, 4-10
- analog trigger, 3-7 to 3-11
 - above-high-level analog triggering mode (figure), 3-9
 - below-low-level analog triggering mode (figure), 3-9
 - high-hysteresis analog triggering mode (figure), 3-10
 - inside-region analog triggering mode (figure), 3-10
 - low-hysteresis analog triggering mode (figure), 3-10
 - specifications, A-8
- antialias filtering, 6-3 to 6-7
 - alias rejection at oversample rate (figure), 6-6
 - clipped or overranged, 6-6 to 6-7
 - comparison of clipped signal to proper signal (figure), 6-7
 - frequency response, 6-5
 - input frequency response (figure), 6-4
 - input frequency response near cutoff (figure), 6-5
 - Nyquist frequency example, 6-3

- anti-image filtering
 - signal spectra in DAC (figure), 6-11
 - theory of operation, 6-10 to 6-11
- AOGND signal
 - analog I/O pin assignments (table), 4-3
 - analog I/O signal summary (table), 4-4
 - analog output signal connections, 4-8 to 4-9

B

- bipolar input, 3-4
- bipolar output, 3-6
- block diagrams
 - analog function, 3-3
 - digital function, 3-2
- bulletin board support, C-1
- bus interface specifications, A-9

C

- cables. *See also* I/O connectors.
 - custom cabling, 1-5
 - field wiring considerations, 4-12 to 4-13
 - optional equipment, 1-4
- calibration, 5-1 to 5-3
 - external calibration, 5-2 to 5-3
 - loading calibration constants, 5-1 to 5-2
 - onboard calibration reference
 - specifications, A-5
 - self-calibration, 5-2
 - specifications, A-10
 - theory of operation
 - analog input circuitry, 6-2
 - analog output circuitry, 6-12
 - traceable recalibration, 5-3
- channel characteristic specifications
 - analog input, A-1 to A-2
 - analog output, A-5 to A-6
- ComponentWorks software, 1-4

configuration
 effect of sampling and update rates,
 3-12 to 3-13
 instrument configuration, 2-2
 connectors. *See* I/O connectors.
 customer communication, *xi*, C-1 to C-2

D

DAC

mute feature, 6-12 to 6-13
 signal spectra in DAC (figure), 6-11
 theory of operation, 6-11 to 6-12

+DAC0OUT signal

analog I/O pin assignments (table), 4-3
 analog I/O signal summary (table), 4-4
 analog output signal connections,
 4-8 to 4-9

-DAC0OUT signal

analog I/O pin assignments (table), 4-3
 analog I/O signal summary (table), 4-4
 analog output signal connections,
 4-8 to 4-9

+DAC1OUT signal

analog I/O pin assignments (table), 4-3
 analog I/O signal summary (table), 4-4
 analog output signal connections,
 4-8 to 4-9

-DAC1OUT signal

analog I/O pin assignments (table), 4-3
 analog I/O signal summary (table), 4-4
 analog output signal connections,
 4-8 to 4-9

DC input coupling, 3-4

DDS (direct digital synthesis)
 technology, 3-12

delta-sigma modulation

analog input circuitry, 6-7
 analog output circuitry, 6-9
 overview, 1-1
 triggering effect, 3-8

DGND signal

analog I/O pin assignments (table), 4-3
 analog I/O signal summary (table), 4-4
 digital I/O pin assignments (table), 4-6
 digital I/O signal connections, 4-10
 digital I/O signal summary (table), 4-6

DIFF configuration, 4-7

DIFF input mode, 3-4

DIFF output mode, 3-6

digital function (block diagram), 3-2

digital I/O

high impedance state (note), 3-11
 overview, 3-11

pin connections (figure), 4-5

signal connections, 4-10 to 4-11

signal descriptions

pin assignments (table), 4-6

pin connections (figure), 4-5

signal summary (table), 4-6

specifications, A-7 to A-8

digital power connections, 4-12

digital trigger specifications, A-9

DIO<0..31> signal

digital I/O pin assignments (table), 4-6

digital I/O signal connections,
 4-10 to 4-11

digital I/O signal summary (table), 4-6

direct digital synthesis (DDS)

technology, 3-12

documentation

conventions used in manual, *x*

National Instruments documentation, *xi*

organization of manual, *ix-x*

related documentation, *xi*

dynamic characteristic specifications

analog input, A-4 to A-5

analog output, A-7

E

- EEPROM storage of calibration constants, 5-1
- electronic support services, C-1 to C-2
- e-mail support, C-2
- environment specifications, A-10
- environmental noise, avoiding, 4-12 to 4-13, 6-8
- equipment, optional, 1-4
- EXT_TRIG signal
 - digital I/O pin assignments (table), 4-6
 - digital I/O signal summary (table), 4-6

F

- fax and telephone support numbers, C-2
- Fax-on-Demand support, C-2
- field wiring considerations, 4-12 to 4-13
- floating signal sources, 4-8
- FTP support, C-1
- fuse, self-resetting
 - analog power connections, 4-10
 - digital power connections, 4-12

G

- ground-referenced signal sources, 4-8

H

- hardware installation, 2-1 to 2-2
- hardware overview
 - analog input, 3-4 to 3-6
 - input mode, 3-4
 - input polarity and range, 3-4 to 3-5
 - input range selection considerations, 3-5 to 3-6
 - analog output, 3-6 to 3-7
 - analog trigger, 3-7 to 3-11
 - block diagrams
 - analog function, 3-3
 - digital function, 3-2
 - digital I/O, 3-11
 - LEDs, 3-13
 - timing signal routing, 3-11

I

- input coupling
 - analog input, 3-4
 - theory of operation, 6-2
- input mode, 3-4
- input polarity and range, 3-4 to 3-5
 - actual range and measurement precision (table), 3-5
 - exceeding rated input voltages (caution), 3-5
 - selection considerations, 3-5 to 3-6
- installation
 - hardware, 2-1 to 2-2
 - software, 2-1
 - unpacking NI 4451/4552, 1-2 to 1-3
- instrument configuration, 2-2
- I/O connectors, 4-1 to 4-6
 - developing cable connectors, 1-5
 - exceeding maximum ratings (warning), 4-1
 - pin assignments (table)
 - analog I/O, 4-3
 - digital I/O, 4-6
 - pin connections
 - 68-pin digital connector (figure), B-2
 - analog (figure), 4-2
 - digital (figure), 4-5
 - signal summary (table)
 - analog I/O, 4-4
 - digital I/O, 4-6

J

- jitter, with triggering, 3-8

L

LabVIEW and LabWindows/CVI application software, 1-3
LEDs, 3-13

M

manual. *See* documentation.
Measure software, 1-4
mute feature, 6-12 to 6-13

N

NI 4451/4552. *See also* hardware overview.
custom cabling, 1-5
optional equipment, 1-4
overview, 1-1
requirements for getting started, 1-2
software programming choices, 1-3 to 1-4
 ComponentWorks, 1-4
 LabVIEW and LabWindows/CVI, 1-3
 Measure, 1-4
 National Instruments application software, 1-3 to 1-4
 VirtualBench, 1-3 to 1-4
unpacking, 1-2 to 1-3
noise, avoiding, 4-12 to 4-13, 6-8
Nyquist frequency, 6-2

O

onboard calibration reference specifications, A-5
optional equipment, 1-4
output mode, 3-6
output polarity and range, 3-6 to 3-7
 actual range and measurement precision (table), 3-7
 power-up modes (note), 3-7

P

physical specifications, A-10
pin assignments
 analog I/O (table), 4-3
 digital I/O (table), 4-6
pin connections (figure)
 68-pin digital connector (figure), B-2
 analog I/O, 4-2
 digital I/O, 4-5
polarity selection
 analog input, 3-4 to 3-5
 actual range and measurement precision (table), 3-5
 selection considerations, 3-5 to 3-6
 analog output, 3-6 to 3-7
 actual range and measurement precision, 3-7
power connections
 analog power connections, 4-10
 digital power connections, 4-12
power requirement specifications, A-9

R

recalibration, traceable, 5-3
requirements for getting started, 1-2
RTSI trigger lines, 3-11

S

sample/update clock frequency, selecting, 3-12 to 3-13
signal connections
 analog input, 4-6 to 4-7
 digital I/O, 4-10 to 4-11
 field wiring considerations, 4-12 to 4-13
 I/O connectors, 4-1 to 4-6
 68-pin digital connector pin connections (figure), B-2
 analog I/O pin assignments (table), 4-3

- analog I/O pin connections
 - (figure), 4-2
 - analog I/O signal summary
 - (table), 4-4
 - digital I/O pin assignments
 - (table), 4-6
 - digital I/O signal summary
 - (table), 4-6
 - digital pin connections (figure), 4-5
 - exceeding maximum ratings
 - (warning), 4-1
 - power connections, 4-10
 - types of signal sources, 4-8
 - floating, 4-8
 - ground-referenced, 4-8
 - software installation, 2-1
 - software programming choices, 1-3 to 1-4
 - ComponentWorks, 1-4
 - LabVIEW and LabWindows/CVI, 1-3
 - Measure, 1-4
 - National Instruments application
 - software, 1-3 to 1-4
 - VirtualBench, 1-3 to 1-4
 - specifications
 - analog input, A-1 to A-5
 - amplifier characteristics, A-3 to A-4
 - channel characteristics, A-1 to A-2
 - dynamic characteristics, A-4 to A-5
 - transfer characteristics, A-2
 - analog output, A-5 to A-7
 - channel characteristics, A-5 to A-6
 - dynamic characteristics, A-7
 - transfer characteristics, A-6
 - voltage output, A-6
 - analog trigger, A-8
 - bus interface, A-9
 - calibration, A-10
 - digital I/O, A-7 to A-8
 - digital trigger, A-9
 - environment, A-10
 - onboard calibration reference, A-5
 - physical, A-10
 - power requirements, A-9
- ## T
- technical support, C-1 to C-2
 - telephone and fax support numbers, C-2
 - theory of operation. *See* analog operation theory.
 - timing signal routing, 3-11
 - traceable recalibration, 5-3
 - transfer characteristic specifications
 - analog input, A-2
 - analog output, A-6
 - triggers
 - analog, 3-7 to 3-11
 - above-high-level triggering mode
 - (figure), 3-9
 - below-low-level triggering mode
 - (figure), 3-9
 - high-hysteresis triggering
 - mode, 3-10
 - inside-region triggering mode
 - (figure), 3-10
 - low-hysteresis triggering mode, 3-10
 - specifications, A-8
 - digital, specifications, A-9
 - RTSI triggers, 3-11
 - specifications, A-8 to A-9
- ## U
- unipolar input/output. *See* polarity selection.
 - unpacking NI 4451/4552, 1-2 to 1-3
 - update clock frequency, selecting,
 - 3-12 to 3-13

V

VirtualBench software, 1-3 to 1-4
voltage output specifications, A-6

W

wiring considerations, 4-12 to 4-13