

# SPECIFICATION CORRECTION FOR THE NI 6036E

In the *NI 6034E/6035E/6036E User Manual*, the *NI 6036E Accuracy Information* table in Appendix A, *Specifications*, incorrectly labels the unit for Offset as mV. The correct unit of measure for Offset is  $\mu\text{V}$ . To view the manual, refer to [ni.com/manuals](http://ni.com/manuals). The corrected specification should read as follows (change shown in *italics*):

## Accuracy Information

### NI 6036E Accuracy Information

Nominal Range (V)		Absolute Accuracy				
		% of Reading			Offset ( $\mu\text{V}$ )	Temp Drift (%/°C)
Positive FS	Negative FS	24 Hours	90 Days	1 Year		
10	-10	0.0089	0.0109	0.0131	1102.94	0.0005